

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP013023\
 Data File : BP013641.D
 Acq On : 30 Jan 2023 12:24
 Operator : CG/JU
 Sample : PB150519BL
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 PB150519BL

Quant Time: Jan 31 04:21:14 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP012623.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Jan 31 04:17:33 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.158	152	236389	20.000	ng	0.00
21) Naphthalene-d8	10.987	136	891191	20.000	ng	0.00
39) Acenaphthene-d10	14.787	164	585042	20.000	ng	0.00
64) Phenanthrene-d10	17.540	188	1239052	20.000	ng	0.00
76) Chrysene-d12	21.616	240	998899	20.000	ng	0.00
86) Perylene-d12	24.204	264	1067659	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.693	112	1868908	141.674	ng	0.01
7) Phenol-d6	7.305	99	2402881	134.839	ng	0.00
23) Nitrobenzene-d5	9.334	82	1469015	90.574	ng	0.01
42) 2,4,6-Tribromophenol	16.281	330	1173342	131.730	ng	0.00
45) 2-Fluorobiphenyl	13.416	172	3676614	89.754	ng	0.00
79) Terphenyl-d14	20.075	244	5268085	93.065	ng	0.00

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP013023\
Data File : BP013641.D
Acq On : 30 Jan 2023 12:24
Operator : CG/JU
Sample : PB150519BL
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
PB150519BL

Quant Time: Jan 31 04:21:14 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP012623.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Jan 31 04:17:33 2023
Response via : Initial Calibration

