

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP013023\
 Data File : BP013657.D
 Acq On : 30 Jan 2023 22:01
 Operator : CG/JU
 Sample : 01329-02
 Misc :
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 COMP-6-BORING-2-7-14

Quant Time: Jan 31 04:44:46 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP012623.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Jan 31 04:38:43 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.146	152	229118	20.000	ng	0.00
21) Naphthalene-d8	10.975	136	813768	20.000	ng	0.00
39) Acenaphthene-d10	14.781	164	478950	20.000	ng	0.00
64) Phenanthrene-d10	17.534	188	1031314	20.000	ng	0.00
76) Chrysene-d12	21.610	240	1037176	20.000	ng	-0.01
86) Perylene-d12	24.186	264	1308727	20.000	ng	-0.02
System Monitoring Compounds						
5) 2-Fluorophenol	5.681	112	1924305	150.502	ng	0.00
7) Phenol-d6	7.293	99	2231827	129.215	ng	0.00
23) Nitrobenzene-d5	9.322	82	1422111	96.024	ng	0.00
42) 2,4,6-Tribromophenol	16.269	330	1025645	140.654	ng	0.00
45) 2-Fluorobiphenyl	13.404	172	3567948	106.395	ng	0.00
79) Terphenyl-d14	20.063	244	5748838	97.810	ng	0.00
Target Compounds						
						Qvalue
74) Di-n-butylphthalate	18.463	149	12606	3.836	ng	# 98
84) Bis(2-ethylhexyl)phtha...	21.492	149	21161	4.707	ng	# 97
85) Di-n-octyl phthalate	22.480	149	45938	3.223	ng	# 93
88) Benzo(b)fluoranthene	23.392	252	200450	2.386	ng	# 98
89) Benzo(k)fluoranthene	23.445	252	208850	2.504	ng	# 98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP013023\
Data File : BP013657.D
Acq On : 30 Jan 2023 22:01
Operator : CG/JU
Sample : 01329-02
Misc :
ALS Vial : 11 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
COMP-6-BORING-2-7-14

Quant Time: Jan 31 04:44:46 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP012623.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Jan 31 04:38:43 2023
Response via : Initial Calibration

