

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP013024\
 Data File : BP019489.D
 Acq On : 30 Jan 2024 19:17
 Operator : MA/JU
 Sample : SSTDICV040
 Misc :
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 ICVBP013024

Quant Time: Jan 31 01:57:11 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP013024.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Jan 31 01:55:36 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	115	0.00
2	1,4-Dioxane	40.000	35.907	10.2	110	0.00
3	Pyridine	40.000	37.408	6.5	110	0.00
4	n-Nitrosodimethylamine	40.000	37.771	5.6	113	0.00
5 S	2-Fluorophenol	80.000	73.335	8.3	112	0.00
6	Aniline	40.000	39.470	1.3	118	0.00
7 S	Phenol-d6	80.000	78.171	2.3	119	0.00
8	2-Chlorophenol	40.000	38.311	4.2	119	0.00
9	Benzaldehyde	40.000	40.359	-0.9	118	0.00
10 C	Phenol	40.000	38.902	2.7	119	0.00
11	bis(2-Chloroethyl)ether	40.000	38.325	4.2	117	0.00
12	1,3-Dichlorobenzene	40.000	37.343	6.6	117	0.00
13 C	1,4-Dichlorobenzene	40.000	37.255	6.9	115	0.00
14	1,2-Dichlorobenzene	40.000	37.549	6.1	117	0.00
15	Benzyl Alcohol	40.000	40.477	-1.2	123	0.00
16	2,2'-oxybis(1-Chloropropane	40.000	39.126	2.2	123	0.00
17	2-Methylphenol	40.000	39.655	0.9	121	0.00
18	Hexachloroethane	40.000	37.574	6.1	117	0.00
19 P	n-Nitroso-di-n-propylamine	40.000	41.029	-2.6	126	0.00
20	3+4-Methylphenols	40.000	40.341	-0.9	123	0.00
21 I	Naphthalene-d8	20.000	20.000	0.0	124	0.00
22	Acetophenone	40.000	38.273	4.3	125	0.00
23 S	Nitrobenzene-d5	80.000	76.664	4.2	123	0.00
24	Nitrobenzene	40.000	37.888	5.3	122	0.00
25	Isophorone	40.000	38.935	2.7	124	0.00
26 C	2-Nitrophenol	40.000	40.103	-0.3	127	0.00
27	2,4-Dimethylphenol	40.000	38.883	2.8	126	0.00
28	bis(2-Chloroethoxy)methane	40.000	38.910	2.7	127	0.00
29 C	2,4-Dichlorophenol	40.000	38.604	3.5	124	0.00
30	1,2,4-Trichlorobenzene	40.000	37.709	5.7	124	0.00
31	Naphthalene	40.000	37.963	5.1	124	0.00
32	Benzoic acid	40.000	42.047	-5.1	128	0.01
33	4-Chloroaniline	40.000	39.069	2.3	123	0.00
34 C	Hexachlorobutadiene	40.000	37.190	7.0	126	0.00
35	Caprolactam	40.000	36.646	8.4	110	0.00
36 C	4-Chloro-3-methylphenol	40.000	39.418	1.5	123	0.00
37	2-Methylnaphthalene	40.000	39.241	1.9	127	0.00
38	1-Methylnaphthalene	40.000	39.451	1.4	128	0.00
39 I	Acenaphthene-d10	20.000	20.000	0.0	124	0.00
40	1,2,4,5-Tetrachlorobenzene	40.000	37.795	5.5	126	0.00
41 P	Hexachlorocyclopentadiene	40.000	36.738	8.2	124	0.00
42 S	2,4,6-Tribromophenol	80.000	74.719	6.6	116	0.00
43 C	2,4,6-Trichlorophenol	40.000	38.344	4.1	123	0.00
44	2,4,5-Trichlorophenol	40.000	38.844	2.9	124	0.00
45 S	2-Fluorobiphenyl	80.000	76.214	4.7	126	0.00
46	1,1'-Biphenyl	40.000	38.033	4.9	126	0.00
47	2-Chloronaphthalene	40.000	38.461	3.8	125	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP013024\
 Data File : BP019489.D
 Acq On : 30 Jan 2024 19:17
 Operator : MA/JU
 Sample : SSTDICV040
 Misc :
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 ICVBP013024

Quant Time: Jan 31 01:57:11 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP013024.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Jan 31 01:55:36 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48	2-Nitroaniline	40.000	38.703	3.2	118	0.00
49	Acenaphthylene	40.000	38.146	4.6	123	0.00
50	Dimethylphthalate	40.000	37.373	6.6	119	0.00
51	2,6-Dinitrotoluene	40.000	38.244	4.4	120	0.00
52 C	Acenaphthene	40.000	37.997	5.0	122	0.00
53	3-Nitroaniline	40.000	37.338	6.7	113	0.00
54 P	2,4-Dinitrophenol	40.000	38.986	2.5	116	0.00
55	Dibenzofuran	40.000	37.644	5.9	122	0.00
56 P	4-Nitrophenol	40.000	36.221	9.4	106	0.00
57	2,4-Dinitrotoluene	40.000	37.542	6.1	112	0.00
58	Fluorene	40.000	37.490	6.3	119	0.00
59	2,3,4,6-Tetrachlorophenol	40.000	37.501	6.2	119	0.00
60	Diethylphthalate	40.000	36.686	8.3	115	0.00
61	4-Chlorophenyl-phenylether	40.000	38.000	5.0	121	0.00
62	4-Nitroaniline	40.000	35.147	12.1	104	0.00
63	Azobenzene	40.000	37.452	6.4	118	0.00
64 I	Phenanthrene-d10	20.000	20.000	0.0	106	0.00
65	4,6-Dinitro-2-methylphenol	40.000	41.700	-4.3	111	0.00
66 c	n-Nitrosodiphenylamine	40.000	40.594	-1.5	116	0.00
67	4-Bromophenyl-phenylether	40.000	40.733	-1.8	118	0.00
68	Hexachlorobenzene	40.000	39.722	0.7	114	0.00
69	Atrazine	40.000	36.803	8.0	101	0.00
70 C	Pentachlorophenol	40.000	39.319	1.7	106	0.00
71	Phenanthrene	40.000	38.111	4.7	105	0.00
72	Anthracene	40.000	37.837	5.4	104	0.00
73	Carbazole	40.000	35.476	11.3	96	0.00
74	Di-n-butylphthalate	40.000	35.304	11.7	97	0.00
75 C	Fluoranthene	40.000	32.809	18.0	88	0.00
76 I	Chrysene-d12	20.000	20.000	0.0	75	0.00
77	Benzydine	40.000	35.993	10.0	70	0.00
78	Pyrene	40.000	41.094	-2.7	86	0.00
79 S	Terphenyl-d14	80.000	81.767	-2.2	85	0.00
80	Butylbenzylphthalate	40.000	38.097	4.8	77	0.00
81	Benzo(a)anthracene	40.000	38.044	4.9	75	0.00
82	3,3'-Dichlorobenzidine	40.000	37.021	7.4	72	0.00
83	Chrysene	40.000	37.883	5.3	75	0.00
84	Bis(2-ethylhexyl)phthalate	40.000	36.923	7.7	73	0.00
85 c	Di-n-octyl phthalate	40.000	37.438	6.4	71	0.00
86 I	Perylene-d12	20.000	20.000	0.0	76	0.00
87	Indeno(1,2,3-cd)pyrene	40.000	40.431	-1.1	80	0.00
88	Benzo(b)fluoranthene	40.000	37.487	6.3	74	0.00
89	Benzo(k)fluoranthene	40.000	36.565	8.6	74	0.00
90 C	Benzo(a)pyrene	40.000	37.908	5.2	76	0.00
91	Dibenzo(a,h)anthracene	40.000	40.234	-0.6	79	0.00
92	Benzo(g,h,i)perylene	40.000	40.521	-1.3	81	-0.02

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP013024\
Data File : BP019489.D
Acq On : 30 Jan 2024 19:17
Operator : MA/JU
Sample : SSTDICV040
Misc :
ALS Vial : 11 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
ICVBP013024

Quant Time: Jan 31 01:57:11 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP013024.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Jan 31 01:55:36 2024
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range SPCC's out = 0 CCC's out = 0