

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP013122\
 Data File : BP008970.D
 Acq On : 31 Jan 2022 18:03
 Operator : CG/JU
 Sample : N1303-05
 Misc :
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 CC-7

Quant Time: Feb 01 04:21:15 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP011422.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Feb 01 04:16:03 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.828	152	194060	20.000	ng	0.00
21) Naphthalene-d8	10.628	136	722165	20.000	ng	0.00
39) Acenaphthene-d10	14.469	164	471936	20.000	ng	0.00
64) Phenanthrene-d10	17.222	188	1025676	20.000	ng	-0.01
76) Chrysene-d12	21.316	240	1161554	20.000	ng	-0.01
86) Perylene-d12	23.721	264	1302500	20.000	ng	-0.02
System Monitoring Compounds						
5) 2-Fluorophenol	5.422	112	27237	2.170	ng	0.00
7) Phenol-d6	7.011	99	463091	30.474	ng	0.00
23) Nitrobenzene-d5	8.987	82	944440	74.248	ng	-0.01
42) 2,4,6-Tribromophenol	15.981	330	1209	0.176	ng	0.00
45) 2-Fluorobiphenyl	13.087	172	2367798	66.164	ng	-0.01
79) Terphenyl-d14	19.786	244	4128222	59.990	ng	-0.01
Target Compounds						Qvalue
9) Benzaldehyde	6.975	77	138155	13.603	ng	# 1
18) Hexachloroethane	8.934	117	230096	42.027	ng	# 1
19) n-Nitroso-di-n-propyla...	8.599	70	38076	4.286	ng	# 51
22) Acetophenone	8.634	105	525743	28.578	ng	# 62
24) Nitrobenzene	9.028	77	36137	2.813	ng	# 35
25) Isophorone	9.552	82	114029	4.960	ng	# 1
31) Naphthalene	10.675	128	2257655	57.559	ng	99
33) 4-Chloroaniline	10.675	127	297496	18.612	ng	# 34
37) 2-Methylnaphthalene	12.293	142	303837	10.596	ng	98
38) 1-Methylnaphthalene	12.510	142	168942	6.419	ng	97
52) Acenaphthene	14.534	154	97971	3.693	ng	98
55) Dibenzofuran	14.869	168	126163	2.918	ng	96
58) Fluorene	15.516	166	162100	4.747	ng	100
71) Phenanthrene	17.263	178	1848843	32.015	ng	100
72) Anthracene	17.357	178	340267	5.874	ng	99
73) Carbazole	17.634	167	147921	2.970	ng	99
75) Fluoranthene	19.239	202	2267495	35.308	ng	96
78) Pyrene	19.592	202	2139084	26.393	ng	99
81) Benzo(a)anthracene	21.298	228	1230276	15.744	ng	99
83) Chrysene	21.351	228	1220686	16.114	ng	97
84) Bis(2-ethylhexyl)phtha...	21.227	149	178904	3.544	ng	98
87) Indeno(1,2,3-cd)pyrene	26.233	276	763970	7.118	ng	96
88) Benzo(b)fluoranthene	22.992	252	2072240	23.823	ng	99
89) Benzo(k)fluoranthene	22.992	252	2072240	24.611	ng	99
90) Benzo(a)pyrene	23.616	252	1052871	12.540	ng	98
91) Dibenzo(a,h)anthracene	26.239	278	219774	2.408	ng	# 90
92) Benzo(g,h,i)perylene	27.010	276	758545	8.513	ng	100


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP013122\
Data File : BP008970.D
Acq On : 31 Jan 2022 18:03
Operator : CG/JU
Sample : N1303-05
Misc :
ALS Vial : 12 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
CC-7

Quant Time: Feb 01 04:21:15 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP011422.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Feb 01 04:16:03 2022
Response via : Initial Calibration









#1
1,4-Dichlorobenzene-d4
Concen: 20.000 ng
RT: 7.828 min Scan# 81
Delta R.T. -0.006 min
Lab File: BP008970.D
Acq: 31 Jan 2022 18:03




Instrument :
BNA_P
ClientSampleId :
CC-7


Tgt Ion:152 Resp: 194060
Ion Ratio Lower Upper
152 100
150 149.9 127.0 190.6
115 53.9 43.7 65.5



#5
2-Fluorophenol
Concen: 2.170 ng
RT: 5.422 min Scan# 414
Delta R.T. 0.000 min
Lab File: BP008970.D
Acq: 31 Jan 2022 18:03

Tgt Ion:112 Resp: 27237
Ion Ratio Lower Upper
112 100
64 50.0 39.7 59.5
63 27.8 20.0 30.0









#7
Phenol-d6
Concen: 30.474 ng
RT: 7.011 min Scan# 61
Delta R.T. -0.006 min
Lab File: BP008970.D
Acq: 31 Jan 2022 18:03




Instrument :
BNA_P
ClientSampleId :
CC-7


Tgt Ion: 99 Resp: 463091
Ion Ratio Lower Upper
99 100
42 17.7 10.8 16.2#
71 40.8 23.8 35.6#



#9
Benzaldehyde
Concen: 13.603 ng
RT: 6.975 min Scan# 678
Delta R.T. 0.000 min
Lab File: BP008970.D
Acq: 31 Jan 2022 18:03

Tgt Ion: 77 Resp: 138155
Ion Ratio Lower Upper
77 100
105 813.1 84.2 124.2#
106 74.7 80.6 120.6#









#18
Hexachloroethane
Concen: 42.027 ng
RT: 8.934 min Scan# 1012
Delta R.T. 0.024 min
Lab File: BP008970.D
Acq: 31 Jan 2022 18:03




Instrument :
BNA_P
ClientSampleId :
CC-7


Tgt Ion:117 Resp: 230096
Ion Ratio Lower Upper
117 100
119 533.2 76.9 115.3#
201 0.0 93.7 140.5#



#19
n-Nitroso-di-n-propylamine
Concen: 4.286 ng
RT: 8.599 min Scan# 954
Delta R.T. -0.047 min
Lab File: BP008970.D
Acq: 31 Jan 2022 18:03

Tgt Ion: 70 Resp: 38076
Ion Ratio Lower Upper
70 100
42 82.6 38.2 57.2#
101 0.0 8.7 13.1#
130 0.3 21.0 31.6#









#21
Naphthalene-d8
Concen: 20.000 ng
RT: 10.628 min Scan# 11
Delta R.T. -0.006 min
Lab File: BP008970.D
Acq: 31 Jan 2022 18:03




Instrument :
BNA_P
ClientSampleId :
CC-7


Tgt Ion:	136	Resp:	722165
Ion Ratio	Lower	Upper	
136	100		
137	11.0	9.0	13.4
54	6.8	4.9	7.3
68	5.8	4.3	6.5



#22
Acetophenone
Concen: 28.578 ng
RT: 8.634 min Scan# 960
Delta R.T. -0.029 min
Lab File: BP008970.D
Acq: 31 Jan 2022 18:03

Tgt Ion:	105	Resp:	525743
Ion Ratio	Lower	Upper	
105	100		
71	5.3	5.3	7.9#
51	3.6	16.3	24.5#
120	0.3	19.2	28.8#





#23

Nitrobenzene-d5

Concen: 74.248 ng

RT: 8.987 min Scan# 1020

Delta R.T. -0.011 min

Lab File: BP008970.D

Acq: 31 Jan 2022 18:03

Instrument :

BNA_P

ClientSampleId :

CC-7





Tgt Ion: 82 Resp: 944440

Ion Ratio Lower Upper

82 100

128 45.4 39.4 59.2

54 42.9 32.6 48.8



#24

Nitrobenzene

Concen: 2.813 ng

RT: 9.028 min Scan# 1027

Delta R.T. -0.011 min

Lab File: BP008970.D

Acq: 31 Jan 2022 18:03




Tgt Ion: 77 Resp: 36137


Ion Ratio Lower Upper

77 100

123 18.8 41.4 62.2#

65 66.8 9.8 14.8#









#25
Isophorone
Concen: 4.960 ng
RT: 9.552 min Scan# 1116
Delta R.T. -0.017 min
Lab File: BP008970.D
Acq: 31 Jan 2022 18:03




Instrument :
BNA_P
ClientSampleId :
CC-7


Tgt Ion: 82 Resp: 114029
Ion Ratio Lower Upper
82 100
95 124.7 5.4 8.2#
138 11.3 16.6 24.8#



#31
Naphthalene
Concen: 57.559 ng
RT: 10.675 min Scan# 1307
Delta R.T. -0.012 min
Lab File: BP008970.D
Acq: 31 Jan 2022 18:03

Tgt Ion: 128 Resp: 2257655
Ion Ratio Lower Upper
128 100
129 11.7 8.8 13.2
127 13.2 10.4 15.6









#33
4-Chloroaniline
Concen: 18.612 ng
RT: 10.675 min Scan# 11
Delta R.T. -0.123 min
Lab File: BP008970.D
Acq: 31 Jan 2022 18:03




Instrument :
BNA_P
ClientSampleId :
CC-7


Tgt Ion:127	Resp:	297496	
Ion	Ratio	Lower	Upper
127	100		
129	88.2	26.0	39.0#
65	4.2	20.5	30.7#
92	0.9	14.1	21.1#



#37
2-Methylnaphthalene
Concen: 10.596 ng
RT: 12.293 min Scan# 1582
Delta R.T. -0.006 min
Lab File: BP008970.D
Acq: 31 Jan 2022 18:03

Tgt Ion:142	Resp:	303837	
Ion	Ratio	Lower	Upper
142	100		
141	86.5	70.5	105.7
115	32.5	24.6	36.8









#38
1-Methylnaphthalene
Concen: 6.419 ng
RT: 12.510 min Scan# 1
Delta R.T. -0.006 min
Lab File: BP008970.D
Acq: 31 Jan 2022 18:03




Instrument :
BNA_P
ClientSampleId :
CC-7


Tgt Ion:142 Resp: 168942
Ion Ratio Lower Upper
142 100
141 94.4 73.2 109.8
116 3.9 2.7 4.1



#39
Acenaphthene-d10
Concen: 20.000 ng
RT: 14.469 min Scan# 1952
Delta R.T. -0.006 min
Lab File: BP008970.D
Acq: 31 Jan 2022 18:03

Tgt Ion:164 Resp: 471936
Ion Ratio Lower Upper
164 100
162 99.8 79.9 119.9
160 45.6 35.5 53.3









#42
2,4,6-Tribromophenol
Concen: 0.176 ng
RT: 15.981 min Scan# 21
Delta R.T. 0.006 min
Lab File: BP008970.D
Acq: 31 Jan 2022 18:03




Instrument :
BNA_P
ClientSampleId :
CC-7


Tgt Ion:330 Resp: 1209
Ion Ratio Lower Upper
330 100
332 97.3 77.0 115.4
141 48.1 21.3 31.9#



#45
2-Fluorobiphenyl
Concen: 66.164 ng
RT: 13.087 min Scan# 1717
Delta R.T. -0.011 min
Lab File: BP008970.D
Acq: 31 Jan 2022 18:03

Tgt Ion:172 Resp: 2367798
Ion Ratio Lower Upper
172 100
171 35.0 29.3 43.9
170 23.7 19.5 29.3









#52
Acenaphthene
Concen: 3.693 ng
RT: 14.534 min Scan# 1963
Delta R.T. -0.006 min
Lab File: BP008970.D
Acq: 31 Jan 2022 18:03




Instrument :
BNA_P
ClientSampleId :
CC-7


Tgt Ion:154 Resp: 97971
Ion Ratio Lower Upper
154 100
153 113.2 89.4 134.2
152 56.0 42.6 63.8



#55
Dibenzofuran
Concen: 2.918 ng
RT: 14.869 min Scan# 2020
Delta R.T. -0.006 min
Lab File: BP008970.D
Acq: 31 Jan 2022 18:03

Tgt Ion:168 Resp: 126163
Ion Ratio Lower Upper
168 100
139 36.2 27.4 41.0
169 15.4 10.9 16.3









#58
Fluorene
Concen: 4.747 ng
RT: 15.516 min Scan# 2135
Delta R.T. -0.011 min
Lab File: BP008970.D
Acq: 31 Jan 2022 18:03




Instrument :
BNA_P
ClientSampleId :
CC-7


Tgt Ion:166 Resp: 162100
Ion Ratio Lower Upper
166 100
165 96.7 77.2 115.8
167 14.7 11.2 16.8



#64
Phenanthrene-d10
Concen: 20.000 ng
RT: 17.222 min Scan# 2420
Delta R.T. -0.012 min
Lab File: BP008970.D
Acq: 31 Jan 2022 18:03

Tgt Ion:188 Resp: 1025676
Ion Ratio Lower Upper
188 100
94 7.9 6.8 10.2
80 7.9 7.2 10.8









#71
Phenanthrene
Concen: 32.015 ng
RT: 17.263 min Scan# 2432
Delta R.T. -0.012 min
Lab File: BP008970.D
Acq: 31 Jan 2022 18:03




Instrument :
BNA_P
ClientSampleId :
CC-7


Tgt Ion:178 Resp: 1848843
Ion Ratio Lower Upper
178 100
176 19.9 16.0 24.0
179 15.9 13.0 19.4



#72
Anthracene
Concen: 5.874 ng
RT: 17.357 min Scan# 2443
Delta R.T. -0.012 min
Lab File: BP008970.D
Acq: 31 Jan 2022 18:03

Tgt Ion:178 Resp: 340267
Ion Ratio Lower Upper
178 100
176 18.3 15.1 22.7
179 16.2 12.8 19.2









#73
 Carbazole
 Concen: 2.970 ng
 RT: 17.634 min Scan# 2490
 Delta R.T. -0.006 min
 Lab File: BP008970.D
 Acq: 31 Jan 2022 18:03




Instrument :
 BNA_P
 ClientSampleId :
 CC-7


Tgt Ion:167 Resp: 147921
 Ion Ratio Lower Upper
 167 100
 166 22.1 17.7 26.5
 139 13.3 9.8 14.6



#75
 Fluoranthene
 Concen: 35.308 ng
 RT: 19.239 min Scan# 2763
 Delta R.T. -0.006 min
 Lab File: BP008970.D
 Acq: 31 Jan 2022 18:03

Tgt Ion:202 Resp: 2267495
 Ion Ratio Lower Upper
 202 100
 101 8.5 0.0 30.7
 203 17.6 0.0 38.9









#76
Chrysene-d12
Concen: 20.000 ng
RT: 21.316 min Scan# 31
Delta R.T. -0.011 min
Lab File: BP008970.D
Acq: 31 Jan 2022 18:03




Instrument :
BNA_P
ClientSampleId :
CC-7


Tgt Ion:240 Resp: 1161554
Ion Ratio Lower Upper
240 100
120 8.9 7.7 11.5
236 26.9 21.2 31.8



#78
Pyrene
Concen: 26.393 ng
RT: 19.592 min Scan# 2823
Delta R.T. -0.006 min
Lab File: BP008970.D
Acq: 31 Jan 2022 18:03

Tgt Ion:202 Resp: 2139084
Ion Ratio Lower Upper
202 100
200 22.1 18.0 27.0
203 17.9 15.0 22.6









#79
 Terphenyl-d14
 Concen: 59.990 ng
 RT: 19.786 min Scan# 21
 Delta R.T. -0.012 min
 Lab File: BP008970.D
 Acq: 31 Jan 2022 18:03




Instrument :
 BNA_P
 ClientSampleId :
 CC-7


Tgt Ion:244 Resp: 4128222
 Ion Ratio Lower Upper
 244 100
 212 7.4 6.5 9.7
 122 9.0 9.2 13.8#



#81
 Benzo(a)anthracene
 Concen: 15.744 ng
 RT: 21.298 min Scan# 3113
 Delta R.T. -0.011 min
 Lab File: BP008970.D
 Acq: 31 Jan 2022 18:03

Tgt Ion:228 Resp: 1230276
 Ion Ratio Lower Upper
 228 100
 226 28.5 23.3 34.9
 229 21.7 17.0 25.6









#83
Chrysene
Concen: 16.114 ng
RT: 21.351 min Scan# 3101
Delta R.T. -0.011 min
Lab File: BP008970.D
Acq: 31 Jan 2022 18:03




Instrument :
BNA_P
ClientSampleId :
CC-7


Tgt Ion:228 Resp: 1220686
Ion Ratio Lower Upper
228 100
226 30.1 26.0 39.0
229 21.4 16.9 25.3



#84
Bis(2-ethylhexyl)phthalate
Concen: 3.544 ng
RT: 21.227 min Scan# 3101
Delta R.T. -0.012 min
Lab File: BP008970.D
Acq: 31 Jan 2022 18:03

Tgt Ion:149 Resp: 178904
Ion Ratio Lower Upper
149 100
167 29.9 25.0 37.4
279 5.3 4.8 7.2









#86
 Perylene-d12
 Concen: 20.000 ng
 RT: 23.721 min Scan# 31
 Delta R.T. -0.017 min
 Lab File: BP008970.D
 Acq: 31 Jan 2022 18:03




Instrument :
 BNA_P
 ClientSampleId :
 CC-7


Tgt Ion:264 Resp: 1302500
 Ion Ratio Lower Upper
 264 100
 260 25.0 19.1 28.7
 265 22.9 17.1 25.7



#87
 Indeno(1,2,3-cd)pyrene
 Concen: 7.118 ng
 RT: 26.233 min Scan# 3952
 Delta R.T. -0.029 min
 Lab File: BP008970.D
 Acq: 31 Jan 2022 18:03

Tgt Ion:276 Resp: 763970
 Ion Ratio Lower Upper
 276 100
 138 18.7 17.8 26.6
 277 24.8 20.4 30.6









#88
Benzo(b)fluoranthene
Concen: 23.823 ng
RT: 22.992 min Scan# 3401
Delta R.T. -0.011 min
Lab File: BP008970.D
Acq: 31 Jan 2022 18:03




Instrument :
BNA_P
ClientSampleId :
CC-7


Tgt Ion:252 Resp: 2072240
Ion Ratio Lower Upper
252 100
253 23.0 18.0 27.0
125 8.9 6.9 10.3



#89
Benzo(k)fluoranthene
Concen: 24.611 ng
RT: 22.992 min Scan# 3401
Delta R.T. -0.059 min
Lab File: BP008970.D
Acq: 31 Jan 2022 18:03

Tgt Ion:252 Resp: 2072240
Ion Ratio Lower Upper
252 100
253 23.0 18.2 27.2
125 8.9 7.3 10.9









#90
Benzo(a)pyrene
Concen: 12.540 ng
RT: 23.616 min Scan# 31
Delta R.T. -0.017 min
Lab File: BP008970.D
Acq: 31 Jan 2022 18:03




Instrument :
BNA_P
ClientSampleId :
CC-7


Tgt Ion:252 Resp: 1052871
Ion Ratio Lower Upper
252 100
253 23.4 17.8 26.8
125 9.2 7.8 11.6



#91
Dibenzo(a,h)anthracene
Concen: 2.408 ng
RT: 26.239 min Scan# 3953
Delta R.T. -0.035 min
Lab File: BP008970.D
Acq: 31 Jan 2022 18:03

Tgt Ion:278 Resp: 219774
Ion Ratio Lower Upper
278 100
139 18.7 11.5 17.3#
279 28.9 19.1 28.7#





#92

Benzo(g,h,i)perylene

Concen: 8.513 ng

RT: 27.010 min Scan# 40

Delta R.T. -0.023 min

Lab File: BP008970.D


Acq: 31 Jan 2022 18:03

Instrument :

BNA_P

ClientSampleId :

CC-7



Tgt Ion:276 Resp: 758545

Ion Ratio Lower Upper

276 100

277 24.3 19.2 28.8

138 18.8 15.0 22.4

