

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP020222\
 Data File : BP009000.D
 Acq On : 02 Feb 2022 10:47
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 LabSampleId :
 SSTDCCC040

Quant Time: Feb 03 06:22:02 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP011422.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Feb 03 06:20:22 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	89	0.00
2	1,4-Dioxane	40.000	32.223	19.4	85	0.00
3	Pyridine	40.000	38.965	2.6	97	0.00
4	n-Nitrosodimethylamine	40.000	38.157	4.6	95	0.00
5 S	2-Fluorophenol	80.000	77.831	2.7	98	0.00
6	Aniline	40.000	41.801	-4.5	103	0.00
7 S	Phenol-d6	80.000	82.373	-3.0	101	0.00
8	2-Chlorophenol	40.000	41.152	-2.9	102	0.00
9	Benzaldehyde	40.000	38.813	3.0	97	0.00
10 C	Phenol	40.000	41.900	-4.7	103	0.00
11	bis(2-Chloroethyl)ether	40.000	41.229	-3.1	102	0.00
12	1,3-Dichlorobenzene	40.000	38.825	2.9	97	0.00
13 C	1,4-Dichlorobenzene	40.000	39.310	1.7	99	0.00
14	1,2-Dichlorobenzene	40.000	39.627	0.9	100	0.00
15	Benzyl Alcohol	40.000	44.452	-11.1	108	0.00
16	2,2'-oxybis(1-Chloropropane	40.000	42.313	-5.8	104	0.00
17	2-Methylphenol	40.000	42.801	-7.0	104	0.00
18	Hexachloroethane	40.000	40.366	-0.9	100	0.00
19 P	n-Nitroso-di-n-propylamine	40.000	44.616	-11.5	107	0.00
20	3+4-Methylphenols	40.000	44.224	-10.6	107	0.00
21 I	Naphthalene-d8	20.000	20.000	0.0	93	0.00
22	Acetophenone	40.000	40.329	-0.8	104	0.00
23 S	Nitrobenzene-d5	80.000	82.268	-2.8	106	0.00
24	Nitrobenzene	40.000	41.081	-2.7	106	0.00
25	Isophorone	40.000	43.898	-9.7	112	0.00
26 C	2-Nitrophenol	40.000	43.001	-7.5	109	0.00
27	2,4-Dimethylphenol	40.000	41.606	-4.0	106	0.00
28	bis(2-Chloroethoxy)methane	40.000	41.621	-4.1	107	0.00
29 C	2,4-Dichlorophenol	40.000	42.000	-5.0	108	0.00
30	1,2,4-Trichlorobenzene	40.000	39.250	1.9	104	0.00
31	Naphthalene	40.000	39.843	0.4	105	0.00
32	Benzoic acid	40.000	47.648	-19.1	114	0.00
33	4-Chloroaniline	40.000	42.779	-6.9	110	0.00
34 C	Hexachlorobutadiene	40.000	39.117	2.2	103	0.00
35	Caprolactam	40.000	49.307	-23.3	125	0.00
36 C	4-Chloro-3-methylphenol	40.000	45.102	-12.8	115	0.00
37	2-Methylnaphthalene	40.000	41.373	-3.4	108	0.00
38	1-Methylnaphthalene	40.000	42.047	-5.1	108	0.00
39 I	Acenaphthene-d10	20.000	20.000	0.0	101	0.00
40	1,2,4,5-Tetrachlorobenzene	40.000	38.103	4.7	109	0.00
41 P	Hexachlorocyclopentadiene	40.000	36.557	8.6	96	0.00
42 S	2,4,6-Tribromophenol	80.000	87.914	-9.9	124	0.00
43 C	2,4,6-Trichlorophenol	40.000	40.734	-1.8	113	0.00
44	2,4,5-Trichlorophenol	40.000	40.901	-2.3	114	0.00
45 S	2-Fluorobiphenyl	80.000	74.954	6.3	108	0.00
46	1,1'-Biphenyl	40.000	38.441	3.9	110	0.00
47	2-Chloronaphthalene	40.000	38.895	2.8	111	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP020222\
 Data File : BP009000.D
 Acq On : 02 Feb 2022 10:47
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 LabSampleId :
 SSTDCCC040

Quant Time: Feb 03 06:22:02 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP011422.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Feb 03 06:20:22 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48	2-Nitroaniline	40.000	45.248	-13.1	125	0.00
49	Acenaphthylene	40.000	40.633	-1.6	115	0.00
50	Dimethylphthalate	40.000	41.735	-4.3	120	0.00
51	2,6-Dinitrotoluene	40.000	43.910	-9.8	122	0.00
52 C	Acenaphthene	40.000	40.622	-1.6	116	0.00
53	3-Nitroaniline	40.000	45.809	-14.5	128	0.00
54 P	2,4-Dinitrophenol	40.000	47.933	-19.8	129	0.00
55	Dibenzofuran	40.000	40.428	-1.1	116	0.00
56 P	4-Nitrophenol	40.000	44.762	-11.9	122	0.00
57	2,4-Dinitrotoluene	40.000	46.749	-16.9	128	0.00
58	Fluorene	40.000	40.355	-0.9	114	0.00
59	2,3,4,6-Tetrachlorophenol	40.000	43.370	-8.4	121	0.00
60	Diethylphthalate	40.000	42.941	-7.4	122	0.00
61	4-Chlorophenyl-phenylether	40.000	40.448	-1.1	114	0.00
62	4-Nitroaniline	40.000	47.092	-17.7	131	0.00
63	Azobenzene	40.000	43.259	-8.1	120	0.00
64 I	Phenanthrene-d10	20.000	20.000	0.0	114	0.00
65	4,6-Dinitro-2-methylphenol	40.000	45.945	-14.9	137	0.00
66 c	n-Nitrosodiphenylamine	40.000	38.315	4.2	121	0.00
67	4-Bromophenyl-phenylether	40.000	38.576	3.6	120	0.00
68	Hexachlorobenzene	40.000	39.163	2.1	124	0.00
69	Atrazine	40.000	42.134	-5.3	133	0.00
70 C	Pentachlorophenol	40.000	42.197	-5.5	131	0.00
71	Phenanthrene	40.000	39.395	1.5	126	0.00
72	Anthracene	40.000	40.130	-0.3	127	0.00
73	Carbazole	40.000	41.559	-3.9	134	0.00
74	Di-n-butylphthalate	40.000	42.262	-5.7	132	0.00
75 C	Fluoranthene	40.000	41.394	-3.5	135	0.00
76 I	Chrysene-d12	20.000	20.000	0.0	123	0.00
77	Benzydine	40.000	42.217	-5.5	129	0.00
78	Pyrene	40.000	40.746	-1.9	134	0.00
79 S	Terphenyl-d14	80.000	78.673	1.7	130	0.00
80	Butylbenzylphthalate	40.000	43.120	-7.8	143	0.00
81	Benzo(a)anthracene	40.000	40.612	-1.5	140	0.00
82	3,3'-Dichlorobenzidine	40.000	38.350	4.1	130	0.00
83	Chrysene	40.000	39.738	0.7	136	0.00
84	Bis(2-ethylhexyl)phthalate	40.000	40.930	-2.3	135	0.00
85 c	Di-n-octyl phthalate	40.000	40.880	-2.2	149	0.00
86 I	Perylene-d12	20.000	20.000	0.0	108	0.00
87	Indeno(1,2,3-cd)pyrene	40.000	35.446	11.4	107	0.00
88	Benzo(b)fluoranthene	40.000	44.468	-11.2	136	0.00
89	Benzo(k)fluoranthene	40.000	44.079	-10.2	138	0.00
90 C	Benzo(a)pyrene	40.000	42.827	-7.1	131	0.00
91	Dibenzo(a,h)anthracene	40.000	35.009	12.5	106	0.00
92	Benzo(g,h,i)perylene	40.000	34.446	13.9	103	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP020222\
Data File : BP009000.D
Acq On : 02 Feb 2022 10:47
Operator : CG/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_P
LabSampleId :
SSTDCCC040

Quant Time: Feb 03 06:22:02 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP011422.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Feb 03 06:20:22 2022
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range

SPCC's out = 0 CCC's out = 0