

DDT Breakdown

Date	Instrument Name	DFTPP Data File
2/2/2024	BNA_P	<u>BP019517.D</u>
Compound Name	Response	Retention Time
DDT	529862	20.704
DDD	11722	20.274
DDE	570	19.774
SUM(DDD+DDE)	SUM(DDT+DDD+DDE)	% Breakdown Of DDT
12292	542154	2.27