

DDT Breakdown

Date	Instrument Name	DFTPP Data File
2/2/2024	BNA_P	<u>BP019534.D</u>
Compound Name	Response	Retention Time
DDT	731956	20.698
DDD	15947	20.269
DDE	819	19.769
SUM(DDD+DDE)	SUM(DDT+DDD+DDE)	% Breakdown Of DDT
16766	748722	2.24