




Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP020224\
Data File : BP019534.D
Acq On : 02 Feb 2024 19:02
Operator : MA/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Quant Time: Feb 03 03:51:08 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP013124.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Feb 01 00:51:48 2024
Response via : Initial Calibration






TIC: BP019534.D\data.ms

(70) Pentachlorophenol (C)**16.939min (-0.006) 1204269.16 ng****response 273706**

Ion	Exp%	Act%
265.70	100.00	100.00
268.00	62.90	65.45
264.00	62.10	64.99
0.00	0.00	0.00

Quant Time: Feb 03 03:51:08 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP013124.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Feb 01 00:51:48 2024
Response via : Initial Calibration



TIC: BP019534.D\data.ms

(77) Benzidine

19.498min (+ 0.000) 2680457.45 ng

response **1232868**

Ion	Exp%	Act%
184.00	100.00	100.00
185.00	13.90	13.97
183.00	12.10	11.92
0.00	0.00	0.00