

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP020322\
 Data File : BP009024.D
 Acq On : 03 Feb 2022 12:16
 Operator : CG/JU
 Sample : PB142402TB
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 PB142402TB

Quant Time: Feb 03 13:31:19 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP011422.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Feb 03 06:20:22 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.822	152	249601	20.000	ng	0.00
21) Naphthalene-d8	10.616	136	926728	20.000	ng	-0.01
39) Acenaphthene-d10	14.463	164	566696	20.000	ng	0.00
64) Phenanthrene-d10	17.222	188	1077815	20.000	ng	0.00
76) Chrysene-d12	21.316	240	900629	20.000	ng	0.00
86) Perylene-d12	23.721	264	974469	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.411	112	1847625	114.471	ng	0.00
7) Phenol-d6	7.005	99	2213942	113.272	ng	0.00
23) Nitrobenzene-d5	8.981	82	1399968	85.765	ng	0.00
42) 2,4,6-Tribromophenol	15.963	330	928114	112.514	ng	0.00
45) 2-Fluorobiphenyl	13.081	172	3543010	82.448	ng	-0.01
79) Terphenyl-d14	19.781	244	4268544	80.000	ng	0.00

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP020322\
Data File : BP009024.D
Acq On : 03 Feb 2022 12:16
Operator : CG/JU
Sample : PB142402TB
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
PB142402TB

Quant Time: Feb 03 13:31:19 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP011422.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Feb 03 06:20:22 2022
Response via : Initial Calibration

