




Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP020322\
 Data File : BP009020.D
 Acq On : 03 Feb 2022 09:56
 Operator : CG/JU
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Quant Time: Feb 03 13:27:29 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP011422.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Feb 03 06:20:22 2022
 Response via : Initial Calibration



TIC: BP009020.D\data.ms

(70) Pentachlorophenol (C)




16.881min (-0.006) 2170969.71 ng

response 628780

Ion	Exp%	Act%
265.70	100.00	100.00
268.00	63.60	63.48
264.00	62.90	64.05
0.00	0.00	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP020322\
Data File : BP009020.D
Acq On : 03 Feb 2022 09:56
Operator : CG/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Quant Time: Feb 03 13:27:29 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP011422.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Feb 03 06:20:22 2022
Response via : Initial Calibration



TIC: BP009020.D\data.ms

(77) Benzidine

19.416min (+ 0.000) 550520.77 ng

response 3053443

Ion	Exp%	Act%
184.00	100.00	100.00
185.00	15.20	14.10
183.00	11.90	11.56
0.00	0.00	0.00