

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP020325\
 Data File : BP023885.D
 Acq On : 03 Feb 2025 18:35
 Operator : RC/JU
 Sample : PB166441BL
 Misc :
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SBLK441

Quant Time: Feb 04 07:33:59 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP012925.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Mon Feb 03 11:19:36 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.775	152	460240	20.000	ng/ul	0.00
20) Naphthalene-d8	10.552	136	1980260	20.000	ng/ul	0.00
38) Acenaphthene-d10	14.398	164	1301749	20.000	ng/ul	0.01
64) Phenanthrene-d10	17.186	188	2631804	20.000	ng/ul	0.00
79) Chrysene-d12	21.627	240	2428276	20.000	ng/ul	-0.01
88) Perylene-d12	24.998	264	2504453	20.000	ng/ul	-0.01

System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.293	96	70771	6.389	ng/uL	0.00
4) Pyridine-d5	3.705	84	1003627	31.399	ng/ul	0.00
7) Phenol-d5	6.963	99	1364286	33.496	ng/ul	0.00
9) Bis-(2-Chloroethyl)eth...	7.111	67	702353	32.578	ng/ul	0.00
11) 2-Chlorophenol-d4	7.322	132	1075302	33.620	ng/ul	0.00
15) 4-Methylphenol-d8	8.481	113	1079934	32.458	ng/ul	0.00
21) Nitrobenzene-d5	8.922	128	487729	30.784	ng/ul	0.00
24) 2-Nitrophenol-d4	9.634	143	502209	29.377	ng/ul	0.00
28) 2,4-Dichlorophenol-d3	10.181	165	977201	31.159	ng/ul	0.01
31) 4-Chloroaniline-d4	10.681	131	1437699	30.654	ng/ul	0.00
46) Dimethylphthalate-d6	13.804	166	2997132	30.867	ng/ul	0.01
49) Acenaphthylene-d8	14.087	160	3455765	31.556	ng/ul	0.00
54) 4-Nitrophenol-d4	14.610	143	490711	25.448	ng/ul	0.00
60) Fluorene-d10	15.404	176	2627007	32.128	ng/ul	0.01
65) 4,6-Dinitro-2-methylph...	15.516	200	319972	19.707	ng/ul	0.00
73) Anthracene-d10	17.292	188	4258705	32.957	ng/ul	0.00
81) Pyrene-d10	19.651	212	4624199	35.540	ng/ul	-0.01
92) Benzo(a)pyrene-d12	24.768	264	4382003	33.535	ng/ul	0.00

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP020325\
Data File : BP023885.D
Acq On : 03 Feb 2025 18:35
Operator : RC/JU
Sample : PB166441BL
Misc :
ALS Vial : 14 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
SBLK441

Quant Time: Feb 04 07:33:59 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP012925.MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Mon Feb 03 11:19:36 2025
Response via : Initial Calibration

