

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP020521\
 Data File : BP004684.D
 Acq On : 05 Feb 2021 17:32
 Operator : CG/JU
 Sample : SSTDICV040
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 ICVBP020521

Quant Time: Feb 05 18:07:47 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_P\METHODS\8270E-BP020521.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Feb 05 18:04:48 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	100	0.00
2	1,4-Dioxane	0.554	0.523	5.6	100	0.00
3	Pyridine	1.404	1.344	4.3	98	0.00
4	n-Nitrosodimethylamine	0.584	0.559	4.3	101	0.00
5 S	2-Fluorophenol	1.295	1.213	6.3	103	0.00
6	Aniline	1.987	1.908	4.0	104	0.00
7 S	Phenol-d6	1.781	1.720	3.4	104	0.00
8	2-Chlorophenol	1.353	1.260	6.9	104	0.00
9	Benzaldehyde	0.748	0.824	-10.2	106	0.00
10 C	Phenol	1.724	1.649	4.4	103	0.00
11	bis(2-Chloroethyl)ether	1.319	1.224	7.2	101	0.00
12	1,3-Dichlorobenzene	1.647	1.518	7.8	103	0.00
13 C	1,4-Dichlorobenzene	1.673	1.558	6.9	102	0.00
14	1,2-Dichlorobenzene	1.604	1.490	7.1	103	0.00
15	Benzyl Alcohol	1.338	1.311	2.0	102	0.00
16	2,2'-oxybis(1-Chloropropane	1.067	1.004	5.9	103	0.00
17	2-Methylphenol	1.165	1.117	4.1	103	0.00
18	Hexachloroethane	0.627	0.591	5.7	100	0.00
19 P	n-Nitroso-di-n-propylamine	1.076	1.040	3.3	101	0.00
20	3+4-Methylphenols	1.546	1.526	1.3	105	0.00
21 I	Naphthalene-d8	1.000	1.000	0.0	103	0.00
22	Acetophenone	0.581	0.541	6.9	102	0.00
23 S	Nitrobenzene-d5	0.459	0.435	5.2	103	0.00
24	Nitrobenzene	0.454	0.422	7.0	102	0.00
25	Isophorone	0.734	0.697	5.0	102	0.00
26 C	2-Nitrophenol	0.180	0.178	1.1	106	0.00
27	2,4-Dimethylphenol	0.280	0.257	8.2	101	0.00
28	bis(2-Chloroethoxy)methane	0.473	0.435	8.0	101	0.00
29 C	2,4-Dichlorophenol	0.338	0.321	5.0	103	0.00
30	1,2,4-Trichlorobenzene	0.417	0.381	8.6	102	0.00
31	Naphthalene	1.167	1.069	8.4	103	0.00
32	Benzoic acid	0.210	0.216	-2.9	107	0.00
33	4-Chloroaniline	0.479	0.454	5.2	103	0.00
34 C	Hexachlorobutadiene	0.280	0.260	7.1	103	0.00
35	Caprolactam	0.107	0.107	0.0	100	0.00
36 C	4-Chloro-3-methylphenol	0.401	0.377	6.0	102	0.00
37	2-Methylnaphthalene	0.830	0.764	8.0	102	0.00
38	1-Methylnaphthalene	0.786	0.728	7.4	103	0.00
39 I	Acenaphthene-d10	1.000	1.000	0.0	103	0.00
40	1,2,4,5-Tetrachlorobenzene	0.719	0.652	9.3	100	0.00
41 P	Hexachlorocyclopentadiene	0.409	0.375	8.3	101	0.00
42 S	2,4,6-Tribromophenol	0.266	0.248	6.8	102	0.00
43 C	2,4,6-Trichlorophenol	0.453	0.429	5.3	103	0.00
44	2,4,5-Trichlorophenol	0.476	0.452	5.0	105	0.00
45 S	2-Fluorobiphenyl	1.636	1.497	8.5	102	0.00
46	1,1'-Biphenyl	1.674	1.526	8.8	102	0.00
47	2-Chloronaphthalene	1.372	1.261	8.1	103	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP020521\
 Data File : BP004684.D
 Acq On : 05 Feb 2021 17:32
 Operator : CG/JU
 Sample : SSTDICV040
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 ICVBP020521

Quant Time: Feb 05 18:07:47 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_P\METHODS\8270E-BP020521.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Feb 05 18:04:48 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
48	2-Nitroaniline	0.405	0.402	0.7	103	0.00
49	Acenaphthylene	2.027	1.882	7.2	102	0.00
50	Dimethylphthalate	1.759	1.618	8.0	102	0.00
51	2,6-Dinitrotoluene	0.355	0.334	5.9	101	0.00
52 C	Acenaphthene	1.285	1.161	9.6	100	0.00
53	3-Nitroaniline	0.369	0.356	3.5	102	0.00
54 P	2,4-Dinitrophenol	0.204	0.202	1.0	102	0.00
55	Dibenzofuran	2.021	1.832	9.4	101	0.00
56 P	4-Nitrophenol	0.289	0.275	4.8	101	0.00
57	2,4-Dinitrotoluene	0.486	0.469	3.5	103	0.00
58	Fluorene	1.661	1.521	8.4	101	0.00
59	2,3,4,6-Tetrachlorophenol	0.453	0.419	7.5	101	0.00
60	Diethylphthalate	1.763	1.644	6.7	103	0.00
61	4-Chlorophenyl-phenylether	0.922	0.828	10.2	102	0.00
62	4-Nitroaniline	0.403	0.392	2.7	104	0.00
63	Azobenzene	1.656	1.535	7.3	101	0.00
64 I	Phenanthrene-d10	1.000	1.000	0.0	100	0.00
65	4,6-Dinitro-2-methylphenol	0.122	0.124	-1.6	106	0.00
66 c	n-Nitrosodiphenylamine	0.650	0.615	5.4	102	0.00
67	4-Bromophenyl-phenylether	0.258	0.235	8.9	100	0.00
68	Hexachlorobenzene	0.281	0.260	7.5	100	0.00
69	Atrazine	0.213	0.209	1.9	103	0.00
70 C	Pentachlorophenol	0.157	0.154	1.9	102	0.00
71	Phenanthrene	1.225	1.110	9.4	100	0.00
72	Anthracene	1.177	1.092	7.2	101	0.00
73	Carbazole	1.080	1.002	7.2	101	0.00
74	Di-n-butylphthalate	1.326	1.276	3.8	104	0.00
75 C	Fluoranthene	1.459	1.339	8.2	100	0.00
76 I	Chrysene-d12	1.000	1.000	0.0	100	0.00
77	Benzydine	0.340	0.404	-18.8	100	0.00
78	Pyrene	1.410	1.335	5.3	102	0.00
79 S	Terphenyl-d14	1.191	1.114	6.5	100	0.00
80	Butylbenzylphthalate	0.554	0.545	1.6	101	0.00
81	Benzo(a)anthracene	1.418	1.306	7.9	99	0.00
82	3,3'-Dichlorobenzidine	0.450	0.438	2.7	101	0.00
83	Chrysene	1.374	1.264	8.0	100	0.00
84	Bis(2-ethylhexyl)phthalate	0.837	0.806	3.7	101	0.00
85 c	Di-n-octyl phthalate	1.357	1.319	2.8	103	0.00
86 I	Perylene-d12	1.000	1.000	0.0	101	0.00
87	Indeno(1,2,3-cd)pyrene	1.597	1.484	7.1	101	0.00
88	Benzo(b)fluoranthene	1.465	1.358	7.3	100	0.00
89	Benzo(k)fluoranthene	1.406	1.298	7.7	102	0.00
90 C	Benzo(a)pyrene	1.324	1.218	8.0	99	0.00
91	Dibenzo(a,h)anthracene	1.332	1.238	7.1	102	0.00
92	Benzo(g,h,i)perylene	1.278	1.181	7.6	100	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP020521\
Data File : BP004684.D
Acq On : 05 Feb 2021 17:32
Operator : CG/JU
Sample : SSTDICV040
Misc :
ALS Vial : 9 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
ICVBP020521

Quant Time: Feb 05 18:07:47 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_P\METHODS\8270E-BP020521.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Feb 05 18:04:48 2021
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)
----------	-------	------	-----------	------------

(#) = Out of Range SPCC's out = 0 CCC's out = 0