

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP020521\
 Data File : BP004684.D
 Acq On : 05 Feb 2021 17:32
 Operator : CG/JU
 Sample : SSTDICV040
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 ICVBP020521

Quant Time: Feb 05 18:07:47 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_P\METHODS\8270E-BP020521.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Feb 05 18:04:48 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	100	0.00
2	1,4-Dioxane	40.000	37.768	5.6	100	0.00
3	Pyridine	40.000	38.303	4.2	98	0.00
4	n-Nitrosodimethylamine	40.000	38.285	4.3	101	0.00
5 S	2-Fluorophenol	80.000	74.930	6.3	103	0.00
6	Aniline	40.000	38.414	4.0	104	0.00
7 S	Phenol-d6	80.000	77.247	3.4	104	0.00
8	2-Chlorophenol	40.000	37.255	6.9	104	0.00
9	Benzaldehyde	40.000	44.031	-10.1	106	0.00
10 C	Phenol	40.000	38.250	4.4	103	0.00
11	bis(2-Chloroethyl)ether	40.000	37.134	7.2	101	0.00
12	1,3-Dichlorobenzene	40.000	36.878	7.8	103	0.00
13 C	1,4-Dichlorobenzene	40.000	37.256	6.9	102	0.00
14	1,2-Dichlorobenzene	40.000	37.159	7.1	103	0.00
15	Benzyl Alcohol	40.000	39.198	2.0	102	0.00
16	2,2'-oxybis(1-Chloropropane	40.000	37.650	5.9	103	0.00
17	2-Methylphenol	40.000	38.343	4.1	103	0.00
18	Hexachloroethane	40.000	37.680	5.8	100	0.00
19 P	n-Nitroso-di-n-propylamine	40.000	38.663	3.3	101	0.00
20	3+4-Methylphenols	40.000	39.467	1.3	105	0.00
21 I	Naphthalene-d8	20.000	20.000	0.0	103	0.00
22	Acetophenone	40.000	37.216	7.0	102	0.00
23 S	Nitrobenzene-d5	80.000	75.865	5.2	103	0.00
24	Nitrobenzene	40.000	37.161	7.1	102	0.00
25	Isophorone	40.000	38.000	5.0	102	0.00
26 C	2-Nitrophenol	40.000	39.705	0.7	106	0.00
27	2,4-Dimethylphenol	40.000	36.836	7.9	101	0.00
28	bis(2-Chloroethoxy)methane	40.000	36.785	8.0	101	0.00
29 C	2,4-Dichlorophenol	40.000	38.050	4.9	103	0.00
30	1,2,4-Trichlorobenzene	40.000	36.518	8.7	102	0.00
31	Naphthalene	40.000	36.625	8.4	103	0.00
32	Benzoic acid	40.000	38.178	4.6	107	0.00
33	4-Chloroaniline	40.000	37.915	5.2	103	0.00
34 C	Hexachlorobutadiene	40.000	37.097	7.3	103	0.00
35	Caprolactam	40.000	40.033	-0.1	100	0.00
36 C	4-Chloro-3-methylphenol	40.000	37.650	5.9	102	0.00
37	2-Methylnaphthalene	40.000	36.821	7.9	102	0.00
38	1-Methylnaphthalene	40.000	37.049	7.4	103	0.00
39 I	Acenaphthene-d10	20.000	20.000	0.0	103	0.00
40	1,2,4,5-Tetrachlorobenzene	40.000	36.296	9.3	100	0.00
41 P	Hexachlorocyclopentadiene	40.000	36.654	8.4	101	0.00
42 S	2,4,6-Tribromophenol	80.000	74.777	6.5	102	0.00
43 C	2,4,6-Trichlorophenol	40.000	37.911	5.2	103	0.00
44	2,4,5-Trichlorophenol	40.000	37.968	5.1	105	0.00
45 S	2-Fluorobiphenyl	80.000	73.185	8.5	102	0.00
46	1,1'-Biphenyl	40.000	36.447	8.9	102	0.00
47	2-Chloronaphthalene	40.000	36.764	8.1	103	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP020521\
 Data File : BP004684.D
 Acq On : 05 Feb 2021 17:32
 Operator : CG/JU
 Sample : SSTDICV040
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 ICVBP020521

Quant Time: Feb 05 18:07:47 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_P\METHODS\8270E-BP020521.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Feb 05 18:04:48 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48	2-Nitroaniline	40.000	39.673	0.8	103	0.00
49	Acenaphthylene	40.000	37.134	7.2	102	0.00
50	Dimethylphthalate	40.000	36.785	8.0	102	0.00
51	2,6-Dinitrotoluene	40.000	37.713	5.7	101	0.00
52 C	Acenaphthene	40.000	36.118	9.7	100	0.00
53	3-Nitroaniline	40.000	38.620	3.5	102	0.00
54 P	2,4-Dinitrophenol	40.000	39.664	0.8	102	0.00
55	Dibenzofuran	40.000	36.245	9.4	101	0.00
56 P	4-Nitrophenol	40.000	38.097	4.8	101	0.00
57	2,4-Dinitrotoluene	40.000	38.604	3.5	103	0.00
58	Fluorene	40.000	36.644	8.4	101	0.00
59	2,3,4,6-Tetrachlorophenol	40.000	37.057	7.4	101	0.00
60	Diethylphthalate	40.000	37.306	6.7	103	0.00
61	4-Chlorophenyl-phenylether	40.000	35.906	10.2	102	0.00
62	4-Nitroaniline	40.000	38.907	2.7	104	0.00
63	Azobenzene	40.000	37.079	7.3	101	0.00
64 I	Phenanthrene-d10	20.000	20.000	0.0	100	0.00
65	4,6-Dinitro-2-methylphenol	40.000	40.698	-1.7	106	0.00
66 c	n-Nitrosodiphenylamine	40.000	37.821	5.4	102	0.00
67	4-Bromophenyl-phenylether	40.000	36.567	8.6	100	0.00
68	Hexachlorobenzene	40.000	37.055	7.4	100	0.00
69	Atrazine	40.000	39.173	2.1	103	0.00
70 C	Pentachlorophenol	40.000	39.154	2.1	102	0.00
71	Phenanthrene	40.000	36.272	9.3	100	0.00
72	Anthracene	40.000	37.111	7.2	101	0.00
73	Carbazole	40.000	37.102	7.2	101	0.00
74	Di-n-butylphthalate	40.000	38.493	3.8	104	0.00
75 C	Fluoranthene	40.000	36.701	8.2	100	0.00
76 I	Chrysene-d12	20.000	20.000	0.0	100	0.00
77	Benzydine	40.000	42.049	-5.1	100	0.00
78	Pyrene	40.000	37.873	5.3	102	0.00
79 S	Terphenyl-d14	80.000	74.868	6.4	100	0.00
80	Butylbenzylphthalate	40.000	39.303	1.7	101	0.00
81	Benzo(a)anthracene	40.000	36.837	7.9	99	0.00
82	3,3'-Dichlorobenzidine	40.000	38.967	2.6	101	0.00
83	Chrysene	40.000	36.801	8.0	100	0.00
84	Bis(2-ethylhexyl)phthalate	40.000	38.526	3.7	101	0.00
85 c	Di-n-octyl phthalate	40.000	38.868	2.8	103	0.00
86 I	Perylene-d12	20.000	20.000	0.0	101	0.00
87	Indeno(1,2,3-cd)pyrene	40.000	37.183	7.0	101	0.00
88	Benzo(b)fluoranthene	40.000	37.071	7.3	100	0.00
89	Benzo(k)fluoranthene	40.000	36.915	7.7	102	0.00
90 C	Benzo(a)pyrene	40.000	36.816	8.0	99	0.00
91	Dibenzo(a,h)anthracene	40.000	37.178	7.1	102	0.00
92	Benzo(g,h,i)perylene	40.000	36.954	7.6	100	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP020521\
Data File : BP004684.D
Acq On : 05 Feb 2021 17:32
Operator : CG/JU
Sample : SSTDICV040
Misc :
ALS Vial : 9 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
ICVBP020521

Quant Time: Feb 05 18:07:47 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_P\METHODS\8270E-BP020521.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Feb 05 18:04:48 2021
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range SPCC's out = 0 CCC's out = 0