

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP020525\
 Data File : BP023990.D
 Acq On : 07 Feb 2025 04:51
 Operator : RC/JU
 Sample : Q1229-14
 Misc :
 ALS Vial : 65 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 E29C0

Quant Time: Feb 07 07:35:48 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP012925.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Mon Feb 03 11:19:36 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.763	152	483673	20.000	ng/ul	-0.01
20) Naphthalene-d8	10.540	136	2114971	20.000	ng/ul	0.00
38) Acenaphthene-d10	14.387	164	1323732	20.000	ng/ul	0.00
64) Phenanthrene-d10	17.192	188	2610382	20.000	ng/ul	0.00
79) Chrysene-d12	21.633	240	2397173	20.000	ng/ul	0.00
88) Perylene-d12	25.010	264	2527088	20.000	ng/ul	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.287	96	60013	5.155	ng/uL	0.00
4) Pyridine-d5	3.705	84	399496	11.893	ng/ul	0.00
7) Phenol-d5	6.952	99	465091	10.866	ng/ul	0.00
9) Bis-(2-Chloroethyl)eth...	7.105	67	853667	37.679	ng/ul	0.00
11) 2-Chlorophenol-d4	7.311	132	1138894	33.883	ng/ul	0.00
15) 4-Methylphenol-d8	8.469	113	911531	26.069	ng/ul	0.00
21) Nitrobenzene-d5	8.910	128	640591	37.858	ng/ul	0.00
24) 2-Nitrophenol-d4	9.622	143	691987	37.900	ng/ul	0.00
28) 2,4-Dichlorophenol-d3	10.169	165	1224117	36.546	ng/ul	0.00
31) 4-Chloroaniline-d4	10.675	131	1546072	30.865	ng/ul	0.00
46) Dimethylphthalate-d6	13.793	166	3973294	40.241	ng/ul	0.00
49) Acenaphthylene-d8	14.075	160	4413788	39.635	ng/ul	0.00
54) 4-Nitrophenol-d4	14.610	143	149392	7.619	ng/ul	0.00
60) Fluorene-d10	15.398	176	3499324	42.085	ng/ul	0.00
65) 4,6-Dinitro-2-methylph...	15.522	200	517800	32.153	ng/ul	0.01
73) Anthracene-d10	17.298	188	5668657	44.228	ng/ul	0.01
81) Pyrene-d10	19.669	212	6473126	50.396	ng/ul	0.00
92) Benzo(a)pyrene-d12	24.763	264	6149384	46.639	ng/ul	-0.01

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP020525\
Data File : BP023990.D
Acq On : 07 Feb 2025 04:51
Operator : RC/JU
Sample : Q1229-14
Misc :
ALS Vial : 65 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
E29C0

Quant Time: Feb 07 07:35:48 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP012925.MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Mon Feb 03 11:19:36 2025
Response via : Initial Calibration

