

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP020525\
 Data File : BP023933.D
 Acq On : 05 Feb 2025 13:02
 Operator : RC/JU
 Sample : Q1226-03DL 5X
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 DCZJ8DL

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 02/06/2025
 Supervised By :mohammad ahmed 02/07/2025

Quant Time: Feb 05 13:26:12 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP012925.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Mon Feb 03 11:19:36 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.758	152	387045	20.000	ng/u1	-0.02
20) Naphthalene-d8	10.528	136	1649490	20.000	ng/u1	-0.02
38) Acenaphthene-d10	14.381	164	1039942	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.192	188	2071577	20.000	ng/u1	0.00
79) Chrysene-d12	21.627	240	2102528	20.000	ng/u1	-0.01
88) Perylene-d12	25.010	264	2082110	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.281	96	14600	1.567	ng/uL	-0.01
4) Pyridine-d5	3.693	84	186093	6.923	ng/u1	-0.01
7) Phenol-d5	6.934	99	253537	7.402	ng/u1	-0.02
9) Bis-(2-Chloroethyl)eth...	7.087	67	153943	8.491	ng/u1	-0.02
11) 2-Chlorophenol-d4	7.299	132	204178	7.591	ng/u1	-0.02
15) 4-Methylphenol-d8	8.458	113	193343	6.910	ng/u1	-0.02
21) Nitrobenzene-d5	8.899	128	99308	7.525	ng/u1	-0.02
24) 2-Nitrophenol-d4	9.616	143	97455	6.844	ng/u1	-0.01
28) 2,4-Dichlorophenol-d3	10.157	165	182916	7.002	ng/u1	-0.01
31) 4-Chloroaniline-d4	10.663	131	275942	7.063	ng/u1	-0.01
46) Dimethylphthalate-d6	13.787	166	625351	8.062	ng/u1	0.00
49) Acenaphthylene-d8	14.075	160	725849	8.297	ng/u1	0.00
54) 4-Nitrophenol-d4	14.604	143	78256	5.080	ng/u1	0.00
60) Fluorene-d10	15.387	176	562234	8.607	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.516	200	49097	3.842	ng/u1	0.00
73) Anthracene-d10	17.292	188	863649	8.491	ng/u1	0.00
81) Pyrene-d10	19.663	212	1001831	8.893	ng/u1	0.00
92) Benzo(a)pyrene-d12	24.763	264	921377	8.481	ng/u1	-0.01
Target Compounds						
					Qvalue	
10) Bis(2-Chloroethyl)ether	7.181	93	886394	33.107	ng/u1	99
14) 2,2'-oxybis(1-Chloropr...	8.281	45	169886	6.150	ng/u1	98
17) N-Nitroso-di-n-propyla...	8.546	70	109409	5.432	ng/u1	95
19) Hexachloroethane	8.828	117	254877	23.261	ng/u1	98
22) Nitrobenzene	8.940	77	305599	10.670	ng/u1	96
23) Isophorone	9.463	82	1769522	30.601	ng/u1	97
27) Bis(2-Chloroethoxy)met...	9.940	93	787222	21.855	ng/u1	100
33) Hexachlorobutadiene	10.869	225	241344	15.049	ng/u1	99
40) Hexachlorocyclopentadiene	12.551	237	507965	29.321	ng/u1	99
48) 2,6-Dinitrotoluene	13.957	165	236382	15.672	ng/u1	94
50) Acenaphthylene	14.098	152	837944	8.994	ng/u1	99
57) 2,4-Dinitrotoluene	14.751	165	729252	32.506	ng/u1	93
59) Diethylphthalate	15.222	149	1549426	20.038	ng/u1	99
61) Fluorene	15.451	166	578636	7.620	ng/u1	100
62) 4-Chlorophenyl-phenyle...	15.451	204	606193	16.127	ng/u1	97
69) Hexachlorobenzene	16.481	284	300773	10.621	ng/u1	96
72) Phenanthrene	17.239	178	886866	7.581	ng/u1	99
78) Di-n-butylphthalate	18.210	149	3137069	24.523	ng/u1	100
82) Pyrene	19.692	202	1231277	9.073	ng/u1	96
83) Butylbenzylphthalate	20.645	149	619211	10.596	ng/u1	95
85) Benzo(a)anthracene	21.610	228	1121326	8.110	ng/u1	99
87) Chrysene	21.674	228	1153289	9.053	ng/u1	99
89) Di-n-octyl phthalate	22.845	149	2375495	19.033	ng/u1	100
90) Benzo(b)fluoranthene	23.933	252	1043726	8.261	ng/u1	100

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP020525\
 Data File : BP023933.D
 Acq On : 05 Feb 2025 13:02
 Operator : RC/JU
 Sample : Q1226-03DL 5X
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 DCZJ8DL

Quant Time: Feb 05 13:26:12 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP012925.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Mon Feb 03 11:19:36 2025
 Response via : Initial Calibration

Manual Integrations APPROVED

Reviewed By :Yogesh Patel 02/06/2025
 Supervised By :mohammad ahmed 02/07/2025

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
93) Benzo(a)pyrene	24.833	252	680288	5.766	ng/ul	99
94) Indeno(1,2,3-cd)pyrene	28.815	276	1223549m	8.001	ng/ul	99
95) Dibenzo(a,h)anthracene	28.904	278	878210	7.079	ng/ul	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

02/07/2025


Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP020525\
 Data File : BP023933.D
 Acq On : 05 Feb 2025 13:02
 Operator : RC/JU
 Sample : Q1226-03DL 5X
 Misc :
 ALS Vial : 6 Sample Multiplier: 1


Instrument :
 BNA_P
 ClientSampleId :
 DCZJ8DL

Quant Time: Feb 05 13:26:12 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP012925.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Mon Feb 03 11:19:36 2025
 Response via : Initial Calibration

Manual Integrations APPROVED


Reviewed By :Yogesh Patel 02/06/2025
 Supervised By :mohammad ahmed 02/07/2025





#10
Bis(2-Chloroethyl)ether
Concen: 33.107 ng/ul
RT: 7.181 min Scan# 699
Delta R.T. -0.018 min
Lab File: BP023933.D
Acq: 05 Feb 2025 13:02



Instrument :
BNA_P
ClientSampleId :
DCZJ8DL



Tgt Ion: 93 Resp: 886394
Ion Ratio Lower Upper
93 100
63 66.5 52.6 79.0
95 32.4 25.8 38.8


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 02/06/2025
Supervised By :mohammad ahmed 02/07/2025






02/06/2025
Supervised By :mohammad
ahmed


02/07/2025



#14
2,2'-oxybis(1-Chloropropane)
Concen: 6.150 ng/ul
RT: 8.281 min Scan# 883
Delta R.T. 0.000 min
Lab File: BP023933.D
Acq: 05 Feb 2025 13:02


Tgt Ion: 45 Resp: 169886
Ion Ratio Lower Upper
45 100
77 17.9 14.7 22.1
79 14.3 10.7 16.1





#17
N-Nitroso-di-n-propylamine
Concen: 5.432 ng/ul
RT: 8.546 min Scan# 911
Delta R.T. -0.024 min
Lab File: BP023933.D
Acq: 05 Feb 2025 13:02



Instrument :
BNA_P
ClientSampleId :
DCZJ8DL



Tgt Ion: 70 Resp: 109409
Ion Ratio Lower Upper
70 100
42 49.7 36.2 54.4
101 9.9 8.6 13.0
130 23.7 20.0 30.0


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 02/06/2025
Supervised By :mohammad ahmed 02/07/2025






02/06/2025
Supervised By :mohammad
ahmed


02/07/2025



#19
Hexachloroethane
Concen: 23.261 ng/ul
RT: 8.828 min Scan# 976
Delta R.T. -0.012 min
Lab File: BP023933.D
Acq: 05 Feb 2025 13:02


Tgt Ion:117 Resp: 254877
Ion Ratio Lower Upper
117 100
201 97.7 79.4 119.2
199 59.8 50.0 75.0





#22
Nitrobenzene
Concen: 10.670 ng/ul
RT: 8.940 min Scan# 998
Delta R.T. -0.018 min
Lab File: BP023933.D
Acq: 05 Feb 2025 13:02



Instrument :
BNA_P
ClientSampleId :
DCZJ8DL



Tgt Ion: 77 Resp: 305599
Ion Ratio Lower Upper
77 100
123 46.9 40.3 60.5
65 12.6 9.8 14.8


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 02/06/2025
Supervised By :mohammad ahmed 02/07/2025






02/06/2025
Supervised By :mohammad
ahmed


02/07/2025



#23
Isophorone
Concen: 30.601 ng/ul
RT: 9.463 min Scan# 1084
Delta R.T. -0.012 min
Lab File: BP023933.D
Acq: 05 Feb 2025 13:02


Tgt Ion: 82 Resp: 1769522
Ion Ratio Lower Upper
82 100
95 6.8 5.5 8.3
138 18.0 15.8 23.6





#27
 Bis(2-Chloroethoxy)methane
 Concen: 21.855 ng/ul
 RT: 9.940 min Scan# 1167
 Delta R.T. -0.012 min
 Lab File: BP023933.D
 Acq: 05 Feb 2025 13:02



Instrument :
 BNA_P
 ClientSampleId :
 DCZJ8DL



Tgt Ion: 93 Resp: 78722
 Ion Ratio Lower Upper
 93 100
 95 32.0 25.8 38.6
 123 11.9 9.7 14.5


Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 02/06/2025
 Supervised By :mohammad ahmed 02/07/2025






02/06/2025
 Supervised By :mohammad
 ahmed


02/07/2025



#33
 Hexachlorobutadiene
 Concen: 15.049 ng/ul
 RT: 10.869 min Scan# 1323
 Delta R.T. -0.018 min
 Lab File: BP023933.D
 Acq: 05 Feb 2025 13:02


Tgt Ion:225 Resp: 241344
 Ion Ratio Lower Upper
 225 100
 223 63.9 49.7 74.5
 227 64.7 51.4 77.0





#40
Hexachlorocyclopentadiene
Concen: 29.321 ng/ul
RT: 12.551 min Scan# 1609
Delta R.T. -0.006 min
Lab File: BP023933.D
Acq: 05 Feb 2025 13:02



Instrument :
BNA_P
ClientSampleId :
DCZJ8DL



Tgt Ion:237 Resp: 50796
Ion Ratio Lower Upper
237 100
235 62.5 50.2 75.4
272 13.3 11.7 17.5


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 02/06/2025
Supervised By :mohammad ahmed 02/07/2025






02/06/2025
Supervised By :mohammad
ahmed


02/07/2025



#48
2,6-Dinitrotoluene
Concen: 15.672 ng/ul
RT: 13.957 min Scan# 1848
Delta R.T. 0.000 min
Lab File: BP023933.D
Acq: 05 Feb 2025 13:02


Tgt Ion:165 Resp: 236382
Ion Ratio Lower Upper
165 100
89 59.2 43.3 64.9
121 22.8 16.8 25.2





#50
Acenaphthylene
Concen: 8.994 ng/ul
RT: 14.098 min Scan# 1872
Delta R.T. -0.012 min
Lab File: BP023933.D
Acq: 05 Feb 2025 13:02



Instrument :
BNA_P
ClientSampleId :
DCZJ8DL



Tgt Ion:152 Resp: 837944
Ion Ratio Lower Upper
152 100
151 19.6 16.1 24.1
153 13.0 10.4 15.6


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 02/06/2025
Supervised By :mohammad ahmed 02/07/2025




02/06/2025
Supervised By :mohammad
ahmed



02/07/2025




#57
2,4-Dinitrotoluene
Concen: 32.506 ng/ul
RT: 14.751 min Scan# 1983
Delta R.T. 0.000 min
Lab File: BP023933.D
Acq: 05 Feb 2025 13:02




Tgt Ion:165 Resp: 729252
Ion Ratio Lower Upper
165 100
63 39.1 27.8 41.6
182 2.5 2.2 3.4





#59
Diethylphthalate
Concen: 20.038 ng/ul
RT: 15.222 min Scan# 2063
Delta R.T. 0.000 min
Lab File: BP023933.D
Acq: 05 Feb 2025 13:02



Instrument :
BNA_P
ClientSampleId :
DCZJ8DL



Tgt Ion:149 Resp: 1549420
Ion Ratio Lower Upper
149 100
177 21.3 17.4 26.0
150 12.3 9.9 14.9


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 02/06/2025
Supervised By :mohammad ahmed 02/07/2025






02/06/2025
Supervised By :mohammad
ahmed


02/07/2025



#61
Fluorene
Concen: 7.620 ng/ul
RT: 15.451 min Scan# 2102
Delta R.T. 0.000 min
Lab File: BP023933.D
Acq: 05 Feb 2025 13:02

Tgt Ion:166 Resp: 578636
Ion Ratio Lower Upper
166 100
165 97.5 77.9 116.9
167 13.7 10.6 16.0







#62
4-Chlorophenyl-phenylether
Concen: 16.127 ng/ul
RT: 15.451 min Scan# 2101
Delta R.T. 0.006 min
Lab File: BP023933.D
Acq: 05 Feb 2025 13:02

Instrument :
BNA_P
ClientSampleId :
DCZJ8DL

Tgt Ion:204 Resp: 60619
Ion Ratio Lower Upper
204 100
206 32.8 26.5 39.7
141 58.5 44.5 66.7



Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 02/06/2025
Supervised By :mohammad ahmed 02/07/2025






02/06/2025
Supervised By :mohammad
ahmed


02/07/2025



#69
Hexachlorobenzene
Concen: 10.621 ng/ul
RT: 16.481 min Scan# 2277
Delta R.T. 0.006 min
Lab File: BP023933.D
Acq: 05 Feb 2025 13:02


Tgt Ion:284 Resp: 300773
Ion Ratio Lower Upper
284 100
142 35.9 25.9 38.9
249 29.6 24.0 36.0





#72
Phenanthrene
Concen: 7.581 ng/ul
RT: 17.239 min Scan# 2404
Delta R.T. 0.012 min
Lab File: BP023933.D
Acq: 05 Feb 2025 13:02



Instrument :
BNA_P
ClientSampleId :
DCZJ8DL



Tgt Ion:178 Resp: 886860
Ion Ratio Lower Upper
178 100
179 15.2 12.2 18.4
176 19.1 15.6 23.4


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 02/06/2025
Supervised By :mohammad ahmed 02/07/2025






02/06/2025
Supervised By :mohammad
ahmed


02/07/2025



#78
Di-n-butylphthalate
Concen: 24.523 ng/ul
RT: 18.210 min Scan# 2571
Delta R.T. 0.018 min
Lab File: BP023933.D
Acq: 05 Feb 2025 13:02

Tgt Ion:149 Resp: 3137069
Ion Ratio Lower Upper
149 100
150 9.1 7.4 11.2
104 5.1 4.0 6.0








#82
 Pyrene
 Concen: 9.073 ng/ul
 RT: 19.692 min Scan# 2823
 Delta R.T. 0.000 min
 Lab File: BP023933.D
 Acq: 05 Feb 2025 13:02

Instrument : BNA_P
 ClientSampleId : DCZJ8DL

Tgt Ion: 202 Resp: 123127


Ion	Ratio	Lower	Upper
202	100		
101	14.9	10.4	15.6
100	11.9	8.6	12.8

Manual Integrations
APPROVED
 Reviewed By :Yogesh Patel 02/06/2025
 Supervised By :mohammad ahmed 02/07/2025



02/06/2025
 Supervised By :mohammad ahmed




02/07/2025




#83
 Butylbenzylphthalate
 Concen: 10.596 ng/ul
 RT: 20.645 min Scan# 2985
 Delta R.T. 0.000 min
 Lab File: BP023933.D
 Acq: 05 Feb 2025 13:02

Tgt Ion: 149 Resp: 619211


Ion	Ratio	Lower	Upper
149	100		
91	73.6	54.9	82.3
206	19.7	16.2	24.4





#85
Benzo(a)anthracene
Concen: 8.110 ng/ul
RT: 21.610 min Scan# 3148
Delta R.T. -0.006 min
Lab File: BP023933.D
Acq: 05 Feb 2025 13:02



Instrument :
BNA_P
ClientSampleId :
DCZJ8DL



Tgt Ion:228 Resp: 1121320
Ion Ratio Lower Upper
228 100
229 19.4 15.9 23.9
226 26.9 21.7 32.5


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 02/06/2025
Supervised By :mohammad ahmed 02/07/2025






02/06/2025
Supervised By :mohammad
ahmed


02/07/2025



#87
Chrysene
Concen: 9.053 ng/ul
RT: 21.674 min Scan# 3160
Delta R.T. -0.012 min
Lab File: BP023933.D
Acq: 05 Feb 2025 13:02

Tgt Ion:228 Resp: 1153289
Ion Ratio Lower Upper
228 100
226 29.2 23.8 35.6
229 19.9 15.7 23.5







#89
Di-n-octyl phthalate
Concen: 19.033 ng/ul
RT: 22.845 min Scan# 31
Delta R.T. -0.006 min
Lab File: BP023933.D
Acq: 05 Feb 2025 13:02

Instrument :
BNA_P
ClientSampleId :
DCZJ8DL

Tgt Ion:149 Resp: 237549



Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 02/06/2025
Supervised By :mohammad ahmed 02/07/2025






02/06/2025
Supervised By :mohammad
ahmed


02/07/2025



#90
Benzo(b)fluoranthene
Concen: 8.261 ng/ul
RT: 23.933 min Scan# 3544
Delta R.T. -0.012 min
Lab File: BP023933.D
Acq: 05 Feb 2025 13:02

Tgt Ion:252 Resp: 1043726
Ion Ratio Lower Upper
252 100
253 21.8 17.5 26.3
125 10.6 8.2 12.2





#93

Benzo(a)pyrene

Concen: 5.766 ng/ul

RT: 24.833 min Scan# 30

Delta R.T. -0.024 min

Lab File: BP023933.D

Acq: 05 Feb 2025 13:02

Instrument :

BNA_P

ClientSampleId :

DCZJ8DL



Tgt Ion:	252	Resp:	68028
Ion Ratio	Lower	Upper	
252	100		
253	21.3	17.8	26.6
125	11.8	9.2	13.8

Manual Integrations

APPROVED

Reviewed By :Yogesh Patel 02/06/2025



Supervised By :mohammad ahmed 02/07/2025



02/06/2025

Supervised By :mohammad
ahmed

02/07/2025



#94

Indeno(1,2,3-cd)pyrene

Concen: 8.001 ng/ul m




RT: 28.815 min Scan# 4374


Delta R.T. -0.012 min

Lab File: BP023933.D

Acq: 05 Feb 2025 13:02


Tgt Ion:	276	Resp:	1223549
Ion Ratio	Lower	Upper	
276	100		
138	18.7	15.9	23.9
277	23.7	19.0	28.4





#95
 Dibenzo(a,h)anthracene
 Concen: 7.079 ng/ul
 RT: 28.904 min Scan# 41
 Delta R.T. -0.024 min
 Lab File: BP023933.D
 Acq: 05 Feb 2025 13:02



Instrument :
 BNA_P
 ClientSampleId :
 DCZJ8DL



Tgt Ion:278 Resp: 878210
 Ion Ratio Lower Upper
 278 100
 139 16.4 12.7 19.1
 279 23.7 19.0 28.6

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 02/06/2025
 Supervised By :mohammad ahmed 02/07/2025



02/06/2025
 Supervised By :mohammad
 ahmed

02/07/2025