

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP020525\
 Data File : BP023962.D
 Acq On : 06 Feb 2025 09:15
 Operator : RC/JU
 Sample : Q1229-04
 Misc :
 ALS Vial : 35 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 E2993

Quant Time: Feb 06 12:41:19 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP012925.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Mon Feb 03 11:19:36 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.769	152	465458	20.000	ng/u1	0.00
20) Naphthalene-d8	10.540	136	2034362	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.392	164	1339028	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.192	188	2693771	20.000	ng/u1	0.00
79) Chrysene-d12	21.616	240	2556287	20.000	ng/u1	-0.02
88) Perylene-d12	24.992	264	2780656	20.000	ng/u1	-0.02

System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.293	96	57282	5.113	ng/uL	0.00
4) Pyridine-d5	3.705	84	260800	8.068	ng/u1	0.00
7) Phenol-d5	6.958	99	427111	10.369	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.105	67	657368	30.150	ng/u1	0.00
11) 2-Chlorophenol-d4	7.310	132	957933	29.614	ng/u1	0.00
15) 4-Methylphenol-d8	8.475	113	769645	22.873	ng/u1	0.00
21) Nitrobenzene-d5	8.910	128	510683	31.376	ng/u1	0.00
24) 2-Nitrophenol-d4	9.628	143	570793	32.501	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.169	165	996235	30.921	ng/u1	0.00
31) 4-Chloroaniline-d4	10.681	131	1329018	27.583	ng/u1	0.00
46) Dimethylphthalate-d6	13.798	166	3198401	32.023	ng/u1	0.00
49) Acenaphthylene-d8	14.087	160	3616649	32.106	ng/u1	0.00
54) 4-Nitrophenol-d4	14.622	143	161836	8.159	ng/u1	0.01
60) Fluorene-d10	15.398	176	2815375	33.473	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.522	200	414455	24.939	ng/u1	0.01
73) Anthracene-d10	17.292	188	4654234	35.189	ng/u1	0.00
81) Pyrene-d10	19.651	212	5193153	37.914	ng/u1	-0.01
92) Benzo(a)pyrene-d12	24.757	264	5313089	36.621	ng/u1	-0.02

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP020525\
Data File : BP023962.D
Acq On : 06 Feb 2025 09:15
Operator : RC/JU
Sample : Q1229-04
Misc :
ALS Vial : 35 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
E2993

Quant Time: Feb 06 12:41:19 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP012925.MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Mon Feb 03 11:19:36 2025
Response via : Initial Calibration

