

Data Path : Z:\svoasrv\HPCHEM1\BNA\_P\Data\BP020624\  
 Data File : BP019578.D  
 Acq On : 06 Feb 2024 17:34  
 Operator : MA/JU  
 Sample : PB158822BL  
 Misc :  
 ALS Vial : 16 Sample Multiplier: 1

Instrument :  
 BNA\_P  
 ClientSampleId :  
 PB158822BL

Quant Time: Feb 07 01:06:07 2024  
 Quant Method : Z:\svoasrv\HPCHEM1\BNA\_P\Methods\8270E-BP013124.M  
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
 QLast Update : Thu Feb 01 00:51:48 2024  
 Response via : Initial Calibration

| Compound                    | R.T.   | QIon | Response | Conc    | Units | Dev(Min) |
|-----------------------------|--------|------|----------|---------|-------|----------|
| -----                       |        |      |          |         |       |          |
| Internal Standards          |        |      |          |         |       |          |
| 1) 1,4-Dichlorobenzene-d4   | 7.910  | 152  | 93673    | 20.000  | ng    | 0.00     |
| 21) Naphthalene-d8          | 10.710 | 136  | 328689   | 20.000  | ng    | 0.00     |
| 39) Acenaphthene-d10        | 14.539 | 164  | 191801   | 20.000  | ng    | -0.01    |
| 64) Phenanthrene-d10        | 17.292 | 188  | 351870   | 20.000  | ng    | -0.01    |
| 76) Chrysene-d12            | 21.386 | 240  | 344196   | 20.000  | ng    | 0.00     |
| 86) Perylene-d12            | 23.804 | 264  | 470681   | 20.000  | ng    | -0.01    |
| System Monitoring Compounds |        |      |          |         |       |          |
| 5) 2-Fluorophenol           | 5.481  | 112  | 693486   | 119.335 | ng    | 0.00     |
| 7) Phenol-d6                | 7.081  | 99   | 877890   | 112.037 | ng    | 0.00     |
| 23) Nitrobenzene-d5         | 9.075  | 82   | 601127   | 79.247  | ng    | -0.01    |
| 42) 2,4,6-Tribromophenol    | 16.033 | 330  | 300019   | 107.133 | ng    | 0.00     |
| 45) 2-Fluorobiphenyl        | 13.169 | 172  | 1286617  | 82.927  | ng    | -0.01    |
| 79) Terphenyl-d14           | 19.863 | 244  | 1490459  | 71.213  | ng    | 0.00     |

| Target Compounds | Qvalue |
|------------------|--------|
| -----            |        |

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA\_P\Data\BP020624\  
Data File : BP019578.D  
Acq On : 06 Feb 2024 17:34  
Operator : MA/JU  
Sample : PB158822BL  
Misc :  
ALS Vial : 16 Sample Multiplier: 1

Instrument :  
BNA\_P  
ClientSampleId :  
PB158822BL

Quant Time: Feb 07 01:06:07 2024  
Quant Method : Z:\svoasrv\HPCHEM1\BNA\_P\Methods\8270E-BP013124.M  
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
QLast Update : Thu Feb 01 00:51:48 2024  
Response via : Initial Calibration

