

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP020723\
 Data File : BP013779.D
 Acq On : 07 Feb 2023 15:50
 Operator : CG/JU
 Sample : SSTD00529
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SSTD005610

Quant Time: Feb 08 00:01:36 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP020723.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue Feb 07 23:57:24 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.122	152	143015	20.000	ng/ul	0.00
20) Naphthalene-d8	10.951	136	636600	20.000	ng/ul	0.00
38) Acenaphthene-d10	14.763	164	457136	20.000	ng/ul	0.00
64) Phenanthrene-d10	17.516	188	1108150	20.000	ng/ul	0.00
79) Chrysene-d12	21.586	240	1311306	20.000	ng/ul	0.00
88) Perylene-d12	24.151	264	1372907	20.000	ng/ul	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.470	96	6529	1.936	ng/uL	0.00
4) Pyridine-d5	3.911	84	47143	4.871	ng/ul	0.00
7) Phenol-d5	7.269	99	54121	4.306	ng/ul	0.00
9) Bis-(2-Chloroethyl)eth...	7.440	67	35106	4.762	ng/ul	0.00
11) 2-Chlorophenol-d4	7.652	132	40820	4.430	ng/ul	0.00
15) 4-Methylphenol-d8	8.828	113	42667	4.258	ng/ul	0.00
21) Nitrobenzene-d5	9.293	128	17618	4.338	ng/ul	0.00
24) 2-Nitrophenol-d4	10.028	143	19571	4.135	ng/ul	0.00
28) 2,4-Dichlorophenol-d3	10.563	165	45343	4.097	ng/ul	0.00
31) 4-Chloroaniline-d4	11.087	131	70091	4.795	ng/ul	0.00
46) Dimethylphthalate-d6	14.163	166	170630	4.735	ng/ul	0.00
49) Acenaphthylene-d8	14.457	160	175586	4.410	ng/ul	0.00
54) 4-Nitrophenol-d4	14.940	143	21369	3.518	ng/ul	0.00
60) Fluorene-d10	15.751	176	148550	4.783	ng/ul	0.00
65) 4,6-Dinitro-2-methylph...	15.857	200	20318	3.272	ng/ul	0.00
73) Anthracene-d10	17.610	188	246715	4.776	ng/ul	0.00
81) Pyrene-d10	19.827	212	363753	4.974	ng/ul	0.00
92) Benzo(a)pyrene-d12	23.986	264	310522	4.431	ng/ul	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.511	88	7345	2.089	ng/uL#	86
5) Pyridine	3.928	79	45176	4.636	ng/ul	94
6) Benzaldehyde	7.258	77	24524	4.866	ng/ul	92
8) Phenol	7.293	94	55481	4.313	ng/ul	94
10) Bis(2-Chloroethyl)ether	7.534	93	48396	4.682	ng/ul	97
12) 2-Chlorophenol	7.681	128	43286	4.637	ng/ul	97
13) 2-Methylphenol	8.563	108	43078	4.386	ng/ul	93
14) 2,2'-oxybis(1-Chloropr...	8.652	45	55543	4.538	ng/ul	97
16) Acetophenone	8.957	105	78250	5.009	ng/ul	98
17) N-Nitroso-di-n-propyla...	8.934	70	36858	4.710	ng/ul	95
18) 4-Methylphenol	8.893	108	46745	4.382	ng/ul	97
19) Hexachloroethane	9.222	117	18980	5.194	ng/ul	96
22) Nitrobenzene	9.340	77	49024	4.469	ng/ul	97
23) Isophorone	9.869	82	108364	4.315	ng/ul	99
25) 2-Nitrophenol	10.057	139	21354	4.078	ng/ul	99
26) 2,4-Dimethylphenol	10.110	107	56095	4.741	ng/ul	97
27) Bis(2-Chloroethoxy)met...	10.346	93	67357	4.426	ng/ul	96
29) 2,4-Dichlorophenol	10.593	162	45843	4.280	ng/ul	95
30) Naphthalene	11.004	128	163717	4.866	ng/ul	99
32) 4-Chloroaniline	11.110	127	71160	4.837	ng/ul	94
33) Hexachlorobutadiene	11.287	225	35268	4.669	ng/ul	99
34) Caprolactam	11.887	113	12625	3.486	ng/ul	95
35) 4-Chloro-3-methylphenol	12.216	107	50767	4.425	ng/ul	96
36) 2-Methylnaphthalene	12.598	142	112443	4.695	ng/ul	97

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP020723\
 Data File : BP013779.D
 Acq On : 07 Feb 2023 15:50
 Operator : CG/JU
 Sample : SST00529
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SST005610

Quant Time: Feb 08 00:01:36 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP020723.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue Feb 07 23:57:24 2023
 Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.816	142	117789	4.870	ng/ul	100
39) 1,2,4,5-Tetrachloroben...	12.957	216	67253	4.454	ng/ul	96
40) Hexachlorocyclopentadiene	12.940	237	32138	3.785	ng/ul	93
41) 2,4,6-Trichlorophenol	13.198	196	41020	4.101	ng/ul	92
42) 2,4,5-Trichlorophenol	13.269	196	42996	3.944	ng/ul	99
43) 1,1'-Biphenyl	13.593	154	165913	4.794	ng/ul	99
44) 2-Chloronaphthalene	13.640	162	129452	4.728	ng/ul	99
45) 2-Nitroaniline	13.840	65	23472	4.503	ng/ul	95
47) Dimethylphthalate	14.204	163	171091	4.819	ng/ul	99
48) 2,6-Dinitrotoluene	14.328	165	22192	4.386	ng/ul	93
50) Acenaphthylene	14.487	152	199314	4.748	ng/ul	100
51) 3-Nitroaniline	14.663	138	21542	4.028	ng/ul#	97
52) Acenaphthene	14.822	153	143972	4.990	ng/ul	96
53) 2,4-Dinitrophenol	14.863	184	11901	3.029	ng/ul#	89
55) 4-Nitrophenol	14.957	109	21575	4.660	ng/ul	90
56) Dibenzofuran	15.157	168	212041	5.060	ng/ul	100
57) 2,4-Dinitrotoluene	15.110	165	33860	4.261	ng/ul	97
58) 2,3,4,6-Tetrachlorophenol	15.381	232	42743	4.172	ng/ul#	95
59) Diethylphthalate	15.563	149	170316	4.997	ng/ul	98
61) Fluorene	15.804	166	176207	5.167	ng/ul	96
62) 4-Chlorophenyl-phenyle...	15.792	204	89544	4.819	ng/ul	97
63) 4-Nitroaniline	15.822	138	17550	3.505	ng/ul	88
66) 4,6-Dinitro-2-methylph...	15.875	198	21435	3.468	ng/ul#	91
67) N-Nitrosodiphenylamine	16.010	169	144576	4.834	ng/ul	98
68) 4-Bromophenyl-phenylether	16.692	248	53393	4.235	ng/ul	95
69) Hexachlorobenzene	16.810	284	65223	4.410	ng/ul	92
70) Atrazine	16.957	200	51459	4.275	ng/ul	98
71) Pentachlorophenol	17.157	266	34106	3.613	ng/ul	97
72) Phenanthrene	17.557	178	291554	4.981	ng/ul	99
74) Anthracene	17.645	178	291784	4.922	ng/ul	99
75) 1,2,3,4-Tetrachloroben...	13.563	216	68981	4.369	ng/ul	97
76) Pentachlorobenzene	15.075	250	72451	4.476	ng/ul	99
77) Carbazole	17.910	167	255444	4.966	ng/ul	99
78) Di-n-butylphthalate	18.445	149	297421	4.894	ng/ul	100
80) Fluoranthene	19.504	202	422797	4.960	ng/ul	99
82) Pyrene	19.857	202	459980	5.278	ng/ul	99
83) Butylbenzylphthalate	20.710	149	122849	4.906	ng/ul	100
84) 3,3'-Dichlorobenzidine	21.498	252	119703	4.019	ng/ul	99
85) Benzo(a)anthracene	21.569	228	427117	4.738	ng/ul	98
86) Bis(2-ethylhexyl)phtha...	21.469	149	193939	4.517	ng/ul	99
87) Chrysene	21.627	228	422900	4.983	ng/ul	97
89) Di-n-octyl phthalate	22.445	149	324336	4.485	ng/ul	100
90) Benzo(b)fluoranthene	23.363	252	422056	4.709	ng/ul	99
91) Benzo(k)fluoranthene	23.410	252	421986	4.927	ng/ul	97
93) Benzo(a)pyrene	24.033	252	363802	4.702	ng/ul	98
94) Indeno(1,2,3-cd)pyrene	26.851	276	438937	4.579	ng/ul	98
95) Dibenzo(a,h)anthracene	26.868	278	365542	4.600	ng/ul	99
96) Benzo(g,h,i)perylene	27.680	276	351819	4.639	ng/ul	97


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP020723\
 Data File : BP013779.D
 Acq On : 07 Feb 2023 15:50
 Operator : CG/JU
 Sample : SSTD00529
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SSTD005610

Quant Time: Feb 08 00:01:36 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP020723.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue Feb 07 23:57:24 2023
 Response via : Initial Calibration









#2
1,4-Dioxane
Concen: 2.089 ng/uL
RT: 3.511 min Scan# 51
Delta R.T. 0.006 min
Lab File: BP013779.D
Acq: 07 Feb 2023 15:50




Instrument :
BNA_P
ClientSampleId :
SSTD005610


Tgt Ion: 88 Resp: 7345
Ion Ratio Lower Upper
88 100
43 22.5 20.1 30.1
58 52.4 53.0 79.6#



#5
Pyridine
Concen: 4.636 ng/uL
RT: 3.928 min Scan# 126
Delta R.T. 0.006 min
Lab File: BP013779.D
Acq: 07 Feb 2023 15:50

Tgt Ion: 79 Resp: 45176
Ion Ratio Lower Upper
79 100
52 75.3 55.1 82.7
51 33.5 25.7 38.5









#6
Benzaldehyde
Concen: 4.866 ng/ul
RT: 7.258 min Scan# 691
Delta R.T. 0.006 min
Lab File: BP013779.D
Acq: 07 Feb 2023 15:50




Instrument :
BNA_P
ClientSampleId :
SSTD005610


Tgt Ion: 77 Resp: 24524
Ion Ratio Lower Upper
77 100
105 93.0 79.5 119.3
106 86.6 76.5 114.7



#8
Phenol
Concen: 4.313 ng/ul
RT: 7.293 min Scan# 698
Delta R.T. -0.006 min
Lab File: BP013779.D
Acq: 07 Feb 2023 15:50

Tgt Ion: 94 Resp: 55481
Ion Ratio Lower Upper
94 100
65 30.3 22.2 33.4
66 40.7 29.6 44.4





#10

Bis(2-Chloroethyl)ether

Concen: 4.682 ng/ul

RT: 7.534 min Scan# 739

Delta R.T. -0.000 min

Lab File: BP013779.D

Acq: 07 Feb 2023 15:50

Instrument :

BNA_P

ClientSampleId :

SSTD005610





Tgt Ion: 93 Resp: 48396

Ion Ratio Lower Upper

93 100

63 72.6 59.4 89.0

95 29.0 25.8 38.8



#12

2-Chlorophenol

Concen: 4.637 ng/ul

RT: 7.681 min Scan# 764

Delta R.T. -0.000 min

Lab File: BP013779.D

Acq: 07 Feb 2023 15:50




Tgt Ion:128 Resp: 43286


Ion Ratio Lower Upper

128 100

64 49.3 37.3 55.9

130 32.5 26.3 39.5









#13
2-Methylphenol
Concen: 4.386 ng/ul
RT: 8.563 min Scan# 914
Delta R.T. -0.000 min
Lab File: BP013779.D
Acq: 07 Feb 2023 15:50




Instrument :
BNA_P
ClientSampleId :
SSTD005610


Tgt Ion:108 Resp: 43078
Ion Ratio Lower Upper
108 100
107 83.8 71.9 107.9



#14
2,2'-oxybis(1-Chloropropane)
Concen: 4.538 ng/ul
RT: 8.652 min Scan# 929
Delta R.T. -0.000 min
Lab File: BP013779.D
Acq: 07 Feb 2023 15:50

Tgt Ion: 45 Resp: 55543
Ion Ratio Lower Upper
45 100
77 17.6 13.4 20.2
79 11.6 11.0 16.4









#16
Acetophenone
Concen: 5.009 ng/ul
RT: 8.957 min Scan# 980
Delta R.T. 0.006 min
Lab File: BP013779.D
Acq: 07 Feb 2023 15:50




Instrument :
BNA_P
ClientSampleId :
SSTD005610


Tgt Ion:105 Resp: 78250
Ion Ratio Lower Upper
105 100
77 74.4 61.0 91.4
51 24.2 19.6 29.4



#17
N-Nitroso-di-n-propylamine
Concen: 4.710 ng/ul
RT: 8.934 min Scan# 977
Delta R.T. -0.000 min
Lab File: BP013779.D
Acq: 07 Feb 2023 15:50

Tgt Ion: 70 Resp: 36858
Ion Ratio Lower Upper
70 100
42 53.3 39.8 59.8
101 10.0 7.4 11.2
130 20.9 14.6 22.0









#18
4-Methylphenol
Concen: 4.382 ng/ul
RT: 8.893 min Scan# 91
Delta R.T. -0.000 min
Lab File: BP013779.D
Acq: 07 Feb 2023 15:50




Instrument :
BNA_P
ClientSampleId :
SSTD005610


Tgt Ion:108 Resp: 46745
Ion Ratio Lower Upper
108 100
107 110.4 90.9 136.3



#19
Hexachloroethane
Concen: 5.194 ng/ul
RT: 9.222 min Scan# 1026
Delta R.T. 0.006 min
Lab File: BP013779.D
Acq: 07 Feb 2023 15:50

Tgt Ion:117 Resp: 18980
Ion Ratio Lower Upper
117 100
201 109.8 86.6 129.8
199 62.8 55.2 82.8









#22
Nitrobenzene
Concen: 4.469 ng/ul
RT: 9.340 min Scan# 1046
Delta R.T. -0.000 min
Lab File: BP013779.D
Acq: 07 Feb 2023 15:50




Instrument :
BNA_P
ClientSampleId :
SSTD005610


Tgt Ion: 77 Resp: 49024
Ion Ratio Lower Upper
77 100
123 41.8 35.1 52.7
65 12.4 10.6 16.0



#23
Isophorone
Concen: 4.315 ng/ul
RT: 9.869 min Scan# 1136
Delta R.T. -0.000 min
Lab File: BP013779.D
Acq: 07 Feb 2023 15:50

Tgt Ion: 82 Resp: 108364
Ion Ratio Lower Upper
82 100
95 6.6 6.0 9.0
138 16.8 13.7 20.5









#25
2-Nitrophenol
Concen: 4.078 ng/ul
RT: 10.057 min Scan# 1168
Delta R.T. -0.000 min
Lab File: BP013779.D
Acq: 07 Feb 2023 15:50




Instrument :
BNA_P
ClientSampleId :
SSTD005610


Tgt Ion:139 Resp: 21354
Ion Ratio Lower Upper
139 100
65 50.1 39.9 59.9
109 33.4 27.4 41.0



#26
2,4-Dimethylphenol
Concen: 4.741 ng/ul
RT: 10.110 min Scan# 1177
Delta R.T. 0.006 min
Lab File: BP013779.D
Acq: 07 Feb 2023 15:50

Tgt Ion:107 Resp: 56095
Ion Ratio Lower Upper
107 100
121 49.1 37.8 56.6
122 82.0 63.7 95.5









#27
Bis(2-Chloroethoxy)methane
Concen: 4.426 ng/ul
RT: 10.346 min Scan# 1217
Delta R.T. -0.000 min
Lab File: BP013779.D
Acq: 07 Feb 2023 15:50




Instrument :
BNA_P
ClientSampleId :
SSTD005610


Tgt Ion: 93 Resp: 67357
Ion Ratio Lower Upper
93 100
95 34.0 25.0 37.6
123 10.2 8.2 12.4



#29
2,4-Dichlorophenol
Concen: 4.280 ng/ul
RT: 10.593 min Scan# 1259
Delta R.T. -0.000 min
Lab File: BP013779.D
Acq: 07 Feb 2023 15:50

Tgt Ion: 162 Resp: 45843
Ion Ratio Lower Upper
162 100
164 67.3 51.4 77.2
98 39.3 27.5 41.3









#30
Naphthalene
Concen: 4.866 ng/ul
RT: 11.004 min Scan# 11
Delta R.T. -0.000 min
Lab File: BP013779.D
Acq: 07 Feb 2023 15:50




Instrument :
BNA_P
ClientSampleId :
SSTD005610


Tgt Ion:128 Resp: 163717
Ion Ratio Lower Upper
128 100
129 10.9 8.6 12.8
127 13.7 10.6 15.8



#32
4-Chloroaniline
Concen: 4.837 ng/ul
RT: 11.110 min Scan# 1347
Delta R.T. -0.000 min
Lab File: BP013779.D
Acq: 07 Feb 2023 15:50

Tgt Ion:127 Resp: 71160
Ion Ratio Lower Upper
127 100
129 29.3 26.2 39.4





#33

Hexachlorobutadiene

Concen: 4.669 ng/ul

RT: 11.287 min Scan# 11

Delta R.T. -0.000 min

Lab File: BP013779.D

Acq: 07 Feb 2023 15:50

Instrument :

BNA_P

ClientSampleId :

SSTD005610





Tgt Ion:225 Resp: 35268

Ion Ratio Lower Upper

225 100

223 63.0 50.6 76.0

227 62.8 49.0 73.4



#34

Caprolactam

Concen: 3.486 ng/ul

RT: 11.887 min Scan# 1479

Delta R.T. -0.000 min

Lab File: BP013779.D

Acq: 07 Feb 2023 15:50




Tgt Ion:113 Resp: 12625


Ion Ratio Lower Upper

113 100

55 156.1 120.9 181.3

56 109.3 93.8 140.8









#35
4-Chloro-3-methylphenol
Concen: 4.425 ng/ul
RT: 12.216 min Scan# 1535
Delta R.T. -0.000 min
Lab File: BP013779.D
Acq: 07 Feb 2023 15:50




Instrument :
BNA_P
ClientSampleId :
SSTD005610


Tgt Ion:107 Resp: 50767
Ion Ratio Lower Upper
107 100
144 26.0 20.0 30.0
142 73.5 62.1 93.1



#36
2-Methylnaphthalene
Concen: 4.695 ng/ul
RT: 12.598 min Scan# 1600
Delta R.T. -0.000 min
Lab File: BP013779.D
Acq: 07 Feb 2023 15:50

Tgt Ion:142 Resp: 112443
Ion Ratio Lower Upper
142 100
141 88.3 72.8 109.2









#37
1-Methylnaphthalene
Concen: 4.870 ng/ul
RT: 12.816 min Scan# 1637
Delta R.T. -0.000 min
Lab File: BP013779.D
Acq: 07 Feb 2023 15:50




Instrument :
BNA_P
ClientSampleId :
SSTD005610


Tgt Ion:142 Resp: 117789
Ion Ratio Lower Upper
142 100
141 92.7 74.2 111.2
116 4.2 3.1 4.7



#39
1,2,4,5-Tetrachlorobenzene
Concen: 4.454 ng/ul
RT: 12.957 min Scan# 1661
Delta R.T. -0.006 min
Lab File: BP013779.D
Acq: 07 Feb 2023 15:50

Tgt Ion:216 Resp: 67253
Ion Ratio Lower Upper
216 100
214 74.9 62.6 93.8
179 21.1 15.4 23.2
108 18.0 12.7 19.1









#40
Hexachlorocyclopentadiene
Concen: 3.785 ng/ul
RT: 12.940 min Scan# 1658
Delta R.T. -0.000 min
Lab File: BP013779.D
Acq: 07 Feb 2023 15:50




Instrument :
BNA_P
ClientSampleId :
SSTD005610


Tgt Ion:237	Resp:	32138
Ion Ratio	Lower	Upper
237	100	
235	68.3	49.8 74.8
272	13.0	10.1 15.1



#41
2,4,6-Trichlorophenol
Concen: 4.101 ng/ul
RT: 13.198 min Scan# 1702
Delta R.T. 0.006 min
Lab File: BP013779.D
Acq: 07 Feb 2023 15:50

Tgt Ion:196	Resp:	41020
Ion Ratio	Lower	Upper
196	100	
198	101.9	75.1 112.7
200	33.7	24.3 36.5





#42

2,4,5-Trichlorophenol

Concen: 3.944 ng/ul

RT: 13.269 min Scan# 1714

Delta R.T. -0.000 min

Lab File: BP013779.D

Acq: 07 Feb 2023 15:50

Instrument :

BNA_P

ClientSampleId :

SSTD005610





Tgt Ion:196 Resp: 42996

Ion Ratio Lower Upper

196 100

198 96.7 77.0 115.4

200 33.2 25.8 38.6



#43

1,1'-Biphenyl

Concen: 4.794 ng/ul

RT: 13.593 min Scan# 1769

Delta R.T. -0.006 min

Lab File: BP013779.D

Acq: 07 Feb 2023 15:50




Tgt Ion:154 Resp: 165913


Ion Ratio Lower Upper

154 100

153 39.4 31.9 47.9

76 12.8 9.9 14.9





#44

2-Chloronaphthalene

Concen: 4.728 ng/ul

RT: 13.640 min Scan# 1777

Delta R.T. -0.000 min

Lab File: BP013779.D

Acq: 07 Feb 2023 15:50

Instrument :

BNA_P

ClientSampleId :

SSTD005610





Tgt Ion:162 Resp: 129452

Ion Ratio Lower Upper

162 100

164 32.2 26.1 39.1

127 38.7 31.1 46.7



#45

2-Nitroaniline

Concen: 4.503 ng/ul

RT: 13.840 min Scan# 1811

Delta R.T. -0.000 min

Lab File: BP013779.D

Acq: 07 Feb 2023 15:50




Tgt Ion: 65 Resp: 23472


Ion Ratio Lower Upper

65 100

92 69.1 58.6 88.0

138 95.8 81.2 121.8









#47
Dimethylphthalate
Concen: 4.819 ng/ul
RT: 14.204 min Scan# 1873
Delta R.T. -0.006 min
Lab File: BP013779.D
Acq: 07 Feb 2023 15:50




Instrument :
BNA_P
ClientSampleId :
SSTD005610


Tgt Ion:163 Resp: 171091
Ion Ratio Lower Upper
163 100
194 3.8 3.4 5.2
164 10.1 7.8 11.8



#48
2,6-Dinitrotoluene
Concen: 4.386 ng/ul
RT: 14.328 min Scan# 1894
Delta R.T. -0.006 min
Lab File: BP013779.D
Acq: 07 Feb 2023 15:50

Tgt Ion:165 Resp: 22192
Ion Ratio Lower Upper
165 100
89 64.7 47.5 71.3
121 24.7 16.6 24.8









#50
Acenaphthylene
Concen: 4.748 ng/ul
RT: 14.487 min Scan# 1921
Delta R.T. -0.000 min
Lab File: BP013779.D
Acq: 07 Feb 2023 15:50




Instrument :
BNA_P
ClientSampleId :
SSTD005610


Tgt Ion:152 Resp: 199314
Ion Ratio Lower Upper
152 100
151 19.6 15.8 23.6
153 13.2 10.4 15.6



#51
3-Nitroaniline
Concen: 4.028 ng/ul
RT: 14.663 min Scan# 1951
Delta R.T. -0.000 min
Lab File: BP013779.D
Acq: 07 Feb 2023 15:50

Tgt Ion:138 Resp: 21542
Ion Ratio Lower Upper
138 100
108 15.5 9.4 14.2#
92 119.7 97.4 146.0









#52
Acenaphthene
Concen: 4.990 ng/ul
RT: 14.822 min Scan# 1978
Delta R.T. -0.000 min
Lab File: BP013779.D
Acq: 07 Feb 2023 15:50




Instrument :
BNA_P
ClientSampleId :
SSTD005610


Tgt Ion:153 Resp: 143972
Ion Ratio Lower Upper
153 100
152 49.3 37.4 56.0
154 91.7 70.1 105.1



#53
2,4-Dinitrophenol
Concen: 3.029 ng/ul
RT: 14.863 min Scan# 1985
Delta R.T. -0.000 min
Lab File: BP013779.D
Acq: 07 Feb 2023 15:50

Tgt Ion:184 Resp: 11901
Ion Ratio Lower Upper
184 100
63 58.8 49.5 74.3
154 54.2 54.3 81.5#









#55
4-Nitrophenol
Concen: 4.660 ng/ul
RT: 14.957 min Scan# 2001
Delta R.T. -0.000 min
Lab File: BP013779.D
Acq: 07 Feb 2023 15:50




Instrument :
BNA_P
ClientSampleId :
SSTD005610


Tgt Ion:109 Resp: 21575
Ion Ratio Lower Upper
109 100
139 110.6 96.0 144.0
65 107.1 95.8 143.6



#56
Dibenzofuran
Concen: 5.060 ng/ul
RT: 15.157 min Scan# 2035
Delta R.T. -0.000 min
Lab File: BP013779.D
Acq: 07 Feb 2023 15:50

Tgt Ion:168 Resp: 212041
Ion Ratio Lower Upper
168 100
139 37.1 29.6 44.4









#57
2,4-Dinitrotoluene
Concen: 4.261 ng/ul
RT: 15.110 min Scan# 2028
Delta R.T. -0.006 min
Lab File: BP013779.D
Acq: 07 Feb 2023 15:50




Instrument :
BNA_P
ClientSampleId :
SSTD005610


Tgt Ion:165 Resp: 33860
Ion Ratio Lower Upper
165 100
63 42.0 32.0 48.0
182 3.0 2.4 3.6



#58
2,3,4,6-Tetrachlorophenol
Concen: 4.172 ng/ul
RT: 15.381 min Scan# 2073
Delta R.T. -0.000 min
Lab File: BP013779.D
Acq: 07 Feb 2023 15:50

Tgt Ion:232 Resp: 42743
Ion Ratio Lower Upper
232 100
131 43.2 33.0 49.6
130 2.9 1.5 2.3#
166 33.0 22.6 33.8









#59
Diethylphthalate
Concen: 4.997 ng/ul
RT: 15.563 min Scan# 2105
Delta R.T. -0.006 min
Lab File: BP013779.D
Acq: 07 Feb 2023 15:50




Instrument :
BNA_P
ClientSampleId :
SSTD005610


Tgt Ion:149 Resp: 170316
Ion Ratio Lower Upper
149 100
177 21.2 17.8 26.8
150 12.0 10.1 15.1



#61
Fluorene
Concen: 5.167 ng/ul
RT: 15.804 min Scan# 2145
Delta R.T. -0.000 min
Lab File: BP013779.D
Acq: 07 Feb 2023 15:50

Tgt Ion:166 Resp: 176207
Ion Ratio Lower Upper
166 100
165 95.1 79.1 118.7
167 12.8 10.8 16.2









#62
4-Chlorophenyl-phenylether
Concen: 4.819 ng/ul
RT: 15.792 min Scan# 2144
Delta R.T. -0.006 min
Lab File: BP013779.D
Acq: 07 Feb 2023 15:50




Instrument :
BNA_P
ClientSampleId :
SSTD005610


Tgt Ion:204 Resp: 89544
Ion Ratio Lower Upper
204 100
206 32.4 26.7 40.1
141 58.8 44.6 66.8



#63
4-Nitroaniline
Concen: 3.505 ng/ul
RT: 15.822 min Scan# 2148
Delta R.T. -0.000 min
Lab File: BP013779.D
Acq: 07 Feb 2023 15:50

Tgt Ion:138 Resp: 17550
Ion Ratio Lower Upper
138 100
92 49.7 43.8 65.8
108 72.7 69.8 104.6









#66
4,6-Dinitro-2-methylphenol
Concen: 3.468 ng/ul
RT: 15.875 min Scan# 2157
Delta R.T. -0.000 min
Lab File: BP013779.D
Acq: 07 Feb 2023 15:50




Instrument :
BNA_P
ClientSampleId :
SSTD005610


Tgt Ion:198 Resp: 21435
Ion Ratio Lower Upper
198 100
182 8.8 4.1 6.1#
77 21.4 20.3 30.5



#67
N-Nitrosodiphenylamine
Concen: 4.834 ng/ul
RT: 16.010 min Scan# 2180
Delta R.T. -0.000 min
Lab File: BP013779.D
Acq: 07 Feb 2023 15:50

Tgt Ion:169 Resp: 144576
Ion Ratio Lower Upper
169 100
168 68.9 53.5 80.3
167 37.8 29.8 44.6









#68
4-Bromophenyl-phenylether
Concen: 4.235 ng/ul
RT: 16.692 min Scan# 21
Delta R.T. -0.000 min
Lab File: BP013779.D
Acq: 07 Feb 2023 15:50




Instrument :
BNA_P
ClientSampleId :
SSTD005610


Tgt Ion:248	Resp:	53393
Ion Ratio	Lower	Upper
248	100	
250	94.8	76.0 114.0
141	79.7	56.0 84.0



#69
Hexachlorobenzene
Concen: 4.410 ng/ul
RT: 16.810 min Scan# 2316
Delta R.T. -0.006 min
Lab File: BP013779.D
Acq: 07 Feb 2023 15:50

Tgt Ion:284	Resp:	65223
Ion Ratio	Lower	Upper
284	100	
142	37.3	25.0 37.4
249	33.7	24.9 37.3





#70

Atrazine

Concen: 4.275 ng/ul

RT: 16.957 min Scan# 21

Delta R.T. -0.000 min

Lab File: BP013779.D

Acq: 07 Feb 2023 15:50

Instrument :

BNA_P

ClientSampleId :

SSTD005610





Tgt Ion:200 Resp: 51459

Ion Ratio Lower Upper

200 100

173 26.8 20.4 30.6

215 47.5 39.3 58.9



#71

Pentachlorophenol

Concen: 3.613 ng/ul

RT: 17.157 min Scan# 2375

Delta R.T. -0.000 min

Lab File: BP013779.D

Acq: 07 Feb 2023 15:50




Tgt Ion:266 Resp: 34106


Ion Ratio Lower Upper

266 100

264 60.2 50.5 75.7

268 61.7 51.0 76.6





#72

Phenanthrene

Concen: 4.981 ng/ul

RT: 17.557 min Scan# 2443

Delta R.T. -0.000 min

Lab File: BP013779.D

Acq: 07 Feb 2023 15:50

Instrument :

BNA_P

ClientSampleId :

SSTD005610





Tgt Ion:178 Resp: 291554

Ion Ratio Lower Upper

178 100

179 15.8 12.2 18.2

176 18.4 15.2 22.8



#74

Anthracene

Concen: 4.922 ng/ul

RT: 17.645 min Scan# 2458

Delta R.T. -0.006 min

Lab File: BP013779.D

Acq: 07 Feb 2023 15:50




Tgt Ion:178 Resp: 291784


Ion Ratio Lower Upper

178 100

179 15.6 12.2 18.4

176 18.2 15.0 22.4









#75
1,2,3,4-Tetrachlorobenzene
Concen: 4.369 ng/uL
RT: 13.563 min Scan# 1764
Delta R.T. -0.000 min
Lab File: BP013779.D
Acq: 07 Feb 2023 15:50




Instrument :
BNA_P
ClientSampleId :
SSTD005610


Tgt Ion:216 Resp: 68981
Ion Ratio Lower Upper
216 100
214 80.5 62.6 94.0
218 50.8 39.0 58.4



#76
Pentachlorobenzene
Concen: 4.476 ng/uL
RT: 15.075 min Scan# 2021
Delta R.T. -0.000 min
Lab File: BP013779.D
Acq: 07 Feb 2023 15:50

Tgt Ion:250 Resp: 72451
Ion Ratio Lower Upper
250 100
248 63.9 51.1 76.7
252 64.6 51.0 76.6









#77
 Carbazole
 Concen: 4.966 ng/ul
 RT: 17.910 min Scan# 2503
 Delta R.T. -0.000 min
 Lab File: BP013779.D
 Acq: 07 Feb 2023 15:50




Instrument :
 BNA_P
 ClientSampleId :
 SSTD005610


Tgt Ion:167 Resp: 255444
 Ion Ratio Lower Upper
 167 100
 166 22.1 17.3 25.9
 139 12.7 10.4 15.6



#78
 Di-n-butylphthalate
 Concen: 4.894 ng/ul
 RT: 18.445 min Scan# 2594
 Delta R.T. -0.000 min
 Lab File: BP013779.D
 Acq: 07 Feb 2023 15:50

Tgt Ion:149 Resp: 297421
 Ion Ratio Lower Upper
 149 100
 150 9.0 7.1 10.7
 104 5.4 4.1 6.1





#80

Fluoranthene

Concen: 4.960 ng/ul

RT: 19.504 min Scan# 21

Delta R.T. -0.000 min

Lab File: BP013779.D

Acq: 07 Feb 2023 15:50

Instrument :

BNA_P

ClientSampleId :

SSTD005610





Tgt Ion:202 Resp: 422797

Ion Ratio Lower Upper

202 100

101 11.3 8.6 12.8

100 8.3 6.3 9.5



#82

Pyrene

Concen: 5.278 ng/ul

RT: 19.857 min Scan# 2834

Delta R.T. -0.000 min

Lab File: BP013779.D

Acq: 07 Feb 2023 15:50




Tgt Ion:202 Resp: 459980


Ion Ratio Lower Upper

202 100

101 12.2 10.3 15.5

100 10.5 8.5 12.7





#83

Butylbenzylphthalate

Concen: 4.906 ng/ul

RT: 20.710 min Scan# 2979

Delta R.T. -0.000 min

Lab File: BP013779.D

Acq: 07 Feb 2023 15:50

Instrument :

BNA_P

ClientSampleId :

SSTD005610





Tgt Ion:149 Resp: 122849

Ion Ratio Lower Upper

149 100

91 72.1 57.7 86.5

206 21.2 16.7 25.1



#84

3,3'-Dichlorobenzidine

Concen: 4.019 ng/ul

RT: 21.498 min Scan# 3113

Delta R.T. -0.000 min

Lab File: BP013779.D

Acq: 07 Feb 2023 15:50




Tgt Ion:252 Resp: 119703


Ion Ratio Lower Upper

252 100

254 63.6 51.8 77.6

126 9.8 7.9 11.9





#85

Benzo(a)anthracene

Concen: 4.738 ng/ul

RT: 21.569 min Scan# 3126

Delta R.T. -0.006 min

Lab File: BP013779.D

Acq: 07 Feb 2023 15:50

Instrument :

BNA_P

ClientSampleId :

SSTD005610





Tgt Ion:228 Resp: 427117

Ion Ratio Lower Upper

228 100

229 19.9 15.4 23.2

226 26.3 21.8 32.6



#86

Bis(2-ethylhexyl)phthalate

Concen: 4.517 ng/ul

RT: 21.469 min Scan# 3108

Delta R.T. -0.000 min

Lab File: BP013779.D

Acq: 07 Feb 2023 15:50




Tgt Ion:149 Resp: 193939


Ion Ratio Lower Upper

149 100

167 27.4 21.5 32.3

279 4.4 3.2 4.8





#87

Chrysene

Concen: 4.983 ng/ul

RT: 21.627 min Scan# 31

Delta R.T. -0.000 min

Lab File: BP013779.D

Acq: 07 Feb 2023 15:50

Instrument :

BNA_P

ClientSampleId :

SSTD005610





Tgt Ion:228 Resp: 422900

Ion Ratio Lower Upper

228 100

226 29.8 25.0 37.6

229 19.3 16.4 24.6



#89

Di-n-octyl phthalate

Concen: 4.485 ng/ul




RT: 22.445 min Scan# 3274


Delta R.T. -0.006 min

Lab File: BP013779.D

Acq: 07 Feb 2023 15:50

Tgt Ion:149 Resp: 324336









#90
Benzo(b)fluoranthene
Concen: 4.709 ng/ul
RT: 23.363 min Scan# 3430
Delta R.T. -0.000 min
Lab File: BP013779.D
Acq: 07 Feb 2023 15:50




Instrument :
BNA_P
ClientSampleId :
SSTD005610


Tgt Ion:252 Resp: 422056
Ion Ratio Lower Upper
252 100
253 21.2 17.5 26.3
125 9.6 7.8 11.6



#91
Benzo(k)fluoranthene
Concen: 4.927 ng/ul
RT: 23.410 min Scan# 3438
Delta R.T. -0.006 min
Lab File: BP013779.D
Acq: 07 Feb 2023 15:50

Tgt Ion:252 Resp: 421986
Ion Ratio Lower Upper
252 100
253 22.4 17.1 25.7
125 10.2 7.2 10.8





#93

Benzo(a)pyrene

Concen: 4.702 ng/ul

RT: 24.033 min Scan# 3

Delta R.T. -0.006 min

Lab File: BP013779.D

Acq: 07 Feb 2023 15:50

Instrument :

BNA_P

ClientSampleId :

SSTD005610





Tgt Ion:252 Resp: 363802

Ion Ratio Lower Upper

252 100

253 21.3 17.4 26.0

125 11.4 8.2 12.2



#94

Indeno(1,2,3-cd)pyrene

Concen: 4.579 ng/ul

RT: 26.851 min Scan# 4023

Delta R.T. -0.006 min

Lab File: BP013779.D

Acq: 07 Feb 2023 15:50




Tgt Ion:276 Resp: 438937


Ion Ratio Lower Upper

276 100

138 22.9 18.4 27.6

277 26.5 19.7 29.5









#95
Dibenzo(a,h)anthracene
Concen: 4.600 ng/ul
RT: 26.868 min Scan# 4027
Delta R.T. -0.006 min
Lab File: BP013779.D
Acq: 07 Feb 2023 15:50

Instrument :
BNA_P
ClientSampleId :
SSTD005610

Tgt Ion:278 Resp: 365542
Ion Ratio Lower Upper
278 100
139 15.2 12.0 18.0
279 23.4 18.9 28.3



#96
Benzo(g,h,i)perylene
Concen: 4.639 ng/ul
RT: 27.680 min Scan# 4164
Delta R.T. -0.012 min
Lab File: BP013779.D
Acq: 07 Feb 2023 15:50

Tgt Ion:276 Resp: 351819
Ion Ratio Lower Upper
276 100
138 22.1 15.9 23.9
277 23.0 19.0 28.6

