

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP020723\
 Data File : BP013783.D
 Acq On : 07 Feb 2023 18:07
 Operator : CG/JU
 Sample : SST08033
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SST080614

Quant Time: Feb 08 00:04:10 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP020723.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue Feb 07 23:57:24 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.122	152	191208	20.000	ng/u1	0.00
20) Naphthalene-d8	10.957	136	847044	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.763	164	571925	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.516	188	1329901	20.000	ng/u1	0.00
79) Chrysene-d12	21.598	240	1262388	20.000	ng/u1	0.00
88) Perylene-d12	24.156	264	1165908	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.469	96	140803	31.221	ng/uL	0.00
4) Pyridine-d5	3.899	84	1088278	84.101	ng/u1	0.00
7) Phenol-d5	7.275	99	1394295	82.969	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.446	67	787003	79.849	ng/u1	0.00
11) 2-Chlorophenol-d4	7.651	132	1052241	85.417	ng/u1	0.00
15) 4-Methylphenol-d8	8.840	113	1143869	85.384	ng/u1	0.01
21) Nitrobenzene-d5	9.304	128	535982	99.191	ng/u1	0.00
24) 2-Nitrophenol-d4	10.028	143	633772	100.643	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.569	165	1158761	78.688	ng/u1	0.00
31) 4-Chloroaniline-d4	11.092	131	1533668	78.855	ng/u1	0.00
46) Dimethylphthalate-d6	14.169	166	3534195	78.394	ng/u1	0.00
49) Acenaphthylene-d8	14.463	160	3888461	78.066	ng/u1	0.00
54) 4-Nitrophenol-d4	14.963	143	669642	88.124	ng/u1	0.02
60) Fluorene-d10	15.757	176	2917274	75.074	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.880	200	714849	95.919	ng/u1	0.02
73) Anthracene-d10	17.622	188	4638133	74.820	ng/u1	0.00
81) Pyrene-d10	19.839	212	5604352	79.598	ng/u1	0.01
92) Benzo(a)pyrene-d12	23.998	264	4769282	80.131	ng/u1	0.01
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.505	88	148031	31.497	ng/uL	98
5) Pyridine	3.922	79	1075897	82.572	ng/u1	97
6) Benzaldehyde	7.257	77	627571	93.138	ng/u1	100
8) Phenol	7.304	94	1422873	82.735	ng/u1	98
10) Bis(2-Chloroethyl)ether	7.540	93	1108204	80.189	ng/u1	98
12) 2-Chlorophenol	7.687	128	1063991	85.252	ng/u1	99
13) 2-Methylphenol	8.569	108	1089704	82.987	ng/u1	100
14) 2,2'-oxybis(1-Chloropr...	8.657	45	1250633	76.433	ng/u1	99
16) Acetophenone	8.963	105	1721730	82.440	ng/u1	99
17) N-Nitroso-di-n-propyla...	8.957	70	873126	83.457	ng/u1	97
18) 4-Methylphenol	8.910	108	1195175	83.807	ng/u1	98
19) Hexachloroethane	9.222	117	430464	88.117	ng/u1	98
22) Nitrobenzene	9.345	77	1327512	90.960	ng/u1	99
23) Isophorone	9.887	82	2662709	79.691	ng/u1	99
25) 2-Nitrophenol	10.063	139	648107	93.017	ng/u1	99
26) 2,4-Dimethylphenol	10.116	107	1322675	84.018	ng/u1	99
27) Bis(2-Chloroethoxy)met...	10.357	93	1552213	76.654	ng/u1	99
29) 2,4-Dichlorophenol	10.598	162	1122144	78.738	ng/u1	97
30) Naphthalene	11.010	128	3493077	78.026	ng/u1	100
32) 4-Chloroaniline	11.116	127	1544492	78.905	ng/u1	98
33) Hexachlorobutadiene	11.287	225	761373	75.750	ng/u1	98
34) Caprolactam	11.922	113	224443	46.577	ng/u1	99
35) 4-Chloro-3-methylphenol	12.228	107	1286027	84.250	ng/u1	99
36) 2-Methylnaphthalene	12.604	142	2419974	75.941	ng/u1	98

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP020723\
 Data File : BP013783.D
 Acq On : 07 Feb 2023 18:07
 Operator : CG/JU
 Sample : SST08033
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SST080614


Quant Time: Feb 08 00:04:10 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP020723.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue Feb 07 23:57:24 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.822	142	2477978	76.993	ng/ul	100
39) 1,2,4,5-Tetrachloroben...	12.963	216	1464683	77.529	ng/ul	98
40) Hexachlorocyclopentadiene	12.945	237	1002944	94.421	ng/ul	99
41) 2,4,6-Trichlorophenol	13.198	196	1006295	80.407	ng/ul	98
42) 2,4,5-Trichlorophenol	13.275	196	1108264	81.257	ng/ul	99
43) 1,1'-Biphenyl	13.598	154	3318901	76.645	ng/ul	100
44) 2-Chloronaphthalene	13.645	162	2657907	77.583	ng/ul	99
45) 2-Nitroaniline	13.851	65	785243	120.412	ng/ul	95
47) Dimethylphthalate	14.216	163	3459655	77.895	ng/ul	99
48) 2,6-Dinitrotoluene	14.345	165	743937	117.532	ng/ul	98
50) Acenaphthylene	14.492	152	4071116	77.512	ng/ul	99
51) 3-Nitroaniline	14.675	138	667030	99.701	ng/ul	96
52) Acenaphthene	14.828	153	2802435	77.642	ng/ul	99
53) 2,4-Dinitrophenol	14.875	184	514102	104.572	ng/ul	96
55) 4-Nitrophenol	14.980	109	580256	100.172	ng/ul	96
56) Dibenzofuran	15.163	168	3943552	75.217	ng/ul	98
57) 2,4-Dinitrotoluene	15.128	165	1067869	107.414	ng/ul	99
58) 2,3,4,6-Tetrachlorophenol	15.386	232	990133	77.243	ng/ul	100
59) Diethylphthalate	15.580	149	3467593	81.324	ng/ul	99
61) Fluorene	15.816	166	3114735	73.001	ng/ul	99
62) 4-Chlorophenyl-phenyle...	15.798	204	1661875	71.492	ng/ul	99
63) 4-Nitroaniline	15.845	138	641913	102.469	ng/ul	97
66) 4,6-Dinitro-2-methylph...	15.892	198	681548	91.890	ng/ul#	98
67) N-Nitrosodiphenylamine	16.016	169	2767770	77.105	ng/ul	99
68) 4-Bromophenyl-phenylether	16.698	248	1131132	74.758	ng/ul	98
69) Hexachlorobenzene	16.822	284	1303909	73.455	ng/ul	98
70) Atrazine	16.969	200	1140358	78.934	ng/ul	99
71) Pentachlorophenol	17.163	266	882065	77.850	ng/ul	99
72) Phenanthrene	17.563	178	5278455	75.139	ng/ul	99
74) Anthracene	17.657	178	5240741	73.657	ng/ul	99
75) 1,2,3,4-Tetrachloroben...	13.569	216	1468617	77.514	ng/ul	100
76) Pentachlorobenzene	15.081	250	1467207	75.537	ng/ul	99
77) Carbazole	17.916	167	4792362	77.639	ng/ul	99
78) Di-n-butylphthalate	18.445	149	5881164	80.638	ng/ul	99
80) Fluoranthene	19.510	202	6528001	79.546	ng/ul	98
82) Pyrene	19.868	202	6544362	77.998	ng/ul	94
83) Butylbenzylphthalate	20.715	149	2642634	109.625	ng/ul	98
84) 3,3'-Dichlorobenzidine	21.504	252	2347010	81.856	ng/ul	100
85) Benzo(a)anthracene	21.580	228	6713611	77.357	ng/ul	99
86) Bis(2-ethylhexyl)phtha...	21.474	149	4024904	97.385	ng/ul#	98
87) Chrysene	21.639	228	6248100	76.468	ng/ul	99
89) Di-n-octyl phthalate	22.457	149	6967030	113.439	ng/ul	100
90) Benzo(b)fluoranthene	23.374	252	6207273	81.557	ng/ul	98
91) Benzo(k)fluoranthene	23.427	252	6211237	85.388	ng/ul	98
93) Benzo(a)pyrene	24.057	252	5231621	79.620	ng/ul	99
94) Indeno(1,2,3-cd)pyrene	26.880	276	5878063	72.204	ng/ul	98
95) Dibenzo(a,h)anthracene	26.892	278	4860782	72.023	ng/ul	98
96) Benzo(g,h,i)perylene	27.715	276	4746651	73.702	ng/ul	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Quant Time: Feb 08 00:04:10 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP020723.M
Quant Title : SVOA CALIBRATION
QLast Update : Tue Feb 07 23:57:24 2023
Response via : Initial Calibration









#2
1,4-Dioxane
Concen: 31.497 ng/uL
RT: 3.505 min Scan# 54
Delta R.T. -0.000 min
Lab File: BP013783.D
Acq: 07 Feb 2023 18:07




Instrument :
BNA_P
ClientSampleId :
SSTD080614


Tgt Ion: 88 Resp: 148031
Ion Ratio Lower Upper
88 100
43 25.7 20.1 30.1
58 68.0 53.0 79.6



#5
Pyridine
Concen: 82.572 ng/uL
RT: 3.922 min Scan# 125
Delta R.T. -0.000 min
Lab File: BP013783.D
Acq: 07 Feb 2023 18:07

Tgt Ion: 79 Resp: 1075897
Ion Ratio Lower Upper
79 100
52 66.0 55.1 82.7
51 32.3 25.7 38.5









#6
Benzaldehyde
Concen: 93.138 ng/ul
RT: 7.257 min Scan# 691
Delta R.T. 0.006 min
Lab File: BP013783.D
Acq: 07 Feb 2023 18:07




Instrument :
BNA_P
ClientSampleId :
SSTD080614


Tgt Ion: 77 Resp: 627571
Ion Ratio Lower Upper
77 100
105 99.3 79.5 119.3
106 95.9 76.5 114.7



#8
Phenol
Concen: 82.735 ng/ul
RT: 7.304 min Scan# 700
Delta R.T. 0.006 min
Lab File: BP013783.D
Acq: 07 Feb 2023 18:07

Tgt Ion: 94 Resp: 1422873
Ion Ratio Lower Upper
94 100
65 28.1 22.2 33.4
66 38.8 29.6 44.4









#10
Bis(2-Chloroethyl)ether
Concen: 80.189 ng/ul
RT: 7.540 min Scan# 740
Delta R.T. 0.006 min
Lab File: BP013783.D
Acq: 07 Feb 2023 18:07




Instrument :
BNA_P
ClientSampleId :
SSTD080614


Tgt Ion: 93 Resp: 1108204
Ion Ratio Lower Upper
93 100
63 72.3 59.4 89.0
95 32.5 25.8 38.8



#12
2-Chlorophenol
Concen: 85.252 ng/ul
RT: 7.687 min Scan# 765
Delta R.T. 0.006 min
Lab File: BP013783.D
Acq: 07 Feb 2023 18:07

Tgt Ion:128 Resp: 1063991
Ion Ratio Lower Upper
128 100
64 47.0 37.3 55.9
130 32.6 26.3 39.5









#13
2-Methylphenol
Concen: 82.987 ng/ul
RT: 8.569 min Scan# 914
Delta R.T. 0.006 min
Lab File: BP013783.D
Acq: 07 Feb 2023 18:07




Instrument :
BNA_P
ClientSampleId :
SSTD080614


Tgt Ion:108 Resp: 1089704
Ion Ratio Lower Upper
108 100
107 89.8 71.9 107.9



#14
2,2'-oxybis(1-Chloropropane)
Concen: 76.433 ng/ul
RT: 8.657 min Scan# 930
Delta R.T. 0.006 min
Lab File: BP013783.D
Acq: 07 Feb 2023 18:07

Tgt Ion: 45 Resp: 1250633
Ion Ratio Lower Upper
45 100
77 17.6 13.4 20.2
79 13.5 11.0 16.4









#16
 Acetophenone
 Concen: 82.440 ng/ul
 RT: 8.963 min Scan# 91
 Delta R.T. 0.011 min
 Lab File: BP013783.D
 Acq: 07 Feb 2023 18:07




Instrument :
 BNA_P
 ClientSampleId :
 SSTD080614


Tgt Ion:105 Resp: 1721730
 Ion Ratio Lower Upper
 105 100
 77 76.8 61.0 91.4
 51 25.1 19.6 29.4



#17
 N-Nitroso-di-n-propylamine
 Concen: 83.457 ng/ul
 RT: 8.957 min Scan# 981
 Delta R.T. 0.023 min
 Lab File: BP013783.D
 Acq: 07 Feb 2023 18:07

Tgt Ion: 70 Resp: 873126
 Ion Ratio Lower Upper
 70 100
 42 51.2 39.8 59.8
 101 10.2 7.4 11.2
 130 20.0 14.6 22.0









#18
4-Methylphenol
Concen: 83.807 ng/ul
RT: 8.910 min Scan# 91
Delta R.T. 0.017 min
Lab File: BP013783.D
Acq: 07 Feb 2023 18:07




Instrument :
BNA_P
ClientSampleId :
SSTD080614


Tgt Ion:108 Resp: 1195175
Ion Ratio Lower Upper
108 100
107 111.5 90.9 136.3



#19
Hexachloroethane
Concen: 88.117 ng/ul
RT: 9.222 min Scan# 1026
Delta R.T. 0.006 min
Lab File: BP013783.D
Acq: 07 Feb 2023 18:07

Tgt Ion:117 Resp: 430464
Ion Ratio Lower Upper
117 100
201 111.6 86.6 129.8
199 69.7 55.2 82.8









#22
Nitrobenzene
Concen: 90.960 ng/ul
RT: 9.345 min Scan# 1046
Delta R.T. 0.006 min
Lab File: BP013783.D
Acq: 07 Feb 2023 18:07




Instrument :
BNA_P
ClientSampleId :
SSTD080614


Tgt Ion: 77 Resp: 1327512
Ion Ratio Lower Upper
77 100
123 44.2 35.1 52.7
65 13.7 10.6 16.0



#23
Isophorone
Concen: 79.691 ng/ul
RT: 9.887 min Scan# 1139
Delta R.T. 0.017 min
Lab File: BP013783.D
Acq: 07 Feb 2023 18:07

Tgt Ion: 82 Resp: 2662709
Ion Ratio Lower Upper
82 100
95 7.5 6.0 9.0
138 16.8 13.7 20.5









#25
2-Nitrophenol
Concen: 93.017 ng/ul
RT: 10.063 min Scan# 1169
Delta R.T. 0.006 min
Lab File: BP013783.D
Acq: 07 Feb 2023 18:07




Instrument :
BNA_P
ClientSampleId :
SSTD080614


Tgt Ion:139 Resp: 648107
Ion Ratio Lower Upper
139 100
65 50.6 39.9 59.9
109 34.4 27.4 41.0



#26
2,4-Dimethylphenol
Concen: 84.018 ng/ul
RT: 10.116 min Scan# 1178
Delta R.T. 0.011 min
Lab File: BP013783.D
Acq: 07 Feb 2023 18:07

Tgt Ion:107 Resp: 1322675
Ion Ratio Lower Upper
107 100
121 48.1 37.8 56.6
122 80.1 63.7 95.5









#27
 Bis(2-Chloroethoxy)methane
 Concen: 76.654 ng/ul
 RT: 10.357 min Scan# 1219
 Delta R.T. 0.011 min
 Lab File: BP013783.D
 Acq: 07 Feb 2023 18:07

Instrument :
 BNA_P
 ClientSampleId :
 SSTD080614

Tgt Ion: 93 Resp: 1552213




Ion	Ratio	Lower	Upper
93	100		
95	32.0	25.0	37.6
123	10.9	8.2	12.4




#29
 2,4-Dichlorophenol
 Concen: 78.738 ng/ul
 RT: 10.598 min Scan# 1260
 Delta R.T. 0.006 min
 Lab File: BP013783.D
 Acq: 07 Feb 2023 18:07

Tgt Ion: 162 Resp: 1122144

Ion	Ratio	Lower	Upper
162	100		
164	66.1	51.4	77.2
98	36.6	27.5	41.3









#30
Naphthalene
Concen: 78.026 ng/ul
RT: 11.010 min Scan# 11
Delta R.T. 0.006 min
Lab File: BP013783.D
Acq: 07 Feb 2023 18:07




Instrument :
BNA_P
ClientSampleId :
SSTD080614


Tgt Ion:128 Resp: 3493077
Ion Ratio Lower Upper
128 100
129 10.9 8.6 12.8
127 13.3 10.6 15.8



#32
4-Chloroaniline
Concen: 78.905 ng/ul
RT: 11.116 min Scan# 1348
Delta R.T. 0.006 min
Lab File: BP013783.D
Acq: 07 Feb 2023 18:07

Tgt Ion:127 Resp: 1544492
Ion Ratio Lower Upper
127 100
129 31.9 26.2 39.4









#33
Hexachlorobutadiene
Concen: 75.750 ng/ul
RT: 11.287 min Scan# 11
Delta R.T. -0.000 min
Lab File: BP013783.D
Acq: 07 Feb 2023 18:07




Instrument :
BNA_P
ClientSampleId :
SSTD080614


Tgt Ion:225 Resp: 761373
Ion Ratio Lower Upper
225 100
223 61.7 50.6 76.0
227 63.4 49.0 73.4



#34
Caprolactam
Concen: 46.577 ng/ul
RT: 11.922 min Scan# 1485
Delta R.T. 0.035 min
Lab File: BP013783.D
Acq: 07 Feb 2023 18:07

Tgt Ion:113 Resp: 224443
Ion Ratio Lower Upper
113 100
55 149.5 120.9 181.3
56 115.8 93.8 140.8









#35
4-Chloro-3-methylphenol
Concen: 84.250 ng/ul
RT: 12.228 min Scan# 1537
Delta R.T. 0.011 min
Lab File: BP013783.D
Acq: 07 Feb 2023 18:07




Instrument :
BNA_P
ClientSampleId :
SSTD080614


Tgt Ion:107 Resp: 1286027
Ion Ratio Lower Upper
107 100
144 25.1 20.0 30.0
142 76.0 62.1 93.1



#36
2-Methylnaphthalene
Concen: 75.941 ng/ul
RT: 12.604 min Scan# 1601
Delta R.T. 0.006 min
Lab File: BP013783.D
Acq: 07 Feb 2023 18:07

Tgt Ion:142 Resp: 2419974
Ion Ratio Lower Upper
142 100
141 89.0 72.8 109.2









#37
1-Methylnaphthalene
Concen: 76.993 ng/ul
RT: 12.822 min Scan# 1638
Delta R.T. 0.006 min
Lab File: BP013783.D
Acq: 07 Feb 2023 18:07




Instrument :
BNA_P
ClientSampleId :
SSTD080614


Tgt Ion:142 Resp: 2477978
Ion Ratio Lower Upper
142 100
141 92.3 74.2 111.2
116 4.1 3.1 4.7



#39
1,2,4,5-Tetrachlorobenzene
Concen: 77.529 ng/ul
RT: 12.963 min Scan# 1662
Delta R.T. -0.000 min
Lab File: BP013783.D
Acq: 07 Feb 2023 18:07

Tgt Ion:216 Resp: 1464683
Ion Ratio Lower Upper
216 100
214 79.2 62.6 93.8
179 21.2 15.4 23.2
108 16.7 12.7 19.1









#40
Hexachlorocyclopentadiene
Concen: 94.421 ng/ul
RT: 12.945 min Scan# 1658
Delta R.T. 0.006 min
Lab File: BP013783.D
Acq: 07 Feb 2023 18:07




Instrument :
BNA_P
ClientSampleId :
SSTD080614


Tgt Ion:237 Resp: 1002944
Ion Ratio Lower Upper
237 100
235 62.8 49.8 74.8
272 12.8 10.1 15.1



#41
2,4,6-Trichlorophenol
Concen: 80.407 ng/ul
RT: 13.198 min Scan# 1702
Delta R.T. 0.006 min
Lab File: BP013783.D
Acq: 07 Feb 2023 18:07

Tgt Ion:196 Resp: 1006295
Ion Ratio Lower Upper
196 100
198 95.8 75.1 112.7
200 30.5 24.3 36.5









#42
2,4,5-Trichlorophenol
Concen: 81.257 ng/ul
RT: 13.275 min Scan# 1715
Delta R.T. 0.006 min
Lab File: BP013783.D
Acq: 07 Feb 2023 18:07




Instrument :
BNA_P
ClientSampleId :
SSTD080614


Tgt Ion:196 Resp: 1108264
Ion Ratio Lower Upper
196 100
198 95.4 77.0 115.4
200 30.8 25.8 38.6



#43
1,1'-Biphenyl
Concen: 76.645 ng/ul
RT: 13.598 min Scan# 1770
Delta R.T. -0.000 min
Lab File: BP013783.D
Acq: 07 Feb 2023 18:07

Tgt Ion:154 Resp: 3318901
Ion Ratio Lower Upper
154 100
153 40.3 31.9 47.9
76 12.4 9.9 14.9





#44

2-Chloronaphthalene

Concen: 77.583 ng/ul

RT: 13.645 min Scan# 1

Delta R.T. 0.006 min

Lab File: BP013783.D

Acq: 07 Feb 2023 18:07

Instrument :

BNA_P

ClientSampleId :

SSTD080614





Tgt Ion:162 Resp: 2657907

Ion Ratio Lower Upper

162 100

164 33.1 26.1 39.1

127 39.3 31.1 46.7



#45

2-Nitroaniline

Concen: 120.412 ng/ul

RT: 13.851 min Scan# 1813

Delta R.T. 0.011 min

Lab File: BP013783.D

Acq: 07 Feb 2023 18:07




Tgt Ion: 65 Resp: 785243


Ion Ratio Lower Upper

65 100

92 73.4 58.6 88.0

138 109.9 81.2 121.8









#47
Dimethylphthalate
Concen: 77.895 ng/ul
RT: 14.216 min Scan# 1875
Delta R.T. 0.006 min
Lab File: BP013783.D
Acq: 07 Feb 2023 18:07




Instrument :
BNA_P
ClientSampleId :
SSTD080614


Tgt Ion:163 Resp: 3459655
Ion Ratio Lower Upper
163 100
194 4.1 3.4 5.2
164 10.3 7.8 11.8



#48
2,6-Dinitrotoluene
Concen: 117.532 ng/ul
RT: 14.345 min Scan# 1897
Delta R.T. 0.011 min
Lab File: BP013783.D
Acq: 07 Feb 2023 18:07

Tgt Ion:165 Resp: 743937
Ion Ratio Lower Upper
165 100
89 57.0 47.5 71.3
121 20.4 16.6 24.8





#50

Acenaphthylene

Concen: 77.512 ng/ul

RT: 14.492 min Scan# 1922

Delta R.T. 0.006 min

Lab File: BP013783.D

Acq: 07 Feb 2023 18:07

Instrument :

BNA_P

ClientSampleId :

SSTD080614





Tgt Ion:152 Resp: 4071116

Ion Ratio Lower Upper

152 100

151 20.2 15.8 23.6

153 13.2 10.4 15.6



#51

3-Nitroaniline

Concen: 99.701 ng/ul

RT: 14.675 min Scan# 1953

Delta R.T. 0.011 min

Lab File: BP013783.D

Acq: 07 Feb 2023 18:07




Tgt Ion:138 Resp: 667030


Ion Ratio Lower Upper

138 100

108 11.0 9.4 14.2

92 117.5 97.4 146.0





#52

Acenaphthene

Concen: 77.642 ng/ul

RT: 14.828 min Scan# 1978

Delta R.T. 0.006 min

Lab File: BP013783.D

Acq: 07 Feb 2023 18:07

Instrument :

BNA_P

ClientSampleId :

SSTD080614





Tgt Ion:153 Resp: 2802435

Ion Ratio Lower Upper

153 100

152 47.8 37.4 56.0

154 88.7 70.1 105.1



#53

2,4-Dinitrophenol

Concen: 104.572 ng/ul

RT: 14.875 min Scan# 1987

Delta R.T. 0.011 min

Lab File: BP013783.D

Acq: 07 Feb 2023 18:07




Tgt Ion:184 Resp: 514102


Ion Ratio Lower Upper

184 100

63 58.4 49.5 74.3

154 65.6 54.3 81.5









#55
4-Nitrophenol
Concen: 100.172 ng/ul
RT: 14.980 min Scan# 2005
Delta R.T. 0.023 min
Lab File: BP013783.D
Acq: 07 Feb 2023 18:07




Instrument :
BNA_P
ClientSampleId :
SSTD080614


Tgt Ion:109 Resp: 580256
Ion Ratio Lower Upper
109 100
139 118.2 96.0 144.0
65 113.5 95.8 143.6



#56
Dibenzofuran
Concen: 75.217 ng/ul
RT: 15.163 min Scan# 2036
Delta R.T. 0.006 min
Lab File: BP013783.D
Acq: 07 Feb 2023 18:07

Tgt Ion:168 Resp: 3943552
Ion Ratio Lower Upper
168 100
139 38.3 29.6 44.4









#57
2,4-Dinitrotoluene
Concen: 107.414 ng/ul
RT: 15.128 min Scan# 2028
Delta R.T. 0.011 min
Lab File: BP013783.D
Acq: 07 Feb 2023 18:07




Instrument :
BNA_P
ClientSampleId :
SSTD080614


Tgt Ion:165 Resp: 1067869
Ion Ratio Lower Upper
165 100
63 40.4 32.0 48.0
182 2.8 2.4 3.6



#58
2,3,4,6-Tetrachlorophenol
Concen: 77.243 ng/ul
RT: 15.386 min Scan# 2074
Delta R.T. 0.006 min
Lab File: BP013783.D
Acq: 07 Feb 2023 18:07

Tgt Ion:232 Resp: 990133
Ion Ratio Lower Upper
232 100
131 41.3 33.0 49.6
130 2.3 1.5 2.3
166 28.7 22.6 33.8









#59
Diethylphthalate
Concen: 81.324 ng/ul
RT: 15.580 min Scan# 2107
Delta R.T. 0.011 min
Lab File: BP013783.D
Acq: 07 Feb 2023 18:07




Instrument :
BNA_P
ClientSampleId :
SSTD080614


Tgt Ion:149 Resp: 3467593
Ion Ratio Lower Upper
149 100
177 23.0 17.8 26.8
150 12.5 10.1 15.1



#61
Fluorene
Concen: 73.001 ng/ul
RT: 15.816 min Scan# 2147
Delta R.T. 0.011 min
Lab File: BP013783.D
Acq: 07 Feb 2023 18:07

Tgt Ion:166 Resp: 3114735
Ion Ratio Lower Upper
166 100
165 97.6 79.1 118.7
167 13.4 10.8 16.2









#62
4-Chlorophenyl-phenylether
Concen: 71.492 ng/ul
RT: 15.798 min Scan# 2144
Delta R.T. -0.000 min
Lab File: BP013783.D
Acq: 07 Feb 2023 18:07




Instrument :
BNA_P
ClientSampleId :
SSTD080614


Tgt Ion:204 Resp: 1661875
Ion Ratio Lower Upper
204 100
206 33.2 26.7 40.1
141 57.2 44.6 66.8



#63
4-Nitroaniline
Concen: 102.469 ng/ul
RT: 15.845 min Scan# 2152
Delta R.T. 0.023 min
Lab File: BP013783.D
Acq: 07 Feb 2023 18:07

Tgt Ion:138 Resp: 641913
Ion Ratio Lower Upper
138 100
92 52.5 43.8 65.8
108 89.3 69.8 104.6









#66
4,6-Dinitro-2-methylphenol
Concen: 91.890 ng/ul
RT: 15.892 min Scan# 2157
Delta R.T. 0.017 min
Lab File: BP013783.D
Acq: 07 Feb 2023 18:07




Instrument :
BNA_P
ClientSampleId :
SSTD080614


Tgt Ion:198 Resp: 681548
Ion Ratio Lower Upper
198 100
182 3.7 4.1 6.1#
77 24.6 20.3 30.5



#67
N-Nitrosodiphenylamine
Concen: 77.105 ng/ul
RT: 16.016 min Scan# 2181
Delta R.T. 0.006 min
Lab File: BP013783.D
Acq: 07 Feb 2023 18:07

Tgt Ion:169 Resp: 2767770
Ion Ratio Lower Upper
169 100
168 67.4 53.5 80.3
167 36.5 29.8 44.6









#68
4-Bromophenyl-phenylether
Concen: 74.758 ng/ul
RT: 16.698 min Scan# 211
Delta R.T. 0.006 min
Lab File: BP013783.D
Acq: 07 Feb 2023 18:07




Instrument :
BNA_P
ClientSampleId :
SSTD080614


Tgt Ion:248 Resp: 1131132
Ion Ratio Lower Upper
248 100
250 97.6 76.0 114.0
141 68.9 56.0 84.0



#69
Hexachlorobenzene
Concen: 73.455 ng/ul
RT: 16.822 min Scan# 2318
Delta R.T. 0.006 min
Lab File: BP013783.D
Acq: 07 Feb 2023 18:07

Tgt Ion:284 Resp: 1303909
Ion Ratio Lower Upper
284 100
142 30.3 25.0 37.4
249 31.9 24.9 37.3





#70

Atrazine

Concen: 78.934 ng/ul

RT: 16.969 min Scan# 21

Delta R.T. 0.011 min

Lab File: BP013783.D

Acq: 07 Feb 2023 18:07

Instrument :

BNA_P

ClientSampleId :

SSTD080614





Tgt Ion:200 Resp: 1140358

Ion Ratio Lower Upper

200 100

173 25.0 20.4 30.6

215 48.1 39.3 58.9



#71

Pentachlorophenol

Concen: 77.850 ng/ul

RT: 17.163 min Scan# 2376

Delta R.T. 0.006 min

Lab File: BP013783.D

Acq: 07 Feb 2023 18:07




Tgt Ion:266 Resp: 882065


Ion Ratio Lower Upper

266 100

264 63.2 50.5 75.7

268 62.7 51.0 76.6





#72

Phenanthrene

Concen: 75.139 ng/ul

RT: 17.563 min Scan# 2443

Delta R.T. 0.006 min

Lab File: BP013783.D

Acq: 07 Feb 2023 18:07

Instrument :

BNA_P

ClientSampleId :

SSTD080614





Tgt Ion:178 Resp: 5278455

Ion Ratio Lower Upper

178 100

179 15.5 12.2 18.2

176 19.5 15.2 22.8



#74

Anthracene

Concen: 73.657 ng/ul

RT: 17.657 min Scan# 2460

Delta R.T. 0.006 min

Lab File: BP013783.D

Acq: 07 Feb 2023 18:07




Tgt Ion:178 Resp: 5240741


Ion Ratio Lower Upper

178 100

179 15.7 12.2 18.4

176 19.1 15.0 22.4





#75

1,2,3,4-Tetrachlorobenzene

Concen: 77.514 ng/uL

RT: 13.569 min Scan# 1

Delta R.T. 0.006 min

Lab File: BP013783.D

Acq: 07 Feb 2023 18:07

Instrument :

BNA_P

ClientSampleId :

SSTD080614





Tgt Ion:216 Resp: 1468617

Ion Ratio Lower Upper

216 100

214 78.6 62.6 94.0

218 48.8 39.0 58.4



#76

Pentachlorobenzene

Concen: 75.537 ng/uL

RT: 15.081 min Scan# 2022

Delta R.T. 0.006 min

Lab File: BP013783.D

Acq: 07 Feb 2023 18:07




Tgt Ion:250 Resp: 1467207


Ion Ratio Lower Upper

250 100

248 63.1 51.1 76.7

252 64.1 51.0 76.6





#77

Carbazole

Concen: 77.639 ng/ul

RT: 17.916 min Scan# 2503

Delta R.T. 0.006 min

Lab File: BP013783.D

Acq: 07 Feb 2023 18:07

Instrument :

BNA_P

ClientSampleId :

SSTD080614





Tgt Ion:167 Resp: 4792362

Ion Ratio Lower Upper

167 100

166 22.0 17.3 25.9

139 13.2 10.4 15.6



#78

Di-n-butylphthalate

Concen: 80.638 ng/ul

RT: 18.445 min Scan# 2594

Delta R.T. -0.000 min

Lab File: BP013783.D

Acq: 07 Feb 2023 18:07




Tgt Ion:149 Resp: 5881164


Ion Ratio Lower Upper

149 100

150 9.4 7.1 10.7

104 5.5 4.1 6.1









#80
Fluoranthene
Concen: 79.546 ng/ul
RT: 19.510 min Scan# 21
Delta R.T. 0.006 min
Lab File: BP013783.D
Acq: 07 Feb 2023 18:07




Instrument :
BNA_P
ClientSampleId :
SSTD080614


Tgt Ion:202 Resp: 6528001
Ion Ratio Lower Upper
202 100
101 9.6 8.6 12.8
100 7.3 6.3 9.5



#82
Pyrene
Concen: 77.998 ng/ul
RT: 19.868 min Scan# 2836
Delta R.T. 0.011 min
Lab File: BP013783.D
Acq: 07 Feb 2023 18:07

Tgt Ion:202 Resp: 6544362
Ion Ratio Lower Upper
202 100
101 10.6 10.3 15.5
100 8.6 8.5 12.7





#83

Butylbenzylphthalate

Concen: 109.625 ng/ul

RT: 20.715 min Scan# 2979

Delta R.T. 0.006 min

Lab File: BP013783.D

Acq: 07 Feb 2023 18:07

Instrument :

BNA_P

ClientSampleId :

SSTD080614





Tgt Ion:149 Resp: 2642634

Ion Ratio Lower Upper

149 100

91 70.7 57.7 86.5

206 23.0 16.7 25.1



#84

3,3'-Dichlorobenzidine

Concen: 81.856 ng/ul

RT: 21.504 min Scan# 3114

Delta R.T. 0.006 min

Lab File: BP013783.D

Acq: 07 Feb 2023 18:07




Tgt Ion:252 Resp: 2347010


Ion Ratio Lower Upper

252 100

254 64.5 51.8 77.6

126 9.2 7.9 11.9





#85

Benzo(a)anthracene

Concen: 77.357 ng/ul

RT: 21.580 min Scan# 3126

Delta R.T. 0.006 min

Lab File: BP013783.D

Acq: 07 Feb 2023 18:07

Instrument :

BNA_P

ClientSampleId :

SSTD080614





Tgt Ion:228 Resp: 6713611

Ion Ratio Lower Upper

228 100

229 20.1 15.4 23.2

226 27.7 21.8 32.6



#86

Bis(2-ethylhexyl)phthalate

Concen: 97.385 ng/ul

RT: 21.474 min Scan# 3109

Delta R.T. 0.006 min

Lab File: BP013783.D

Acq: 07 Feb 2023 18:07




Tgt Ion:149 Resp: 4024904


Ion Ratio Lower Upper

149 100

167 28.0 21.5 32.3

279 5.0 3.2 4.8#





#87

Chrysene

Concen: 76.468 ng/ul

RT: 21.639 min Scan# 3135

Delta R.T. 0.011 min

Lab File: BP013783.D

Acq: 07 Feb 2023 18:07

Instrument :

BNA_P

ClientSampleId :

SSTD080614





Tgt Ion:228 Resp: 6248100

Ion Ratio Lower Upper

228 100

226 31.0 25.0 37.6

229 20.1 16.4 24.6



#89

Di-n-octyl phthalate

Concen: 113.439 ng/ul




RT: 22.457 min Scan# 3276


Delta R.T. 0.006 min

Lab File: BP013783.D

Acq: 07 Feb 2023 18:07

Tgt Ion:149 Resp: 6967030









#90
Benzo(b)fluoranthene
Concen: 81.557 ng/ul
RT: 23.374 min Scan# 3430
Delta R.T. 0.011 min
Lab File: BP013783.D
Acq: 07 Feb 2023 18:07




Instrument :
BNA_P
ClientSampleId :
SSTD080614


Tgt Ion:252 Resp: 6207273
Ion Ratio Lower Upper
252 100
253 22.2 17.5 26.3
125 8.4 7.8 11.6



#91
Benzo(k)fluoranthene
Concen: 85.388 ng/ul
RT: 23.427 min Scan# 3441
Delta R.T. 0.011 min
Lab File: BP013783.D
Acq: 07 Feb 2023 18:07

Tgt Ion:252 Resp: 6211237
Ion Ratio Lower Upper
252 100
253 22.4 17.1 25.7
125 8.3 7.2 10.8





#93

Benzo(a)pyrene

Concen: 79.620 ng/ul

RT: 24.057 min Scan# 3

Delta R.T. 0.017 min

Lab File: BP013783.D

Acq: 07 Feb 2023 18:07

Instrument :

BNA_P

ClientSampleId :

SSTD080614





Tgt Ion:252 Resp: 5231621

Ion Ratio Lower Upper

252 100

253 22.1 17.4 26.0

125 9.3 8.2 12.2



#94

Indeno(1,2,3-cd)pyrene

Concen: 72.204 ng/ul

RT: 26.880 min Scan# 4028

Delta R.T. 0.023 min

Lab File: BP013783.D

Acq: 07 Feb 2023 18:07




Tgt Ion:276 Resp: 5878063


Ion Ratio Lower Upper

276 100

138 21.8 18.4 27.6

277 25.3 19.7 29.5









#95
Dibenzo(a,h)anthracene
Concen: 72.023 ng/ul
RT: 26.892 min Scan# 4170
Delta R.T. 0.017 min
Lab File: BP013783.D
Acq: 07 Feb 2023 18:07

Instrument :
BNA_P
ClientSampleId :
SSTD080614

Tgt Ion:278 Resp: 4860782
Ion Ratio Lower Upper
278 100
139 13.3 12.0 18.0
279 23.6 18.9 28.3



#96
Benzo(g,h,i)perylene
Concen: 73.702 ng/ul
RT: 27.715 min Scan# 4170
Delta R.T. 0.023 min
Lab File: BP013783.D
Acq: 07 Feb 2023 18:07

Tgt Ion:276 Resp: 4746651
Ion Ratio Lower Upper
276 100
138 19.5 15.9 23.9
277 23.6 19.0 28.6

