

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP020723\
 Data File : BP013785.D
 Acq On : 07 Feb 2023 19:15
 Operator : CG/JU
 Sample : SSTDICV020
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SICV616

Quant Time: Feb 08 00:40:05 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP020723.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed Feb 08 00:37:14 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.122	152	171257	20.000	ng/u1	0.00
20) Naphthalene-d8	10.952	136	795034	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.763	164	566725	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.516	188	1361025	20.000	ng/u1	0.00
79) Chrysene-d12	21.592	240	1422859	20.000	ng/u1	0.00
88) Perylene-d12	24.157	264	1540619	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.470	96	28003	6.777	ng/uL	0.00
4) Pyridine-d5	3.905	84	201891	16.481	ng/u1	0.00
7) Phenol-d5	7.269	99	255941	16.773	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.440	67	171579	19.189	ng/u1	0.00
11) 2-Chlorophenol-d4	7.652	132	198902	17.708	ng/u1	0.00
15) 4-Methylphenol-d8	8.828	113	218014	17.481	ng/u1	0.00
21) Nitrobenzene-d5	9.299	128	99766	18.036	ng/u1	0.00
24) 2-Nitrophenol-d4	10.022	143	115013	18.008	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.563	165	229033	17.597	ng/u1	0.00
31) 4-Chloroaniline-d4	11.087	131	307458	16.665	ng/u1	0.00
46) Dimethylphthalate-d6	14.163	166	780012	17.475	ng/u1	0.00
49) Acenaphthylene-d8	14.457	160	861962	17.839	ng/u1	0.00
54) 4-Nitrophenol-d4	14.945	143	138379	17.401	ng/u1	0.00
60) Fluorene-d10	15.751	176	656738	17.273	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.863	200	142330	17.046	ng/u1	0.00
73) Anthracene-d10	17.610	188	1072486	17.233	ng/u1	0.00
81) Pyrene-d10	19.833	212	1345571	16.676	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.986	264	1343358	17.402	ng/u1	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.505	88	29314	6.609	ng/uL	97
5) Pyridine	3.922	79	195422	16.143	ng/u1	100
6) Benzaldehyde	7.252	77	190519	27.687	ng/u1	100
8) Phenol	7.299	94	277681	17.728	ng/u1	98
10) Bis(2-Chloroethyl)ether	7.534	93	220618	17.577	ng/u1	97
12) 2-Chlorophenol	7.681	128	209508	18.178	ng/u1	99
13) 2-Methylphenol	8.563	108	212719	17.792	ng/u1	99
14) 2,2'-oxybis(1-Chloropr...	8.658	45	257689	18.049	ng/u1	99
16) Acetophenone	8.952	105	375624	18.780	ng/u1	98
17) N-Nitroso-di-n-propyla...	8.934	70	186884	18.815	ng/u1	99
18) 4-Methylphenol	8.893	108	237871	18.021	ng/u1	99
19) Hexachloroethane	9.216	117	85242	18.062	ng/u1	99
22) Nitrobenzene	9.340	77	263640	18.116	ng/u1	99
23) Isophorone	9.869	82	536724	17.481	ng/u1	100
25) 2-Nitrophenol	10.057	139	128247	18.526	ng/u1	99
26) 2,4-Dimethylphenol	10.105	107	273782	17.763	ng/u1	97
27) Bis(2-Chloroethoxy)met...	10.346	93	324246	17.487	ng/u1	97
29) 2,4-Dichlorophenol	10.593	162	231045	18.031	ng/u1	98
30) Naphthalene	11.004	128	734217	17.348	ng/u1	100
32) 4-Chloroaniline	11.110	127	327495	17.608	ng/u1	100
33) Hexachlorobutadiene	11.287	225	159523	17.673	ng/u1	98
34) Caprolactam	11.887	113	82330	18.311	ng/u1	92
35) 4-Chloro-3-methylphenol	12.216	107	265726	17.963	ng/u1	99
36) 2-Methylnaphthalene	12.599	142	544146	18.287	ng/u1	99

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP020723\
 Data File : BP013785.D
 Acq On : 07 Feb 2023 19:15
 Operator : CG/JU
 Sample : SSTDICV020
 Misc :
 ALS Vial : 8 Sample Multiplier: 1


Instrument :
 BNA_P
 ClientSampleId :
 SICV616


Quant Time: Feb 08 00:40:05 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP020723.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed Feb 08 00:37:14 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.816	142	506551	16.476	ng/ul	99
39) 1,2,4,5-Tetrachloroben...	12.963	216	327280	18.326	ng/ul	99
40) Hexachlorocyclopentadiene	12.940	237	192359	16.775	ng/ul	98
41) 2,4,6-Trichlorophenol	13.193	196	212451	18.128	ng/ul	99
42) 2,4,5-Trichlorophenol	13.263	196	234151	18.458	ng/ul	98
43) 1,1'-Biphenyl	13.598	154	756738	17.796	ng/ul	100
44) 2-Chloronaphthalene	13.640	162	581699	17.388	ng/ul	99
45) 2-Nitroaniline	13.840	65	155591	19.310	ng/ul	94
47) Dimethylphthalate	14.210	163	779943	17.617	ng/ul	99
48) 2,6-Dinitrotoluene	14.328	165	161891	20.930	ng/ul	99
50) Acenaphthylene	14.487	152	938123	17.998	ng/ul	99
51) 3-Nitroaniline	14.663	138	166724	21.802	ng/ul	96
52) Acenaphthene	14.822	153	634007	17.378	ng/ul	99
53) 2,4-Dinitrophenol	14.863	184	95006	17.224	ng/ul	94
55) 4-Nitrophenol	14.957	109	125180	17.940	ng/ul	99
56) Dibenzofuran	15.157	168	920206	17.593	ng/ul	97
57) 2,4-Dinitrotoluene	15.116	165	222999	19.338	ng/ul	98
58) 2,3,4,6-Tetrachlorophenol	15.298	232	226789	18.968	ng/ul	91
59) Diethylphthalate	15.569	149	779085	17.525	ng/ul	99
61) Fluorene	15.810	166	753354	17.655	ng/ul	98
62) 4-Chlorophenyl-phenyle...	15.792	204	391877	17.662	ng/ul	98
63) 4-Nitroaniline	15.822	138	175228	24.197	ng/ul	97
66) 4,6-Dinitro-2-methylph...	15.881	198	147369	18.003	ng/ul#	94
67) N-Nitrosodiphenylamine	16.010	169	663666	17.940	ng/ul	99
68) 4-Bromophenyl-phenylether	16.692	248	253063	17.607	ng/ul	96
69) Hexachlorobenzene	16.816	284	297203	17.746	ng/ul	98
70) Atrazine	16.957	200	261917	17.704	ng/ul	100
71) Pentachlorophenol	17.157	266	188404	17.287	ng/ul	99
72) Phenanthrene	17.557	178	1247518	17.322	ng/ul	100
74) Anthracene	17.645	178	1275238	17.599	ng/ul	100
75) 1,2,3,4-Tetrachloroben...	13.563	216	326141	17.951	ng/uL	99
76) Pentachlorobenzene	15.075	250	329687	17.584	ng/uL	98
77) Carbazole	17.910	167	1176759	18.436	ng/ul	100
78) Di-n-butylphthalate	18.445	149	1366892	17.657	ng/ul	99
80) Fluoranthene	19.504	202	1560694	16.896	ng/ul	99
82) Pyrene	19.857	202	1642041	16.673	ng/ul	99
83) Butylbenzylphthalate	20.710	149	589516	17.490	ng/ul	99
84) 3,3'-Dichlorobenzidine	21.498	252	657437	20.530	ng/ul	99
85) Benzo(a)anthracene	21.575	228	1710444	17.593	ng/ul	99
86) Bis(2-ethylhexyl)phtha...	21.469	149	923461	17.892	ng/ul	99
87) Chrysene	21.633	228	1615366	17.611	ng/ul	98
89) Di-n-octyl phthalate	22.451	149	1600224	15.095	ng/ul	100
90) Benzo(b)fluoranthene	23.363	252	1773126	17.260	ng/ul	99
91) Benzo(k)fluoranthene	23.416	252	1690302	16.585	ng/ul	99
93) Benzo(a)pyrene	24.039	252	1510940	17.353	ng/ul	99
94) Indeno(1,2,3-cd)pyrene	26.857	276	1742777	17.845	ng/ul	97
95) Dibenzo(a,h)anthracene	26.874	278	1453258	17.814	ng/ul	99
96) Benzo(g,h,i)perylene	27.686	276	1368832	17.604	ng/ul	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Quant Time: Feb 08 00:40:05 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP020723.M
Quant Title : SVOA CALIBRATION
QLast Update : Wed Feb 08 00:37:14 2023
Response via : Initial Calibration









#2
 1,4-Dioxane
 Concen: 6.609 ng/uL
 RT: 3.505 min Scan# 54
 Delta R.T. -0.000 min
 Lab File: BP013785.D
 Acq: 07 Feb 2023 19:15




Instrument :
 BNA_P
 ClientSampleId :
 SICV616


Tgt Ion: 88	Resp:	29314
Ion	Ratio	Lower Upper
88	100	
43	25.2	20.1 30.1
58	63.5	53.0 79.6



#5
 Pyridine
 Concen: 16.143 ng/uL
 RT: 3.922 min Scan# 125
 Delta R.T. -0.000 min
 Lab File: BP013785.D
 Acq: 07 Feb 2023 19:15

Tgt Ion: 79	Resp:	195422
Ion	Ratio	Lower Upper
79	100	
52	68.8	55.1 82.7
51	32.3	25.7 38.5









#6
Benzaldehyde
Concen: 27.687 ng/ul
RT: 7.252 min Scan# 691
Delta R.T. -0.000 min
Lab File: BP013785.D
Acq: 07 Feb 2023 19:15




Instrument :
BNA_P
ClientSampleId :
SICV616


Tgt Ion: 77 Resp: 190519
Ion Ratio Lower Upper
77 100
105 99.0 79.5 119.3
106 95.3 76.5 114.7



#8
Phenol
Concen: 17.728 ng/ul
RT: 7.299 min Scan# 699
Delta R.T. -0.000 min
Lab File: BP013785.D
Acq: 07 Feb 2023 19:15

Tgt Ion: 94 Resp: 277681
Ion Ratio Lower Upper
94 100
65 27.5 22.2 33.4
66 39.0 29.6 44.4









#10
 Bis(2-Chloroethyl)ether
 Concen: 17.577 ng/ul
 RT: 7.534 min Scan# 739
 Delta R.T. -0.000 min
 Lab File: BP013785.D
 Acq: 07 Feb 2023 19:15




Instrument :
 BNA_P
 ClientSampleId :
 SICV616


Tgt Ion: 93	Resp: 220618
Ion Ratio	Lower Upper
93 100	
63 70.8	59.4 89.0
95 32.6	25.8 38.8



#12
 2-Chlorophenol
 Concen: 18.178 ng/ul
 RT: 7.681 min Scan# 764
 Delta R.T. -0.000 min
 Lab File: BP013785.D
 Acq: 07 Feb 2023 19:15

Tgt Ion:128	Resp: 209508
Ion Ratio	Lower Upper
128 100	
64 47.0	37.3 55.9
130 32.0	26.3 39.5









#13
2-Methylphenol
Concen: 17.792 ng/ul
RT: 8.563 min Scan# 914
Delta R.T. -0.000 min
Lab File: BP013785.D
Acq: 07 Feb 2023 19:15




Instrument :
BNA_P
ClientSampleId :
SICV616


Tgt Ion:108 Resp: 212719
Ion Ratio Lower Upper
108 100
107 90.8 71.9 107.9



#14
2,2'-oxybis(1-Chloropropane)
Concen: 18.049 ng/ul
RT: 8.658 min Scan# 930
Delta R.T. 0.006 min
Lab File: BP013785.D
Acq: 07 Feb 2023 19:15

Tgt Ion: 45 Resp: 257689
Ion Ratio Lower Upper
45 100
77 16.9 13.4 20.2
79 13.4 11.0 16.4









#16
Acetophenone
Concen: 18.780 ng/ul
RT: 8.952 min Scan# 91
Delta R.T. -0.000 min
Lab File: BP013785.D
Acq: 07 Feb 2023 19:15




Instrument :
BNA_P
ClientSampleId :
SICV616


Tgt Ion:105 Resp: 375624
Ion Ratio Lower Upper
105 100
77 78.0 61.0 91.4
51 24.4 19.6 29.4



#17
N-Nitroso-di-n-propylamine
Concen: 18.815 ng/ul
RT: 8.934 min Scan# 977
Delta R.T. -0.000 min
Lab File: BP013785.D
Acq: 07 Feb 2023 19:15

Tgt Ion: 70 Resp: 186884
Ion Ratio Lower Upper
70 100
42 49.4 39.8 59.8
101 10.1 7.4 11.2
130 18.8 14.6 22.0









#18
4-Methylphenol
Concen: 18.021 ng/ul
RT: 8.893 min Scan# 91
Delta R.T. -0.000 min
Lab File: BP013785.D
Acq: 07 Feb 2023 19:15




Instrument :
BNA_P
ClientSampleId :
SICV616


Tgt Ion:108 Resp: 237871
Ion Ratio Lower Upper
108 100
107 114.9 90.9 136.3



#19
Hexachloroethane
Concen: 18.062 ng/ul
RT: 9.216 min Scan# 1025
Delta R.T. -0.000 min
Lab File: BP013785.D
Acq: 07 Feb 2023 19:15

Tgt Ion:117 Resp: 85242
Ion Ratio Lower Upper
117 100
201 109.0 86.6 129.8
199 68.3 55.2 82.8









#22
Nitrobenzene
Concen: 18.116 ng/ul
RT: 9.340 min Scan# 1046
Delta R.T. -0.000 min
Lab File: BP013785.D
Acq: 07 Feb 2023 19:15




Instrument :
BNA_P
ClientSampleId :
SICV616


Tgt Ion: 77 Resp: 263640
Ion Ratio Lower Upper
77 100
123 43.4 35.1 52.7
65 13.2 10.6 16.0



#23
Isophorone
Concen: 17.481 ng/ul
RT: 9.869 min Scan# 1136
Delta R.T. -0.000 min
Lab File: BP013785.D
Acq: 07 Feb 2023 19:15

Tgt Ion: 82 Resp: 536724
Ion Ratio Lower Upper
82 100
95 7.9 6.0 9.0
138 17.1 13.7 20.5









#25
2-Nitrophenol
Concen: 18.526 ng/ul
RT: 10.057 min Scan# 1168
Delta R.T. -0.000 min
Lab File: BP013785.D
Acq: 07 Feb 2023 19:15




Instrument :
BNA_P
ClientSampleId :
SICV616


Tgt Ion:139 Resp: 128247
Ion Ratio Lower Upper
139 100
65 50.4 39.9 59.9
109 33.9 27.4 41.0



#26
2,4-Dimethylphenol
Concen: 17.763 ng/ul
RT: 10.105 min Scan# 1176
Delta R.T. -0.000 min
Lab File: BP013785.D
Acq: 07 Feb 2023 19:15

Tgt Ion:107 Resp: 273782
Ion Ratio Lower Upper
107 100
121 48.9 37.8 56.6
122 82.4 63.7 95.5









#27
Bis(2-Chloroethoxy)methane
Concen: 17.487 ng/ul
RT: 10.346 min Scan# 1217
Delta R.T. -0.000 min
Lab File: BP013785.D
Acq: 07 Feb 2023 19:15




Instrument :
BNA_P
ClientSampleId :
SICV616


Tgt Ion: 93 Resp: 324246
Ion Ratio Lower Upper
93 100
95 33.0 25.0 37.6
123 11.2 8.2 12.4



#29
2,4-Dichlorophenol
Concen: 18.031 ng/ul
RT: 10.593 min Scan# 1259
Delta R.T. -0.000 min
Lab File: BP013785.D
Acq: 07 Feb 2023 19:15

Tgt Ion: 162 Resp: 231045
Ion Ratio Lower Upper
162 100
164 66.0 51.4 77.2
98 36.1 27.5 41.3









#30
Naphthalene
Concen: 17.348 ng/ul
RT: 11.004 min Scan# 1329
Delta R.T. -0.000 min
Lab File: BP013785.D
Acq: 07 Feb 2023 19:15




Instrument :
BNA_P
ClientSampleId :
SICV616


Tgt Ion:128 Resp: 734217
Ion Ratio Lower Upper
128 100
129 10.6 8.6 12.8
127 13.3 10.6 15.8



#32
4-Chloroaniline
Concen: 17.608 ng/ul
RT: 11.110 min Scan# 1347
Delta R.T. -0.000 min
Lab File: BP013785.D
Acq: 07 Feb 2023 19:15

Tgt Ion:127 Resp: 327495
Ion Ratio Lower Upper
127 100
129 32.7 26.2 39.4









#33
Hexachlorobutadiene
Concen: 17.673 ng/ul
RT: 11.287 min Scan# 11
Delta R.T. -0.000 min
Lab File: BP013785.D
Acq: 07 Feb 2023 19:15




Instrument :
BNA_P
ClientSampleId :
SICV616


Tgt Ion:225 Resp: 159523
Ion Ratio Lower Upper
225 100
223 62.2 50.6 76.0
227 62.6 49.0 73.4



#34
Caprolactam
Concen: 18.311 ng/ul
RT: 11.887 min Scan# 1479
Delta R.T. -0.000 min
Lab File: BP013785.D
Acq: 07 Feb 2023 19:15

Tgt Ion:113 Resp: 82330
Ion Ratio Lower Upper
113 100
55 143.3 120.9 181.3
56 106.9 93.8 140.8









#35
4-Chloro-3-methylphenol
Concen: 17.963 ng/ul
RT: 12.216 min Scan# 1535
Delta R.T. -0.000 min
Lab File: BP013785.D
Acq: 07 Feb 2023 19:15




Instrument :
BNA_P
ClientSampleId :
SICV616


Tgt Ion:107 Resp: 265726
Ion Ratio Lower Upper
107 100
144 25.7 20.0 30.0
142 77.8 62.1 93.1



#36
2-Methylnaphthalene
Concen: 18.287 ng/ul
RT: 12.599 min Scan# 1600
Delta R.T. -0.000 min
Lab File: BP013785.D
Acq: 07 Feb 2023 19:15

Tgt Ion:142 Resp: 544146
Ion Ratio Lower Upper
142 100
141 89.9 72.8 109.2









#37
1-Methylnaphthalene
Concen: 16.476 ng/ul
RT: 12.816 min Scan# 1637
Delta R.T. -0.000 min
Lab File: BP013785.D
Acq: 07 Feb 2023 19:15




Instrument :
BNA_P
ClientSampleId :
SICV616


Tgt Ion:142 Resp: 506551
Ion Ratio Lower Upper
142 100
141 93.3 74.2 111.2
116 4.1 3.1 4.7



#39
1,2,4,5-Tetrachlorobenzene
Concen: 18.326 ng/ul
RT: 12.963 min Scan# 1662
Delta R.T. -0.000 min
Lab File: BP013785.D
Acq: 07 Feb 2023 19:15

Tgt Ion:216 Resp: 327280
Ion Ratio Lower Upper
216 100
214 77.7 62.6 93.8
179 19.8 15.4 23.2
108 15.9 12.7 19.1









#40
Hexachlorocyclopentadiene
Concen: 16.775 ng/ul
RT: 12.940 min Scan# 1658
Delta R.T. -0.000 min
Lab File: BP013785.D
Acq: 07 Feb 2023 19:15




Instrument :
BNA_P
ClientSampleId :
SICV616


Tgt Ion:237 Resp: 192359
Ion Ratio Lower Upper
237 100
235 63.7 49.8 74.8
272 13.8 10.1 15.1



#41
2,4,6-Trichlorophenol
Concen: 18.128 ng/ul
RT: 13.193 min Scan# 1701
Delta R.T. -0.000 min
Lab File: BP013785.D
Acq: 07 Feb 2023 19:15

Tgt Ion:196 Resp: 212451
Ion Ratio Lower Upper
196 100
198 95.3 75.1 112.7
200 31.1 24.3 36.5









#42
2,4,5-Trichlorophenol
Concen: 18.458 ng/ul
RT: 13.263 min Scan# 1713
Delta R.T. -0.006 min
Lab File: BP013785.D
Acq: 07 Feb 2023 19:15




Instrument :
BNA_P
ClientSampleId :
SICV616


Tgt Ion:196 Resp: 234151
Ion Ratio Lower Upper
196 100
198 95.2 77.0 115.4
200 30.0 25.8 38.6



#43
1,1'-Biphenyl
Concen: 17.796 ng/ul
RT: 13.598 min Scan# 1770
Delta R.T. -0.000 min
Lab File: BP013785.D
Acq: 07 Feb 2023 19:15

Tgt Ion:154 Resp: 756738
Ion Ratio Lower Upper
154 100
153 39.8 31.9 47.9
76 12.0 9.9 14.9





#44

2-Chloronaphthalene

Concen: 17.388 ng/ul

RT: 13.640 min Scan# 1777

Delta R.T. -0.000 min

Lab File: BP013785.D

Acq: 07 Feb 2023 19:15

Instrument :

BNA_P

ClientSampleId :

SICV616





Tgt Ion:162 Resp: 581699

Ion Ratio Lower Upper

162 100

164 33.2 26.1 39.1

127 38.8 31.1 46.7



#45

2-Nitroaniline

Concen: 19.310 ng/ul

RT: 13.840 min Scan# 1811

Delta R.T. -0.000 min

Lab File: BP013785.D

Acq: 07 Feb 2023 19:15




Tgt Ion: 65 Resp: 155591


Ion Ratio Lower Upper

65 100

92 72.8 58.6 88.0

138 110.7 81.2 121.8









#47
Dimethylphthalate
Concen: 17.617 ng/ul
RT: 14.210 min Scan# 1874
Delta R.T. -0.000 min
Lab File: BP013785.D
Acq: 07 Feb 2023 19:15




Instrument :
BNA_P
ClientSampleId :
SICV616


Tgt Ion:163 Resp: 779943
Ion Ratio Lower Upper
163 100
194 4.4 3.4 5.2
164 10.2 7.8 11.8



#48
2,6-Dinitrotoluene
Concen: 20.930 ng/ul
RT: 14.328 min Scan# 1894
Delta R.T. -0.006 min
Lab File: BP013785.D
Acq: 07 Feb 2023 19:15

Tgt Ion:165 Resp: 161891
Ion Ratio Lower Upper
165 100
89 58.7 47.5 71.3
121 20.8 16.6 24.8









#50
Acenaphthylene
Concen: 17.998 ng/ul
RT: 14.487 min Scan# 1921
Delta R.T. -0.000 min
Lab File: BP013785.D
Acq: 07 Feb 2023 19:15




Instrument :
BNA_P
ClientSampleId :
SICV616


Tgt Ion:152 Resp: 938123
Ion Ratio Lower Upper
152 100
151 20.1 15.8 23.6
153 13.0 10.4 15.6



#51
3-Nitroaniline
Concen: 21.802 ng/ul
RT: 14.663 min Scan# 1951
Delta R.T. -0.000 min
Lab File: BP013785.D
Acq: 07 Feb 2023 19:15

Tgt Ion:138 Resp: 166724
Ion Ratio Lower Upper
138 100
108 11.1 9.4 14.2
92 116.5 97.4 146.0





#52

Acenaphthene

Concen: 17.378 ng/ul

RT: 14.822 min Scan# 1978

Delta R.T. -0.000 min

Lab File: BP013785.D

Acq: 07 Feb 2023 19:15

Instrument :

BNA_P

ClientSampleId :

SICV616





Tgt Ion:153 Resp: 634007

Ion Ratio Lower Upper

153 100

152 47.7 37.4 56.0

154 87.6 70.1 105.1



#53

2,4-Dinitrophenol

Concen: 17.224 ng/ul

RT: 14.863 min Scan# 1985

Delta R.T. -0.000 min

Lab File: BP013785.D

Acq: 07 Feb 2023 19:15




Tgt Ion:184 Resp: 95006


Ion Ratio Lower Upper

184 100

63 57.9 49.5 74.3

154 62.3 54.3 81.5









#55
4-Nitrophenol
Concen: 17.940 ng/ul
RT: 14.957 min Scan# 2001
Delta R.T. -0.000 min
Lab File: BP013785.D
Acq: 07 Feb 2023 19:15




Instrument :
BNA_P
ClientSampleId :
SICV616


Tgt Ion:109 Resp: 125180
Ion Ratio Lower Upper
109 100
139 120.7 96.0 144.0
65 117.9 95.8 143.6



#56
Dibenzofuran
Concen: 17.593 ng/ul
RT: 15.157 min Scan# 2035
Delta R.T. -0.000 min
Lab File: BP013785.D
Acq: 07 Feb 2023 19:15

Tgt Ion:168 Resp: 920206
Ion Ratio Lower Upper
168 100
139 38.7 29.6 44.4









#57
2,4-Dinitrotoluene
Concen: 19.338 ng/ul
RT: 15.116 min Scan# 2059
Delta R.T. -0.000 min
Lab File: BP013785.D
Acq: 07 Feb 2023 19:15




Instrument :
BNA_P
ClientSampleId :
SICV616


Tgt Ion:165 Resp: 222999
Ion Ratio Lower Upper
165 100
63 41.3 32.0 48.0
182 2.8 2.4 3.6



#58
2,3,4,6-Tetrachlorophenol
Concen: 18.968 ng/ul
RT: 15.298 min Scan# 2059
Delta R.T. -0.082 min
Lab File: BP013785.D
Acq: 07 Feb 2023 19:15

Tgt Ion:232 Resp: 226789
Ion Ratio Lower Upper
232 100
131 35.6 33.0 49.6
130 2.0 1.5 2.3
166 23.1 22.6 33.8









#59
Diethylphthalate
Concen: 17.525 ng/ul
RT: 15.569 min Scan# 2105
Delta R.T. -0.000 min
Lab File: BP013785.D
Acq: 07 Feb 2023 19:15




Instrument :
BNA_P
ClientSampleId :
SICV616


Tgt Ion:149 Resp: 779085
Ion Ratio Lower Upper
149 100
177 22.7 17.8 26.8
150 12.5 10.1 15.1



#61
Fluorene
Concen: 17.655 ng/ul
RT: 15.810 min Scan# 2146
Delta R.T. 0.006 min
Lab File: BP013785.D
Acq: 07 Feb 2023 19:15

Tgt Ion:166 Resp: 753354
Ion Ratio Lower Upper
166 100
165 96.4 79.1 118.7
167 13.0 10.8 16.2









#62
4-Chlorophenyl-phenylether
Concen: 17.662 ng/ul
RT: 15.792 min Scan# 2143
Delta R.T. -0.006 min
Lab File: BP013785.D
Acq: 07 Feb 2023 19:15




Instrument :
BNA_P
ClientSampleId :
SICV616


Tgt Ion:204 Resp: 391877
Ion Ratio Lower Upper
204 100
206 33.7 26.7 40.1
141 57.8 44.6 66.8



#63
4-Nitroaniline
Concen: 24.197 ng/ul
RT: 15.822 min Scan# 2148
Delta R.T. -0.000 min
Lab File: BP013785.D
Acq: 07 Feb 2023 19:15

Tgt Ion:138 Resp: 175228
Ion Ratio Lower Upper
138 100
92 56.0 43.8 65.8
108 91.3 69.8 104.6









#66
4,6-Dinitro-2-methylphenol
Concen: 18.003 ng/ul
RT: 15.881 min Scan# 2158
Delta R.T. 0.006 min
Lab File: BP013785.D
Acq: 07 Feb 2023 19:15




Instrument :
BNA_P
ClientSampleId :
SICV616


Tgt Ion:198 Resp: 147369
Ion Ratio Lower Upper
198 100
182 3.8 4.1 6.1#
77 22.3 20.3 30.5



#67
N-Nitrosodiphenylamine
Concen: 17.940 ng/ul
RT: 16.010 min Scan# 2180
Delta R.T. -0.000 min
Lab File: BP013785.D
Acq: 07 Feb 2023 19:15

Tgt Ion:169 Resp: 663666
Ion Ratio Lower Upper
169 100
168 67.4 53.5 80.3
167 36.4 29.8 44.6









#68
4-Bromophenyl-phenylether
Concen: 17.607 ng/ul
RT: 16.692 min Scan# 2296
Delta R.T. -0.000 min
Lab File: BP013785.D
Acq: 07 Feb 2023 19:15




Instrument :
BNA_P
ClientSampleId :
SICV616


Tgt Ion:248 Resp: 253063
Ion Ratio Lower Upper
248 100
250 100.0 76.0 114.0
141 71.3 56.0 84.0



#69
Hexachlorobenzene
Concen: 17.746 ng/ul
RT: 16.816 min Scan# 2317
Delta R.T. -0.000 min
Lab File: BP013785.D
Acq: 07 Feb 2023 19:15

Tgt Ion:284 Resp: 297203
Ion Ratio Lower Upper
284 100
142 29.4 25.0 37.4
249 31.3 24.9 37.3





#70

Atrazine

Concen: 17.704 ng/ul

RT: 16.957 min Scan# 21

Delta R.T. -0.000 min

Lab File: BP013785.D

Acq: 07 Feb 2023 19:15

Instrument :

BNA_P

ClientSampleId :

SICV616





Tgt Ion:200 Resp: 261917

Ion Ratio Lower Upper

200 100

173 25.6 20.4 30.6

215 49.2 39.3 58.9



#71

Pentachlorophenol

Concen: 17.287 ng/ul

RT: 17.157 min Scan# 2375

Delta R.T. -0.000 min

Lab File: BP013785.D

Acq: 07 Feb 2023 19:15




Tgt Ion:266 Resp: 188404


Ion Ratio Lower Upper

266 100

264 64.4 50.5 75.7

268 64.3 51.0 76.6









#72
Phenanthrene
Concen: 17.322 ng/ul
RT: 17.557 min Scan# 2443
Delta R.T. -0.000 min
Lab File: BP013785.D
Acq: 07 Feb 2023 19:15




Instrument :
BNA_P
ClientSampleId :
SICV616


Tgt Ion:178 Resp: 1247518
Ion Ratio Lower Upper
178 100
179 15.1 12.2 18.2
176 19.0 15.2 22.8



#74
Anthracene
Concen: 17.599 ng/ul
RT: 17.645 min Scan# 2458
Delta R.T. -0.006 min
Lab File: BP013785.D
Acq: 07 Feb 2023 19:15

Tgt Ion:178 Resp: 1275238
Ion Ratio Lower Upper
178 100
179 15.5 12.2 18.4
176 18.8 15.0 22.4









#75
1,2,3,4-Tetrachlorobenzene
Concen: 17.951 ng/uL
RT: 13.563 min Scan# 1764
Delta R.T. -0.000 min
Lab File: BP013785.D
Acq: 07 Feb 2023 19:15




Instrument :
BNA_P
ClientSampleId :
SICV616


Tgt Ion:216 Resp: 326141
Ion Ratio Lower Upper
216 100
214 77.4 62.6 94.0
218 47.8 39.0 58.4



#76
Pentachlorobenzene
Concen: 17.584 ng/uL
RT: 15.075 min Scan# 2021
Delta R.T. -0.000 min
Lab File: BP013785.D
Acq: 07 Feb 2023 19:15

Tgt Ion:250 Resp: 329687
Ion Ratio Lower Upper
250 100
248 61.9 51.1 76.7
252 63.3 51.0 76.6









#77
 Carbazole
 Concen: 18.436 ng/ul
 RT: 17.910 min Scan# 2503
 Delta R.T. -0.000 min
 Lab File: BP013785.D
 Acq: 07 Feb 2023 19:15




Instrument :
 BNA_P
 ClientSampleId :
 SICV616


Tgt Ion:167 Resp: 1176759
 Ion Ratio Lower Upper
 167 100
 166 21.7 17.3 25.9
 139 12.7 10.4 15.6



#78
 Di-n-butylphthalate
 Concen: 17.657 ng/ul
 RT: 18.445 min Scan# 2594
 Delta R.T. -0.000 min
 Lab File: BP013785.D
 Acq: 07 Feb 2023 19:15

Tgt Ion:149 Resp: 1366892
 Ion Ratio Lower Upper
 149 100
 150 9.2 7.1 10.7
 104 5.0 4.1 6.1









#80
Fluoranthene
Concen: 16.896 ng/ul
RT: 19.504 min Scan# 21
Delta R.T. -0.000 min
Lab File: BP013785.D
Acq: 07 Feb 2023 19:15




Instrument :
BNA_P
ClientSampleId :
SICV616


Tgt Ion:202 Resp: 1560694
Ion Ratio Lower Upper
202 100
101 10.4 8.6 12.8
100 8.0 6.3 9.5



#82
Pyrene
Concen: 16.673 ng/ul
RT: 19.857 min Scan# 2834
Delta R.T. -0.000 min
Lab File: BP013785.D
Acq: 07 Feb 2023 19:15

Tgt Ion:202 Resp: 1642041
Ion Ratio Lower Upper
202 100
101 12.8 10.3 15.5
100 9.9 8.5 12.7





#83

Butylbenzylphthalate

Concen: 17.490 ng/ul

RT: 20.710 min Scan# 2979

Delta R.T. -0.000 min

Lab File: BP013785.D

Acq: 07 Feb 2023 19:15

Instrument :

BNA_P

ClientSampleId :

SICV616





Tgt Ion:149 Resp: 589516

Ion Ratio Lower Upper

149 100

91 71.7 57.7 86.5

206 22.3 16.7 25.1



#84

3,3'-Dichlorobenzidine

Concen: 20.530 ng/ul

RT: 21.498 min Scan# 3113

Delta R.T. -0.000 min

Lab File: BP013785.D

Acq: 07 Feb 2023 19:15




Tgt Ion:252 Resp: 657437


Ion Ratio Lower Upper

252 100

254 64.3 51.8 77.6

126 9.5 7.9 11.9





#85

Benzo(a)anthracene

Concen: 17.593 ng/ul

RT: 21.575 min Scan# 3126

Delta R.T. -0.000 min

Lab File: BP013785.D

Acq: 07 Feb 2023 19:15

Instrument :

BNA_P

ClientSampleId :

SICV616





Tgt Ion:228 Resp: 1710444

Ion Ratio Lower Upper

228 100

229 19.4 15.4 23.2

226 26.7 21.8 32.6



#86

Bis(2-ethylhexyl)phthalate

Concen: 17.892 ng/ul

RT: 21.469 min Scan# 3108

Delta R.T. -0.000 min

Lab File: BP013785.D

Acq: 07 Feb 2023 19:15




Tgt Ion:149 Resp: 923461


Ion Ratio Lower Upper

149 100

167 27.3 21.5 32.3

279 4.4 3.2 4.8





#87

Chrysene

Concen: 17.611 ng/ul

RT: 21.633 min Scan# 3135

Delta R.T. 0.006 min

Lab File: BP013785.D

Acq: 07 Feb 2023 19:15

Instrument :

BNA_P

ClientSampleId :

SICV616





Tgt Ion:228 Resp: 1615366

Ion Ratio Lower Upper

228 100

226 29.4 25.0 37.6

229 20.0 16.4 24.6



#89

Di-n-octyl phthalate

Concen: 15.095 ng/ul




RT: 22.451 min Scan# 3275


Delta R.T. -0.000 min

Lab File: BP013785.D

Acq: 07 Feb 2023 19:15

Tgt Ion:149 Resp: 1600224









#90
Benzo(b)fluoranthene
Concen: 17.260 ng/ul
RT: 23.363 min Scan# 3430
Delta R.T. -0.000 min
Lab File: BP013785.D
Acq: 07 Feb 2023 19:15




Instrument :
BNA_P
ClientSampleId :
SICV616


Tgt Ion:252 Resp: 1773126
Ion Ratio Lower Upper
252 100
253 22.2 17.5 26.3
125 9.2 7.8 11.6



#91
Benzo(k)fluoranthene
Concen: 16.585 ng/ul
RT: 23.416 min Scan# 3439
Delta R.T. -0.000 min
Lab File: BP013785.D
Acq: 07 Feb 2023 19:15

Tgt Ion:252 Resp: 1690302
Ion Ratio Lower Upper
252 100
253 21.8 17.1 25.7
125 8.8 7.2 10.8









#93
Benzo(a)pyrene
Concen: 17.353 ng/ul
RT: 24.039 min Scan# 3545
Delta R.T. -0.000 min
Lab File: BP013785.D
Acq: 07 Feb 2023 19:15




Instrument :
BNA_P
ClientSampleId :
SICV616


Tgt Ion:252 Resp: 1510940
Ion Ratio Lower Upper
252 100
253 21.4 17.4 26.0
125 9.5 8.2 12.2



#94
Indeno(1,2,3-cd)pyrene
Concen: 17.845 ng/ul
RT: 26.857 min Scan# 4024
Delta R.T. -0.000 min
Lab File: BP013785.D
Acq: 07 Feb 2023 19:15

Tgt Ion:276 Resp: 1742777
Ion Ratio Lower Upper
276 100
138 21.7 18.4 27.6
277 25.8 19.7 29.5





#95

Dibenzo(a,h)anthracene

Concen: 17.814 ng/ul

RT: 26.874 min Scan# 4027

Delta R.T. -0.000 min

Lab File: BP013785.D

Acq: 07 Feb 2023 19:15

Instrument :

BNA_P

ClientSampleId :

SICV616





Tgt Ion:278 Resp: 1453258

Ion Ratio Lower Upper

278 100

139 14.6 12.0 18.0

279 24.0 18.9 28.3



#96

Benzo(g,h,i)perylene

Concen: 17.604 ng/ul

RT: 27.686 min Scan# 4165

Delta R.T. -0.006 min

Lab File: BP013785.D

Acq: 07 Feb 2023 19:15

Tgt Ion:276 Resp: 1368832

Ion Ratio Lower Upper

276 100

138 20.3 15.9 23.9

277 24.6 19.0 28.6

