

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP020824\
 Data File : BP019619.D
 Acq On : 08 Feb 2024 21:08
 Operator : MA/JU
 Sample : P1391-41
 Misc :
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SP-5

Quant Time: Feb 09 00:43:47 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP013124.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Feb 01 00:51:48 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.911	152	88404	20.000	ng	0.00
21) Naphthalene-d8	10.710	136	339926	20.000	ng	0.00
39) Acenaphthene-d10	14.540	164	224040	20.000	ng	-0.01
64) Phenanthrene-d10	17.298	188	483363	20.000	ng	0.00
76) Chrysene-d12	21.392	240	411395	20.000	ng	0.00
86) Perylene-d12	23.810	264	478569	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.487	112	613517	111.866	ng	0.00
7) Phenol-d6	7.081	99	817227	110.512	ng	0.00
23) Nitrobenzene-d5	9.081	82	538559	68.652	ng	0.00
42) 2,4,6-Tribromophenol	16.040	330	413547	126.423	ng	0.00
45) 2-Fluorobiphenyl	13.169	172	1232696	68.019	ng	-0.01
79) Terphenyl-d14	19.863	244	1848765	73.904	ng	0.00
Target Compounds						
						Qvalue
75) Fluoranthene	19.304	202	98519	3.170	ng	98
78) Pyrene	19.657	202	80576	2.775	ng	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP020824\
Data File : BP019619.D
Acq On : 08 Feb 2024 21:08
Operator : MA/JU
Sample : P1391-41
Misc :
ALS Vial : 16 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
SP-5

Quant Time: Feb 09 00:43:47 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP013124.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Feb 01 00:51:48 2024
Response via : Initial Calibration

