

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP020922\
 Data File : BP009088.D
 Acq On : 09 Feb 2022 20:51
 Operator : CG/JU
 Sample : N1422-01RE 5X
 Misc :
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 BP-2RE

Quant Time: Feb 10 01:46:34 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP020722.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Feb 07 15:13:17 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.805	152	316707	20.000	ng	0.00
21) Naphthalene-d8	10.605	136	1277260	20.000	ng	0.00
39) Acenaphthene-d10	14.445	164	879161	20.000	ng	0.00
64) Phenanthrene-d10	17.204	188	1861679	20.000	ng	0.00
76) Chrysene-d12	21.304	240	1359436	20.000	ng	# 0.00
86) Perylene-d12	23.704	264	1408202	20.000	ng	0.01
System Monitoring Compounds						
5) 2-Fluorophenol	5.405	112	57252	3.225	ng	0.01
7) Phenol-d6	6.999	99	185848	7.944	ng	0.01
23) Nitrobenzene-d5	8.969	82	172138	7.600	ng	0.00
42) 2,4,6-Tribromophenol	15.957	330	8787	0.749	ng	0.01
45) 2-Fluorobiphenyl	13.069	172	548668	8.736	ng	0.00
79) Terphenyl-d14	19.769	244	752504	9.954	ng	0.00
Target Compounds						
						Qvalue
71) Phenanthrene	17.245	178	387642	3.846	ng	98
74) Di-n-butylphthalate	18.169	149	6142865	64.131	ng	99
75) Fluoranthene	19.222	202	448260	3.968	ng	97
78) Pyrene	19.575	202	346473	3.583	ng	99
80) Butylbenzylphthalate	20.445	149	153777	4.625	ng	98
84) Bis(2-ethylhexyl)phtha...	21.210	149	266447	6.058	ng	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP020922\
Data File : BP009088.D
Acq On : 09 Feb 2022 20:51
Operator : CG/JU
Sample : N1422-01RE 5X
Misc :
ALS Vial : 18 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
BP-2RE

Quant Time: Feb 10 01:46:34 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP020722.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Feb 07 15:13:17 2022
Response via : Initial Calibration

