

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP020924\
 Data File : BP019636.D
 Acq On : 09 Feb 2024 17:19
 Operator : MA/JU
 Sample : P1372-05DL3 80X
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 CG-MW-17-ch4DL3

Quant Time: Feb 09 17:49:58 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP013124.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Feb 01 00:51:48 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.904	152	81199	20.000	ng	-0.01
21) Naphthalene-d8	10.710	136	311566	20.000	ng	0.00
39) Acenaphthene-d10	14.539	164	204577	20.000	ng	-0.01
64) Phenanthrene-d10	17.298	188	453264	20.000	ng	0.00
76) Chrysene-d12	21.386	240	473318	20.000	ng	0.00
86) Perylene-d12	23.810	264	525607	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.481	112	3926	0.779	ng	0.00
7) Phenol-d6	7.075	99	2833	0.417	ng	-0.01
23) Nitrobenzene-d5	9.075	82	8921	1.241	ng	-0.01
42) 2,4,6-Tribromophenol	16.039	330	5724	1.916	ng	0.00
45) 2-Fluorobiphenyl	13.169	172	20236	1.223	ng	-0.01
79) Terphenyl-d14	19.863	244	36828	1.280	ng	0.00
Target Compounds						
31) Naphthalene	10.757	128	976146	57.125	ng	99
38) 1-Methylnaphthalene	12.587	142	65311	5.620	ng	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP020924\
Data File : BP019636.D
Acq On : 09 Feb 2024 17:19
Operator : MA/JU
Sample : P1372-05DL3 80X
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
CG-MW-17-ch4DL3

Quant Time: Feb 09 17:49:58 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP013124.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Feb 01 00:51:48 2024
Response via : Initial Calibration

