

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP020924\
 Data File : BP019649.D
 Acq On : 10 Feb 2024 00:40
 Operator : MA/JU
 Sample : P1389-01
 Misc :
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 WC-1

Quant Time: Feb 10 03:13:42 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP013124.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Feb 01 00:51:48 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.904	152	110897	20.000	ng	-0.01
21) Naphthalene-d8	10.704	136	400551	20.000	ng	-0.01
39) Acenaphthene-d10	14.539	164	220702	20.000	ng	-0.01
64) Phenanthrene-d10	17.298	188	420971	20.000	ng	0.00
76) Chrysene-d12	21.386	240	465616	20.000	ng	0.00
86) Perylene-d12	23.809	264	557054	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.481	112	639320	92.927	ng	0.00
7) Phenol-d6	7.081	99	875266	94.353	ng	0.00
23) Nitrobenzene-d5	9.075	82	600239	64.934	ng	-0.01
42) 2,4,6-Tribromophenol	16.033	330	330866	102.677	ng	0.00
45) 2-Fluorobiphenyl	13.169	172	1230625	68.932	ng	-0.01
79) Terphenyl-d14	19.863	244	1651081	58.316	ng	0.00
Target Compounds						
75) Fluoranthene	19.304	202	57205	2.114	ng	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP020924\
Data File : BP019649.D
Acq On : 10 Feb 2024 00:40
Operator : MA/JU
Sample : P1389-01
Misc :
ALS Vial : 17 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
WC-1

Quant Time: Feb 10 03:13:42 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP013124.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Feb 01 00:51:48 2024
Response via : Initial Calibration

