




Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP020924\
Data File : BP019623.D
Acq On : 09 Feb 2024 09:58
Operator : MA/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Quant Time: Feb 09 18:34:18 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP013124.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Feb 01 00:51:48 2024
Response via : Initial Calibration



TIC: BP019623.D\data.ms




(70) Pentachlorophenol (C)

16.951min (+ 0.006) 84534.30 ng

response 221333

Ion	Exp%	Act%
265.70	100.00	100.00
268.00	62.90	64.07
264.00	62.10	61.75
0.00	0.00	0.00

Quant Time: Feb 09 18:34:18 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP013124.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Feb 01 00:51:48 2024
Response via : Initial Calibration



TIC: BP019623.D\data.ms

(77) Benzidine

19.504min (+ 0.006) 88599.52 ng

response **979878**

Ion	Exp%	Act%
184.00	100.00	100.00
185.00	13.90	13.76
183.00	12.10	11.82
0.00	0.00	0.00