

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP021122\
 Data File : BP009128.D
 Acq On : 11 Feb 2022 23:32
 Operator : CG/JU
 Sample : N1505-01 2X
 Misc :
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SU-02-021022

Manual Integrations
 APPROVED

Reviewed By :Jagrut Upadhyay 02/14/2022
 Supervised By :mohammad ahmed 02/15/2022

Quant Time: Feb 12 00:40:58 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP020722.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Feb 07 15:13:17 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.793	152	218222	20.000	ng	-0.01
21) Naphthalene-d8	10.587	136	934297	20.000	ng	-0.02
39) Acenaphthene-d10	14.440	164	672061	20.000	ng	0.00
64) Phenanthrene-d10	17.198	188	1500242	20.000	ng	0.00
76) Chrysene-d12	21.304	240	1418133	20.000	ng	0.00
86) Perylene-d12	23.698	264	1465432	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.393	112	590596	48.277	ng	0.00
7) Phenol-d6	6.987	99	774315	48.035	ng	0.00
23) Nitrobenzene-d5	8.958	82	480541	29.005	ng	-0.01
42) 2,4,6-Tribromophenol	15.940	330	483602	53.894	ng	0.00
45) 2-Fluorobiphenyl	13.057	172	1431367	29.815	ng	-0.01
79) Terphenyl-d14	19.763	244	2331979	29.572	ng	0.00
Target Compounds						Qvalue
31) Naphthalene	10.640	128	608179	12.763	ng	99
37) 2-Methylnaphthalene	12.257	142	456780	13.546	ng	95
38) 1-Methylnaphthalene	12.475	142	566609	17.196	ng	98
46) 1,1'-Biphenyl	13.269	154	107536	2.178	ng	98
49) Acenaphthylene	14.157	152	854155	14.440	ng	97
52) Acenaphthene	14.498	154	710836	19.750	ng	98
55) Dibenzofuran	14.840	168	691575	11.405	ng	# 96
58) Fluorene	15.492	166	1485034	31.877	ng	98
71) Phenanthrene	17.245	178	10489268	129.125	ng	99
72) Anthracene	17.334	178	3619151	45.537	ng	100
73) Carbazole	17.610	167	711249	9.336	ng	99
75) Fluoranthene	19.228	202	14070824	154.547	ng	95
78) Pyrene	19.581	202	13613711	134.949	ng	97
81) Benzo(a)anthracene	21.286	228	7484544	81.899	ng	98
83) Chrysene	21.339	228	6090288	69.130	ng	98
87) Indeno(1,2,3-cd)pyrene	26.215	276	3867092	35.516	ng	# 93
88) Benzo(b)fluoranthene	22.974	252	7727958	85.662	ng	99
89) Benzo(k)fluoranthene	23.016	252	2706447m	30.070	ng	
90) Benzo(a)pyrene	23.604	252	6132789	81.538	ng	96
91) Dibenzo(a,h)anthracene	26.210	278	1042204	11.647	ng	# 91
92) Benzo(g,h,i)perylene	26.980	276	3655093	41.026	ng	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP021122\
Data File : BP009128.D
Acq On : 11 Feb 2022 23:32
Operator : CG/JU
Sample : N1505-01 2X
Misc :
ALS Vial : 19 Sample Multiplier: 1


Instrument :
BNA_P
ClientSampleId :
SU-02-021022

Quant Time: Feb 12 00:40:58 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP020722.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Feb 07 15:13:17 2022
Response via : Initial Calibration

Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 02/14/2022
Supervised By :mohammad ahmed 02/15/2022





#1
1,4-Dichlorobenzene-d4
Concen: 20.000 ng
RT: 7.793 min Scan# 819
Delta R.T. -0.012 min
Lab File: BP009128.D
Acq: 11 Feb 2022 23:32




Instrument :
BNA_P
ClientSampleId :
SU-02-021022




Tgt Ion:152 Resp: 21822
Ion Ratio Lower Upper
152 100
150 149.3 127.0 190.6
115 54.7 43.7 65.5

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 02/14/2022
Supervised By :mohammad ahmed 02/15/2022




#5
2-Fluorophenol
Concen: 48.277 ng
RT: 5.393 min Scan# 409
Delta R.T. 0.000 min
Lab File: BP009128.D
Acq: 11 Feb 2022 23:32




Tgt Ion:112 Resp: 590596
Ion Ratio Lower Upper
112 100
64 53.1 39.7 59.5
63 27.8 20.0 30.0





#7
Phenol-d6
Concen: 48.035 ng
RT: 6.987 min Scan# 680
Delta R.T. 0.000 min
Lab File: BP009128.D
Acq: 11 Feb 2022 23:32




Instrument :
BNA_P
ClientSampleId :
SU-02-021022




Tgt Ion: 99 Resp: 774315
Ion Ratio Lower Upper
99 100
42 15.0 10.8 16.2
71 32.3 23.8 35.6

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 02/14/2022
Supervised By :mohammad ahmed 02/15/2022




#21
Naphthalene-d8
Concen: 20.000 ng
RT: 10.587 min Scan# 1292
Delta R.T. -0.017 min
Lab File: BP009128.D
Acq: 11 Feb 2022 23:32



Tgt Ion:136 Resp: 934297
Ion Ratio Lower Upper
136 100
137 10.7 9.0 13.4
54 6.1 4.9 7.3
68 4.9 4.3 6.5





#23
Nitrobenzene-d5
Concen: 29.005 ng
RT: 8.958 min Scan# 1015
Delta R.T. -0.012 min
Lab File: BP009128.D
Acq: 11 Feb 2022 23:32

Instrument :

BNA_P





ClientSampleId :

SU-02-021022

Tgt Ion: 82 Resp: 48054
Ion Ratio Lower Upper
82 100
128 48.0 39.4 59.2
54 43.1 32.6 48.8




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 02/14/2022
Supervised By :mohammad ahmed 02/15/2022



#31
Naphthalene
Concen: 12.763 ng
RT: 10.640 min Scan# 1301
Delta R.T. -0.018 min
Lab File: BP009128.D
Acq: 11 Feb 2022 23:32


Tgt Ion:128 Resp: 608179
Ion Ratio Lower Upper
128 100
129 10.5 8.8 13.2
127 12.8 10.4 15.6





#37
2-Methylnaphthalene
Concen: 13.546 ng
RT: 12.257 min Scan# 1576
Delta R.T. -0.012 min
Lab File: BP009128.D
Acq: 11 Feb 2022 23:32




Instrument :
BNA_P
ClientSampleId :
SU-02-021022




Tgt Ion:142 Resp: 456780
Ion Ratio Lower Upper
142 100
141 92.6 70.5 105.7
115 33.1 24.6 36.8

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 02/14/2022
Supervised By :mohammad ahmed 02/15/2022




#38
1-Methylnaphthalene
Concen: 17.196 ng
RT: 12.475 min Scan# 1613
Delta R.T. -0.012 min
Lab File: BP009128.D
Acq: 11 Feb 2022 23:32




Tgt Ion:142 Resp: 566609
Ion Ratio Lower Upper
142 100
141 93.0 73.2 109.8
116 3.8 2.7 4.1





#39
Acenaphthene-d10
Concen: 20.000 ng
RT: 14.440 min Scan# 1948
Delta R.T. -0.005 min
Lab File: BP009128.D
Acq: 11 Feb 2022 23:32




Instrument :
BNA_P
ClientSampleId :
SU-02-021022




Tgt Ion:164 Resp: 67206
Ion Ratio Lower Upper
164 100
162 100.6 79.9 119.9
160 44.1 35.5 53.3

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 02/14/2022
Supervised By :mohammad ahmed 02/15/2022




#42
2,4,6-Tribromophenol
Concen: 53.894 ng
RT: 15.940 min Scan# 2202
Delta R.T. -0.006 min
Lab File: BP009128.D
Acq: 11 Feb 2022 23:32



Tgt Ion:330 Resp: 483602
Ion Ratio Lower Upper
330 100
332 96.5 77.0 115.4
141 28.2 21.3 31.9






#45
2-Fluorobiphenyl
Concen: 29.815 ng
RT: 13.057 min Scan# 1712
Delta R.T. -0.012 min
Lab File: BP009128.D
Acq: 11 Feb 2022 23:32

Instrument :

BNA_P

ClientSampleId :

SU-02-021022






Tgt Ion:172 Resp: 143136
Ion Ratio Lower Upper
172 100
171 35.9 29.3 43.9
170 22.4 19.5 29.3


Manual Integrations

APPROVED



Reviewed By :Jagrut Upadhyay 02/14/2022
Supervised By :mohammad ahmed 02/15/2022




#46
1,1'-Biphenyl
Concen: 2.178 ng
RT: 13.269 min Scan# 1748
Delta R.T. -0.012 min
Lab File: BP009128.D
Acq: 11 Feb 2022 23:32



Tgt Ion:154 Resp: 107536
Ion Ratio Lower Upper
154 100
153 39.3 20.3 60.3
76 10.4 0.0 31.1









#49
Acenaphthylene
Concen: 14.440 ng
RT: 14.157 min Scan# 1899
Delta R.T. -0.012 min
Lab File: BP009128.D
Acq: 11 Feb 2022 23:32

Instrument :
BNA_P
ClientSampleId :
SU-02-021022

Tgt Ion:152 Resp: 85415
Ion Ratio Lower Upper
152 100
151 20.3 16.1 24.1
153 15.7 10.6 16.0




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 02/14/2022
Supervised By :mohammad ahmed 02/15/2022



#52
Acenaphthene
Concen: 19.750 ng
RT: 14.498 min Scan# 1957
Delta R.T. -0.012 min
Lab File: BP009128.D
Acq: 11 Feb 2022 23:32

Tgt Ion:154 Resp: 710836
Ion Ratio Lower Upper
154 100
153 113.9 89.4 134.2
152 53.7 42.6 63.8






#55
Dibenzenofuran
Concen: 11.405 ng
RT: 14.840 min Scan# 2017
Delta R.T. -0.012 min
Lab File: BP009128.D
Acq: 11 Feb 2022 23:32

Instrument :

BNA_P

ClientSampleId :

SU-02-021022






Tgt Ion:168 Resp: 691579
Ion Ratio Lower Upper
168 100
139 35.3 27.4 41.0
169 17.9 10.9 16.3

Manual Integrations




APPROVED


Reviewed By :Jagrut Upadhyay 02/14/2022
Supervised By :mohammad ahmed 02/15/2022



#58
Fluorene
Concen: 31.877 ng
RT: 15.492 min Scan# 2126
Delta R.T. -0.006 min
Lab File: BP009128.D
Acq: 11 Feb 2022 23:32

Tgt Ion:166 Resp: 1485034
Ion Ratio Lower Upper
166 100
165 94.7 77.2 115.8
167 13.7 11.2 16.8





#64
Phenanthrene-d10
Concen: 20.000 ng
RT: 17.198 min Scan# 2416
Delta R.T. -0.006 min
Lab File: BP009128.D
Acq: 11 Feb 2022 23:32

Instrument :

BNA_P





ClientSampleId :

SU-02-021022

Tgt Ion:188 Resp: 150024
Ion Ratio Lower Upper
188 100
94 7.6 6.8 10.2
80 8.1 7.2 10.8




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 02/14/2022
Supervised By :mohammad ahmed 02/15/2022



#71
Phenanthrene
Concen: 129.125 ng
RT: 17.245 min Scan# 2424
Delta R.T. 0.000 min
Lab File: BP009128.D
Acq: 11 Feb 2022 23:32


Tgt Ion:178 Resp:10489268
Ion Ratio Lower Upper
178 100
176 20.7 16.0 24.0
179 16.6 13.0 19.4





#72
 Anthracene
 Concen: 45.537 ng
 RT: 17.334 min Scan# 2440
 Delta R.T. -0.006 min
 Lab File: BP009128.D
 Acq: 11 Feb 2022 23:32




Instrument :
 BNA_P
 ClientSampleId :
 SU-02-021022



Tgt Ion:178 Resp: 361915
 Ion Ratio Lower Upper
 178 100
 176 19.0 15.1 22.7
 179 16.0 12.8 19.2




Manual Integrations
 APPROVED


Reviewed By :Jagrut Upadhyay 02/14/2022
 Supervised By :mohammad ahmed 02/15/2022



#73
 Carbazole
 Concen: 9.336 ng
 RT: 17.610 min Scan# 2486
 Delta R.T. -0.006 min
 Lab File: BP009128.D
 Acq: 11 Feb 2022 23:32


Tgt Ion:167 Resp: 711249
 Ion Ratio Lower Upper
 167 100
 166 22.6 17.7 26.5
 139 12.6 9.8 14.6





#75
Fluoranthene
Concen: 154.547 ng
RT: 19.228 min Scan# 21
Delta R.T. 0.006 min
Lab File: BP009128.D
Acq: 11 Feb 2022 23:32




Instrument :
BNA_P
ClientSampleId :
SU-02-021022



Tgt Ion:202 Resp:14070824
Ion Ratio Lower Upper
202 100
101 9.5 0.0 30.7
203 21.3 0.0 38.9




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 02/14/2022
Supervised By :mohammad ahmed 02/15/2022



#76
Chrysene-d12
Concen: 20.000 ng
RT: 21.304 min Scan# 3114
Delta R.T. 0.006 min
Lab File: BP009128.D
Acq: 11 Feb 2022 23:32

Tgt Ion:240 Resp: 1418133
Ion Ratio Lower Upper
240 100
120 7.8 7.7 11.5
236 26.7 21.2 31.8





#78

Pyrene

Concen: 134.949 ng

RT: 19.581 min Scan# 2819

Delta R.T. 0.012 min

Lab File: BP009128.D

Acq: 11 Feb 2022 23:32

Instrument :

BNA_P

ClientSampleId :

SU-02-021022

Tgt Ion:202 Resp:1361371

Ion Ratio Lower Upper

202 100





200 23.8 18.0 27.0

203 20.2 15.0 22.6

Manual Integrations

APPROVED

Reviewed By :Jagrut Upadhyay 02/14/2022
Supervised By :mohammad ahmed 02/15/2022



#79

Terphenyl-d14

Concen: 29.572 ng

RT: 19.763 min Scan# 2852

Delta R.T. -0.006 min

Lab File: BP009128.D

Acq: 11 Feb 2022 23:32




Tgt Ion:244 Resp: 2331979


Ion Ratio Lower Upper

244 100

212 7.0 6.5 9.7


122 7.8 9.2 13.8#





#81
Benzo(a)anthracene
Concen: 81.899 ng
RT: 21.286 min Scan# 3110
Delta R.T. 0.006 min
Lab File: BP009128.D
Acq: 11 Feb 2022 23:32




Instrument :
BNA_P
ClientSampleId :
SU-02-021022



Tgt Ion:228 Resp: 7484544
Ion Ratio Lower Upper
228 100
226 30.2 23.3 34.9
229 22.3 17.0 25.6




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 02/14/2022
Supervised By :mohammad ahmed 02/15/2022



#83
Chrysene
Concen: 69.130 ng
RT: 21.339 min Scan# 3120
Delta R.T. 0.000 min
Lab File: BP009128.D
Acq: 11 Feb 2022 23:32

Tgt Ion:228 Resp: 6090288
Ion Ratio Lower Upper
228 100
226 32.4 26.0 39.0
229 23.2 16.9 25.3









#86
Perylene-d12
Concen: 20.000 ng
RT: 23.698 min Scan# 3521
Delta R.T. 0.006 min
Lab File: BP009128.D
Acq: 11 Feb 2022 23:32

Instrument :
BNA_P
ClientSampleId :
SU-02-021022

Tgt Ion:264 Resp: 146543
Ion Ratio Lower Upper
264 100
260 24.6 19.1 28.7
265 24.6 17.1 25.7




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 02/14/2022
Supervised By :mohammad ahmed 02/15/2022



#87
Indeno(1,2,3-cd)pyrene
Concen: 35.516 ng
RT: 26.215 min Scan# 3949
Delta R.T. 0.024 min
Lab File: BP009128.D
Acq: 11 Feb 2022 23:32


Tgt Ion:276 Resp: 3867092
Ion Ratio Lower Upper
276 100
138 15.7 17.8 26.6#
277 25.0 20.4 30.6





#88
Benzo(b)fluoranthene
Concen: 85.662 ng
RT: 22.974 min Scan# 31
Delta R.T. 0.012 min
Lab File: BP009128.D
Acq: 11 Feb 2022 23:32




Instrument :
BNA_P
ClientSampleId :
SU-02-021022



Tgt Ion:252 Resp: 7727958
Ion Ratio Lower Upper
252 100
253 23.2 18.0 27.0
125 8.3 6.9 10.3




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 02/14/2022
Supervised By :mohammad ahmed 02/15/2022



#89
Benzo(k)fluoranthene
Concen: 30.070 ng m
RT: 23.016 min Scan# 3405
Delta R.T. 0.006 min
Lab File: BP009128.D
Acq: 11 Feb 2022 23:32


Tgt Ion:252 Resp: 2706447
Ion Ratio Lower Upper
252 100
253 24.3 18.2 27.2
125 8.0 7.3 10.9





#90
Benzo(a)pyrene
Concen: 81.538 ng
RT: 23.604 min Scan# 3505
Delta R.T. 0.018 min
Lab File: BP009128.D
Acq: 11 Feb 2022 23:32




Instrument :
BNA_P
ClientSampleId :
SU-02-021022




Tgt Ion:252 Resp: 6132789
Ion Ratio Lower Upper
252 100
253 23.8 17.8 26.8
125 7.9 7.8 11.6

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 02/14/2022
Supervised By :mohammad ahmed 02/15/2022




#91
Dibenzo(a,h)anthracene
Concen: 11.647 ng
RT: 26.210 min Scan# 3948
Delta R.T. 0.006 min
Lab File: BP009128.D
Acq: 11 Feb 2022 23:32



Tgt Ion:278 Resp: 1042204
Ion Ratio Lower Upper
278 100
139 17.2 11.5 17.3
279 28.8 19.1 28.7#





#92

Benzo(g,h,i)perylene

Concen: 41.026 ng

RT: 26.980 min Scan# 4079

Delta R.T. 0.024 min

Lab File: BP009128.D


Acq: 11 Feb 2022 23:32

Instrument :

BNA_P

ClientSampleId :

SU-02-021022



Tgt Ion:276 Resp: 365509

Ion Ratio Lower Upper

276 100

277 24.2 19.2 28.8

138 15.9 15.0 22.4

Manual Integrations

APPROVED

Reviewed By :Jagrut Upadhyay 02/14/2022

Supervised By :mohammad ahmed 02/15/2022

