

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP021423\
 Data File : BP013813.D
 Acq On : 14 Feb 2023 09:29
 Operator : CG/JU
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SSTD020762

Quant Time: Feb 15 01:20:23 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP020723.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed Feb 15 01:19:22 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.128	152	200095	20.000	ng/u1	0.00
20) Naphthalene-d8	10.957	136	753631	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.763	164	494309	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.515	188	1105695	20.000	ng/u1	0.00
79) Chrysene-d12	21.598	240	944831	20.000	ng/u1	0.00
88) Perylene-d12	24.168	264	874346	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.469	96	43883	9.090	ng/uL	0.00
4) Pyridine-d5	3.904	84	297416	20.780	ng/u1	0.00
7) Phenol-d5	7.269	99	323202	18.129	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.445	67	194634	18.630	ng/u1	0.00
11) 2-Chlorophenol-d4	7.651	132	248310	18.920	ng/u1	0.00
15) 4-Methylphenol-d8	8.834	113	250045	17.159	ng/u1	0.00
21) Nitrobenzene-d5	9.298	128	112498	21.454	ng/u1	0.00
24) 2-Nitrophenol-d4	10.028	143	117895	19.473	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.569	165	238510	19.332	ng/u1	0.00
31) 4-Chloroaniline-d4	11.086	131	311803	17.829	ng/u1	0.00
46) Dimethylphthalate-d6	14.163	166	736296	18.912	ng/u1	0.00
49) Acenaphthylene-d8	14.457	160	817367	19.395	ng/u1	0.00
54) 4-Nitrophenol-d4	14.945	143	115313	16.625	ng/u1	0.00
60) Fluorene-d10	15.757	176	605093	18.247	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.869	200	122401	18.044	ng/u1	0.00
73) Anthracene-d10	17.615	188	932480	18.443	ng/u1	0.00
81) Pyrene-d10	19.833	212	1112518	20.764	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.997	264	817307	18.655	ng/u1	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.504	88	45545	8.788	ng/uL	100
5) Pyridine	3.922	79	291092	20.580	ng/u1	100
6) Benzaldehyde	7.257	77	146052	18.166	ng/u1	99
8) Phenol	7.298	94	333322	18.214	ng/u1	97
10) Bis(2-Chloroethyl)ether	7.540	93	272591	18.588	ng/u1	97
12) 2-Chlorophenol	7.681	128	257879	19.151	ng/u1	99
13) 2-Methylphenol	8.563	108	242816	17.383	ng/u1	98
14) 2,2'-oxybis(1-Chloropr...	8.657	45	294576	17.660	ng/u1	98
16) Acetophenone	8.957	105	410704	17.575	ng/u1	100
17) N-Nitroso-di-n-propyla...	8.939	70	199165	17.162	ng/u1	97
18) 4-Methylphenol	8.892	108	267352	17.335	ng/u1	99
19) Hexachloroethane	9.222	117	105632	19.156	ng/u1	96
22) Nitrobenzene	9.339	77	293906	21.305	ng/u1	99
23) Isophorone	9.869	82	531050	18.246	ng/u1	99
25) 2-Nitrophenol	10.057	139	126897	19.338	ng/u1	93
26) 2,4-Dimethylphenol	10.110	107	277035	18.962	ng/u1	99
27) Bis(2-Chloroethoxy)met...	10.351	93	322605	18.354	ng/u1	99
29) 2,4-Dichlorophenol	10.592	162	230605	18.986	ng/u1	99
30) Naphthalene	11.004	128	748344	18.653	ng/u1	99
32) 4-Chloroaniline	11.110	127	316639	17.960	ng/u1	98
33) Hexachlorobutadiene	11.286	225	171155	20.004	ng/u1	96
34) Caprolactam	11.886	113	64375	15.104	ng/u1	94
35) 4-Chloro-3-methylphenol	12.216	107	251140	17.909	ng/u1	97
36) 2-Methylnaphthalene	12.598	142	509315	18.057	ng/u1	100

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP021423\
 Data File : BP013813.D
 Acq On : 14 Feb 2023 09:29
 Operator : CG/JU
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SSTD020762


Quant Time: Feb 15 01:20:23 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP020723.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed Feb 15 01:19:22 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.816	142	518327	17.785	ng/ul	100
39) 1,2,4,5-Tetrachloroben...	12.963	216	311158	19.976	ng/ul	99
40) Hexachlorocyclopentadiene	12.939	237	200255	20.022	ng/ul	98
41) 2,4,6-Trichlorophenol	13.198	196	200351	19.600	ng/ul	100
42) 2,4,5-Trichlorophenol	13.269	196	217142	19.625	ng/ul	99
43) 1,1'-Biphenyl	13.598	154	707903	19.086	ng/ul	99
44) 2-Chloronaphthalene	13.645	162	602121	20.636	ng/ul	99
45) 2-Nitroaniline	13.845	65	138091	19.649	ng/ul	93
47) Dimethylphthalate	14.210	163	742837	19.237	ng/ul	100
48) 2,6-Dinitrotoluene	14.333	165	137966	20.450	ng/ul	99
50) Acenaphthylene	14.486	152	884229	19.449	ng/ul	100
51) 3-Nitroaniline	14.663	138	111588	16.729	ng/ul	98
52) Acenaphthene	14.827	153	588410	18.491	ng/ul	99
53) 2,4-Dinitrophenol	14.869	184	77454	16.099	ng/ul	99
55) 4-Nitrophenol	14.963	109	108016	17.748	ng/ul	99
56) Dibenzofuran	15.163	168	860439	18.860	ng/ul	98
57) 2,4-Dinitrotoluene	15.116	165	198150	19.700	ng/ul	96
58) 2,3,4,6-Tetrachlorophenol	15.386	232	193062	18.512	ng/ul	99
59) Diethylphthalate	15.574	149	746963	19.264	ng/ul	100
61) Fluorene	15.810	166	673522	18.096	ng/ul	99
62) 4-Chlorophenyl-phenyle...	15.798	204	361135	18.661	ng/ul	97
63) 4-Nitroaniline	15.827	138	95520	15.123	ng/ul	93
66) 4,6-Dinitro-2-methylph...	15.880	198	122558	18.429	ng/ul	98
67) N-Nitrosodiphenylamine	16.016	169	581399	19.345	ng/ul	99
68) 4-Bromophenyl-phenylether	16.698	248	226818	19.425	ng/ul	97
69) Hexachlorobenzene	16.816	284	259665	19.085	ng/ul	95
70) Atrazine	16.963	200	227187	18.903	ng/ul	99
71) Pentachlorophenol	17.163	266	159048	17.964	ng/ul	98
72) Phenanthrene	17.563	178	1082779	18.506	ng/ul	100
74) Anthracene	17.651	178	1094875	18.599	ng/ul	100
75) 1,2,3,4-Tetrachloroben...	13.569	216	305918	20.726	ng/uL	98
76) Pentachlorobenzene	15.080	250	308407	20.248	ng/uL	100
77) Carbazole	17.915	167	1030539	19.874	ng/ul	99
78) Di-n-butylphthalate	18.451	149	1174149	18.669	ng/ul	99
80) Fluoranthene	19.509	202	1312378	21.396	ng/ul	97
82) Pyrene	19.862	202	1350423	20.650	ng/ul	95
83) Butylbenzylphthalate	20.715	149	467526	20.888	ng/ul	99
84) 3,3'-Dichlorobenzidine	21.504	252	374268	17.600	ng/ul	99
85) Benzo(a)anthracene	21.580	228	1198345	18.562	ng/ul	100
86) Bis(2-ethylhexyl)phtha...	21.480	149	678090	19.785	ng/ul#	98
87) Chrysene	21.633	228	1141038	18.733	ng/ul	97
89) Di-n-octyl phthalate	22.456	149	1058786	17.598	ng/ul	100
90) Benzo(b)fluoranthene	23.374	252	1064920	18.265	ng/ul	99
91) Benzo(k)fluoranthene	23.427	252	1075356	18.592	ng/ul	99
93) Benzo(a)pyrene	24.050	252	925754	18.735	ng/ul	99
94) Indeno(1,2,3-cd)pyrene	26.874	276	1171637	21.138	ng/ul	98
95) Dibenzo(a,h)anthracene	26.891	278	983120	21.234	ng/ul	99
96) Benzo(g,h,i)perylene	27.709	276	953129	21.599	ng/ul	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Quant Time: Feb 15 01:20:23 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP020723.M
Quant Title : SVOA CALIBRATION
QLast Update : Wed Feb 15 01:19:22 2023
Response via : Initial Calibration









#2
 1,4-Dioxane
 Concen: 8.788 ng/uL
 RT: 3.504 min Scan# 54
 Delta R.T. 0.000 min
 Lab File: BP013813.D
 Acq: 14 Feb 2023 09:29




Instrument :
 BNA_P
 ClientSampleId :
 SSTD020762


Tgt Ion:	88	Resp:	45545
Ion	Ratio	Lower	Upper
88	100		
43	24.8	20.1	30.1
58	66.5	53.0	79.6



#5
 Pyridine
 Concen: 20.580 ng/uL
 RT: 3.922 min Scan# 125
 Delta R.T. 0.000 min
 Lab File: BP013813.D
 Acq: 14 Feb 2023 09:29

Tgt Ion:	79	Resp:	291092
Ion	Ratio	Lower	Upper
79	100		
52	68.7	55.1	82.7
51	32.4	25.7	38.5









#6
Benzaldehyde
Concen: 18.166 ng/ul
RT: 7.257 min Scan# 691
Delta R.T. 0.000 min
Lab File: BP013813.D
Acq: 14 Feb 2023 09:29




Instrument :
BNA_P
ClientSampleId :
SSTD020762


Tgt Ion: 77 Resp: 146052
Ion Ratio Lower Upper
77 100
105 100.2 79.5 119.3
106 93.8 76.5 114.7



#8
Phenol
Concen: 18.214 ng/ul
RT: 7.298 min Scan# 699
Delta R.T. 0.000 min
Lab File: BP013813.D
Acq: 14 Feb 2023 09:29

Tgt Ion: 94 Resp: 333322
Ion Ratio Lower Upper
94 100
65 27.4 22.2 33.4
66 39.7 29.6 44.4









#10
Bis(2-Chloroethyl)ether
Concen: 18.588 ng/ul
RT: 7.540 min Scan# 740
Delta R.T. 0.000 min
Lab File: BP013813.D
Acq: 14 Feb 2023 09:29




Instrument :
BNA_P
ClientSampleId :
SSTD020762


Tgt Ion: 93 Resp: 272591
Ion Ratio Lower Upper
93 100
63 71.3 59.4 89.0
95 32.5 25.8 38.8



#12
2-Chlorophenol
Concen: 19.151 ng/ul
RT: 7.681 min Scan# 764
Delta R.T. 0.000 min
Lab File: BP013813.D
Acq: 14 Feb 2023 09:29

Tgt Ion: 128 Resp: 257879
Ion Ratio Lower Upper
128 100
64 46.0 37.3 55.9
130 31.9 26.3 39.5









#13
2-Methylphenol
Concen: 17.383 ng/ul
RT: 8.563 min Scan# 914
Delta R.T. 0.000 min
Lab File: BP013813.D
Acq: 14 Feb 2023 09:29




Instrument :
BNA_P
ClientSampleId :
SSTD020762


Tgt Ion:108 Resp: 242816
Ion Ratio Lower Upper
108 100
107 88.3 71.9 107.9



#14
2,2'-oxybis(1-Chloropropane)
Concen: 17.660 ng/ul
RT: 8.657 min Scan# 930
Delta R.T. 0.000 min
Lab File: BP013813.D
Acq: 14 Feb 2023 09:29

Tgt Ion: 45 Resp: 294576
Ion Ratio Lower Upper
45 100
77 17.5 13.4 20.2
79 13.1 11.0 16.4









#16
Acetophenone
Concen: 17.575 ng/ul
RT: 8.957 min Scan# 91
Delta R.T. 0.000 min
Lab File: BP013813.D
Acq: 14 Feb 2023 09:29




Instrument :
BNA_P
ClientSampleId :
SSTD020762


Tgt Ion:105 Resp: 410704
Ion Ratio Lower Upper
105 100
77 76.7 61.0 91.4
51 24.5 19.6 29.4



#17
N-Nitroso-di-n-propylamine
Concen: 17.162 ng/ul
RT: 8.939 min Scan# 978
Delta R.T. 0.000 min
Lab File: BP013813.D
Acq: 14 Feb 2023 09:29

Tgt Ion: 70 Resp: 199165
Ion Ratio Lower Upper
70 100
42 51.4 39.8 59.8
101 10.7 7.4 11.2
130 20.4 14.6 22.0









#18
4-Methylphenol
Concen: 17.335 ng/ul
RT: 8.892 min Scan# 91
Delta R.T. 0.000 min
Lab File: BP013813.D
Acq: 14 Feb 2023 09:29




Instrument :
BNA_P
ClientSampleId :
SSTD020762


Tgt Ion:108 Resp: 267352
Ion Ratio Lower Upper
108 100
107 112.3 90.9 136.3



#19
Hexachloroethane
Concen: 19.156 ng/ul
RT: 9.222 min Scan# 1026
Delta R.T. 0.000 min
Lab File: BP013813.D
Acq: 14 Feb 2023 09:29

Tgt Ion:117 Resp: 105632
Ion Ratio Lower Upper
117 100
201 113.9 86.6 129.8
199 70.7 55.2 82.8









#22
Nitrobenzene
Concen: 21.305 ng/ul
RT: 9.339 min Scan# 1046
Delta R.T. 0.000 min
Lab File: BP013813.D
Acq: 14 Feb 2023 09:29




Instrument :
BNA_P
ClientSampleId :
SSTD020762


Tgt Ion: 77 Resp: 293906
Ion Ratio Lower Upper
77 100
123 43.3 35.1 52.7
65 13.5 10.6 16.0



#23
Isophorone
Concen: 18.246 ng/ul
RT: 9.869 min Scan# 1136
Delta R.T. 0.000 min
Lab File: BP013813.D
Acq: 14 Feb 2023 09:29

Tgt Ion: 82 Resp: 531050
Ion Ratio Lower Upper
82 100
95 7.6 6.0 9.0
138 16.4 13.7 20.5









#25
2-Nitrophenol
Concen: 19.338 ng/ul
RT: 10.057 min Scan# 1168
Delta R.T. 0.000 min
Lab File: BP013813.D
Acq: 14 Feb 2023 09:29




Instrument :
BNA_P
ClientSampleId :
SSTD020762


Tgt Ion:139 Resp: 126897
Ion Ratio Lower Upper
139 100
65 55.8 39.9 59.9
109 36.7 27.4 41.0



#26
2,4-Dimethylphenol
Concen: 18.962 ng/ul
RT: 10.110 min Scan# 1177
Delta R.T. 0.000 min
Lab File: BP013813.D
Acq: 14 Feb 2023 09:29

Tgt Ion:107 Resp: 277035
Ion Ratio Lower Upper
107 100
121 48.7 37.8 56.6
122 80.2 63.7 95.5









#27
Bis(2-Chloroethoxy)methane
Concen: 18.354 ng/ul
RT: 10.351 min Scan# 1218
Delta R.T. 0.000 min
Lab File: BP013813.D
Acq: 14 Feb 2023 09:29




Instrument :
BNA_P
ClientSampleId :
SSTD020762


Tgt Ion: 93 Resp: 322605
Ion Ratio Lower Upper
93 100
95 32.0 25.0 37.6
123 10.9 8.2 12.4



#29
2,4-Dichlorophenol
Concen: 18.986 ng/ul
RT: 10.592 min Scan# 1259
Delta R.T. 0.000 min
Lab File: BP013813.D
Acq: 14 Feb 2023 09:29

Tgt Ion:162 Resp: 230605
Ion Ratio Lower Upper
162 100
164 63.9 51.4 77.2
98 35.3 27.5 41.3









#30
Naphthalene
Concen: 18.653 ng/ul
RT: 11.004 min Scan# 1329
Delta R.T. 0.000 min
Lab File: BP013813.D
Acq: 14 Feb 2023 09:29




Instrument :
BNA_P
ClientSampleId :
SSTD020762


Tgt Ion:128 Resp: 748344
Ion Ratio Lower Upper
128 100
129 10.8 8.6 12.8
127 12.9 10.6 15.8



#32
4-Chloroaniline
Concen: 17.960 ng/ul
RT: 11.110 min Scan# 1347
Delta R.T. 0.000 min
Lab File: BP013813.D
Acq: 14 Feb 2023 09:29

Tgt Ion:127 Resp: 316639
Ion Ratio Lower Upper
127 100
129 31.6 26.2 39.4









#33
Hexachlorobutadiene
Concen: 20.004 ng/ul
RT: 11.286 min Scan# 11
Delta R.T. 0.000 min
Lab File: BP013813.D
Acq: 14 Feb 2023 09:29




Instrument :
BNA_P
ClientSampleId :
SSTD020762


Tgt Ion:225 Resp: 171155
Ion Ratio Lower Upper
225 100
223 60.9 50.6 76.0
227 65.5 49.0 73.4



#34
Caprolactam
Concen: 15.104 ng/ul
RT: 11.886 min Scan# 1479
Delta R.T. 0.000 min
Lab File: BP013813.D
Acq: 14 Feb 2023 09:29

Tgt Ion:113 Resp: 64375
Ion Ratio Lower Upper
113 100
55 145.4 120.9 181.3
56 107.8 93.8 140.8









#35
4-Chloro-3-methylphenol
Concen: 17.909 ng/ul
RT: 12.216 min Scan# 1535
Delta R.T. 0.000 min
Lab File: BP013813.D
Acq: 14 Feb 2023 09:29




Instrument :
BNA_P
ClientSampleId :
SSTD020762


Tgt Ion:107 Resp: 251140
Ion Ratio Lower Upper
107 100
144 24.3 20.0 30.0
142 74.5 62.1 93.1



#36
2-Methylnaphthalene
Concen: 18.057 ng/ul
RT: 12.598 min Scan# 1600
Delta R.T. 0.000 min
Lab File: BP013813.D
Acq: 14 Feb 2023 09:29

Tgt Ion:142 Resp: 509315
Ion Ratio Lower Upper
142 100
141 90.9 72.8 109.2









#37
1-Methylnaphthalene
Concen: 17.785 ng/ul
RT: 12.816 min Scan# 1637
Delta R.T. 0.000 min
Lab File: BP013813.D
Acq: 14 Feb 2023 09:29




Instrument :
BNA_P
ClientSampleId :
SSTD020762


Tgt Ion:142 Resp: 518327
Ion Ratio Lower Upper
142 100
141 92.6 74.2 111.2
116 4.0 3.1 4.7



#39
1,2,4,5-Tetrachlorobenzene
Concen: 19.976 ng/ul
RT: 12.963 min Scan# 1662
Delta R.T. 0.000 min
Lab File: BP013813.D
Acq: 14 Feb 2023 09:29

Tgt Ion:216 Resp: 311158
Ion Ratio Lower Upper
216 100
214 78.4 62.6 93.8
179 21.0 15.4 23.2
108 15.8 12.7 19.1





#40

Hexachlorocyclopentadiene

Concen: 20.022 ng/ul

RT: 12.939 min Scan# 1

Delta R.T. 0.000 min

Lab File: BP013813.D

Acq: 14 Feb 2023 09:29

Instrument :

BNA_P

ClientSampleId :

SSTD020762





Tgt Ion:237 Resp: 200255

Ion Ratio Lower Upper

237 100

235 64.4 49.8 74.8

272 12.4 10.1 15.1



#41

2,4,6-Trichlorophenol

Concen: 19.600 ng/ul

RT: 13.198 min Scan# 1702

Delta R.T. 0.000 min

Lab File: BP013813.D

Acq: 14 Feb 2023 09:29




Tgt Ion:196 Resp: 200351


Ion Ratio Lower Upper

196 100

198 94.3 75.1 112.7

200 30.7 24.3 36.5





#42

2,4,5-Trichlorophenol

Concen: 19.625 ng/ul

RT: 13.269 min Scan# 1

Delta R.T. 0.000 min

Lab File: BP013813.D

Acq: 14 Feb 2023 09:29

Instrument :

BNA_P

ClientSampleId :

SSTD020762





Tgt Ion:196 Resp: 217142

Ion Ratio Lower Upper

196 100

198 96.3 77.0 115.4

200 29.5 25.8 38.6



#43

1,1'-Biphenyl

Concen: 19.086 ng/ul

RT: 13.598 min Scan# 1770

Delta R.T. 0.000 min

Lab File: BP013813.D

Acq: 14 Feb 2023 09:29




Tgt Ion:154 Resp: 707903


Ion Ratio Lower Upper

154 100

153 40.1 31.9 47.9

76 11.4 9.9 14.9





#44

2-Chloronaphthalene

Concen: 20.636 ng/ul

RT: 13.645 min Scan# 1

Delta R.T. 0.000 min

Lab File: BP013813.D

Acq: 14 Feb 2023 09:29

Instrument :

BNA_P

ClientSampleId :

SSTD020762





Tgt Ion:162 Resp: 602121

Ion Ratio Lower Upper

162 100

164 33.2 26.1 39.1

127 39.1 31.1 46.7



#45

2-Nitroaniline

Concen: 19.649 ng/ul

RT: 13.845 min Scan# 1812

Delta R.T. 0.000 min

Lab File: BP013813.D

Acq: 14 Feb 2023 09:29




Tgt Ion: 65 Resp: 138091


Ion Ratio Lower Upper

65 100

92 70.8 58.6 88.0

138 112.1 81.2 121.8









#47
Dimethylphthalate
Concen: 19.237 ng/ul
RT: 14.210 min Scan# 1874
Delta R.T. 0.000 min
Lab File: BP013813.D
Acq: 14 Feb 2023 09:29




Instrument :
BNA_P
ClientSampleId :
SSTD020762


Tgt Ion:163 Resp: 742837
Ion Ratio Lower Upper
163 100
194 4.0 3.4 5.2
164 9.8 7.8 11.8



#48
2,6-Dinitrotoluene
Concen: 20.450 ng/ul
RT: 14.333 min Scan# 1895
Delta R.T. 0.000 min
Lab File: BP013813.D
Acq: 14 Feb 2023 09:29

Tgt Ion:165 Resp: 137966
Ion Ratio Lower Upper
165 100
89 59.3 47.5 71.3
121 22.4 16.6 24.8









#50
Acenaphthylene
Concen: 19.449 ng/ul
RT: 14.486 min Scan# 1920
Delta R.T. 0.000 min
Lab File: BP013813.D
Acq: 14 Feb 2023 09:29




Instrument :
BNA_P
ClientSampleId :
SSTD020762


Tgt Ion:152 Resp: 884229
Ion Ratio Lower Upper
152 100
151 19.9 15.8 23.6
153 12.9 10.4 15.6



#51
3-Nitroaniline
Concen: 16.729 ng/ul
RT: 14.663 min Scan# 1951
Delta R.T. 0.000 min
Lab File: BP013813.D
Acq: 14 Feb 2023 09:29

Tgt Ion:138 Resp: 111588
Ion Ratio Lower Upper
138 100
108 11.5 9.4 14.2
92 123.9 97.4 146.0









#52
Acenaphthene
Concen: 18.491 ng/ul
RT: 14.827 min Scan# 1978
Delta R.T. 0.000 min
Lab File: BP013813.D
Acq: 14 Feb 2023 09:29




Instrument :
BNA_P
ClientSampleId :
SSTD020762


Tgt Ion:153 Resp: 588410
Ion Ratio Lower Upper
153 100
152 47.7 37.4 56.0
154 87.6 70.1 105.1



#53
2,4-Dinitrophenol
Concen: 16.099 ng/ul
RT: 14.869 min Scan# 1986
Delta R.T. 0.000 min
Lab File: BP013813.D
Acq: 14 Feb 2023 09:29

Tgt Ion:184 Resp: 77454
Ion Ratio Lower Upper
184 100
63 61.2 49.5 74.3
154 68.9 54.3 81.5









#55
4-Nitrophenol
Concen: 17.748 ng/ul
RT: 14.963 min Scan# 2001
Delta R.T. 0.000 min
Lab File: BP013813.D
Acq: 14 Feb 2023 09:29




Instrument :
BNA_P
ClientSampleId :
SSTD020762


Tgt Ion:109 Resp: 108016
Ion Ratio Lower Upper
109 100
139 120.1 96.0 144.0
65 117.7 95.8 143.6



#56
Dibenzofuran
Concen: 18.860 ng/ul
RT: 15.163 min Scan# 2036
Delta R.T. 0.000 min
Lab File: BP013813.D
Acq: 14 Feb 2023 09:29

Tgt Ion:168 Resp: 860439
Ion Ratio Lower Upper
168 100
139 38.1 29.6 44.4





#57

2,4-Dinitrotoluene

Concen: 19.700 ng/ul

RT: 15.116 min Scan# 2028

Delta R.T. 0.000 min

Lab File: BP013813.D

Acq: 14 Feb 2023 09:29

Instrument :

BNA_P

ClientSampleId :

SSTD020762





Tgt Ion:165 Resp: 198150

Ion Ratio Lower Upper

165 100

63 42.3 32.0 48.0

182 2.8 2.4 3.6



#58

2,3,4,6-Tetrachlorophenol

Concen: 18.512 ng/ul

RT: 15.386 min Scan# 2074

Delta R.T. 0.000 min

Lab File: BP013813.D

Acq: 14 Feb 2023 09:29

Tgt Ion:232 Resp: 193062




Ion Ratio Lower Upper


232 100

131 40.7 33.0 49.6

130 2.2 1.5 2.3

166 29.3 22.6 33.8









#59
Diethylphthalate
Concen: 19.264 ng/ul
RT: 15.574 min Scan# 2106
Delta R.T. 0.000 min
Lab File: BP013813.D
Acq: 14 Feb 2023 09:29




Instrument :
BNA_P
ClientSampleId :
SSTD020762


Tgt Ion:149 Resp: 746963
Ion Ratio Lower Upper
149 100
177 22.4 17.8 26.8
150 12.4 10.1 15.1



#61
Fluorene
Concen: 18.096 ng/ul
RT: 15.810 min Scan# 2146
Delta R.T. 0.000 min
Lab File: BP013813.D
Acq: 14 Feb 2023 09:29

Tgt Ion:166 Resp: 673522
Ion Ratio Lower Upper
166 100
165 98.1 79.1 118.7
167 13.0 10.8 16.2









#62
4-Chlorophenyl-phenylether
Concen: 18.661 ng/ul
RT: 15.798 min Scan# 2143
Delta R.T. 0.000 min
Lab File: BP013813.D
Acq: 14 Feb 2023 09:29




Instrument :
BNA_P
ClientSampleId :
SSTD020762


Tgt Ion:204 Resp: 361135
Ion Ratio Lower Upper
204 100
206 32.7 26.7 40.1
141 59.0 44.6 66.8



#63
4-Nitroaniline
Concen: 15.123 ng/ul
RT: 15.827 min Scan# 2149
Delta R.T. 0.000 min
Lab File: BP013813.D
Acq: 14 Feb 2023 09:29

Tgt Ion:138 Resp: 95520
Ion Ratio Lower Upper
138 100
92 52.4 43.8 65.8
108 95.6 69.8 104.6









#66
4,6-Dinitro-2-methylphenol
Concen: 18.429 ng/ul
RT: 15.880 min Scan# 2158
Delta R.T. 0.000 min
Lab File: BP013813.D
Acq: 14 Feb 2023 09:29




Instrument :
BNA_P
ClientSampleId :
SSTD020762


Tgt Ion:198 Resp: 122558
Ion Ratio Lower Upper
198 100
182 4.3 4.1 6.1
77 24.7 20.3 30.5



#67
N-Nitrosodiphenylamine
Concen: 19.345 ng/ul
RT: 16.016 min Scan# 2181
Delta R.T. 0.000 min
Lab File: BP013813.D
Acq: 14 Feb 2023 09:29

Tgt Ion:169 Resp: 581399
Ion Ratio Lower Upper
169 100
168 67.3 53.5 80.3
167 37.7 29.8 44.6









#68
4-Bromophenyl-phenylether
Concen: 19.425 ng/ul
RT: 16.698 min Scan# 21
Delta R.T. 0.000 min
Lab File: BP013813.D
Acq: 14 Feb 2023 09:29




Instrument :
BNA_P
ClientSampleId :
SSTD020762


Tgt Ion:248 Resp: 226818
Ion Ratio Lower Upper
248 100
250 97.6 76.0 114.0
141 72.4 56.0 84.0



#69
Hexachlorobenzene
Concen: 19.085 ng/ul
RT: 16.816 min Scan# 2317
Delta R.T. 0.000 min
Lab File: BP013813.D
Acq: 14 Feb 2023 09:29

Tgt Ion:284 Resp: 259665
Ion Ratio Lower Upper
284 100
142 34.8 25.0 37.4
249 33.0 24.9 37.3





#70

Atrazine

Concen: 18.903 ng/ul

RT: 16.963 min Scan# 21

Delta R.T. 0.000 min

Lab File: BP013813.D

Acq: 14 Feb 2023 09:29

Instrument :

BNA_P

ClientSampleId :

SSTD020762





Tgt Ion:200 Resp: 227187

Ion Ratio Lower Upper

200 100

173 25.9 20.4 30.6

215 49.4 39.3 58.9



#71

Pentachlorophenol

Concen: 17.964 ng/ul

RT: 17.163 min Scan# 2376

Delta R.T. 0.000 min

Lab File: BP013813.D

Acq: 14 Feb 2023 09:29




Tgt Ion:266 Resp: 159048


Ion Ratio Lower Upper

266 100

264 61.7 50.5 75.7

268 64.8 51.0 76.6





#72

Phenanthrene

Concen: 18.506 ng/ul

RT: 17.563 min Scan# 2443

Delta R.T. 0.000 min

Lab File: BP013813.D

Acq: 14 Feb 2023 09:29

Instrument :

BNA_P

ClientSampleId :

SSTD020762





Tgt Ion:178 Resp: 1082779

Ion Ratio Lower Upper

178 100

179 15.1 12.2 18.2

176 19.3 15.2 22.8



#74

Anthracene

Concen: 18.599 ng/ul

RT: 17.651 min Scan# 2459

Delta R.T. 0.000 min

Lab File: BP013813.D

Acq: 14 Feb 2023 09:29




Tgt Ion:178 Resp: 1094875


Ion Ratio Lower Upper

178 100

179 15.4 12.2 18.4

176 18.8 15.0 22.4









#75
1,2,3,4-Tetrachlorobenzene
Concen: 20.726 ng/uL
RT: 13.569 min Scan# 11
Delta R.T. 0.000 min
Lab File: BP013813.D
Acq: 14 Feb 2023 09:29




Instrument :
BNA_P
ClientSampleId :
SSTD020762


Tgt Ion:216 Resp: 305918
Ion Ratio Lower Upper
216 100
214 76.4 62.6 94.0
218 49.3 39.0 58.4



#76
Pentachlorobenzene
Concen: 20.248 ng/uL
RT: 15.080 min Scan# 2022
Delta R.T. 0.000 min
Lab File: BP013813.D
Acq: 14 Feb 2023 09:29

Tgt Ion:250 Resp: 308407
Ion Ratio Lower Upper
250 100
248 63.5 51.1 76.7
252 63.5 51.0 76.6









#77
Carbazole
Concen: 19.874 ng/ul
RT: 17.915 min Scan# 2503
Delta R.T. 0.000 min
Lab File: BP013813.D
Acq: 14 Feb 2023 09:29




Instrument :
BNA_P
ClientSampleId :
SSTD020762


Tgt Ion:167 Resp: 1030539
Ion Ratio Lower Upper
167 100
166 21.4 17.3 25.9
139 12.3 10.4 15.6



#78
Di-n-butylphthalate
Concen: 18.669 ng/ul
RT: 18.451 min Scan# 2595
Delta R.T. 0.000 min
Lab File: BP013813.D
Acq: 14 Feb 2023 09:29

Tgt Ion:149 Resp: 1174149
Ion Ratio Lower Upper
149 100
150 9.3 7.1 10.7
104 5.4 4.1 6.1





#80

Fluoranthene

Concen: 21.396 ng/ul

RT: 19.509 min Scan# 21

Delta R.T. 0.000 min

Lab File: BP013813.D

Acq: 14 Feb 2023 09:29

Instrument :

BNA_P

ClientSampleId :

SSTD020762





Tgt Ion:202 Resp: 1312378

Ion Ratio Lower Upper

202 100

101 9.4 8.6 12.8

100 7.1 6.3 9.5



#82

Pyrene

Concen: 20.650 ng/ul

RT: 19.862 min Scan# 2835

Delta R.T. 0.000 min

Lab File: BP013813.D

Acq: 14 Feb 2023 09:29




Tgt Ion:202 Resp: 1350423


Ion Ratio Lower Upper

202 100

101 11.0 10.3 15.5

100 9.0 8.5 12.7





#83

Butylbenzylphthalate

Concen: 20.888 ng/ul

RT: 20.715 min Scan# 2979

Delta R.T. 0.000 min

Lab File: BP013813.D

Acq: 14 Feb 2023 09:29

Instrument :

BNA_P

ClientSampleId :

SSTD020762





Tgt Ion:149 Resp: 467526

Ion Ratio Lower Upper

149 100

91 71.4 57.7 86.5

206 21.2 16.7 25.1



#84

3,3'-Dichlorobenzidine

Concen: 17.600 ng/ul

RT: 21.504 min Scan# 3114

Delta R.T. 0.000 min

Lab File: BP013813.D

Acq: 14 Feb 2023 09:29




Tgt Ion:252 Resp: 374268


Ion Ratio Lower Upper

252 100

254 64.4 51.8 77.6

126 10.5 7.9 11.9





#85

Benzo(a)anthracene

Concen: 18.562 ng/ul

RT: 21.580 min Scan# 3110

Delta R.T. 0.000 min

Lab File: BP013813.D

Acq: 14 Feb 2023 09:29

Instrument :

BNA_P

ClientSampleId :

SSTD020762





Tgt Ion:228 Resp: 1198345

Ion Ratio Lower Upper

228 100

229 19.6 15.4 23.2

226 27.2 21.8 32.6



#86

Bis(2-ethylhexyl)phthalate

Concen: 19.785 ng/ul

RT: 21.480 min Scan# 3110

Delta R.T. 0.000 min

Lab File: BP013813.D

Acq: 14 Feb 2023 09:29




Tgt Ion:149 Resp: 678090


Ion Ratio Lower Upper

149 100

167 28.0 21.5 32.3

279 4.9 3.2 4.8#





#87

Chrysene

Concen: 18.733 ng/ul

RT: 21.633 min Scan# 3135

Delta R.T. 0.000 min

Lab File: BP013813.D

Acq: 14 Feb 2023 09:29

Instrument :

BNA_P

ClientSampleId :

SSTD020762





Tgt Ion:228 Resp: 1141038

Ion Ratio Lower Upper

228 100

226 29.8 25.0 37.6

229 19.3 16.4 24.6



#89

Di-n-octyl phthalate

Concen: 17.598 ng/ul




RT: 22.456 min Scan# 3276


Delta R.T. 0.000 min

Lab File: BP013813.D

Acq: 14 Feb 2023 09:29

Tgt Ion:149 Resp: 1058786









#90
Benzo(b)fluoranthene
Concen: 18.265 ng/ul
RT: 23.374 min Scan# 3430
Delta R.T. 0.000 min
Lab File: BP013813.D
Acq: 14 Feb 2023 09:29




Instrument :
BNA_P
ClientSampleId :
SSTD020762


Tgt Ion:252 Resp: 1064920
Ion Ratio Lower Upper
252 100
253 22.0 17.5 26.3
125 9.0 7.8 11.6



#91
Benzo(k)fluoranthene
Concen: 18.592 ng/ul
RT: 23.427 min Scan# 3441
Delta R.T. 0.000 min
Lab File: BP013813.D
Acq: 14 Feb 2023 09:29

Tgt Ion:252 Resp: 1075356
Ion Ratio Lower Upper
252 100
253 21.9 17.1 25.7
125 8.6 7.2 10.8





#93

Benzo(a)pyrene

Concen: 18.735 ng/ul

RT: 24.050 min Scan# 3

Delta R.T. 0.000 min

Lab File: BP013813.D

Acq: 14 Feb 2023 09:29

Instrument :

BNA_P

ClientSampleId :

SSTD020762





Tgt Ion:252 Resp: 925754

Ion Ratio Lower Upper

252 100

253 21.6 17.4 26.0

125 9.4 8.2 12.2



#94

Indeno(1,2,3-cd)pyrene

Concen: 21.138 ng/ul

RT: 26.874 min Scan# 4027

Delta R.T. 0.000 min

Lab File: BP013813.D

Acq: 14 Feb 2023 09:29




Tgt Ion:276 Resp: 1171637


Ion Ratio Lower Upper

276 100

138 21.2 18.4 27.6

277 24.8 19.7 29.5





#95

Dibenzo(a,h)anthracene

Concen: 21.234 ng/ul

RT: 26.891 min Scan# 4026

Delta R.T. 0.000 min

Lab File: BP013813.D

Acq: 14 Feb 2023 09:29

Instrument :

BNA_P

ClientSampleId :

SSTD020762





Tgt Ion:278 Resp: 983120

Ion Ratio Lower Upper

278 100

139 13.7 12.0 18.0

279 23.6 18.9 28.3



#96

Benzo(g,h,i)perylene

Concen: 21.599 ng/ul

RT: 27.709 min Scan# 4169

Delta R.T. 0.000 min

Lab File: BP013813.D

Acq: 14 Feb 2023 09:29

Tgt Ion:276 Resp: 953129

Ion Ratio Lower Upper

276 100

138 18.7 15.9 23.9

277 22.7 19.0 28.6

