

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP021423\
 Data File : BP013816.D
 Acq On : 14 Feb 2023 14:26
 Operator : CG/JU
 Sample : 01512-10
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 C0B84

Quant Time: Feb 15 01:21:41 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP020723.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed Feb 15 01:19:22 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.122	152	199600	20.000	ng/ul	0.00
20) Naphthalene-d8	10.952	136	778254	20.000	ng/ul	0.00
38) Acenaphthene-d10	14.757	164	493306	20.000	ng/ul	0.00
64) Phenanthrene-d10	17.516	188	1087158	20.000	ng/ul	0.00
79) Chrysene-d12	21.592	240	919440	20.000	ng/ul	0.00
88) Perylene-d12	24.157	264	853479	20.000	ng/ul	-0.01
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.470	96	23467	4.873	ng/uL	0.00
4) Pyridine-d5	3.911	84	70202	4.917	ng/ul	0.00
7) Phenol-d5	7.269	99	77643	4.366	ng/ul	0.00
9) Bis-(2-Chloroethyl)eth...	7.440	67	287902	27.626	ng/ul	0.00
11) 2-Chlorophenol-d4	7.652	132	276461	21.118	ng/ul	0.00
15) 4-Methylphenol-d8	8.828	113	161721	11.126	ng/ul	0.00
21) Nitrobenzene-d5	9.299	128	168735	31.161	ng/ul	0.00
24) 2-Nitrophenol-d4	10.022	143	172686	27.621	ng/ul	0.00
28) 2,4-Dichlorophenol-d3	10.563	165	317675	24.934	ng/ul	0.00
31) 4-Chloroaniline-d4	11.087	131	331802	18.372	ng/ul	0.00
46) Dimethylphthalate-d6	14.163	166	1207496	31.078	ng/ul	0.00
49) Acenaphthylene-d8	14.457	160	1300971	30.932	ng/ul	0.00
54) 4-Nitrophenol-d4	14.945	143	16051	2.319	ng/ul	0.00
60) Fluorene-d10	15.751	176	1028698	31.084	ng/ul	0.00
65) 4,6-Dinitro-2-methylph...	15.863	200	130714	19.598	ng/ul	0.00
73) Anthracene-d10	17.616	188	1605018	32.287	ng/ul	0.00
81) Pyrene-d10	19.833	212	1845155	35.388	ng/ul	0.00
92) Benzo(a)pyrene-d12	23.992	264	1389686	32.496	ng/ul	0.00

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP021423\
Data File : BP013816.D
Acq On : 14 Feb 2023 14:26
Operator : CG/JU
Sample : 01512-10
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
C0B84

Quant Time: Feb 15 01:21:41 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP020723.M
Quant Title : SVOA CALIBRATION
QLast Update : Wed Feb 15 01:19:22 2023
Response via : Initial Calibration

