

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP021423\
 Data File : BP013819.D
 Acq On : 14 Feb 2023 16:09
 Operator : CG/JU
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SSTD020763

Quant Time: Feb 15 01:22:42 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP020723.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed Feb 15 01:19:22 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.116	152	181702	20.000	ng/ul	-0.01
20) Naphthalene-d8	10.946	136	691873	20.000	ng/ul	-0.01
38) Acenaphthene-d10	14.751	164	366642	20.000	ng/ul	-0.01
64) Phenanthrene-d10	17.504	188	719396	20.000	ng/ul	-0.01
79) Chrysene-d12	21.580	240	675374	20.000	ng/ul	-0.02
88) Perylene-d12	24.145	264	794723	20.000	ng/ul	-0.02
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.469	96	38813	8.854	ng/uL	0.00
4) Pyridine-d5	3.899	84	256449	19.732	ng/ul	0.00
7) Phenol-d5	7.263	99	289729	17.896	ng/ul	0.00
9) Bis-(2-Chloroethyl)eth...	7.434	67	173538	18.292	ng/ul	-0.01
11) 2-Chlorophenol-d4	7.640	132	231217	19.401	ng/ul	-0.01
15) 4-Methylphenol-d8	8.822	113	217849	16.463	ng/ul	-0.01
21) Nitrobenzene-d5	9.287	128	96873	20.124	ng/ul	-0.01
24) 2-Nitrophenol-d4	10.016	143	109282	19.662	ng/ul	-0.01
28) 2,4-Dichlorophenol-d3	10.557	165	220084	19.431	ng/ul	-0.01
31) 4-Chloroaniline-d4	11.075	131	271045	16.882	ng/ul	-0.01
46) Dimethylphthalate-d6	14.151	166	544589	18.859	ng/ul	-0.01
49) Acenaphthylene-d8	14.445	160	630208	20.161	ng/ul	-0.01
54) 4-Nitrophenol-d4	14.934	143	76638	14.897	ng/ul	-0.01
60) Fluorene-d10	15.739	176	456625	18.564	ng/ul	-0.02
65) 4,6-Dinitro-2-methylph...	15.851	200	76687	17.376	ng/ul	-0.02
73) Anthracene-d10	17.604	188	662290	20.133	ng/ul	-0.01
81) Pyrene-d10	19.822	212	761013	19.870	ng/ul	-0.01
92) Benzo(a)pyrene-d12	23.974	264	777722	19.530	ng/ul	-0.02
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.505	88	41324	8.781	ng/uL	95
5) Pyridine	3.922	79	255828	19.918	ng/ul	96
6) Benzaldehyde	7.246	77	133755	18.321	ng/ul	98
8) Phenol	7.293	94	301155	18.122	ng/ul	100
10) Bis(2-Chloroethyl)ether	7.528	93	242733	18.228	ng/ul	99
12) 2-Chlorophenol	7.675	128	239098	19.553	ng/ul	98
13) 2-Methylphenol	8.557	108	216298	17.052	ng/ul	98
14) 2,2'-oxybis(1-Chloropr...	8.646	45	264540	17.464	ng/ul	98
16) Acetophenone	8.946	105	359855	16.958	ng/ul	99
17) N-Nitroso-di-n-propyla...	8.928	70	164964	15.654	ng/ul	99
18) 4-Methylphenol	8.887	108	236692	16.901	ng/ul	95
19) Hexachloroethane	9.210	117	100688	20.108	ng/ul	98
22) Nitrobenzene	9.334	77	255820	20.199	ng/ul	98
23) Isophorone	9.857	82	451666	16.904	ng/ul	100
25) 2-Nitrophenol	10.046	139	119941	19.910	ng/ul	98
26) 2,4-Dimethylphenol	10.099	107	256942	19.156	ng/ul	96
27) Bis(2-Chloroethoxy)met...	10.340	93	303259	18.794	ng/ul	99
29) 2,4-Dichlorophenol	10.581	162	216889	19.450	ng/ul	99
30) Naphthalene	10.993	128	735876	19.980	ng/ul	99
32) 4-Chloroaniline	11.098	127	274767	16.976	ng/ul	100
33) Hexachlorobutadiene	11.275	225	169163	21.536	ng/ul	98
34) Caprolactam	11.881	113	44330	11.329	ng/ul	98
35) 4-Chloro-3-methylphenol	12.204	107	212326	16.493	ng/ul	97
36) 2-Methylnaphthalene	12.587	142	465916	17.993	ng/ul	98

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP021423\
 Data File : BP013819.D
 Acq On : 14 Feb 2023 16:09
 Operator : CG/JU
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 2 Sample Multiplier: 1


Instrument :
 BNA_P
 ClientSampleId :
 SSTD020763


Quant Time: Feb 15 01:22:42 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP020723.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed Feb 15 01:19:22 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.804	142	474011	17.716	ng/ul	99
39) 1,2,4,5-Tetrachloroben...	12.951	216	281729	24.384	ng/ul	98
40) Hexachlorocyclopentadiene	12.934	237	178635	24.079	ng/ul	97
41) 2,4,6-Trichlorophenol	13.187	196	163807	21.605	ng/ul	99
42) 2,4,5-Trichlorophenol	13.257	196	175845	21.427	ng/ul	98
43) 1,1'-Biphenyl	13.587	154	617555	22.448	ng/ul	100
44) 2-Chloronaphthalene	13.634	162	492672	22.764	ng/ul	99
45) 2-Nitroaniline	13.834	65	95776	18.373	ng/ul	94
47) Dimethylphthalate	14.198	163	533324	18.621	ng/ul	100
48) 2,6-Dinitrotoluene	14.322	165	87689	17.523	ng/ul	98
50) Acenaphthylene	14.475	152	680421	20.177	ng/ul	99
51) 3-Nitroaniline	14.651	138	78330	15.832	ng/ul	96
52) Acenaphthene	14.816	153	472085	20.001	ng/ul	99
53) 2,4-Dinitrophenol	14.857	184	68369	19.159	ng/ul	92
55) 4-Nitrophenol	14.945	109	66942	14.829	ng/ul	97
56) Dibenzofuran	15.145	168	666355	19.692	ng/ul	99
57) 2,4-Dinitrotoluene	15.104	165	125437	16.814	ng/ul	98
58) 2,3,4,6-Tetrachlorophenol	15.369	232	139991	18.098	ng/ul#	98
59) Diethylphthalate	15.557	149	485821	16.892	ng/ul	100
61) Fluorene	15.798	166	521724	18.899	ng/ul	100
62) 4-Chlorophenyl-phenyle...	15.786	204	277115	19.305	ng/ul	97
63) 4-Nitroaniline	15.816	138	64729	13.816	ng/ul	100
66) 4,6-Dinitro-2-methylph...	15.869	198	78805	18.213	ng/ul	97
67) N-Nitrosodiphenylamine	15.998	169	423784	21.673	ng/ul	99
68) 4-Bromophenyl-phenylether	16.686	248	164226	21.617	ng/ul	98
69) Hexachlorobenzene	16.804	284	189038	21.355	ng/ul	99
70) Atrazine	16.951	200	145507	18.608	ng/ul	98
71) Pentachlorophenol	17.145	266	101778	17.668	ng/ul	98
72) Phenanthrene	17.545	178	780612	20.506	ng/ul	100
74) Anthracene	17.639	178	776538	20.275	ng/ul	100
75) 1,2,3,4-Tetrachloroben...	13.557	216	269975	28.112	ng/uL	99
76) Pentachlorobenzene	15.069	250	243183	24.539	ng/uL	98
77) Carbazole	17.904	167	655702	19.435	ng/ul	99
78) Di-n-butylphthalate	18.433	149	710186	17.356	ng/ul	99
80) Fluoranthene	19.498	202	875916	19.977	ng/ul	99
82) Pyrene	19.851	202	933678	19.973	ng/ul	96
83) Butylbenzylphthalate	20.704	149	276501	17.283	ng/ul	99
84) 3,3'-Dichlorobenzidine	21.492	252	293921	19.336	ng/ul	99
85) Benzo(a)anthracene	21.569	228	924022	20.023	ng/ul	99
86) Bis(2-ethylhexyl)phtha...	21.463	149	414266	16.910	ng/ul	99
87) Chrysene	21.621	228	902881	20.737	ng/ul	97
89) Di-n-octyl phthalate	22.439	149	697913	12.762	ng/ul	100
90) Benzo(b)fluoranthene	23.357	252	992776	18.734	ng/ul	99
91) Benzo(k)fluoranthene	23.404	252	992501	18.878	ng/ul	99
93) Benzo(a)pyrene	24.027	252	897263	19.977	ng/ul	99
94) Indeno(1,2,3-cd)pyrene	26.839	276	1259780	25.006	ng/ul	98
95) Dibenzo(a,h)anthracene	26.862	278	1054113	25.048	ng/ul	99
96) Benzo(g,h,i)perylene	27.674	276	1043179	26.008	ng/ul	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Quant Time: Feb 15 01:22:42 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP020723.M
Quant Title : SVOA CALIBRATION
QLast Update : Wed Feb 15 01:19:22 2023
Response via : Initial Calibration









#2
 1,4-Dioxane
 Concen: 8.781 ng/uL
 RT: 3.505 min Scan# 54
 Delta R.T. 0.000 min
 Lab File: BP013819.D
 Acq: 14 Feb 2023 16:09




Instrument :
 BNA_P
 ClientSampleId :
 SSTD020763


Tgt Ion:	88	Resp:	41324
Ion	Ratio	Lower	Upper
88	100		
43	24.7	20.1	30.1
58	71.6	53.0	79.6



#5
 Pyridine
 Concen: 19.918 ng/uL
 RT: 3.922 min Scan# 125
 Delta R.T. 0.000 min
 Lab File: BP013819.D
 Acq: 14 Feb 2023 16:09

Tgt Ion:	79	Resp:	255828
Ion	Ratio	Lower	Upper
79	100		
52	64.7	55.1	82.7
51	31.4	25.7	38.5









#6
Benzaldehyde
Concen: 18.321 ng/ul
RT: 7.246 min Scan# 691
Delta R.T. -0.011 min
Lab File: BP013819.D
Acq: 14 Feb 2023 16:09




Instrument :
BNA_P
ClientSampleId :
SSTD020763


Tgt Ion: 77 Resp: 133755
Ion Ratio Lower Upper
77 100
105 101.5 79.5 119.3
106 97.3 76.5 114.7



#8
Phenol
Concen: 18.122 ng/ul
RT: 7.293 min Scan# 698
Delta R.T. -0.006 min
Lab File: BP013819.D
Acq: 14 Feb 2023 16:09

Tgt Ion: 94 Resp: 301155
Ion Ratio Lower Upper
94 100
65 27.6 22.2 33.4
66 37.2 29.6 44.4









#10
Bis(2-Chloroethyl)ether
Concen: 18.228 ng/ul
RT: 7.528 min Scan# 71
Delta R.T. -0.011 min
Lab File: BP013819.D
Acq: 14 Feb 2023 16:09




Instrument :
BNA_P
ClientSampleId :
SSTD020763


Tgt Ion: 93 Resp: 242733
Ion Ratio Lower Upper
93 100
63 73.2 59.4 89.0
95 33.0 25.8 38.8



#12
2-Chlorophenol
Concen: 19.553 ng/ul
RT: 7.675 min Scan# 763
Delta R.T. -0.006 min
Lab File: BP013819.D
Acq: 14 Feb 2023 16:09

Tgt Ion:128 Resp: 239098
Ion Ratio Lower Upper
128 100
64 44.9 37.3 55.9
130 31.6 26.3 39.5









#13
2-Methylphenol
Concen: 17.052 ng/ul
RT: 8.557 min Scan# 913
Delta R.T. -0.006 min
Lab File: BP013819.D
Acq: 14 Feb 2023 16:09




Instrument :
BNA_P
ClientSampleId :
SSTD020763


Tgt Ion:108 Resp: 216298
Ion Ratio Lower Upper
108 100
107 91.4 71.9 107.9



#14
2,2'-oxybis(1-Chloropropane)
Concen: 17.464 ng/ul
RT: 8.646 min Scan# 928
Delta R.T. -0.011 min
Lab File: BP013819.D
Acq: 14 Feb 2023 16:09

Tgt Ion: 45 Resp: 264540
Ion Ratio Lower Upper
45 100
77 18.5 13.4 20.2
79 13.5 11.0 16.4









#16
Acetophenone
Concen: 16.958 ng/ul
RT: 8.946 min Scan# 91
Delta R.T. -0.011 min
Lab File: BP013819.D
Acq: 14 Feb 2023 16:09




Instrument :
BNA_P
ClientSampleId :
SSTD020763


Tgt Ion:105 Resp: 359855
Ion Ratio Lower Upper
105 100
77 76.6 61.0 91.4
51 23.9 19.6 29.4



#17
N-Nitroso-di-n-propylamine
Concen: 15.654 ng/ul
RT: 8.928 min Scan# 976
Delta R.T. -0.011 min
Lab File: BP013819.D
Acq: 14 Feb 2023 16:09

Tgt Ion: 70 Resp: 164964
Ion Ratio Lower Upper
70 100
42 49.5 39.8 59.8
101 10.1 7.4 11.2
130 19.5 14.6 22.0









#18
4-Methylphenol
Concen: 16.901 ng/ul
RT: 8.887 min Scan# 90
Delta R.T. -0.006 min
Lab File: BP013819.D
Acq: 14 Feb 2023 16:09




Instrument :
BNA_P
ClientSampleId :
SSTD020763


Tgt Ion:108 Resp: 236692
Ion Ratio Lower Upper
108 100
107 108.4 90.9 136.3



#19
Hexachloroethane
Concen: 20.108 ng/ul
RT: 9.210 min Scan# 1024
Delta R.T. -0.011 min
Lab File: BP013819.D
Acq: 14 Feb 2023 16:09

Tgt Ion:117 Resp: 100688
Ion Ratio Lower Upper
117 100
201 111.5 86.6 129.8
199 69.3 55.2 82.8









#22
Nitrobenzene
Concen: 20.199 ng/ul
RT: 9.334 min Scan# 1045
Delta R.T. -0.006 min
Lab File: BP013819.D
Acq: 14 Feb 2023 16:09




Instrument :
BNA_P
ClientSampleId :
SSTD020763


Tgt Ion: 77 Resp: 255820
Ion Ratio Lower Upper
77 100
123 44.9 35.1 52.7
65 13.9 10.6 16.0



#23
Isophorone
Concen: 16.904 ng/ul
RT: 9.857 min Scan# 1134
Delta R.T. -0.011 min
Lab File: BP013819.D
Acq: 14 Feb 2023 16:09

Tgt Ion: 82 Resp: 451666
Ion Ratio Lower Upper
82 100
95 7.4 6.0 9.0
138 17.2 13.7 20.5









#25
2-Nitrophenol
Concen: 19.910 ng/ul
RT: 10.046 min Scan# 1166
Delta R.T. -0.011 min
Lab File: BP013819.D
Acq: 14 Feb 2023 16:09




Instrument :
BNA_P
ClientSampleId :
SSTD020763


Tgt Ion:139 Resp: 119941
Ion Ratio Lower Upper
139 100
65 52.1 39.9 59.9
109 33.8 27.4 41.0



#26
2,4-Dimethylphenol
Concen: 19.156 ng/ul
RT: 10.099 min Scan# 1175
Delta R.T. -0.011 min
Lab File: BP013819.D
Acq: 14 Feb 2023 16:09

Tgt Ion:107 Resp: 256942
Ion Ratio Lower Upper
107 100
121 48.7 37.8 56.6
122 83.5 63.7 95.5









#27
Bis(2-Chloroethoxy)methane
Concen: 18.794 ng/ul
RT: 10.340 min Scan# 1216
Delta R.T. -0.011 min
Lab File: BP013819.D
Acq: 14 Feb 2023 16:09




Instrument :
BNA_P
ClientSampleId :
SSTD020763


Tgt Ion: 93 Resp: 303259
Ion Ratio Lower Upper
93 100
95 31.6 25.0 37.6
123 10.6 8.2 12.4



#29
2,4-Dichlorophenol
Concen: 19.450 ng/ul
RT: 10.581 min Scan# 1257
Delta R.T. -0.011 min
Lab File: BP013819.D
Acq: 14 Feb 2023 16:09

Tgt Ion: 162 Resp: 216889
Ion Ratio Lower Upper
162 100
164 65.3 51.4 77.2
98 34.8 27.5 41.3









#30
Naphthalene
Concen: 19.980 ng/ul
RT: 10.993 min Scan# 11
Delta R.T. -0.011 min
Lab File: BP013819.D
Acq: 14 Feb 2023 16:09




Instrument :
BNA_P
ClientSampleId :
SSTD020763


Tgt Ion:128 Resp: 735876
Ion Ratio Lower Upper
128 100
129 10.9 8.6 12.8
127 13.0 10.6 15.8



#32
4-Chloroaniline
Concen: 16.976 ng/ul
RT: 11.098 min Scan# 1345
Delta R.T. -0.011 min
Lab File: BP013819.D
Acq: 14 Feb 2023 16:09

Tgt Ion:127 Resp: 274767
Ion Ratio Lower Upper
127 100
129 33.0 26.2 39.4









#33
Hexachlorobutadiene
Concen: 21.536 ng/ul
RT: 11.275 min Scan# 11
Delta R.T. -0.011 min
Lab File: BP013819.D
Acq: 14 Feb 2023 16:09




Instrument :
BNA_P
ClientSampleId :
SSTD020763


Tgt Ion:225 Resp: 169163
Ion Ratio Lower Upper
225 100
223 61.9 50.6 76.0
227 62.9 49.0 73.4



#34
Caprolactam
Concen: 11.329 ng/ul
RT: 11.881 min Scan# 1478
Delta R.T. -0.006 min
Lab File: BP013819.D
Acq: 14 Feb 2023 16:09

Tgt Ion:113 Resp: 44330
Ion Ratio Lower Upper
113 100
55 152.9 120.9 181.3
56 115.1 93.8 140.8









#35
4-Chloro-3-methylphenol
Concen: 16.493 ng/ul
RT: 12.204 min Scan# 1533
Delta R.T. -0.011 min
Lab File: BP013819.D
Acq: 14 Feb 2023 16:09




Instrument :
BNA_P
ClientSampleId :
SSTD020763


Tgt Ion:107 Resp: 212326
Ion Ratio Lower Upper
107 100
144 23.6 20.0 30.0
142 75.0 62.1 93.1



#36
2-Methylnaphthalene
Concen: 17.993 ng/ul
RT: 12.587 min Scan# 1598
Delta R.T. -0.011 min
Lab File: BP013819.D
Acq: 14 Feb 2023 16:09

Tgt Ion:142 Resp: 465916
Ion Ratio Lower Upper
142 100
141 89.5 72.8 109.2





#37

1-Methylnaphthalene

Concen: 17.716 ng/ul

RT: 12.804 min Scan# 1

Delta R.T. -0.011 min

Lab File: BP013819.D

Acq: 14 Feb 2023 16:09

Instrument :

BNA_P

ClientSampleId :

SSTD020763





Tgt Ion:142 Resp: 474011

Ion Ratio Lower Upper

142 100

141 91.6 74.2 111.2

116 4.4 3.1 4.7



#39

1,2,4,5-Tetrachlorobenzene

Concen: 24.384 ng/ul

RT: 12.951 min Scan# 1660

Delta R.T. -0.011 min

Lab File: BP013819.D

Acq: 14 Feb 2023 16:09

Tgt Ion:216 Resp: 281729




Ion Ratio Lower Upper


216 100

214 80.0 62.6 93.8

179 19.9 15.4 23.2

108 16.1 12.7 19.1









#40
Hexachlorocyclopentadiene
Concen: 24.079 ng/ul
RT: 12.934 min Scan# 1
Delta R.T. -0.006 min
Lab File: BP013819.D
Acq: 14 Feb 2023 16:09




Instrument :
BNA_P
ClientSampleId :
SSTD020763


Tgt Ion:237 Resp: 178635
Ion Ratio Lower Upper
237 100
235 59.5 49.8 74.8
272 12.4 10.1 15.1



#41
2,4,6-Trichlorophenol
Concen: 21.605 ng/ul
RT: 13.187 min Scan# 1700
Delta R.T. -0.011 min
Lab File: BP013819.D
Acq: 14 Feb 2023 16:09

Tgt Ion:196 Resp: 163807
Ion Ratio Lower Upper
196 100
198 94.5 75.1 112.7
200 31.3 24.3 36.5









#42
2,4,5-Trichlorophenol
Concen: 21.427 ng/ul
RT: 13.257 min Scan# 1712
Delta R.T. -0.011 min
Lab File: BP013819.D
Acq: 14 Feb 2023 16:09




Instrument :
BNA_P
ClientSampleId :
SSTD020763


Tgt Ion:196 Resp: 175845
Ion Ratio Lower Upper
196 100
198 94.8 77.0 115.4
200 29.6 25.8 38.6



#43
1,1'-Biphenyl
Concen: 22.448 ng/ul
RT: 13.587 min Scan# 1768
Delta R.T. -0.011 min
Lab File: BP013819.D
Acq: 14 Feb 2023 16:09

Tgt Ion:154 Resp: 617555
Ion Ratio Lower Upper
154 100
153 39.9 31.9 47.9
76 13.0 9.9 14.9





#44

2-Chloronaphthalene

Concen: 22.764 ng/ul

RT: 13.634 min Scan# 1

Delta R.T. -0.011 min

Lab File: BP013819.D

Acq: 14 Feb 2023 16:09

Instrument :

BNA_P

ClientSampleId :

SSTD020763





Tgt Ion:162 Resp: 492672

Ion Ratio Lower Upper

162 100

164 33.3 26.1 39.1

127 38.3 31.1 46.7



#45

2-Nitroaniline

Concen: 18.373 ng/ul

RT: 13.834 min Scan# 1810

Delta R.T. -0.011 min

Lab File: BP013819.D

Acq: 14 Feb 2023 16:09




Tgt Ion: 65 Resp: 95776


Ion Ratio Lower Upper

65 100

92 76.2 58.6 88.0

138 109.6 81.2 121.8









#47
Dimethylphthalate
Concen: 18.621 ng/ul
RT: 14.198 min Scan# 1872
Delta R.T. -0.011 min
Lab File: BP013819.D
Acq: 14 Feb 2023 16:09




Instrument :
BNA_P
ClientSampleId :
SSTD020763


Tgt Ion:163 Resp: 533324
Ion Ratio Lower Upper
163 100
194 4.2 3.4 5.2
164 9.8 7.8 11.8



#48
2,6-Dinitrotoluene
Concen: 17.523 ng/ul
RT: 14.322 min Scan# 1893
Delta R.T. -0.011 min
Lab File: BP013819.D
Acq: 14 Feb 2023 16:09

Tgt Ion:165 Resp: 87689
Ion Ratio Lower Upper
165 100
89 60.8 47.5 71.3
121 21.4 16.6 24.8









#50
Acenaphthylene
Concen: 20.177 ng/ul
RT: 14.475 min Scan# 1919
Delta R.T. -0.011 min
Lab File: BP013819.D
Acq: 14 Feb 2023 16:09




Instrument :
BNA_P
ClientSampleId :
SSTD020763


Tgt Ion:152 Resp: 680421
Ion Ratio Lower Upper
152 100
151 19.9 15.8 23.6
153 13.4 10.4 15.6



#51
3-Nitroaniline
Concen: 15.832 ng/ul
RT: 14.651 min Scan# 1949
Delta R.T. -0.011 min
Lab File: BP013819.D
Acq: 14 Feb 2023 16:09

Tgt Ion:138 Resp: 78330
Ion Ratio Lower Upper
138 100
108 11.8 9.4 14.2
92 116.8 97.4 146.0









#52
Acenaphthene
Concen: 20.001 ng/ul
RT: 14.816 min Scan# 1977
Delta R.T. -0.011 min
Lab File: BP013819.D
Acq: 14 Feb 2023 16:09




Instrument :
BNA_P
ClientSampleId :
SSTD020763


Tgt Ion:153 Resp: 472085
Ion Ratio Lower Upper
153 100
152 47.4 37.4 56.0
154 87.1 70.1 105.1



#53
2,4-Dinitrophenol
Concen: 19.159 ng/ul
RT: 14.857 min Scan# 1984
Delta R.T. -0.011 min
Lab File: BP013819.D
Acq: 14 Feb 2023 16:09

Tgt Ion:184 Resp: 68369
Ion Ratio Lower Upper
184 100
63 54.9 49.5 74.3
154 62.1 54.3 81.5









#55
4-Nitrophenol
Concen: 14.829 ng/ul
RT: 14.945 min Scan# 1999
Delta R.T. -0.017 min
Lab File: BP013819.D
Acq: 14 Feb 2023 16:09




Instrument :
BNA_P
ClientSampleId :
SSTD020763


Tgt Ion:109 Resp: 66942
Ion Ratio Lower Upper
109 100
139 122.5 96.0 144.0
65 123.2 95.8 143.6



#56
Dibenzofuran
Concen: 19.692 ng/ul
RT: 15.145 min Scan# 2033
Delta R.T. -0.017 min
Lab File: BP013819.D
Acq: 14 Feb 2023 16:09

Tgt Ion:168 Resp: 666355
Ion Ratio Lower Upper
168 100
139 37.4 29.6 44.4





#57

2,4-Dinitrotoluene

Concen: 16.814 ng/ul

RT: 15.104 min Scan# 2026

Delta R.T. -0.011 min

Lab File: BP013819.D

Acq: 14 Feb 2023 16:09

Instrument :

BNA_P

ClientSampleId :

SSTD020763





Tgt Ion:165 Resp: 125437

Ion Ratio Lower Upper

165 100

63 41.5 32.0 48.0

182 3.0 2.4 3.6



#58

2,3,4,6-Tetrachlorophenol

Concen: 18.098 ng/ul

RT: 15.369 min Scan# 2071

Delta R.T. -0.017 min

Lab File: BP013819.D

Acq: 14 Feb 2023 16:09

Tgt Ion:232 Resp: 139991




Ion Ratio Lower Upper


232 100

131 41.5 33.0 49.6

130 2.3 1.5 2.3#

166 30.3 22.6 33.8









#59
Diethylphthalate
Concen: 16.892 ng/ul
RT: 15.557 min Scan# 2103
Delta R.T. -0.017 min
Lab File: BP013819.D
Acq: 14 Feb 2023 16:09




Instrument :
BNA_P
ClientSampleId :
SSTD020763


Tgt Ion:149 Resp: 485821
Ion Ratio Lower Upper
149 100
177 22.4 17.8 26.8
150 12.3 10.1 15.1



#61
Fluorene
Concen: 18.899 ng/ul
RT: 15.798 min Scan# 2144
Delta R.T. -0.011 min
Lab File: BP013819.D
Acq: 14 Feb 2023 16:09

Tgt Ion:166 Resp: 521724
Ion Ratio Lower Upper
166 100
165 98.6 79.1 118.7
167 13.5 10.8 16.2









#62
4-Chlorophenyl-phenylether
Concen: 19.305 ng/ul
RT: 15.786 min Scan# 2142
Delta R.T. -0.011 min
Lab File: BP013819.D
Acq: 14 Feb 2023 16:09




Instrument :
BNA_P
ClientSampleId :
SSTD020763


Tgt Ion:204 Resp: 277115
Ion Ratio Lower Upper
204 100
206 34.8 26.7 40.1
141 57.7 44.6 66.8



#63
4-Nitroaniline
Concen: 13.816 ng/ul
RT: 15.816 min Scan# 2147
Delta R.T. -0.011 min
Lab File: BP013819.D
Acq: 14 Feb 2023 16:09

Tgt Ion:138 Resp: 64729
Ion Ratio Lower Upper
138 100
92 54.5 43.8 65.8
108 86.9 69.8 104.6









#66
4,6-Dinitro-2-methylphenol
Concen: 18.213 ng/ul
RT: 15.869 min Scan# 2157
Delta R.T. -0.011 min
Lab File: BP013819.D
Acq: 14 Feb 2023 16:09




Instrument :
BNA_P
ClientSampleId :
SSTD020763


Tgt Ion:198 Resp: 78805
Ion Ratio Lower Upper
198 100
182 5.4 4.1 6.1
77 23.9 20.3 30.5



#67
N-Nitrosodiphenylamine
Concen: 21.673 ng/ul
RT: 15.998 min Scan# 2178
Delta R.T. -0.017 min
Lab File: BP013819.D
Acq: 14 Feb 2023 16:09

Tgt Ion:169 Resp: 423784
Ion Ratio Lower Upper
169 100
168 66.3 53.5 80.3
167 36.5 29.8 44.6









#68
4-Bromophenyl-phenylether
Concen: 21.617 ng/ul
RT: 16.686 min Scan# 21
Delta R.T. -0.011 min
Lab File: BP013819.D
Acq: 14 Feb 2023 16:09




Instrument :
BNA_P
ClientSampleId :
SSTD020763


Tgt Ion:248 Resp: 164226
Ion Ratio Lower Upper
248 100
250 98.0 76.0 114.0
141 69.1 56.0 84.0



#69
Hexachlorobenzene
Concen: 21.355 ng/ul
RT: 16.804 min Scan# 2315
Delta R.T. -0.011 min
Lab File: BP013819.D
Acq: 14 Feb 2023 16:09

Tgt Ion:284 Resp: 189038
Ion Ratio Lower Upper
284 100
142 32.0 25.0 37.4
249 31.1 24.9 37.3









#70
Atrazine
Concen: 18.608 ng/ul
RT: 16.951 min Scan# 21
Delta R.T. -0.011 min
Lab File: BP013819.D
Acq: 14 Feb 2023 16:09




Instrument :
BNA_P
ClientSampleId :
SSTD020763


Tgt Ion:200 Resp: 145507
Ion Ratio Lower Upper
200 100
173 24.3 20.4 30.6
215 50.6 39.3 58.9



#71
Pentachlorophenol
Concen: 17.668 ng/ul
RT: 17.145 min Scan# 2373
Delta R.T. -0.017 min
Lab File: BP013819.D
Acq: 14 Feb 2023 16:09

Tgt Ion:266 Resp: 101778
Ion Ratio Lower Upper
266 100
264 63.7 50.5 75.7
268 61.9 51.0 76.6









#72
Phenanthrene
Concen: 20.506 ng/ul
RT: 17.545 min Scan# 2441
Delta R.T. -0.017 min
Lab File: BP013819.D
Acq: 14 Feb 2023 16:09




Instrument :
BNA_P
ClientSampleId :
SSTD020763


Tgt Ion:178 Resp: 780612
Ion Ratio Lower Upper
178 100
179 15.2 12.2 18.2
176 18.8 15.2 22.8



#74
Anthracene
Concen: 20.275 ng/ul
RT: 17.639 min Scan# 2457
Delta R.T. -0.011 min
Lab File: BP013819.D
Acq: 14 Feb 2023 16:09

Tgt Ion:178 Resp: 776538
Ion Ratio Lower Upper
178 100
179 15.2 12.2 18.4
176 18.6 15.0 22.4





#75

1,2,3,4-Tetrachlorobenzene

Concen: 28.112 ng/uL

RT: 13.557 min Scan# 1

Delta R.T. -0.011 min

Lab File: BP013819.D

Acq: 14 Feb 2023 16:09

Instrument :

BNA_P

ClientSampleId :

SSTD020763





Tgt Ion:216 Resp: 269975

Ion Ratio Lower Upper

216 100

214 78.8 62.6 94.0

218 49.1 39.0 58.4



#76

Pentachlorobenzene

Concen: 24.539 ng/uL

RT: 15.069 min Scan# 2020

Delta R.T. -0.011 min

Lab File: BP013819.D

Acq: 14 Feb 2023 16:09




Tgt Ion:250 Resp: 243183


Ion Ratio Lower Upper

250 100

248 61.1 51.1 76.7

252 63.0 51.0 76.6









#77
 Carbazole
 Concen: 19.435 ng/ul
 RT: 17.904 min Scan# 2502
 Delta R.T. -0.011 min
 Lab File: BP013819.D
 Acq: 14 Feb 2023 16:09




Instrument :
 BNA_P
 ClientSampleId :
 SSTD020763


Tgt Ion:167 Resp: 655702
 Ion Ratio Lower Upper
 167 100
 166 21.4 17.3 25.9
 139 12.5 10.4 15.6



#78
 Di-n-butylphthalate
 Concen: 17.356 ng/ul
 RT: 18.433 min Scan# 2592
 Delta R.T. -0.017 min
 Lab File: BP013819.D
 Acq: 14 Feb 2023 16:09

Tgt Ion:149 Resp: 710186
 Ion Ratio Lower Upper
 149 100
 150 9.3 7.1 10.7
 104 5.2 4.1 6.1









#80
Fluoranthene
Concen: 19.977 ng/ul
RT: 19.498 min Scan# 21
Delta R.T. -0.011 min
Lab File: BP013819.D
Acq: 14 Feb 2023 16:09




Instrument :
BNA_P
ClientSampleId :
SSTD020763


Tgt Ion:202 Resp: 875916
Ion Ratio Lower Upper
202 100
101 10.3 8.6 12.8
100 7.7 6.3 9.5



#82
Pyrene
Concen: 19.973 ng/ul
RT: 19.851 min Scan# 2833
Delta R.T. -0.011 min
Lab File: BP013819.D
Acq: 14 Feb 2023 16:09

Tgt Ion:202 Resp: 933678
Ion Ratio Lower Upper
202 100
101 11.2 10.3 15.5
100 9.4 8.5 12.7





#83

Butylbenzylphthalate

Concen: 17.283 ng/ul

RT: 20.704 min Scan# 2978

Delta R.T. -0.011 min

Lab File: BP013819.D

Acq: 14 Feb 2023 16:09

Instrument :

BNA_P

ClientSampleId :

SSTD020763





Tgt Ion:149 Resp: 276501

Ion Ratio Lower Upper

149 100

91 72.9 57.7 86.5

206 22.3 16.7 25.1



#84

3,3'-Dichlorobenzidine

Concen: 19.336 ng/ul

RT: 21.492 min Scan# 3112

Delta R.T. -0.011 min

Lab File: BP013819.D

Acq: 14 Feb 2023 16:09




Tgt Ion:252 Resp: 293921


Ion Ratio Lower Upper

252 100

254 63.7 51.8 77.6

126 9.3 7.9 11.9





#85

Benzo(a)anthracene

Concen: 20.023 ng/ul

RT: 21.569 min Scan# 3125

Delta R.T. -0.011 min

Lab File: BP013819.D

Acq: 14 Feb 2023 16:09

Instrument :

BNA_P

ClientSampleId :

SSTD020763





Tgt Ion:228 Resp: 924022

Ion Ratio Lower Upper

228 100

229 19.5 15.4 23.2

226 26.7 21.8 32.6



#86

Bis(2-ethylhexyl)phthalate

Concen: 16.910 ng/ul

RT: 21.463 min Scan# 3107

Delta R.T. -0.017 min

Lab File: BP013819.D

Acq: 14 Feb 2023 16:09




Tgt Ion:149 Resp: 414266


Ion Ratio Lower Upper

149 100

167 27.2 21.5 32.3

279 4.6 3.2 4.8





#87

Chrysene

Concen: 20.737 ng/ul

RT: 21.621 min Scan# 3135

Delta R.T. -0.011 min

Lab File: BP013819.D

Acq: 14 Feb 2023 16:09

Instrument :

BNA_P

ClientSampleId :

SSTD020763





Tgt Ion:228 Resp: 902881

Ion Ratio Lower Upper

228 100

226 29.5 25.0 37.6

229 19.5 16.4 24.6



#89

Di-n-octyl phthalate

Concen: 12.762 ng/ul




RT: 22.439 min Scan# 3273


Delta R.T. -0.017 min

Lab File: BP013819.D

Acq: 14 Feb 2023 16:09

Tgt Ion:149 Resp: 697913









#90
Benzo(b)fluoranthene
Concen: 18.734 ng/ul
RT: 23.357 min Scan# 3429
Delta R.T. -0.017 min
Lab File: BP013819.D
Acq: 14 Feb 2023 16:09




Instrument :
BNA_P
ClientSampleId :
SSTD020763


Tgt Ion:252 Resp: 992776
Ion Ratio Lower Upper
252 100
253 21.8 17.5 26.3
125 8.7 7.8 11.6



#91
Benzo(k)fluoranthene
Concen: 18.878 ng/ul
RT: 23.404 min Scan# 3437
Delta R.T. -0.023 min
Lab File: BP013819.D
Acq: 14 Feb 2023 16:09

Tgt Ion:252 Resp: 992501
Ion Ratio Lower Upper
252 100
253 21.8 17.1 25.7
125 9.1 7.2 10.8





#93

Benzo(a)pyrene

Concen: 19.977 ng/ul

RT: 24.027 min Scan# 3

Delta R.T. -0.023 min

Lab File: BP013819.D

Acq: 14 Feb 2023 16:09

Instrument :

BNA_P

ClientSampleId :

SSTD020763





Tgt Ion:252 Resp: 897263

Ion Ratio Lower Upper

252 100

253 21.7 17.4 26.0

125 11.2 8.2 12.2



#94

Indeno(1,2,3-cd)pyrene

Concen: 25.006 ng/ul

RT: 26.839 min Scan# 4021

Delta R.T. -0.035 min

Lab File: BP013819.D

Acq: 14 Feb 2023 16:09




Tgt Ion:276 Resp: 1259780


Ion Ratio Lower Upper

276 100

138 21.5 18.4 27.6

277 24.6 19.7 29.5









#95
Dibenzo(a,h)anthracene
Concen: 25.048 ng/ul
RT: 26.862 min Scan# 4026
Delta R.T. -0.029 min
Lab File: BP013819.D
Acq: 14 Feb 2023 16:09

Instrument :
BNA_P
ClientSampleId :
SSTD020763

Tgt Ion:278 Resp: 1054113
Ion Ratio Lower Upper
278 100
139 14.7 12.0 18.0
279 24.4 18.9 28.3



#96
Benzo(g,h,i)perylene
Concen: 26.008 ng/ul
RT: 27.674 min Scan# 4163
Delta R.T. -0.035 min
Lab File: BP013819.D
Acq: 14 Feb 2023 16:09

Tgt Ion:276 Resp: 1043179
Ion Ratio Lower Upper
276 100
138 19.3 15.9 23.9
277 24.0 19.0 28.6

