

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP021424\
 Data File : BP019711.D
 Acq On : 14 Feb 2024 20:36
 Operator : MA/JU
 Sample : P1395-02
 Misc :
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 B-3(10-12)

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 02/15/2024
 Supervised By :mohammad ahmed 02/15/2024

Quant Time: Feb 14 23:28:23 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP013124.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Feb 01 00:51:48 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.904	152	82328	20.000	ng	-0.01
21) Naphthalene-d8	10.704	136	287930	20.000	ng	-0.01
39) Acenaphthene-d10	14.539	164	191336	20.000	ng	-0.01
64) Phenanthrene-d10	17.292	188	428306	20.000	ng	-0.01
76) Chrysene-d12	21.392	240	431300	20.000	ng	0.00
86) Perylene-d12	23.815	264	505631	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.493	112	377501	73.912	ng	0.00
7) Phenol-d6	7.116	99	476698	69.220	ng	0.03
23) Nitrobenzene-d5	9.075	82	329973	49.659	ng	-0.01
42) 2,4,6-Tribromophenol	16.033	330	262344	93.908	ng	0.00
45) 2-Fluorobiphenyl	13.169	172	723340	46.735	ng	-0.01
79) Terphenyl-d14	19.869	244	1328617	50.660	ng	0.00
Target Compounds						Qvalue
6) Aniline	7.293	93	2793883m	417.518	ng	
10) Phenol	7.163	94	11476426	1672.016	ng	94
17) 2-Methylphenol	8.357	107	280239	60.645	ng	99
20) 3+4-Methylphenols	8.693	107	1416202	224.561	ng	93
27) 2,4-Dimethylphenol	9.898	122	577029	176.898	ng	100
31) Naphthalene	10.757	128	1236651	78.310	ng	99
32) Benzoic acid	10.122	122	308882m	97.190	ng	
37) 2-Methylnaphthalene	12.363	142	121926	11.157	ng	100
38) 1-Methylnaphthalene	12.587	142	73690	6.861	ng	92
46) 1,1'-Biphenyl	13.375	154	180443	11.816	ng	98
52) Acenaphthene	14.598	154	70518	6.460	ng	96
55) Dibenzofuran	14.939	168	60448	3.400	ng	96
58) Fluorene	15.586	166	49231	3.477	ng	# 98
71) Phenanthrene	17.333	178	190455	8.450	ng	98
73) Carbazole	17.704	167	54248	2.656	ng	# 90
74) Di-n-butylphthalate	18.274	149	6001879	243.007	ng	99
75) Fluoranthene	19.316	202	296437	10.766	ng	100
78) Pyrene	19.663	202	245383	8.061	ng	# 91
81) Benzo(a)anthracene	21.374	228	142676m	4.699	ng	
83) Chrysene	21.427	228	135162	4.689	ng	96
84) Bis(2-ethylhexyl)phtha...	21.327	149	7585165	438.202	ng	# 94
88) Benzo(b)fluoranthene	23.080	252	161146	5.057	ng	# 83
90) Benzo(a)pyrene	23.704	252	103839	3.662	ng	# 77

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP021424\
Data File : BP019711.D
Acq On : 14 Feb 2024 20:36
Operator : MA/JU
Sample : P1395-02
Misc :
ALS Vial : 20 Sample Multiplier: 1


Instrument :
BNA_P
ClientSampleId :
B-3(10-12)

Quant Time: Feb 14 23:28:23 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP013124.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Feb 01 00:51:48 2024
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 02/15/2024
Supervised By :mohammad ahmed 02/15/2024









#1
1,4-Dichlorobenzene-d4
Concen: 20.000 ng
RT: 7.904 min Scan# 81
Delta R.T. -0.012 min
Lab File: BP019711.D
Acq: 14 Feb 2024 20:36

Instrument :
BNA_P
ClientSampleId :
B-3(10-12)

Tgt Ion:152 Resp: 82328
Ion Ratio Lower Upper
152 100
150 150.8 121.8 182.8
115 58.5 49.9 74.9




Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 02/15/2024
Supervised By :mohammad ahmed 02/15/2024



#5
2-Fluorophenol
Concen: 73.912 ng
RT: 5.493 min Scan# 409
Delta R.T. 0.006 min
Lab File: BP019711.D
Acq: 14 Feb 2024 20:36


Tgt Ion:112 Resp: 377501
Ion Ratio Lower Upper
112 100
64 56.4 47.4 71.0
63 33.4 27.0 40.6





#6
Aniline
Concen: 417.518 ng m
RT: 7.293 min Scan# 715
Delta R.T. 0.041 min
Lab File: BP019711.D
Acq: 14 Feb 2024 20:36




Instrument :
BNA_P
ClientSampleId :
B-3(10-12)



Tgt Ion: 93 Resp: 279388
Ion Ratio Lower Upper
93 100
66 41.7 32.3 48.5
65 21.3 16.4 24.6




Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 02/15/2024
Supervised By :mohammad ahmed 02/15/2024



#7
Phenol-d6
Concen: 69.220 ng
RT: 7.116 min Scan# 685
Delta R.T. 0.030 min
Lab File: BP019711.D
Acq: 14 Feb 2024 20:36


Tgt Ion: 99 Resp: 476698
Ion Ratio Lower Upper
99 100
42 25.5 14.7 22.1#
71 42.3 34.8 52.2





#10
Phenol
Concen: 1672.016 ng
RT: 7.163 min Scan# 693
Delta R.T. 0.047 min
Lab File: BP019711.D
Acq: 14 Feb 2024 20:36




Instrument :
BNA_P
ClientSampleId :
B-3(10-12)



Tgt Ion: 94 Resp: 11476420
Ion Ratio Lower Upper
94 100
65 27.7 11.3 51.3
66 41.8 25.1 65.1




Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 02/15/2024
Supervised By :mohammad ahmed 02/15/2024



#17
2-Methylphenol
Concen: 60.645 ng
RT: 8.357 min Scan# 896
Delta R.T. -0.006 min
Lab File: BP019711.D
Acq: 14 Feb 2024 20:36


Tgt Ion: 107 Resp: 280239
Ion Ratio Lower Upper
107 100
108 110.9 87.0 130.6
77 54.0 43.5 65.3
79 54.0 43.6 65.4





#20
3+4-Methylphenols
Concen: 224.561 ng
RT: 8.693 min Scan# 91
Delta R.T. 0.000 min
Lab File: BP019711.D
Acq: 14 Feb 2024 20:36




Instrument :
BNA_P
ClientSampleId :
B-3(10-12)




Tgt Ion:107 Resp: 141620
Ion Ratio Lower Upper
107 100
108 81.4 65.5 105.5
77 36.7 21.8 61.8
79 28.5 14.1 54.1

Manual Integrations
APPROVED



Reviewed By :Yogesh Patel 02/15/2024
Supervised By :mohammad ahmed 02/15/2024




#21
Naphthalene-d8
Concen: 20.000 ng
RT: 10.704 min Scan# 1295
Delta R.T. -0.012 min
Lab File: BP019711.D
Acq: 14 Feb 2024 20:36



Tgt Ion:136 Resp: 287930
Ion Ratio Lower Upper
136 100
137 10.5 8.8 13.2
54 6.8 5.3 7.9
68 5.5 4.1 6.1





#23

Nitrobenzene-d5

Concen: 49.659 ng

RT: 9.075 min Scan# 1018

Delta R.T. -0.012 min

Lab File: BP019711.D

Acq: 14 Feb 2024 20:36

Instrument :

BNA_P

ClientSampleId :

B-3(10-12)

Tgt Ion: 82 Resp: 32997

Ion Ratio Lower Upper

82 100

128 38.6 29.7 44.5





54 47.4 36.8 55.2

Manual Integrations

APPROVED

Reviewed By :Yogesh Patel 02/15/2024

Supervised By :mohammad ahmed 02/15/2024



#27

2,4-Dimethylphenol

Concen: 176.898 ng

RT: 9.898 min Scan# 1158

Delta R.T. 0.000 min

Lab File: BP019711.D

Acq: 14 Feb 2024 20:36




Tgt Ion:122 Resp: 577029


Ion Ratio Lower Upper

122 100

107 132.5 105.9 158.9


121 59.5 47.5 71.3





#31
Naphthalene
Concen: 78.310 ng
RT: 10.757 min Scan# 1196
Delta R.T. -0.012 min
Lab File: BP019711.D
Acq: 14 Feb 2024 20:36




Instrument :
BNA_P
ClientSampleId :
B-3(10-12)



Tgt Ion:128 Resp: 123665
Ion Ratio Lower Upper
128 100
129 10.9 8.6 13.0
127 12.7 10.8 16.2




Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 02/15/2024
Supervised By :mohammad ahmed 02/15/2024



#32
Benzoic acid
Concen: 97.190 ng m
RT: 10.122 min Scan# 1196
Delta R.T. 0.077 min
Lab File: BP019711.D
Acq: 14 Feb 2024 20:36


Tgt Ion:122 Resp: 308882
Ion Ratio Lower Upper
122 100
105 110.9 118.8 158.8#
77 285.6 91.7 131.7#





#37
2-Methylnaphthalene
Concen: 11.157 ng
RT: 12.363 min Scan# 11
Delta R.T. -0.012 min
Lab File: BP019711.D
Acq: 14 Feb 2024 20:36




Instrument :
BNA_P
ClientSampleId :
B-3(10-12)



Tgt Ion:142 Resp: 121920
Ion Ratio Lower Upper
142 100
141 90.7 72.2 108.4
115 42.5 34.2 51.2




Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 02/15/2024
Supervised By :mohammad ahmed 02/15/2024



#38
1-Methylnaphthalene
Concen: 6.861 ng
RT: 12.587 min Scan# 1615
Delta R.T. -0.006 min
Lab File: BP019711.D
Acq: 14 Feb 2024 20:36

Tgt Ion:142 Resp: 73690
Ion Ratio Lower Upper
142 100
141 84.2 74.0 111.0
116 5.5 3.7 5.5





#39

Acenaphthene-d10

Concen: 20.000 ng

RT: 14.539 min Scan# 1947

Delta R.T. -0.012 min

Lab File: BP019711.D

Acq: 14 Feb 2024 20:36

Instrument :

BNA_P

ClientSampleId :

B-3(10-12)

Tgt Ion:164 Resp: 191330

Ion Ratio Lower Upper

164 100

162 101.1 81.2 121.8





160 44.0 35.1 52.7

Manual Integrations

APPROVED

Reviewed By :Yogesh Patel 02/15/2024

Supervised By :mohammad ahmed 02/15/2024



#42

2,4,6-Tribromophenol

Concen: 93.908 ng

RT: 16.033 min Scan# 2201

Delta R.T. -0.006 min

Lab File: BP019711.D

Acq: 14 Feb 2024 20:36




Tgt Ion:330 Resp: 262344


Ion Ratio Lower Upper

330 100

332 97.9 77.2 115.8

141 29.7 27.0 40.4









#45
2-Fluorobiphenyl
Concen: 46.735 ng
RT: 13.169 min Scan# 1716
Delta R.T. -0.012 min
Lab File: BP019711.D
Acq: 14 Feb 2024 20:36

Instrument :
BNA_P
ClientSampleId :
B-3(10-12)

Tgt Ion:172 Resp: 723340
Ion Ratio Lower Upper
172 100
171 35.0 28.2 42.4
170 23.7 18.9 28.3




Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 02/15/2024
Supervised By :mohammad ahmed 02/15/2024



#46
1,1'-Biphenyl
Concen: 11.816 ng
RT: 13.375 min Scan# 1749
Delta R.T. -0.012 min
Lab File: BP019711.D
Acq: 14 Feb 2024 20:36


Tgt Ion:154 Resp: 180443
Ion Ratio Lower Upper
154 100
153 39.0 20.2 60.2
76 12.3 0.0 31.6





#52
Acenaphthene
Concen: 6.460 ng
RT: 14.598 min Scan# 1957
Delta R.T. -0.017 min
Lab File: BP019711.D
Acq: 14 Feb 2024 20:36




Instrument :
BNA_P
ClientSampleId :
B-3(10-12)




Tgt Ion:154 Resp: 70518
Ion Ratio Lower Upper
154 100
153 107.9 90.5 135.7
152 53.6 42.1 63.1

Manual Integrations
APPROVED



Reviewed By :Yogesh Patel 02/15/2024
Supervised By :mohammad ahmed 02/15/2024




#55
Dibenzofuran
Concen: 3.400 ng
RT: 14.939 min Scan# 2015
Delta R.T. -0.012 min
Lab File: BP019711.D
Acq: 14 Feb 2024 20:36



Tgt Ion:168 Resp: 60448
Ion Ratio Lower Upper
168 100
139 36.9 31.5 47.3
169 15.1 10.3 15.5





#58

Fluorene

Concen: 3.477 ng

RT: 15.586 min Scan# 2125

Delta R.T. -0.012 min

Lab File: BP019711.D

Acq: 14 Feb 2024 20:36

Instrument :

BNA_P

ClientSampleId :

B-3(10-12)





Tgt Ion:	166	Resp:	4923
Ion Ratio	Lower	Upper	
166	100		
165	96.1	77.8	116.8
167	18.1	10.7	16.1

Manual Integrations

APPROVED

Reviewed By :Yogesh Patel 02/15/2024

Supervised By :mohammad ahmed 02/15/2024



#64

Phenanthrene-d10

Concen: 20.000 ng




RT: 17.292 min Scan# 2415


Delta R.T. -0.012 min

Lab File: BP019711.D

Acq: 14 Feb 2024 20:36

Tgt Ion:	188	Resp:	428306
Ion Ratio	Lower	Upper	
188	100		
94	6.8	5.4	8.2
80	7.4	6.3	9.5





#71

Phenanthrene

Concen: 8.450 ng

RT: 17.333 min Scan# 2422

Delta R.T. -0.012 min

Lab File: BP019711.D

Acq: 14 Feb 2024 20:36

Instrument :

BNA_P

ClientSampleId :

B-3(10-12)

Tgt Ion:178 Resp: 19045





Ion	Ratio	Lower	Upper
178	100		
176	20.5	15.5	23.3
179	16.6	12.6	19.0

Manual Integrations

APPROVED

Reviewed By :Yogesh Patel 02/15/2024

Supervised By :mohammad ahmed 02/15/2024



#73

Carbazole

Concen: 2.656 ng

RT: 17.704 min Scan# 2485




Delta R.T. -0.006 min


Lab File: BP019711.D

Acq: 14 Feb 2024 20:36

Tgt Ion:167 Resp: 54248

Ion	Ratio	Lower	Upper
167	100		
166	25.7	17.2	25.8
139	17.7	10.6	15.8









#74
Di-n-butylphthalate
Concen: 243.007 ng
RT: 18.274 min Scan# 21
Delta R.T. 0.000 min
Lab File: BP019711.D
Acq: 14 Feb 2024 20:36

Instrument :
BNA_P
ClientSampleId :
B-3(10-12)

Tgt Ion:149 Resp: 6001879
Ion Ratio Lower Upper
149 100
150 9.3 7.3 10.9
104 5.5 4.8 7.2




Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 02/15/2024
Supervised By :mohammad ahmed 02/15/2024



#75
Fluoranthene
Concen: 10.766 ng
RT: 19.316 min Scan# 2759
Delta R.T. 0.006 min
Lab File: BP019711.D
Acq: 14 Feb 2024 20:36

Tgt Ion:202 Resp: 296437
Ion Ratio Lower Upper
202 100
101 7.6 0.0 27.7
203 17.6 0.0 37.5





#76

Chrysene-d12

Concen: 20.000 ng

RT: 21.392 min Scan# 3112

Delta R.T. 0.000 min

Lab File: BP019711.D

Acq: 14 Feb 2024 20:36

Instrument :

BNA_P

ClientSampleId :

B-3(10-12)

Tgt Ion:240 Resp: 431300

Ion Ratio Lower Upper

240 100

120 7.4 6.1 9.1





236 26.2 20.7 31.1

Manual Integrations

APPROVED

Reviewed By :Yogesh Patel 02/15/2024

Supervised By :mohammad ahmed 02/15/2024



#78

Pyrene

Concen: 8.061 ng

RT: 19.663 min Scan# 2818

Delta R.T. 0.000 min

Lab File: BP019711.D

Acq: 14 Feb 2024 20:36




Tgt Ion:202 Resp: 245383


Ion Ratio Lower Upper

202 100

200 28.1 16.9 25.3


203 18.6 14.2 21.2





#79
 Terphenyl-d14
 Concen: 50.660 ng
 RT: 19.869 min Scan# 21
 Delta R.T. 0.000 min
 Lab File: BP019711.D
 Acq: 14 Feb 2024 20:36




Instrument :
 BNA_P
 ClientSampleId :
 B-3(10-12)



Tgt Ion:244 Resp: 132861
 Ion Ratio Lower Upper
 244 100
 212 7.6 6.1 9.1
 122 6.8 5.9 8.9




Manual Integrations
 APPROVED


Reviewed By :Yogesh Patel 02/15/2024
 Supervised By :mohammad ahmed 02/15/2024



#81
 Benzo(a)anthracene
 Concen: 4.699 ng m
 RT: 21.374 min Scan# 3109
 Delta R.T. -0.006 min
 Lab File: BP019711.D
 Acq: 14 Feb 2024 20:36


Tgt Ion:228 Resp: 142676
 Ion Ratio Lower Upper
 228 100
 226 29.8 21.7 32.5
 229 24.9 15.6 23.4#





#83
Chrysene
Concen: 4.689 ng
RT: 21.427 min Scan# 3119
Delta R.T. -0.006 min
Lab File: BP019711.D
Acq: 14 Feb 2024 20:36




Instrument :
BNA_P
ClientSampleId :
B-3(10-12)



Tgt Ion:228 Resp: 135162
Ion Ratio Lower Upper
228 100
226 30.1 23.4 35.2
229 23.3 15.6 23.4




Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 02/15/2024
Supervised By :mohammad ahmed 02/15/2024



#84
Bis(2-ethylhexyl)phthalate
Concen: 438.202 ng
RT: 21.327 min Scan# 3101
Delta R.T. 0.000 min
Lab File: BP019711.D
Acq: 14 Feb 2024 20:36

Tgt Ion:149 Resp: 7585165
Ion Ratio Lower Upper
149 100
167 31.6 22.7 34.1
279 7.4 4.2 6.2#





#86

Perylene-d12

Concen: 20.000 ng

RT: 23.815 min Scan# 3524

Delta R.T. 0.000 min

Lab File: BP019711.D

Acq: 14 Feb 2024 20:36

Instrument :

BNA_P

ClientSampleId :

B-3(10-12)





Tgt Ion:264	Resp: 50563
Ion Ratio	Lower Upper
264 100	
260 23.3	19.2 28.8
265 22.6	17.0 25.4

Manual Integrations

APPROVED

Reviewed By :Yogesh Patel 02/15/2024

Supervised By :mohammad ahmed 02/15/2024



#88

Benzo(b)fluoranthene

Concen: 5.057 ng




RT: 23.080 min Scan# 3399


Delta R.T. 0.000 min

Lab File: BP019711.D

Acq: 14 Feb 2024 20:36


Tgt Ion:252	Resp: 161146
Ion Ratio	Lower Upper
252 100	
253 27.8	17.3 25.9#
125 16.2	4.9 7.3#





#90
Benzo(a)pyrene
Concen: 3.662 ng
RT: 23.704 min Scan# 3506
Delta R.T. -0.006 min
Lab File: BP019711.D
Acq: 14 Feb 2024 20:36

Instrument :
BNA_P
ClientSampleId :
B-3(10-12)



Tgt Ion:252 Resp: 103839
Ion Ratio Lower Upper
252 100
253 29.0 17.1 25.7
125 22.8 5.6 8.4

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 02/15/2024
Supervised By :mohammad ahmed 02/15/2024

