




Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP021522\
Data File : BP009150.D
Acq On : 15 Feb 2022 12:03
Operator : CG/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Quant Time: Feb 16 02:32:44 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP020722.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Feb 16 01:29:44 2022
Response via : Initial Calibration



TIC: BP009150.D\data.ms

(70) Pentachlorophenol (C)




16.869min (+ 0.000) 669658.57 ng

response 485167

Ion	Exp%	Act%
265.70	100.00	100.00
268.00	63.60	63.90
264.00	62.90	61.84
0.00	0.00	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP021522\
 Data File : BP009150.D
 Acq On : 15 Feb 2022 12:03
 Operator : CG/JU
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Quant Time: Feb 16 02:32:44 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP020722.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Feb 16 01:29:44 2022
 Response via : Initial Calibration



TIC: BP009150.D\data.ms

(77) Benzidine

19.392min (-0.006) 920768.28 ng

response 1924039

Ion	Exp%	Act%
184.00	100.00	100.00
185.00	15.20	14.20
183.00	11.90	11.81
0.00	0.00	0.00