

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP021524\
 Data File : BP019715.D
 Acq On : 15 Feb 2024 10:35
 Operator : MA/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 LabSampleId :
 SSTDCCC040

Quant Time: Feb 15 11:04:25 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP013124.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Feb 01 00:51:48 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	77	0.00
2	1,4-Dioxane	40.000	35.368	11.6	77	0.00
3	Pyridine	40.000	37.782	5.5	78	0.00
4	n-Nitrosodimethylamine	40.000	39.230	1.9	80	0.00
5 S	2-Fluorophenol	80.000	80.218	-0.3	84	0.00
6	Aniline	40.000	39.219	2.0	79	-0.01
7 S	Phenol-d6	80.000	79.838	0.2	81	0.00
8	2-Chlorophenol	40.000	40.418	-1.0	82	-0.01
9	Benzaldehyde	40.000	42.314	-5.8	71	0.00
10 C	Phenol	40.000	39.340	1.6	79	0.00
11	bis(2-Chloroethyl)ether	40.000	37.327	6.7	74	0.00
12	1,3-Dichlorobenzene	40.000	40.396	-1.0	83	-0.01
13 C	1,4-Dichlorobenzene	40.000	40.290	-0.7	84	0.00
14	1,2-Dichlorobenzene	40.000	40.627	-1.6	83	0.00
15	Benzyl Alcohol	40.000	39.453	1.4	77	0.00
16	2,2'-oxybis(1-Chloropropane	40.000	35.547	11.1	69	0.00
17	2-Methylphenol	40.000	39.641	0.9	78	0.00
18	Hexachloroethane	40.000	39.257	1.9	79	-0.01
19 P	n-Nitroso-di-n-propylamine	40.000	36.128	9.7	70	0.00
20	3+4-Methylphenols	40.000	39.789	0.5	79	0.00
21 I	Naphthalene-d8	20.000	20.000	0.0	73	0.00
22	Acetophenone	40.000	38.837	2.9	74	0.00
23 S	Nitrobenzene-d5	80.000	79.510	0.6	76	0.00
24	Nitrobenzene	40.000	39.626	0.9	76	0.00
25	Isophorone	40.000	38.992	2.5	73	0.00
26 C	2-Nitrophenol	40.000	45.764	-14.4	85	0.00
27	2,4-Dimethylphenol	40.000	40.795	-2.0	77	0.00
28	bis(2-Chloroethoxy)methane	40.000	37.981	5.0	70	-0.01
29 C	2,4-Dichlorophenol	40.000	42.280	-5.7	80	0.00
30	1,2,4-Trichlorobenzene	40.000	41.799	-4.5	80	0.00
31	Naphthalene	40.000	40.307	-0.8	77	0.00
32	Benzoic acid	40.000	48.244	-20.6	87	0.00
33	4-Chloroaniline	40.000	40.324	-0.8	76	0.00
34 C	Hexachlorobutadiene	40.000	42.077	-5.2	80	-0.01
35	Caprolactam	40.000	46.672	-16.7	94	0.02
36 C	4-Chloro-3-methylphenol	40.000	42.545	-6.4	81	0.00
37	2-Methylnaphthalene	40.000	40.762	-1.9	77	0.00
38	1-Methylnaphthalene	40.000	40.824	-2.1	77	0.00
39 I	Acenaphthene-d10	20.000	20.000	0.0	78	0.00
40	1,2,4,5-Tetrachlorobenzene	40.000	40.492	-1.2	83	0.00
41 P	Hexachlorocyclopentadiene	40.000	43.541	-8.9	84	-0.01
42 S	2,4,6-Tribromophenol	80.000	103.284	-29.1#	109	0.00
43 C	2,4,6-Trichlorophenol	40.000	42.145	-5.4	84	0.00
44	2,4,5-Trichlorophenol	40.000	43.325	-8.3	87	0.00
45 S	2-Fluorobiphenyl	80.000	78.187	2.3	79	-0.01
46	1,1'-Biphenyl	40.000	38.070	4.8	77	0.00
47	2-Chloronaphthalene	40.000	39.039	2.4	80	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP021524\
 Data File : BP019715.D
 Acq On : 15 Feb 2024 10:35
 Operator : MA/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 LabSampleId :
 SSTDCCC040

Quant Time: Feb 15 11:04:25 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP013124.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Feb 01 00:51:48 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48	2-Nitroaniline	40.000	42.630	-6.6	87	0.00
49	Acenaphthylene	40.000	39.989	0.0	83	0.00
50	Dimethylphthalate	40.000	41.905	-4.8	88	0.00
51	2,6-Dinitrotoluene	40.000	44.266	-10.7	91	0.00
52 C	Acenaphthene	40.000	39.872	0.3	82	0.00
53	3-Nitroaniline	40.000	45.356	-13.4	98	0.00
54 P	2,4-Dinitrophenol	40.000	52.835	-32.1#	113	0.01
55	Dibenzofuran	40.000	40.827	-2.1	85	0.00
56 P	4-Nitrophenol	40.000	46.327	-15.8	105	0.00
57	2,4-Dinitrotoluene	40.000	48.198	-20.5	102	0.00
58	Fluorene	40.000	42.644	-6.6	90	0.00
59	2,3,4,6-Tetrachlorophenol	40.000	46.860	-17.1	98	0.00
60	Diethylphthalate	40.000	42.440	-6.1	92	0.00
61	4-Chlorophenyl-phenylether	40.000	42.702	-6.8	89	0.00
62	4-Nitroaniline	40.000	46.527	-16.3	106	0.00
63	Azobenzene	40.000	38.851	2.9	82	0.00
64 I	Phenanthrene-d10	20.000	20.000	0.0	100	0.00
65	4,6-Dinitro-2-methylphenol	40.000	50.309	-25.8#	124	0.00
66 c	n-Nitrosodiphenylamine	40.000	37.570	6.1	93	0.00
67	4-Bromophenyl-phenylether	40.000	40.159	-0.4	98	0.00
68	Hexachlorobenzene	40.000	41.044	-2.6	103	0.00
69	Atrazine	40.000	39.518	1.2	103	0.00
70 C	Pentachlorophenol	40.000	45.948	-14.9	116	0.00
71	Phenanthrene	40.000	40.052	-0.1	106	0.00
72	Anthracene	40.000	39.892	0.3	104	0.00
73	Carbazole	40.000	42.002	-5.0	117	0.00
74	Di-n-butylphthalate	40.000	39.943	0.1	107	0.00
75 C	Fluoranthene	40.000	42.360	-5.9	121	0.00
76 I	Chrysene-d12	20.000	20.000	0.0	117	0.00
77	Benzydine	40.000	29.717	25.7#	97	0.00
78	Pyrene	40.000	41.469	-3.7	123	0.00
79 S	Terphenyl-d14	80.000	85.569	-7.0	126	0.00
80	Butylbenzylphthalate	40.000	39.878	0.3	118	0.00
81	Benzo(a)anthracene	40.000	40.360	-0.9	124	0.00
82	3,3'-Dichlorobenzidine	40.000	41.579	-3.9	127	0.00
83	Chrysene	40.000	40.309	-0.8	124	0.00
84	Bis(2-ethylhexyl)phthalate	40.000	36.755	8.1	108	-0.01
85 c	Di-n-octyl phthalate	40.000	35.719	10.7	104	-0.02
86 I	Perylene-d12	20.000	20.000	0.0	114	0.00
87	Indeno(1,2,3-cd)pyrene	40.000	39.174	2.1	115	0.01
88	Benzo(b)fluoranthene	40.000	40.434	-1.1	118	0.00
89	Benzo(k)fluoranthene	40.000	39.810	0.5	120	0.00
90 C	Benzo(a)pyrene	40.000	40.218	-0.5	119	0.00
91	Dibenzo(a,h)anthracene	40.000	39.553	1.1	115	0.01
92	Benzo(g,h,i)perylene	40.000	39.159	2.1	115	0.02

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP021524\
Data File : BP019715.D
Acq On : 15 Feb 2024 10:35
Operator : MA/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_P
LabSampleId :
SSTDCCC040

Quant Time: Feb 15 11:04:25 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP013124.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Feb 01 00:51:48 2024
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range SPCC's out = 0 CCC's out = 0