

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP021622\
 Data File : BP009172.D
 Acq On : 16 Feb 2022 11:11
 Operator : CG/JU
 Sample : PB142581BL
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 PB142581BL

Quant Time: Feb 16 14:31:39 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP020722.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Feb 16 14:30:24 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.781	152	194744	20.000	ng	0.00
21) Naphthalene-d8	10.581	136	764531	20.000	ng	0.00
39) Acenaphthene-d10	14.428	164	559290	20.000	ng	0.00
64) Phenanthrene-d10	17.186	188	1310414	20.000	ng	0.00
76) Chrysene-d12	21.286	240	1482773	20.000	ng	# 0.00
86) Perylene-d12	23.674	264	1595163	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.381	112	1478423	135.420	ng	0.00
7) Phenol-d6	6.975	99	1872971	130.199	ng	0.00
23) Nitrobenzene-d5	8.946	82	1164185	85.873	ng	0.00
42) 2,4,6-Tribromophenol	15.928	330	1162127	155.624	ng	0.00
45) 2-Fluorobiphenyl	13.046	172	3342958	83.674	ng	0.00
79) Terphenyl-d14	19.751	244	6640564	80.538	ng	0.00

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP021622\
Data File : BP009172.D
Acq On : 16 Feb 2022 11:11
Operator : CG/JU
Sample : PB142581BL
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
PB142581BL

Quant Time: Feb 16 14:31:39 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP020722.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Feb 16 14:30:24 2022
Response via : Initial Calibration

