

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP021622\
 Data File : BP009187.D
 Acq On : 16 Feb 2022 19:58
 Operator : CG/JU
 Sample : N1554-01
 Misc :
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 BU-03-021522

Quant Time: Feb 17 02:03:11 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP020722.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Feb 16 14:30:24 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.781	152	187931	20.000	ng	0.00
21) Naphthalene-d8	10.581	136	711269	20.000	ng	0.00
39) Acenaphthene-d10	14.428	164	478344	20.000	ng	0.00
64) Phenanthrene-d10	17.187	188	1030788	20.000	ng	0.00
76) Chrysene-d12	21.281	240	1095822	20.000	ng	0.00
86) Perylene-d12	23.669	264	1197282	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.381	112	1084678	102.956	ng	0.00
7) Phenol-d6	6.975	99	1377712	99.243	ng	0.00
23) Nitrobenzene-d5	8.946	82	876818	69.519	ng	0.00
42) 2,4,6-Tribromophenol	15.928	330	715205	111.983	ng	0.00
45) 2-Fluorobiphenyl	13.046	172	2331079	68.220	ng	0.00
79) Terphenyl-d14	19.745	244	3634804	59.650	ng	0.00
Target Compounds						
10) Phenol	7.005	94	28319	2.029	ng	Qvalue # 50

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP021622\
Data File : BP009187.D
Acq On : 16 Feb 2022 19:58
Operator : CG/JU
Sample : N1554-01
Misc :
ALS Vial : 18 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
BU-03-021522

Quant Time: Feb 17 02:03:11 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP020722.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Feb 16 14:30:24 2022
Response via : Initial Calibration

