

Data Path : Z:\svoasrv\HPCHEM1\BNA\_P\Data\BP021623\  
 Data File : BP013834.D  
 Acq On : 16 Feb 2023 18:43  
 Operator : CG/JU  
 Sample : PB150875BL  
 Misc :  
 ALS Vial : 14 Sample Multiplier: 1

Instrument :  
 BNA\_P  
 ClientSampleId :  
 PB150875BL

Quant Time: Feb 17 05:10:44 2023  
 Quant Method : Z:\svoasrv\HPCHEM1\BNA\_P\Methods\8270E-BP021623.M  
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
 QLast Update : Fri Feb 17 01:08:52 2023  
 Response via : Initial Calibration

| Compound                    | R.T.   | QIon | Response | Conc    | Units | Dev(Min) |
|-----------------------------|--------|------|----------|---------|-------|----------|
| -----                       |        |      |          |         |       |          |
| Internal Standards          |        |      |          |         |       |          |
| 1) 1,4-Dichlorobenzene-d4   | 8.110  | 152  | 161391   | 20.000  | ng    | 0.00     |
| 21) Naphthalene-d8          | 10.934 | 136  | 600592   | 20.000  | ng    | 0.00     |
| 39) Acenaphthene-d10        | 14.745 | 164  | 353865   | 20.000  | ng    | 0.00     |
| 64) Phenanthrene-d10        | 17.498 | 188  | 749874   | 20.000  | ng    | 0.00     |
| 76) Chrysene-d12            | 21.568 | 240  | 691829   | 20.000  | ng    | 0.00     |
| 86) Perylene-d12            | 24.121 | 264  | 776304   | 20.000  | ng    | -0.01    |
| System Monitoring Compounds |        |      |          |         |       |          |
| 5) 2-Fluorophenol           | 5.646  | 112  | 1484374  | 151.583 | ng    | 0.00     |
| 7) Phenol-d6                | 7.263  | 99   | 1827532  | 145.719 | ng    | 0.00     |
| 23) Nitrobenzene-d5         | 9.281  | 82   | 1118463  | 100.562 | ng    | 0.00     |
| 42) 2,4,6-Tribromophenol    | 16.233 | 330  | 714163   | 138.643 | ng    | 0.00     |
| 45) 2-Fluorobiphenyl        | 13.369 | 172  | 2600620  | 96.529  | ng    | 0.00     |
| 79) Terphenyl-d14           | 20.027 | 244  | 3853194  | 107.158 | ng    | 0.00     |

Target Compounds Qvalue

-----  
 (#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA\_P\Data\BP021623\  
Data File : BP013834.D  
Acq On : 16 Feb 2023 18:43  
Operator : CG/JU  
Sample : PB150875BL  
Misc :  
ALS Vial : 14 Sample Multiplier: 1

Instrument :  
BNA\_P  
ClientSampleId :  
PB150875BL

Quant Time: Feb 17 05:10:44 2023  
Quant Method : Z:\svoasrv\HPCHEM1\BNA\_P\Methods\8270E-BP021623.M  
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
QLast Update : Fri Feb 17 01:08:52 2023  
Response via : Initial Calibration

