

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP021722\
 Data File : BP009192.D
 Acq On : 17 Feb 2022 13:16
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 LabSampleId :
 SSTDCCC040

Quant Time: Feb 18 01:49:50 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP020722.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Feb 16 14:30:24 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	106	0.00
2	1,4-Dioxane	40.000	42.695	-6.7	116	0.00
3	Pyridine	40.000	38.221	4.4	101	0.00
4	n-Nitrosodimethylamine	40.000	35.923	10.2	91	0.00
5 S	2-Fluorophenol	80.000	74.253	7.2	100	0.00
6	Aniline	40.000	34.250	14.4	94	0.00
7 S	Phenol-d6	80.000	69.893	12.6	95	0.00
8	2-Chlorophenol	40.000	36.949	7.6	101	0.00
9	Benzaldehyde	40.000	33.674	15.8	95	0.00
10 C	Phenol	40.000	35.113	12.2	96	0.00
11	bis(2-Chloroethyl)ether	40.000	35.842	10.4	99	0.00
12	1,3-Dichlorobenzene	40.000	37.695	5.8	104	0.00
13 C	1,4-Dichlorobenzene	40.000	36.904	7.7	101	0.00
14	1,2-Dichlorobenzene	40.000	36.725	8.2	101	0.00
15	Benzyl Alcohol	40.000	35.182	12.0	95	0.00
16	2,2'-oxybis(1-Chloropropane	40.000	33.422	16.4	94	0.00
17	2-Methylphenol	40.000	34.818	13.0	96	0.00
18	Hexachloroethane	40.000	37.355	6.6	103	-0.01
19 P	n-Nitroso-di-n-propylamine	40.000	33.786	15.5	94	0.00
20	3+4-Methylphenols	40.000	34.610	13.5	95	0.00
21 I	Naphthalene-d8	20.000	20.000	0.0	101	0.00
22	Acetophenone	40.000	37.643	5.9	95	0.00
23 S	Nitrobenzene-d5	80.000	74.452	6.9	94	0.00
24	Nitrobenzene	40.000	36.627	8.4	93	0.00
25	Isophorone	40.000	36.733	8.2	97	0.00
26 C	2-Nitrophenol	40.000	39.943	0.1	99	0.00
27	2,4-Dimethylphenol	40.000	37.015	7.5	94	0.00
28	bis(2-Chloroethoxy)methane	40.000	36.206	9.5	94	0.00
29 C	2,4-Dichlorophenol	40.000	38.547	3.6	97	0.00
30	1,2,4-Trichlorobenzene	40.000	39.440	1.4	101	0.00
31	Naphthalene	40.000	36.929	7.7	95	0.00
32	Benzoic acid	40.000	35.909	10.2	95	-0.01
33	4-Chloroaniline	40.000	36.031	9.9	93	0.00
34 C	Hexachlorobutadiene	40.000	40.875	-2.2	105	0.00
35	Caprolactam	40.000	37.540	6.2	104	0.00
36 C	4-Chloro-3-methylphenol	40.000	36.369	9.1	97	0.00
37	2-Methylnaphthalene	40.000	36.584	8.5	97	0.00
38	1-Methylnaphthalene	40.000	36.158	9.6	96	0.00
39 I	Acenaphthene-d10	20.000	20.000	0.0	104	0.00
40	1,2,4,5-Tetrachlorobenzene	40.000	40.702	-1.8	104	0.00
41 P	Hexachlorocyclopentadiene	40.000	41.081	-2.7	118	0.00
42 S	2,4,6-Tribromophenol	80.000	84.505	-5.6	116	0.00
43 C	2,4,6-Trichlorophenol	40.000	40.108	-0.3	101	0.00
44	2,4,5-Trichlorophenol	40.000	40.423	-1.1	103	0.00
45 S	2-Fluorobiphenyl	80.000	78.901	1.4	101	0.00
46	1,1'-Biphenyl	40.000	38.415	4.0	99	0.00
47	2-Chloronaphthalene	40.000	38.238	4.4	98	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP021722\
 Data File : BP009192.D
 Acq On : 17 Feb 2022 13:16
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 LabSampleId :
 SSTDCCC040

Quant Time: Feb 18 01:49:50 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP020722.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Feb 16 14:30:24 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48	2-Nitroaniline	40.000	37.962	5.1	98	0.00
49	Acenaphthylene	40.000	38.386	4.0	100	0.00
50	Dimethylphthalate	40.000	38.865	2.8	106	0.00
51	2,6-Dinitrotoluene	40.000	39.882	0.3	106	0.00
52 C	Acenaphthene	40.000	37.624	5.9	99	0.00
53	3-Nitroaniline	40.000	38.622	3.4	102	0.00
54 P	2,4-Dinitrophenol	40.000	37.344	6.6	99	0.00
55	Dibenzofuran	40.000	38.891	2.8	104	0.00
56 P	4-Nitrophenol	40.000	49.213	-23.0	141	0.00
57	2,4-Dinitrotoluene	40.000	42.287	-5.7	113	0.00
58	Fluorene	40.000	39.185	2.0	105	0.00
59	2,3,4,6-Tetrachlorophenol	40.000	40.144	-0.4	108	0.00
60	Diethylphthalate	40.000	39.741	0.6	111	0.00
61	4-Chlorophenyl-phenylether	40.000	40.530	-1.3	109	0.00
62	4-Nitroaniline	40.000	40.804	-2.0	107	0.00
63	Azobenzene	40.000	37.362	6.6	101	0.00
64 I	Phenanthrene-d10	20.000	20.000	0.0	119	0.00
65	4,6-Dinitro-2-methylphenol	40.000	38.733	3.2	117	0.00
66 c	n-Nitrosodiphenylamine	40.000	36.333	9.2	106	0.00
67	4-Bromophenyl-phenylether	40.000	38.458	3.9	114	0.00
68	Hexachlorobenzene	40.000	38.878	2.8	116	0.00
69	Atrazine	40.000	39.017	2.5	117	0.00
70 C	Pentachlorophenol	40.000	41.131	-2.8	121	0.00
71	Phenanthrene	40.000	36.639	8.4	109	0.00
72	Anthracene	40.000	37.253	6.9	110	0.00
73	Carbazole	40.000	37.908	5.2	112	0.00
74	Di-n-butylphthalate	40.000	42.114	-5.3	129	0.00
75 C	Fluoranthene	40.000	40.154	-0.4	119	0.00
76 I	Chrysene-d12	20.000	20.000	0.0	129	0.00
77	Benzydine	40.000	35.112	12.2	95	0.00
78	Pyrene	40.000	32.606	18.5	117	0.00
79 S	Terphenyl-d14	80.000	69.254	13.4	125	0.00
80	Butylbenzylphthalate	40.000	37.712	5.7	133	0.00
81	Benzo(a)anthracene	40.000	38.065	4.8	123	0.00
82	3,3'-Dichlorobenzidine	40.000	42.532	-6.3	128	0.00
83	Chrysene	40.000	38.071	4.8	125	0.00
84	Bis(2-ethylhexyl)phthalate	40.000	42.260	-5.6	146	0.00
85 c	Di-n-octyl phthalate	40.000	46.572	-16.4	144	0.00
86 I	Perylene-d12	20.000	20.000	0.0	132	0.00
87	Indeno(1,2,3-cd)pyrene	40.000	39.268	1.8	120	0.00
88	Benzo(b)fluoranthene	40.000	36.619	8.5	124	0.00
89	Benzo(k)fluoranthene	40.000	36.566	8.6	124	0.00
90 C	Benzo(a)pyrene	40.000	38.277	4.3	125	0.00
91	Dibenzo(a,h)anthracene	40.000	39.582	1.0	120	0.00
92	Benzo(g,h,i)perylene	40.000	39.082	2.3	119	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP021722\
Data File : BP009192.D
Acq On : 17 Feb 2022 13:16
Operator : CG/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_P
LabSampleId :
SSTDCCC040

Quant Time: Feb 18 01:49:50 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP020722.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Feb 16 14:30:24 2022
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range

SPCC's out = 0 CCC's out = 0