




Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP021725\
Data File : BP024068.D
Acq On : 17 Feb 2025 10:47
Operator : RC/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Quant Time: Feb 18 02:37:53 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP021725.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Feb 18 02:25:07 2025
Response via : Initial Calibration



TIC: BP024068.D\data.ms

(70) Pentachlorophenol (C)




16.839min (+ 0.005) 214492.84 ng

response 758401

Ion	Exp%	Act%
265.70	100.00	100.00
268.00	64.10	63.87
264.00	63.10	63.46
0.00	0.00	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP021725\
Data File : BP024068.D
Acq On : 17 Feb 2025 10:47
Operator : RC/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Quant Time: Feb 18 02:37:53 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP021725.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Feb 18 02:25:07 2025
Response via : Initial Calibration



TIC: BP024068.D\data.ms

(77) Benzidine

19.522min (+ 0.006) 903747.62 ng

response 4372031

Ion	Exp%	Act%
184.00	100.00	100.00
185.00	14.80	15.09
183.00	11.60	11.74
0.00	0.00	0.00