

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP021921\
 Data File : BP004825.D
 Acq On : 19 Feb 2021 16:49
 Operator : CG/JU
 Sample : M1433-03
 Misc :
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 FR-E-BUS

Quant Time: Feb 19 23:52:04 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP020521.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Feb 19 12:04:44 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.763	152	49714	20.00	ng	0.00
21) Naphthalene-d8	10.552	136	184290	20.00	ng	# 0.00
39) Acenaphthene-d10	14.410	164	112606	20.00	ng	0.00
64) Phenanthrene-d10	17.163	188	236462	20.00	ng	0.00
76) Chrysene-d12	21.251	240	285145	20.00	ng	# 0.00
86) Perylene-d12	23.551	264	308351	20.00	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.358	112	12175	3.78	ng	0.00
7) Phenol-d6	6.940	99	152660	34.48	ng	0.00
23) Nitrobenzene-d5	8.916	82	222949	52.71	ng	0.00
42) 2,4,6-Tribromophenol	15.904	330	625	0.42	ng	0.00
45) 2-Fluorobiphenyl	13.028	172	518050	56.23	ng	0.00
79) Terphenyl-d14	19.727	244	855071	50.37	ng	0.00
Target Compounds						
10) Phenol	6.963	94	10682	2.49	ng	Qvalue 75
50) Dimethylphthalate	13.863	163	34127	3.45	ng	# 96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP021921\
Data File : BP004825.D
Acq On : 19 Feb 2021 16:49
Operator : CG/JU
Sample : M1433-03
Misc :
ALS Vial : 12 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
FR-E-BUS

Quant Time: Feb 19 23:52:04 2021
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP020521.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Feb 19 12:04:44 2021
Response via : Initial Calibration

