




Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP022025\
Data File : BP024100.D
Acq On : 20 Feb 2025 09:01
Operator : RC/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Quant Time: Feb 20 13:17:26 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP021725.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Feb 18 02:25:07 2025
Response via : Initial Calibration



TIC: BP024100.D\data.ms

(70) Pentachlorophenol (C)




16.851min (-0.003) 87238.15 ng

response 677554

Ion	Exp%	Act%
265.70	100.00	100.00
268.00	64.10	63.22
264.00	63.10	62.56
0.00	0.00	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP022025\
Data File : BP024100.D
Acq On : 20 Feb 2025 09:01
Operator : RC/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Quant Time: Feb 20 13:17:26 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP021725.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Feb 18 02:25:07 2025
Response via : Initial Calibration



TIC: BP024100.D\data.ms

(77) Benzidine**19.528min (-0.008) 164259.39 ng****response 4303425**

Ion	Exp%	Act%
184.00	100.00	100.00
185.00	14.80	14.85
183.00	11.60	11.58
0.00	0.00	0.00