

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP022223\
 Data File : BP013868.D
 Acq On : 22 Feb 2023 09:31
 Operator : CG/JU
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SSTD020766

Quant Time: Feb 23 03:13:47 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP020723.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed Feb 22 01:41:33 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.105	152	178974	20.000	ng/ul	0.00
20) Naphthalene-d8	10.928	136	829312	20.000	ng/ul	0.00
38) Acenaphthene-d10	14.739	164	580598	20.000	ng/ul	0.00
64) Phenanthrene-d10	17.498	188	1294324	20.000	ng/ul	0.00
79) Chrysene-d12	21.574	240	1194944	20.000	ng/ul	# 0.00
88) Perylene-d12	24.127	264	1524816	20.000	ng/ul	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.452	96	29509	6.834	ng/uL	0.00
4) Pyridine-d5	3.887	84	212342	16.587	ng/ul	0.00
7) Phenol-d5	7.252	99	264863	16.610	ng/ul	0.00
9) Bis-(2-Chloroethyl)eth...	7.422	67	151244	16.186	ng/ul	0.00
11) 2-Chlorophenol-d4	7.628	132	206551	17.596	ng/ul	0.00
15) 4-Methylphenol-d8	8.810	113	248965	19.101	ng/ul	0.00
21) Nitrobenzene-d5	9.275	128	114996	19.930	ng/ul	0.00
24) 2-Nitrophenol-d4	10.004	143	136373	20.470	ng/ul	0.00
28) 2,4-Dichlorophenol-d3	10.546	165	267956	19.736	ng/ul	0.00
31) 4-Chloroaniline-d4	11.063	131	353553	18.371	ng/ul	0.00
46) Dimethylphthalate-d6	14.140	166	852643	18.645	ng/ul	0.00
49) Acenaphthylene-d8	14.434	160	937199	18.933	ng/ul	0.00
54) 4-Nitrophenol-d4	14.928	143	141290	17.343	ng/ul	0.00
60) Fluorene-d10	15.734	176	721473	18.523	ng/ul	0.00
65) 4,6-Dinitro-2-methylph...	15.845	200	158750	19.992	ng/ul	0.00
73) Anthracene-d10	17.592	188	1101018	18.603	ng/ul	0.00
81) Pyrene-d10	19.816	212	1283018	18.934	ng/ul	0.00
92) Benzo(a)pyrene-d12	23.957	264	1445218	18.916	ng/ul	0.00
Target Compounds						
						Qvalue
2) 1,4-Dioxane	3.487	88	31246	6.741	ng/uL	98
5) Pyridine	3.905	79	205745	16.263	ng/ul	99
6) Benzaldehyde	7.234	77	119672	16.642	ng/ul	97
8) Phenol	7.275	94	267275	16.328	ng/ul	93
10) Bis(2-Chloroethyl)ether	7.516	93	214058	16.319	ng/ul	96
12) 2-Chlorophenol	7.663	128	217153	18.029	ng/ul	99
13) 2-Methylphenol	8.540	108	237164	18.982	ng/ul	98
14) 2,2'-oxybis(1-Chloropr...	8.634	45	270173	18.108	ng/ul	96
16) Acetophenone	8.934	105	399492	19.113	ng/ul	99
17) N-Nitroso-di-n-propyla...	8.916	70	201287	19.392	ng/ul	97
18) 4-Methylphenol	8.875	108	261199	18.935	ng/ul	98
19) Hexachloroethane	9.199	117	103118	20.907	ng/ul	93
22) Nitrobenzene	9.322	77	291680	19.214	ng/ul	99
23) Isophorone	9.846	82	589429	18.404	ng/ul	98
25) 2-Nitrophenol	10.034	139	142969	19.799	ng/ul	94
26) 2,4-Dimethylphenol	10.087	107	307117	19.102	ng/ul	99
27) Bis(2-Chloroethoxy)met...	10.328	93	352475	18.224	ng/ul	98
29) 2,4-Dichlorophenol	10.569	162	261036	19.530	ng/ul	99
30) Naphthalene	10.981	128	820293	18.581	ng/ul	100
32) 4-Chloroaniline	11.087	127	355929	18.346	ng/ul	98
33) Hexachlorobutadiene	11.263	225	196205	20.839	ng/ul	98
34) Caprolactam	11.863	113	83173	17.734	ng/ul	95
35) 4-Chloro-3-methylphenol	12.198	107	296011	19.183	ng/ul	99
36) 2-Methylnaphthalene	12.581	142	556661	17.934	ng/ul	98

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP022223\
 Data File : BP013868.D
 Acq On : 22 Feb 2023 09:31
 Operator : CG/JU
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SSTD020766

Quant Time: Feb 23 03:13:47 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP020723.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed Feb 22 01:41:33 2023
 Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.798	142	571116	17.808	ng/ul	99
39) 1,2,4,5-Tetrachloroben...	12.940	216	358516	19.595	ng/ul	99
40) Hexachlorocyclopentadiene	12.916	237	235142	20.016	ng/ul	98
41) 2,4,6-Trichlorophenol	13.175	196	234785	19.555	ng/ul	99
42) 2,4,5-Trichlorophenol	13.245	196	254996	19.621	ng/ul	98
43) 1,1'-Biphenyl	13.575	154	802527	18.421	ng/ul	99
44) 2-Chloronaphthalene	13.622	162	639908	18.671	ng/ul	99
45) 2-Nitroaniline	13.822	65	168490	20.411	ng/ul	98
47) Dimethylphthalate	14.187	163	850682	18.756	ng/ul	99
48) 2,6-Dinitrotoluene	14.316	165	167081	21.085	ng/ul	97
50) Acenaphthylene	14.463	152	998062	18.690	ng/ul	98
51) 3-Nitroaniline	14.645	138	144816	18.484	ng/ul	96
52) Acenaphthene	14.804	153	685559	18.342	ng/ul	100
53) 2,4-Dinitrophenol	14.851	184	98329	17.401	ng/ul	93
55) 4-Nitrophenol	14.939	109	131144	18.346	ng/ul	97
56) Dibenzofuran	15.139	168	985538	18.392	ng/ul	98
57) 2,4-Dinitrotoluene	15.098	165	238464	20.185	ng/ul	99
58) 2,3,4,6-Tetrachlorophenol	15.363	232	239685	19.567	ng/ul	97
59) Diethylphthalate	15.551	149	836277	18.362	ng/ul	99
61) Fluorene	15.786	166	800906	18.320	ng/ul	99
62) 4-Chlorophenyl-phenyle...	15.775	204	431047	18.963	ng/ul	98
63) 4-Nitroaniline	15.804	138	126612	17.066	ng/ul	94
66) 4,6-Dinitro-2-methylph...	15.863	198	154843	19.891	ng/ul	93
67) N-Nitrosodiphenylamine	15.992	169	682245	19.393	ng/ul	99
68) 4-Bromophenyl-phenylether	16.675	248	275901	20.185	ng/ul	98
69) Hexachlorobenzene	16.798	284	326042	20.471	ng/ul	96
70) Atrazine	16.939	200	278726	19.811	ng/ul	98
71) Pentachlorophenol	17.139	266	196856	18.994	ng/ul	99
72) Phenanthrene	17.539	178	1253671	18.304	ng/ul	100
74) Anthracene	17.628	178	1267963	18.401	ng/ul	100
75) 1,2,3,4-Tetrachloroben...	13.545	216	361265	20.909	ng/uL	99
76) Pentachlorobenzene	15.057	250	375660	21.069	ng/uL	98
77) Carbazole	17.892	167	1115384	18.375	ng/ul	100
78) Di-n-butylphthalate	18.422	149	1348536	18.317	ng/ul	99
80) Fluoranthene	19.486	202	1505857	19.411	ng/ul	97
82) Pyrene	19.839	202	1533692	18.543	ng/ul	95
83) Butylbenzylphthalate	20.692	149	571494	20.189	ng/ul	97
84) 3,3'-Dichlorobenzidine	21.474	252	522032	19.411	ng/ul	98
85) Benzo(a)anthracene	21.557	228	1520320	18.620	ng/ul	99
86) Bis(2-ethylhexyl)phtha...	21.451	149	854627	19.716	ng/ul#	98
87) Chrysene	21.610	228	1435605	18.636	ng/ul	98
89) Di-n-octyl phthalate	22.427	149	1426456	13.595	ng/ul	100
90) Benzo(b)fluoranthene	23.339	252	1838689	18.084	ng/ul#	97
91) Benzo(k)fluoranthene	23.386	252	1791463	17.760	ng/ul	98
93) Benzo(a)pyrene	24.010	252	1583485	18.375	ng/ul#	98
94) Indeno(1,2,3-cd)pyrene	26.809	276	1954511	20.220	ng/ul#	94
95) Dibenzo(a,h)anthracene	26.833	278	1631030	20.200	ng/ul#	97
96) Benzo(g,h,i)perylene	27.639	276	1558209	20.248	ng/ul#	96


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP022223\
 Data File : BP013868.D
 Acq On : 22 Feb 2023 09:31
 Operator : CG/JU
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SSTD020766

Quant Time: Feb 23 03:13:47 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP020723.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed Feb 22 01:41:33 2023
 Response via : Initial Calibration









#2
1,4-Dioxane
Concen: 6.741 ng/uL
RT: 3.487 min Scan# 51
Delta R.T. -0.000 min
Lab File: BP013868.D
Acq: 22 Feb 2023 09:31




Instrument :
BNA_P
ClientSampleId :
SSTD020766


Tgt Ion: 88 Resp: 31246
Ion Ratio Lower Upper
88 100
43 26.2 20.1 30.1
58 65.1 53.0 79.6



#5
Pyridine
Concen: 16.263 ng/uL
RT: 3.905 min Scan# 122
Delta R.T. -0.000 min
Lab File: BP013868.D
Acq: 22 Feb 2023 09:31

Tgt Ion: 79 Resp: 205745
Ion Ratio Lower Upper
79 100
52 69.6 55.1 82.7
51 33.2 25.7 38.5









#6
Benzaldehyde
Concen: 16.642 ng/ul
RT: 7.234 min Scan# 611
Delta R.T. -0.000 min
Lab File: BP013868.D
Acq: 22 Feb 2023 09:31




Instrument :
BNA_P
ClientSampleId :
SSTD020766


Tgt Ion: 77 Resp: 119672
Ion Ratio Lower Upper
77 100
105 98.3 79.5 119.3
106 90.2 76.5 114.7



#8
Phenol
Concen: 16.328 ng/ul
RT: 7.275 min Scan# 695
Delta R.T. -0.006 min
Lab File: BP013868.D
Acq: 22 Feb 2023 09:31

Tgt Ion: 94 Resp: 267275
Ion Ratio Lower Upper
94 100
65 30.3 22.2 33.4
66 42.5 29.6 44.4









#10
 Bis(2-Chloroethyl)ether
 Concen: 16.319 ng/ul
 RT: 7.516 min Scan# 736
 Delta R.T. -0.000 min
 Lab File: BP013868.D
 Acq: 22 Feb 2023 09:31

Instrument :
 BNA_P
 ClientSampleId :
 SSTD020766

Tgt Ion: 93 Resp: 214058




Ion	Ratio	Lower	Upper
93	100		
63	69.3	59.4	89.0
95	32.4	25.8	38.8




#12
 2-Chlorophenol
 Concen: 18.029 ng/ul
 RT: 7.663 min Scan# 761
 Delta R.T. -0.000 min
 Lab File: BP013868.D
 Acq: 22 Feb 2023 09:31

Tgt Ion:128 Resp: 217153

Ion	Ratio	Lower	Upper
128	100		
64	45.3	37.3	55.9
130	32.5	26.3	39.5









#13
2-Methylphenol
Concen: 18.982 ng/ul
RT: 8.540 min Scan# 910
Delta R.T. -0.006 min
Lab File: BP013868.D
Acq: 22 Feb 2023 09:31




Instrument :
BNA_P
ClientSampleId :
SSTD020766


Tgt Ion:108 Resp: 237164
Ion Ratio Lower Upper
108 100
107 91.6 71.9 107.9



#14
2,2'-oxybis(1-Chloropropane)
Concen: 18.108 ng/ul
RT: 8.634 min Scan# 926
Delta R.T. -0.000 min
Lab File: BP013868.D
Acq: 22 Feb 2023 09:31

Tgt Ion: 45 Resp: 270173
Ion Ratio Lower Upper
45 100
77 18.7 13.4 20.2
79 15.4 11.0 16.4









#16
Acetophenone
Concen: 19.113 ng/ul
RT: 8.934 min Scan# 91
Delta R.T. -0.000 min
Lab File: BP013868.D
Acq: 22 Feb 2023 09:31




Instrument :
BNA_P
ClientSampleId :
SSTD020766


Tgt Ion:105 Resp: 399492
Ion Ratio Lower Upper
105 100
77 77.3 61.0 91.4
51 25.0 19.6 29.4



#17
N-Nitroso-di-n-propylamine
Concen: 19.392 ng/ul
RT: 8.916 min Scan# 974
Delta R.T. -0.000 min
Lab File: BP013868.D
Acq: 22 Feb 2023 09:31

Tgt Ion: 70 Resp: 201287
Ion Ratio Lower Upper
70 100
42 50.8 39.8 59.8
101 10.0 7.4 11.2
130 21.6 14.6 22.0









#18
4-Methylphenol
Concen: 18.935 ng/ul
RT: 8.875 min Scan# 967
Delta R.T. -0.000 min
Lab File: BP013868.D
Acq: 22 Feb 2023 09:31




Instrument :
BNA_P
ClientSampleId :
SSTD020766


Tgt Ion:108 Resp: 261199
Ion Ratio Lower Upper
108 100
107 112.0 90.9 136.3



#19
Hexachloroethane
Concen: 20.907 ng/ul
RT: 9.199 min Scan# 1022
Delta R.T. -0.000 min
Lab File: BP013868.D
Acq: 22 Feb 2023 09:31

Tgt Ion:117 Resp: 103118
Ion Ratio Lower Upper
117 100
201 116.4 86.6 129.8
199 73.7 55.2 82.8









#22
Nitrobenzene
Concen: 19.214 ng/ul
RT: 9.322 min Scan# 1043
Delta R.T. -0.000 min
Lab File: BP013868.D
Acq: 22 Feb 2023 09:31




Instrument :
BNA_P
ClientSampleId :
SSTD020766


Tgt Ion: 77 Resp: 291680
Ion Ratio Lower Upper
77 100
123 44.6 35.1 52.7
65 13.3 10.6 16.0



#23
Isophorone
Concen: 18.404 ng/ul
RT: 9.846 min Scan# 1132
Delta R.T. -0.006 min
Lab File: BP013868.D
Acq: 22 Feb 2023 09:31

Tgt Ion: 82 Resp: 589429
Ion Ratio Lower Upper
82 100
95 8.0 6.0 9.0
138 16.2 13.7 20.5









#25
2-Nitrophenol
Concen: 19.799 ng/ul
RT: 10.034 min Scan# 1164
Delta R.T. -0.000 min
Lab File: BP013868.D
Acq: 22 Feb 2023 09:31




Instrument :
BNA_P
ClientSampleId :
SSTD020766


Tgt Ion:139 Resp: 142969
Ion Ratio Lower Upper
139 100
65 54.9 39.9 59.9
109 36.0 27.4 41.0



#26
2,4-Dimethylphenol
Concen: 19.102 ng/ul
RT: 10.087 min Scan# 1173
Delta R.T. -0.000 min
Lab File: BP013868.D
Acq: 22 Feb 2023 09:31

Tgt Ion:107 Resp: 307117
Ion Ratio Lower Upper
107 100
121 47.9 37.8 56.6
122 80.5 63.7 95.5









#27
 Bis(2-Chloroethoxy)methane
 Concen: 18.224 ng/ul
 RT: 10.328 min Scan# 1214
 Delta R.T. -0.000 min
 Lab File: BP013868.D
 Acq: 22 Feb 2023 09:31




Instrument :
 BNA_P
 ClientSampleId :
 SSTD020766


Tgt Ion: 93 Resp: 352475
 Ion Ratio Lower Upper
 93 100
 95 32.4 25.0 37.6
 123 10.9 8.2 12.4



#29
 2,4-Dichlorophenol
 Concen: 19.530 ng/ul
 RT: 10.569 min Scan# 1255
 Delta R.T. -0.006 min
 Lab File: BP013868.D
 Acq: 22 Feb 2023 09:31

Tgt Ion: 162 Resp: 261036
 Ion Ratio Lower Upper
 162 100
 164 64.9 51.4 77.2
 98 35.7 27.5 41.3









#30
Naphthalene
Concen: 18.581 ng/ul
RT: 10.981 min Scan# 1325
Delta R.T. -0.000 min
Lab File: BP013868.D
Acq: 22 Feb 2023 09:31




Instrument :
BNA_P
ClientSampleId :
SSTD020766


Tgt Ion:128 Resp: 820293
Ion Ratio Lower Upper
128 100
129 10.9 8.6 12.8
127 13.2 10.6 15.8



#32
4-Chloroaniline
Concen: 18.346 ng/ul
RT: 11.087 min Scan# 1343
Delta R.T. -0.000 min
Lab File: BP013868.D
Acq: 22 Feb 2023 09:31

Tgt Ion:127 Resp: 355929
Ion Ratio Lower Upper
127 100
129 31.5 26.2 39.4





#33

Hexachlorobutadiene

Concen: 20.839 ng/ul

RT: 11.263 min Scan# 11

Delta R.T. -0.000 min

Lab File: BP013868.D

Acq: 22 Feb 2023 09:31

Instrument :

BNA_P

ClientSampleId :

SSTD020766





Tgt Ion:225 Resp: 196205

Ion Ratio Lower Upper

225 100

223 63.3 50.6 76.0

227 63.6 49.0 73.4



#34

Caprolactam

Concen: 17.734 ng/ul

RT: 11.863 min Scan# 1475

Delta R.T. -0.000 min

Lab File: BP013868.D

Acq: 22 Feb 2023 09:31




Tgt Ion:113 Resp: 83173


Ion Ratio Lower Upper

113 100

55 142.6 120.9 181.3

56 120.4 93.8 140.8









#35
4-Chloro-3-methylphenol
Concen: 19.183 ng/ul
RT: 12.198 min Scan# 1532
Delta R.T. -0.000 min
Lab File: BP013868.D
Acq: 22 Feb 2023 09:31




Instrument :
BNA_P
ClientSampleId :
SSTD020766


Tgt Ion:107 Resp: 296011
Ion Ratio Lower Upper
107 100
144 24.3 20.0 30.0
142 76.4 62.1 93.1



#36
2-Methylnaphthalene
Concen: 17.934 ng/ul
RT: 12.581 min Scan# 1597
Delta R.T. -0.000 min
Lab File: BP013868.D
Acq: 22 Feb 2023 09:31

Tgt Ion:142 Resp: 556661
Ion Ratio Lower Upper
142 100
141 92.7 72.8 109.2









#37
1-Methylnaphthalene
Concen: 17.808 ng/ul
RT: 12.798 min Scan# 1634
Delta R.T. -0.000 min
Lab File: BP013868.D
Acq: 22 Feb 2023 09:31




Instrument :
BNA_P
ClientSampleId :
SSTD020766


Tgt Ion:142 Resp: 571116
Ion Ratio Lower Upper
142 100
141 93.4 74.2 111.2
116 4.1 3.1 4.7



#39
1,2,4,5-Tetrachlorobenzene
Concen: 19.595 ng/ul
RT: 12.940 min Scan# 1658
Delta R.T. -0.006 min
Lab File: BP013868.D
Acq: 22 Feb 2023 09:31

Tgt Ion:216 Resp: 358516
Ion Ratio Lower Upper
216 100
214 78.9 62.6 93.8
179 20.9 15.4 23.2
108 15.6 12.7 19.1





#40

Hexachlorocyclopentadiene

Concen: 20.016 ng/ul

RT: 12.916 min Scan# 1

Delta R.T. -0.006 min

Lab File: BP013868.D

Acq: 22 Feb 2023 09:31

Instrument :

BNA_P

ClientSampleId :

SSTD020766





Tgt Ion:237 Resp: 235142

Ion Ratio Lower Upper

237 100

235 63.9 49.8 74.8

272 12.0 10.1 15.1



#41

2,4,6-Trichlorophenol

Concen: 19.555 ng/ul

RT: 13.175 min Scan# 1698

Delta R.T. -0.006 min

Lab File: BP013868.D

Acq: 22 Feb 2023 09:31




Tgt Ion:196 Resp: 234785


Ion Ratio Lower Upper

196 100

198 95.2 75.1 112.7

200 30.9 24.3 36.5









#42
2,4,5-Trichlorophenol
Concen: 19.621 ng/ul
RT: 13.245 min Scan# 1710
Delta R.T. -0.006 min
Lab File: BP013868.D
Acq: 22 Feb 2023 09:31




Instrument :
BNA_P
ClientSampleId :
SSTD020766


Tgt Ion:196 Resp: 254996
Ion Ratio Lower Upper
196 100
198 95.2 77.0 115.4
200 30.0 25.8 38.6



#43
1,1'-Biphenyl
Concen: 18.421 ng/ul
RT: 13.575 min Scan# 1766
Delta R.T. -0.006 min
Lab File: BP013868.D
Acq: 22 Feb 2023 09:31

Tgt Ion:154 Resp: 802527
Ion Ratio Lower Upper
154 100
153 40.6 31.9 47.9
76 12.0 9.9 14.9









#44
2-Chloronaphthalene
Concen: 18.671 ng/ul
RT: 13.622 min Scan# 1774
Delta R.T. -0.000 min
Lab File: BP013868.D
Acq: 22 Feb 2023 09:31




Instrument :
BNA_P
ClientSampleId :
SSTD020766


Tgt Ion:162 Resp: 639908
Ion Ratio Lower Upper
162 100
164 33.4 26.1 39.1
127 38.2 31.1 46.7



#45
2-Nitroaniline
Concen: 20.411 ng/ul
RT: 13.822 min Scan# 1808
Delta R.T. -0.000 min
Lab File: BP013868.D
Acq: 22 Feb 2023 09:31

Tgt Ion: 65 Resp: 168490
Ion Ratio Lower Upper
65 100
92 72.7 58.6 88.0
138 104.7 81.2 121.8





#47

Dimethylphthalate

Concen: 18.756 ng/ul

RT: 14.187 min Scan# 1870

Delta R.T. -0.006 min

Lab File: BP013868.D

Acq: 22 Feb 2023 09:31

Instrument :

BNA_P

ClientSampleId :

SSTD020766





Tgt Ion:163 Resp: 850682

Ion Ratio Lower Upper

163 100

194 4.1 3.4 5.2

164 10.3 7.8 11.8



#48

2,6-Dinitrotoluene

Concen: 21.085 ng/ul

RT: 14.316 min Scan# 1892

Delta R.T. -0.000 min

Lab File: BP013868.D

Acq: 22 Feb 2023 09:31




Tgt Ion:165 Resp: 167081


Ion Ratio Lower Upper

165 100

89 57.0 47.5 71.3

121 21.3 16.6 24.8









#50
Acenaphthylene
Concen: 18.690 ng/ul
RT: 14.463 min Scan# 1917
Delta R.T. -0.006 min
Lab File: BP013868.D
Acq: 22 Feb 2023 09:31




Instrument :
BNA_P
ClientSampleId :
SSTD020766


Tgt Ion:152 Resp: 998062
Ion Ratio Lower Upper
152 100
151 20.9 15.8 23.6
153 13.1 10.4 15.6



#51
3-Nitroaniline
Concen: 18.484 ng/ul
RT: 14.645 min Scan# 1948
Delta R.T. -0.000 min
Lab File: BP013868.D
Acq: 22 Feb 2023 09:31

Tgt Ion:138 Resp: 144816
Ion Ratio Lower Upper
138 100
108 11.6 9.4 14.2
92 116.6 97.4 146.0









#52
Acenaphthene
Concen: 18.342 ng/ul
RT: 14.804 min Scan# 1975
Delta R.T. -0.006 min
Lab File: BP013868.D
Acq: 22 Feb 2023 09:31




Instrument :
BNA_P
ClientSampleId :
SSTD020766


Tgt Ion:153 Resp: 685559
Ion Ratio Lower Upper
153 100
152 47.3 37.4 56.0
154 87.7 70.1 105.1



#53
2,4-Dinitrophenol
Concen: 17.401 ng/ul
RT: 14.851 min Scan# 1983
Delta R.T. -0.000 min
Lab File: BP013868.D
Acq: 22 Feb 2023 09:31

Tgt Ion:184 Resp: 98329
Ion Ratio Lower Upper
184 100
63 53.6 49.5 74.3
154 65.2 54.3 81.5









#55
4-Nitrophenol
Concen: 18.346 ng/ul
RT: 14.939 min Scan# 1998
Delta R.T. -0.006 min
Lab File: BP013868.D
Acq: 22 Feb 2023 09:31




Instrument :
BNA_P
ClientSampleId :
SSTD020766


Tgt Ion:109 Resp: 131144
Ion Ratio Lower Upper
109 100
139 115.1 96.0 144.0
65 118.5 95.8 143.6



#56
Dibenzofuran
Concen: 18.392 ng/ul
RT: 15.139 min Scan# 2032
Delta R.T. -0.000 min
Lab File: BP013868.D
Acq: 22 Feb 2023 09:31

Tgt Ion:168 Resp: 985538
Ion Ratio Lower Upper
168 100
139 37.9 29.6 44.4









#57
2,4-Dinitrotoluene
Concen: 20.185 ng/ul
RT: 15.098 min Scan# 2025
Delta R.T. -0.000 min
Lab File: BP013868.D
Acq: 22 Feb 2023 09:31




Instrument :
BNA_P
ClientSampleId :
SSTD020766


Tgt Ion:165 Resp: 238464
Ion Ratio Lower Upper
165 100
63 39.5 32.0 48.0
182 2.8 2.4 3.6



#58
2,3,4,6-Tetrachlorophenol
Concen: 19.567 ng/ul
RT: 15.363 min Scan# 2070
Delta R.T. -0.000 min
Lab File: BP013868.D
Acq: 22 Feb 2023 09:31

Tgt Ion:232 Resp: 239685
Ion Ratio Lower Upper
232 100
131 38.8 33.0 49.6
130 2.0 1.5 2.3
166 28.9 22.6 33.8









#59
Diethylphthalate
Concen: 18.362 ng/ul
RT: 15.551 min Scan# 2102
Delta R.T. -0.000 min
Lab File: BP013868.D
Acq: 22 Feb 2023 09:31




Instrument :
BNA_P
ClientSampleId :
SSTD020766


Tgt Ion:149 Resp: 836277
Ion Ratio Lower Upper
149 100
177 22.8 17.8 26.8
150 12.5 10.1 15.1



#61
Fluorene
Concen: 18.320 ng/ul
RT: 15.786 min Scan# 2142
Delta R.T. -0.006 min
Lab File: BP013868.D
Acq: 22 Feb 2023 09:31

Tgt Ion:166 Resp: 800906
Ion Ratio Lower Upper
166 100
165 97.4 79.1 118.7
167 13.6 10.8 16.2









#62
4-Chlorophenyl-phenylether
Concen: 18.963 ng/ul
RT: 15.775 min Scan# 2140
Delta R.T. -0.000 min
Lab File: BP013868.D
Acq: 22 Feb 2023 09:31




Instrument :
BNA_P
ClientSampleId :
SSTD020766


Tgt Ion:204 Resp: 431047
Ion Ratio Lower Upper
204 100
206 33.1 26.7 40.1
141 58.3 44.6 66.8



#63
4-Nitroaniline
Concen: 17.066 ng/ul
RT: 15.804 min Scan# 2145
Delta R.T. -0.000 min
Lab File: BP013868.D
Acq: 22 Feb 2023 09:31

Tgt Ion:138 Resp: 126612
Ion Ratio Lower Upper
138 100
92 54.2 43.8 65.8
108 95.7 69.8 104.6









#66
4,6-Dinitro-2-methylphenol
Concen: 19.891 ng/ul
RT: 15.863 min Scan# 2155
Delta R.T. -0.000 min
Lab File: BP013868.D
Acq: 22 Feb 2023 09:31




Instrument :
BNA_P
ClientSampleId :
SSTD020766


Tgt Ion:198 Resp: 154843
Ion Ratio Lower Upper
198 100
182 4.2 4.1 6.1
77 21.7 20.3 30.5



#67
N-Nitrosodiphenylamine
Concen: 19.393 ng/ul
RT: 15.992 min Scan# 2177
Delta R.T. -0.000 min
Lab File: BP013868.D
Acq: 22 Feb 2023 09:31

Tgt Ion:169 Resp: 682245
Ion Ratio Lower Upper
169 100
168 67.3 53.5 80.3
167 36.6 29.8 44.6









#68
4-Bromophenyl-phenylether
Concen: 20.185 ng/ul
RT: 16.675 min Scan# 2293
Delta R.T. -0.000 min
Lab File: BP013868.D
Acq: 22 Feb 2023 09:31




Instrument :
BNA_P
ClientSampleId :
SSTD020766


Tgt Ion:	248	Resp:	275901
Ion	Ratio	Lower	Upper
248	100		
250	97.1	76.0	114.0
141	68.9	56.0	84.0



#69
Hexachlorobenzene
Concen: 20.471 ng/ul
RT: 16.798 min Scan# 2314
Delta R.T. -0.000 min
Lab File: BP013868.D
Acq: 22 Feb 2023 09:31

Tgt Ion:	284	Resp:	326042
Ion	Ratio	Lower	Upper
284	100		
142	27.8	25.0	37.4
249	31.9	24.9	37.3









#70
Atrazine
Concen: 19.811 ng/ul
RT: 16.939 min Scan# 21
Delta R.T. -0.006 min
Lab File: BP013868.D
Acq: 22 Feb 2023 09:31




Instrument :
BNA_P
ClientSampleId :
SSTD020766


Tgt Ion:200 Resp: 278726
Ion Ratio Lower Upper
200 100
173 24.0 20.4 30.6
215 48.6 39.3 58.9



#71
Pentachlorophenol
Concen: 18.994 ng/ul
RT: 17.139 min Scan# 2372
Delta R.T. -0.000 min
Lab File: BP013868.D
Acq: 22 Feb 2023 09:31

Tgt Ion:266 Resp: 196856
Ion Ratio Lower Upper
266 100
264 63.6 50.5 75.7
268 65.2 51.0 76.6





#72

Phenanthrene

Concen: 18.304 ng/ul

RT: 17.539 min Scan# 2440

Delta R.T. -0.000 min

Lab File: BP013868.D

Acq: 22 Feb 2023 09:31

Instrument :

BNA_P

ClientSampleId :

SSTD020766





Tgt Ion:178 Resp: 1253671

Ion Ratio Lower Upper

178 100

179 15.2 12.2 18.2

176 19.2 15.2 22.8



#74

Anthracene

Concen: 18.401 ng/ul

RT: 17.628 min Scan# 2455

Delta R.T. -0.006 min

Lab File: BP013868.D

Acq: 22 Feb 2023 09:31




Tgt Ion:178 Resp: 1267963


Ion Ratio Lower Upper

178 100

179 15.7 12.2 18.4

176 18.7 15.0 22.4









#75
1,2,3,4-Tetrachlorobenzene
Concen: 20.909 ng/uL
RT: 13.545 min Scan# 1761
Delta R.T. -0.000 min
Lab File: BP013868.D
Acq: 22 Feb 2023 09:31




Instrument :
BNA_P
ClientSampleId :
SSTD020766


Tgt Ion:216 Resp: 361265
Ion Ratio Lower Upper
216 100
214 76.8 62.6 94.0
218 49.3 39.0 58.4



#76
Pentachlorobenzene
Concen: 21.069 ng/uL
RT: 15.057 min Scan# 2018
Delta R.T. -0.006 min
Lab File: BP013868.D
Acq: 22 Feb 2023 09:31

Tgt Ion:250 Resp: 375660
Ion Ratio Lower Upper
250 100
248 61.4 51.1 76.7
252 63.7 51.0 76.6









#77
 Carbazole
 Concen: 18.375 ng/ul
 RT: 17.892 min Scan# 2500
 Delta R.T. -0.006 min
 Lab File: BP013868.D
 Acq: 22 Feb 2023 09:31




Instrument :
 BNA_P
 ClientSampleId :
 SSTD020766


Tgt Ion:167 Resp: 1115384
 Ion Ratio Lower Upper
 167 100
 166 21.4 17.3 25.9
 139 12.7 10.4 15.6



#78
 Di-n-butylphthalate
 Concen: 18.317 ng/ul
 RT: 18.422 min Scan# 2590
 Delta R.T. -0.006 min
 Lab File: BP013868.D
 Acq: 22 Feb 2023 09:31

Tgt Ion:149 Resp: 1348536
 Ion Ratio Lower Upper
 149 100
 150 9.2 7.1 10.7
 104 5.4 4.1 6.1





#80

Fluoranthene

Concen: 19.411 ng/ul

RT: 19.486 min Scan# 2771

Delta R.T. -0.006 min

Lab File: BP013868.D

Acq: 22 Feb 2023 09:31

Instrument :

BNA_P

ClientSampleId :

SSTD020766





Tgt Ion:202 Resp: 1505857

Ion Ratio Lower Upper

202 100

101 9.1 8.6 12.8

100 7.3 6.3 9.5



#82

Pyrene

Concen: 18.543 ng/ul

RT: 19.839 min Scan# 2831

Delta R.T. -0.006 min

Lab File: BP013868.D

Acq: 22 Feb 2023 09:31




Tgt Ion:202 Resp: 1533692


Ion Ratio Lower Upper

202 100

101 10.9 10.3 15.5

100 8.6 8.5 12.7





#83

Butylbenzylphthalate

Concen: 20.189 ng/ul

RT: 20.692 min Scan# 2976

Delta R.T. -0.006 min

Lab File: BP013868.D

Acq: 22 Feb 2023 09:31

Instrument :

BNA_P

ClientSampleId :

SSTD020766





Tgt Ion:149 Resp: 571494

Ion Ratio Lower Upper

149 100

91 69.9 57.7 86.5

206 22.4 16.7 25.1



#84

3,3'-Dichlorobenzidine

Concen: 19.411 ng/ul

RT: 21.474 min Scan# 3109

Delta R.T. -0.006 min

Lab File: BP013868.D

Acq: 22 Feb 2023 09:31




Tgt Ion:252 Resp: 522032


Ion Ratio Lower Upper

252 100

254 62.9 51.8 77.6

126 9.8 7.9 11.9









#85
Benzo(a)anthracene
Concen: 18.620 ng/ul
RT: 21.557 min Scan# 3123
Delta R.T. -0.000 min
Lab File: BP013868.D
Acq: 22 Feb 2023 09:31




Instrument :
BNA_P
ClientSampleId :
SSTD020766


Tgt Ion:228 Resp: 1520320
Ion Ratio Lower Upper
228 100
229 19.8 15.4 23.2
226 26.9 21.8 32.6



#86
Bis(2-ethylhexyl)phthalate
Concen: 19.716 ng/ul
RT: 21.451 min Scan# 3105
Delta R.T. -0.000 min
Lab File: BP013868.D
Acq: 22 Feb 2023 09:31

Tgt Ion:149 Resp: 854627
Ion Ratio Lower Upper
149 100
167 27.7 21.5 32.3
279 4.9 3.2 4.8





#87

Chrysene

Concen: 18.636 ng/ul

RT: 21.610 min Scan# 3132

Delta R.T. -0.006 min

Lab File: BP013868.D

Acq: 22 Feb 2023 09:31

Instrument :

BNA_P

ClientSampleId :

SSTD020766





Tgt Ion:228 Resp: 1435605

Ion Ratio Lower Upper

228 100

226 30.4 25.0 37.6

229 19.6 16.4 24.6



#89

Di-n-octyl phthalate

Concen: 13.595 ng/ul




RT: 22.427 min Scan# 3271


Delta R.T. -0.000 min

Lab File: BP013868.D

Acq: 22 Feb 2023 09:31

Tgt Ion:149 Resp: 1426456









#90
Benzo(b)fluoranthene
Concen: 18.084 ng/ul
RT: 23.339 min Scan# 3426
Delta R.T. -0.000 min
Lab File: BP013868.D
Acq: 22 Feb 2023 09:31




Instrument :
BNA_P
ClientSampleId :
SSTD020766


Tgt Ion:252 Resp: 1838689
Ion Ratio Lower Upper
252 100
253 22.0 17.5 26.3
125 6.8 7.8 11.6#



#91
Benzo(k)fluoranthene
Concen: 17.760 ng/ul
RT: 23.386 min Scan# 3434
Delta R.T. -0.006 min
Lab File: BP013868.D
Acq: 22 Feb 2023 09:31

Tgt Ion:252 Resp: 1791463
Ion Ratio Lower Upper
252 100
253 21.8 17.1 25.7
125 7.7 7.2 10.8





#93

Benzo(a)pyrene

Concen: 18.375 ng/ul

RT: 24.010 min Scan# 3540

Delta R.T. -0.006 min

Lab File: BP013868.D

Acq: 22 Feb 2023 09:31

Instrument :

BNA_P

ClientSampleId :

SSTD020766





Tgt Ion:252 Resp: 1583485

Ion Ratio Lower Upper

252 100

253 21.8 17.4 26.0

125 8.0 8.2 12.2#



#94

Indeno(1,2,3-cd)pyrene

Concen: 20.220 ng/ul

RT: 26.809 min Scan# 4016

Delta R.T. -0.012 min

Lab File: BP013868.D

Acq: 22 Feb 2023 09:31




Tgt Ion:276 Resp: 1954511


Ion Ratio Lower Upper

276 100

138 17.2 18.4 27.6#

277 24.6 19.7 29.5





#95

Dibenzo(a,h)anthracene

Concen: 20.200 ng/ul

RT: 26.833 min Scan# 4020

Delta R.T. -0.006 min

Lab File: BP013868.D

Acq: 22 Feb 2023 09:31

Instrument :

BNA_P

ClientSampleId :

SSTD020766





Tgt Ion:278 Resp: 1631030

Ion Ratio Lower Upper

278 100

139 11.9 12.0 18.0#

279 23.7 18.9 28.3



#96

Benzo(g,h,i)perylene

Concen: 20.248 ng/ul

RT: 27.639 min Scan# 4157

Delta R.T. -0.018 min

Lab File: BP013868.D

Acq: 22 Feb 2023 09:31

Tgt Ion:276 Resp: 1558209

Ion Ratio Lower Upper

276 100

138 15.8 15.9 23.9#

277 23.6 19.0 28.6

