

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP022223\
 Data File : BP013874.D
 Acq On : 22 Feb 2023 14:59
 Operator : CG/JU
 Sample : PB150982BL
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SBLK982

Quant Time: Feb 23 03:17:31 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP020723.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed Feb 22 01:41:33 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.110	152	190312	20.000	ng/ul	0.00
20) Naphthalene-d8	10.940	136	742792	20.000	ng/ul	0.00
38) Acenaphthene-d10	14.751	164	490828	20.000	ng/ul	0.00
64) Phenanthrene-d10	17.504	188	1120275	20.000	ng/ul	0.00
79) Chrysene-d12	21.580	240	908772	20.000	ng/ul	0.00
88) Perylene-d12	24.133	264	900803	20.000	ng/ul	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.458	96	25543	5.563	ng/uL	0.00
4) Pyridine-d5	3.887	84	323715	23.780	ng/ul	0.00
7) Phenol-d5	7.257	99	394718	23.278	ng/ul	0.00
9) Bis-(2-Chloroethyl)eth...	7.428	67	237779	23.930	ng/ul	0.00
11) 2-Chlorophenol-d4	7.634	132	316422	25.350	ng/ul	0.00
15) 4-Methylphenol-d8	8.816	113	312922	22.578	ng/ul	0.00
21) Nitrobenzene-d5	9.281	128	149046	28.839	ng/ul	0.00
24) 2-Nitrophenol-d4	10.010	143	173779	29.122	ng/ul	0.00
28) 2,4-Dichlorophenol-d3	10.551	165	311841	25.644	ng/ul	0.00
31) 4-Chloroaniline-d4	11.069	131	413096	23.966	ng/ul	0.00
46) Dimethylphthalate-d6	14.151	166	1047133	27.087	ng/ul	0.00
49) Acenaphthylene-d8	14.445	160	1128049	26.956	ng/ul	0.00
54) 4-Nitrophenol-d4	14.934	143	153214	22.246	ng/ul	0.00
60) Fluorene-d10	15.739	176	894252	27.157	ng/ul	0.00
65) 4,6-Dinitro-2-methylph...	15.851	200	170711	24.839	ng/ul	0.00
73) Anthracene-d10	17.604	188	1400374	27.337	ng/ul	0.00
81) Pyrene-d10	19.822	212	1590514	30.863	ng/ul	0.00
92) Benzo(a)pyrene-d12	23.962	264	1237949	27.427	ng/ul	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.493	88	99	0.020	ng/ul#	33
5) Pyridine	3.911	79	130	0.010	ng/ul#	1
6) Benzaldehyde	7.234	77	151	0.020	ng/ul#	12
8) Phenol	7.287	94	222	0.013	ng/ul#	1
10) Bis(2-Chloroethyl)ether	7.516	93	85	0.006	ng/ul#	75
12) 2-Chlorophenol	7.663	128	120	0.009	ng/ul#	1
13) 2-Methylphenol	8.669	108	14	0.001	ng/ul	87
14) 2,2'-oxybis(1-Chloropr...	8.622	45	169	0.011	ng/ul#	1
16) Acetophenone	8.946	105	129	0.006	ng/ul#	1
17) N-Nitroso-di-n-propyla...	8.940	70	32	0.003	ng/ul#	1
18) 4-Methylphenol	8.834	108	29	0.002	ng/ul	96
19) Hexachloroethane	9.210	117	106	0.020	ng/ul#	8
22) Nitrobenzene	9.334	77	138	0.010	ng/ul#	52
23) Isophorone	9.857	82	43	0.001	ng/ul#	16
25) 2-Nitrophenol	10.034	139	8	0.001	ng/ul#	1
26) 2,4-Dimethylphenol	10.093	107	21	0.001	ng/ul#	49
27) Bis(2-Chloroethoxy)met...	10.334	93	54	0.003	ng/ul#	79
29) 2,4-Dichlorophenol	10.581	162	20	0.002	ng/ul#	1
30) Naphthalene	10.987	128	72	0.002	ng/ul#	63
32) 4-Chloroaniline	11.104	127	33	0.002	ng/ul#	1
33) Hexachlorobutadiene	11.257	225	25	0.003	ng/ul#	55
34) Caprolactam	11.869	113	6	0.001	ng/ul#	1
35) 4-Chloro-3-methylphenol	12.216	107	52	0.004	ng/ul#	58
36) 2-Methylnaphthalene	12.551	142	8	0.000	ng/ul	83

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP022223\
 Data File : BP013874.D
 Acq On : 22 Feb 2023 14:59
 Operator : CG/JU
 Sample : PB150982BL
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SBLK982

Quant Time: Feb 23 03:17:31 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP020723.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed Feb 22 01:41:33 2023
 Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.804	142	59	0.002	ng/ul#	18
39) 1,2,4,5-Tetrachloroben...	12.957	216	7	0.000	ng/ul#	7
40) Hexachlorocyclopentadiene	12.916	237	18	0.002	ng/ul#	76
41) 2,4,6-Trichlorophenol	13.192	196	5	0.000	ng/ul#	1
42) 2,4,5-Trichlorophenol	13.263	196	9	0.001	ng/ul#	28
43) 1,1'-Biphenyl	13.581	154	39	0.001	ng/ul#	70
44) 2-Chloronaphthalene	13.622	162	40	0.001	ng/ul#	71
45) 2-Nitroaniline	13.822	65	41	0.006	ng/ul#	25
47) Dimethylphthalate	14.198	163	158	0.004	ng/ul#	83
48) 2,6-Dinitrotoluene	14.387	165	29	0.004	ng/ul	90
50) Acenaphthylene	14.469	152	107	0.002	ng/ul#	44
51) 3-Nitroaniline	14.651	138	14	0.002	ng/ul#	45
52) Acenaphthene	14.828	153	6	0.000	ng/ul#	79
53) 2,4-Dinitrophenol	14.863	184	7	0.001	ng/ul#	1
55) 4-Nitrophenol	14.957	109	71	0.012	ng/ul#	1
56) Dibenzofuran	15.151	168	317	0.007	ng/ul#	86
57) 2,4-Dinitrotoluene	15.116	165	67	0.007	ng/ul#	53
58) 2,3,4,6-Tetrachlorophenol	15.375	232	10	0.001	ng/ul#	1
59) Diethylphthalate	15.545	149	297	0.008	ng/ul#	37
61) Fluorene	15.792	166	149	0.004	ng/ul#	73
62) 4-Chlorophenyl-phenyle...	15.786	204	58	0.003	ng/ul#	59
63) 4-Nitroaniline	15.816	138	17	0.003	ng/ul#	1
66) 4,6-Dinitro-2-methylph...	15.857	198	715	0.106	ng/ul#	1
67) N-Nitrosodiphenylamine	15.992	169	12	0.000	ng/ul#	57
68) 4-Bromophenyl-phenylether	16.681	248	4	0.000	ng/ul#	1
69) Hexachlorobenzene	16.804	284	61	0.004	ng/ul#	59
70) Atrazine	16.939	200	13	0.001	ng/ul#	1
71) Pentachlorophenol	17.157	266	47	0.005	ng/ul	95
72) Phenanthrene	17.545	178	343	0.006	ng/ul#	85
74) Anthracene	17.639	178	56	0.001	ng/ul#	1
75) 1,2,3,4-Tetrachloroben...	13.557	216	17	0.001	ng/uL	86
76) Pentachlorobenzene	15.039	250	12	0.001	ng/uL	96
77) Carbazole	17.916	167	144	0.003	ng/ul#	80
78) Di-n-butylphthalate	18.427	149	674	0.011	ng/ul#	76
80) Fluoranthene	19.498	202	614	0.010	ng/ul#	92
82) Pyrene	19.845	202	707	0.011	ng/ul#	1
83) Butylbenzylphthalate	20.716	149	292	0.014	ng/ul#	47
84) 3,3'-Dichlorobenzidine	21.480	252	112	0.005	ng/ul#	47
85) Benzo(a)anthracene	21.616	228	1658	0.027	ng/ul	98
86) Bis(2-ethylhexyl)phtha...	21.457	149	2495	0.076	ng/ul#	75
87) Chrysene	21.616	228	1700	0.029	ng/ul	96
89) Di-n-octyl phthalate	22.433	149	387	0.006	ng/ul	100
90) Benzo(b)fluoranthene	23.345	252	1526	0.025	ng/ul#	1
91) Benzo(k)fluoranthene	23.392	252	1117	0.019	ng/ul#	1
93) Benzo(a)pyrene	24.051	252	258	0.005	ng/ul#	1
94) Indeno(1,2,3-cd)pyrene	26.827	276	180	0.003	ng/ul#	1
95) Dibenzo(a,h)anthracene	26.862	278	259	0.005	ng/ul#	1
96) Benzo(g,h,i)perylene	27.656	276	203	0.004	ng/ul#	20


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP022223\
 Data File : BP013874.D
 Acq On : 22 Feb 2023 14:59
 Operator : CG/JU
 Sample : PB150982BL
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SBLK982

Quant Time: Feb 23 03:17:31 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP020723.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed Feb 22 01:41:33 2023
 Response via : Initial Calibration









#2
1,4-Dioxane
Concen: 0.020 ng/uL
RT: 3.493 min Scan# 51
Delta R.T. 0.006 min
Lab File: BP013874.D
Acq: 22 Feb 2023 14:59




Instrument :
BNA_P
ClientSampleId :
SBLK982


Tgt Ion: 88 Resp: 99
Ion Ratio Lower Upper
88 100
43 70.8 20.1 30.1#
58 20.6 53.0 79.6#



#5
Pyridine
Concen: 0.010 ng/uL
RT: 3.911 min Scan# 123
Delta R.T. 0.006 min
Lab File: BP013874.D
Acq: 22 Feb 2023 14:59

Tgt Ion: 79 Resp: 130
Ion Ratio Lower Upper
79 100
52 3255.3 55.1 82.7#
51 156.2 25.7 38.5#









#6
Benzaldehyde
Concen: 0.020 ng/ul
RT: 7.234 min Scan# 611
Delta R.T. -0.000 min
Lab File: BP013874.D
Acq: 22 Feb 2023 14:59




Instrument :
BNA_P
ClientSampleId :
SBLK982


Tgt Ion: 77 Resp: 151
Ion Ratio Lower Upper
77 100
105 19.9 79.5 119.3#
106 1.0 76.5 114.7#



#8
Phenol
Concen: 0.013 ng/ul
RT: 7.287 min Scan# 697
Delta R.T. 0.006 min
Lab File: BP013874.D
Acq: 22 Feb 2023 14:59

Tgt Ion: 94 Resp: 222
Ion Ratio Lower Upper
94 100
65 175.8 22.2 33.4#
66 313.3 29.6 44.4#









#10
Bis(2-Chloroethyl)ether
Concen: 0.006 ng/ul
RT: 7.516 min Scan# 710
Delta R.T. -0.000 min
Lab File: BP013874.D
Acq: 22 Feb 2023 14:59




Instrument :
BNA_P
ClientSampleId :
SBLK982


Tgt Ion: 93 Resp: 85
Ion Ratio Lower Upper
93 100
63 48.5 59.4 89.0#
95 40.0 25.8 38.8#



#12
2-Chlorophenol
Concen: 0.009 ng/ul
RT: 7.663 min Scan# 761
Delta R.T. -0.000 min
Lab File: BP013874.D
Acq: 22 Feb 2023 14:59

Tgt Ion: 128 Resp: 120
Ion Ratio Lower Upper
128 100
64 324.3 37.3 55.9#
130 43.2 26.3 39.5#









#13
2-Methylphenol
Concen: 0.001 ng/ul
RT: 8.669 min Scan# 910
Delta R.T. 0.123 min
Lab File: BP013874.D
Acq: 22 Feb 2023 14:59




Instrument :
BNA_P
ClientSampleId :
SBLK982


Tgt Ion:108 Resp: 14
Ion Ratio Lower Upper
108 100
107 77.3 71.9 107.9



#14
2,2'-oxybis(1-Chloropropane)
Concen: 0.011 ng/ul
RT: 8.622 min Scan# 924
Delta R.T. -0.012 min
Lab File: BP013874.D
Acq: 22 Feb 2023 14:59

Tgt Ion: 45 Resp: 169
Ion Ratio Lower Upper
45 100
77 98.1 13.4 20.2#
79 45.3 11.0 16.4#









#16
Acetophenone
Concen: 0.006 ng/ul
RT: 8.946 min Scan# 91
Delta R.T. 0.012 min
Lab File: BP013874.D
Acq: 22 Feb 2023 14:59




Instrument :
BNA_P
ClientSampleId :
SBLK982


Tgt Ion:105 Resp: 129
Ion Ratio Lower Upper
105 100
77 292.9 61.0 91.4#
51 98.4 19.6 29.4#



#17
N-Nitroso-di-n-propylamine
Concen: 0.003 ng/ul
RT: 8.940 min Scan# 978
Delta R.T. 0.023 min
Lab File: BP013874.D
Acq: 22 Feb 2023 14:59

Tgt Ion: 70 Resp: 32
Ion Ratio Lower Upper
70 100
42 510.0 39.8 59.8#
101 56.7 7.4 11.2#
130 46.7 14.6 22.0#









#18
4-Methylphenol
Concen: 0.002 ng/ul
RT: 8.834 min Scan# 90
Delta R.T. -0.041 min
Lab File: BP013874.D
Acq: 22 Feb 2023 14:59




Instrument :
BNA_P
ClientSampleId :
SBLK982


Tgt Ion:108 Resp: 29
Ion Ratio Lower Upper
108 100
107 109.1 90.9 136.3



#19
Hexachloroethane
Concen: 0.020 ng/ul
RT: 9.210 min Scan# 1024
Delta R.T. 0.012 min
Lab File: BP013874.D
Acq: 22 Feb 2023 14:59

Tgt Ion:117 Resp: 106
Ion Ratio Lower Upper
117 100
201 0.0 86.6 129.8#
199 8.2 55.2 82.8#









#22
Nitrobenzene
Concen: 0.010 ng/ul
RT: 9.334 min Scan# 110
Delta R.T. 0.012 min
Lab File: BP013874.D
Acq: 22 Feb 2023 14:59




Instrument :
BNA_P
ClientSampleId :
SBLK982


Tgt Ion: 77 Resp: 138
Ion Ratio Lower Upper
77 100
123 6.0 35.1 52.7#
65 18.1 10.6 16.0#



#23
Isophorone
Concen: 0.001 ng/ul
RT: 9.857 min Scan# 1134
Delta R.T. 0.006 min
Lab File: BP013874.D
Acq: 22 Feb 2023 14:59

Tgt Ion: 82 Resp: 43
Ion Ratio Lower Upper
82 100
95 89.6 6.0 9.0#
138 9.4 13.7 20.5#





#25

2-Nitrophenol

Concen: 0.001 ng/ul

RT: 10.034 min Scan# 1164

Delta R.T. -0.000 min

Lab File: BP013874.D

Acq: 22 Feb 2023 14:59

Instrument :

BNA_P

ClientSampleId :

SBLK982





Tgt Ion:139 Resp: 8

Ion Ratio Lower Upper

139 100

65 2543.5 39.9 59.9#

109 530.4 27.4 41.0#



#26

2,4-Dimethylphenol

Concen: 0.001 ng/ul

RT: 10.093 min Scan# 1174

Delta R.T. 0.006 min

Lab File: BP013874.D

Acq: 22 Feb 2023 14:59




Tgt Ion:107 Resp: 21


Ion Ratio Lower Upper

107 100

121 0.0 37.8 56.6#

122 114.3 63.7 95.5#









#27
 Bis(2-Chloroethoxy)methane
 Concen: 0.003 ng/ul
 RT: 10.334 min Scan# 11
 Delta R.T. 0.006 min
 Lab File: BP013874.D
 Acq: 22 Feb 2023 14:59




Instrument :
 BNA_P
 ClientSampleId :
 SBLK982


Tgt Ion: 93 Resp: 54
 Ion Ratio Lower Upper
 93 100
 95 20.2 25.0 37.6#
 123 19.3 8.2 12.4#



#29
 2,4-Dichlorophenol
 Concen: 0.002 ng/ul
 RT: 10.581 min Scan# 1257
 Delta R.T. 0.006 min
 Lab File: BP013874.D
 Acq: 22 Feb 2023 14:59

Tgt Ion:162 Resp: 20
 Ion Ratio Lower Upper
 162 100
 164 6626.3 51.4 77.2#
 98 357.9 27.5 41.3#









#30
Naphthalene
Concen: 0.002 ng/ul
RT: 10.987 min Scan# 11
Delta R.T. 0.006 min
Lab File: BP013874.D
Acq: 22 Feb 2023 14:59




Instrument :
BNA_P
ClientSampleId :
SBLK982


Tgt Ion:128 Resp: 72
Ion Ratio Lower Upper
128 100
129 38.3 8.6 12.8#
127 10.2 10.6 15.8#



#32
4-Chloroaniline
Concen: 0.002 ng/ul
RT: 11.104 min Scan# 1346
Delta R.T. 0.018 min
Lab File: BP013874.D
Acq: 22 Feb 2023 14:59

Tgt Ion:127 Resp: 33
Ion Ratio Lower Upper
127 100
129 121.1 26.2 39.4#





#33

Hexachlorobutadiene

Concen: 0.003 ng/ul

RT: 11.257 min Scan# 11

Delta R.T. -0.006 min

Lab File: BP013874.D

Acq: 22 Feb 2023 14:59

Instrument :

BNA_P

ClientSampleId :

SBLK982





Tgt Ion:225 Resp: 25

Ion Ratio Lower Upper

225 100

223 72.4 50.6 76.0

227 0.0 49.0 73.4#



#34

Caprolactam

Concen: 0.001 ng/ul

RT: 11.869 min Scan# 1476

Delta R.T. 0.006 min

Lab File: BP013874.D

Acq: 22 Feb 2023 14:59




Tgt Ion:113 Resp: 6


Ion Ratio Lower Upper

113 100

55 1244.4 120.9 181.3#

56 422.2 93.8 140.8#









#35
4-Chloro-3-methylphenol
Concen: 0.004 ng/ul
RT: 12.216 min Scan# 1532
Delta R.T. 0.018 min
Lab File: BP013874.D
Acq: 22 Feb 2023 14:59




Instrument :
BNA_P
ClientSampleId :
SBLK982


Tgt Ion:107 Resp: 52
Ion Ratio Lower Upper
107 100
144 52.4 20.0 30.0#
142 45.2 62.1 93.1#



#36
2-Methylnaphthalene
Concen: 0.000 ng/ul
RT: 12.551 min Scan# 1592
Delta R.T. -0.030 min
Lab File: BP013874.D
Acq: 22 Feb 2023 14:59

Tgt Ion:142 Resp: 8
Ion Ratio Lower Upper
142 100
141 75.0 72.8 109.2









#37
1-Methylnaphthalene
Concen: 0.002 ng/ul
RT: 12.804 min Scan# 1661
Delta R.T. 0.006 min
Lab File: BP013874.D
Acq: 22 Feb 2023 14:59




Instrument :
BNA_P
ClientSampleId :
SBLK982


Tgt Ion:142 Resp: 59
Ion Ratio Lower Upper
142 100
141 12.1 74.2 111.2#
116 18.2 3.1 4.7#



#39
1,2,4,5-Tetrachlorobenzene
Concen: 0.000 ng/ul
RT: 12.957 min Scan# 1661
Delta R.T. 0.012 min
Lab File: BP013874.D
Acq: 22 Feb 2023 14:59

Tgt Ion:216 Resp: 7
Ion Ratio Lower Upper
216 100
214 35.0 62.6 93.8#
179 30.0 15.4 23.2#
108 180.0 12.7 19.1#





#40

Hexachlorocyclopentadiene

Concen: 0.002 ng/ul

RT: 12.916 min Scan# 1

Delta R.T. -0.006 min

Lab File: BP013874.D

Acq: 22 Feb 2023 14:59





Instrument :

BNA_P

ClientSampleId :

SBLK982

Tgt Ion:	237	Resp:	18
Ion	Ratio	Lower	Upper
237	100		
235	44.8	49.8	74.8#
272	0.0	10.1	15.1#



#41

2,4,6-Trichlorophenol

Concen: 0.000 ng/ul




RT: 13.192 min Scan# 1701


Delta R.T. 0.012 min

Lab File: BP013874.D

Acq: 22 Feb 2023 14:59

Tgt Ion:	196	Resp:	5
Ion	Ratio	Lower	Upper
196	100		
198	1285.7	75.1	112.7#
200	542.9	24.3	36.5#





#42

2,4,5-Trichlorophenol

Concen: 0.001 ng/ul

RT: 13.263 min Scan# 1710

Delta R.T. 0.012 min

Lab File: BP013874.D

Acq: 22 Feb 2023 14:59





Instrument :

BNA_P

ClientSampleId :

SBLK982

Tgt Ion:196	Resp:	9
Ion Ratio	Lower	Upper
196	100	
198	4.5	77.0 115.4#
200	36.4	25.8 38.6



#43

1,1'-Biphenyl

Concen: 0.001 ng/ul




RT: 13.581 min Scan# 1767


Delta R.T. -0.000 min

Lab File: BP013874.D

Acq: 22 Feb 2023 14:59

Tgt Ion:154	Resp:	39
Ion Ratio	Lower	Upper
154	100	
153	32.4	31.9 47.9
76	47.1	9.9 14.9#









#44
2-Chloronaphthalene
Concen: 0.001 ng/ul
RT: 13.622 min Scan# 1774
Delta R.T. -0.000 min
Lab File: BP013874.D
Acq: 22 Feb 2023 14:59




Instrument :
BNA_P
ClientSampleId :
SBLK982


Tgt Ion:	162	Resp:	40
Ion	Ratio	Lower	Upper
162	100		
164	27.8	26.1	39.1
127	66.7	31.1	46.7#



#45
2-Nitroaniline
Concen: 0.006 ng/ul
RT: 13.822 min Scan# 1808
Delta R.T. -0.000 min
Lab File: BP013874.D
Acq: 22 Feb 2023 14:59

Tgt Ion:	65	Resp:	41
Ion	Ratio	Lower	Upper
65	100		
92	10.9	58.6	88.0#
138	25.5	81.2	121.8#









#47
Dimethylphthalate
Concen: 0.004 ng/ul
RT: 14.198 min Scan# 11
Delta R.T. 0.006 min
Lab File: BP013874.D
Acq: 22 Feb 2023 14:59




Instrument :
BNA_P
ClientSampleId :
SBLK982


Tgt Ion:163 Resp: 158
Ion Ratio Lower Upper
163 100
194 0.0 3.4 5.2#
164 3.0 7.8 11.8#



#48
2,6-Dinitrotoluene
Concen: 0.004 ng/ul
RT: 14.387 min Scan# 1904
Delta R.T. 0.070 min
Lab File: BP013874.D
Acq: 22 Feb 2023 14:59

Tgt Ion:165 Resp: 29
Ion Ratio Lower Upper
165 100
89 50.0 47.5 71.3
121 19.6 16.6 24.8





#50

Acenaphthylene

Concen: 0.002 ng/ul

RT: 14.469 min Scan# 1917

Delta R.T. -0.000 min

Lab File: BP013874.D

Acq: 22 Feb 2023 14:59

Instrument :

BNA_P

ClientSampleId :

SBLK982





Tgt Ion:152 Resp: 107

Ion Ratio Lower Upper

152 100

151 39.0 15.8 23.6#

153 44.0 10.4 15.6#



#51

3-Nitroaniline

Concen: 0.002 ng/ul

RT: 14.651 min Scan# 1949

Delta R.T. 0.006 min

Lab File: BP013874.D

Acq: 22 Feb 2023 14:59




Tgt Ion:138 Resp: 14


Ion Ratio Lower Upper

138 100

108 152.9 9.4 14.2#

92 94.1 97.4 146.0#





#52

Acenaphthene

Concen: 0.000 ng/ul

RT: 14.828 min Scan# 1979

Delta R.T. 0.018 min

Lab File: BP013874.D

Acq: 22 Feb 2023 14:59

Instrument :

BNA_P

ClientSampleId :

SBLK982





Tgt Ion:153 Resp: 6

Ion Ratio Lower Upper

153 100

152 27.8 37.4 56.0#

154 72.2 70.1 105.1



#53

2,4-Dinitrophenol

Concen: 0.001 ng/ul

RT: 14.863 min Scan# 1985

Delta R.T. 0.012 min

Lab File: BP013874.D

Acq: 22 Feb 2023 14:59




Tgt Ion:184 Resp: 7


Ion Ratio Lower Upper

184 100

63 164.3 49.5 74.3#

154 157.1 54.3 81.5#









#55
4-Nitrophenol
Concen: 0.012 ng/ul
RT: 14.957 min Scan# 2001
Delta R.T. 0.012 min
Lab File: BP013874.D
Acq: 22 Feb 2023 14:59




Instrument :
BNA_P
ClientSampleId :
SBLK982


Tgt Ion:109 Resp: 71
Ion Ratio Lower Upper
109 100
139 20.0 96.0 144.0#
65 814.7 95.8 143.6#



#56
Dibenzofuran
Concen: 0.007 ng/ul
RT: 15.151 min Scan# 2034
Delta R.T. 0.012 min
Lab File: BP013874.D
Acq: 22 Feb 2023 14:59

Tgt Ion:168 Resp: 317
Ion Ratio Lower Upper
168 100
139 28.8 29.6 44.4#









#57
2,4-Dinitrotoluene
Concen: 0.007 ng/ul
RT: 15.116 min Scan# 2072
Delta R.T. 0.018 min
Lab File: BP013874.D
Acq: 22 Feb 2023 14:59




Instrument :
BNA_P
ClientSampleId :
SBLK982


Tgt Ion:165 Resp: 67
Ion Ratio Lower Upper
165 100
63 68.2 32.0 48.0#
182 22.7 2.4 3.6#



#58
2,3,4,6-Tetrachlorophenol
Concen: 0.001 ng/ul
RT: 15.375 min Scan# 2072
Delta R.T. 0.012 min
Lab File: BP013874.D
Acq: 22 Feb 2023 14:59

Tgt Ion:232 Resp: 10
Ion Ratio Lower Upper
232 100
131 125.9 33.0 49.6#
130 74.1 1.5 2.3#
166 92.6 22.6 33.8#









#59
Diethylphthalate
Concen: 0.008 ng/ul
RT: 15.545 min Scan# 2101
Delta R.T. -0.006 min
Lab File: BP013874.D
Acq: 22 Feb 2023 14:59




Instrument :
BNA_P
ClientSampleId :
SBLK982


Tgt Ion:149 Resp: 297
Ion Ratio Lower Upper
149 100
177 67.2 17.8 26.8#
150 9.4 10.1 15.1#



#61
Fluorene
Concen: 0.004 ng/ul
RT: 15.792 min Scan# 2143
Delta R.T. -0.000 min
Lab File: BP013874.D
Acq: 22 Feb 2023 14:59

Tgt Ion:166 Resp: 149
Ion Ratio Lower Upper
166 100
165 69.0 79.1 118.7#
167 14.7 10.8 16.2









#62
4-Chlorophenyl-phenylether
Concen: 0.003 ng/ul
RT: 15.786 min Scan# 2140
Delta R.T. 0.012 min
Lab File: BP013874.D
Acq: 22 Feb 2023 14:59




Instrument :
BNA_P
ClientSampleId :
SBLK982


Tgt Ion:204 Resp: 58
Ion Ratio Lower Upper
204 100
206 79.0 26.7 40.1#
141 43.5 44.6 66.8#



#63
4-Nitroaniline
Concen: 0.003 ng/ul
RT: 15.816 min Scan# 2147
Delta R.T. 0.012 min
Lab File: BP013874.D
Acq: 22 Feb 2023 14:59

Tgt Ion:138 Resp: 17
Ion Ratio Lower Upper
138 100
92 235.0 43.8 65.8#
108 260.0 69.8 104.6#









#66
4,6-Dinitro-2-methylphenol
Concen: 0.106 ng/ul
RT: 15.857 min Scan# 2154
Delta R.T. -0.006 min
Lab File: BP013874.D
Acq: 22 Feb 2023 14:59




Instrument :
BNA_P
ClientSampleId :
SBLK982


Tgt Ion:198 Resp: 715
Ion Ratio Lower Upper
198 100
182 468.6 4.1 6.1#
77 1255.2 20.3 30.5#



#67
N-Nitrosodiphenylamine
Concen: 0.000 ng/ul
RT: 15.992 min Scan# 2177
Delta R.T. -0.000 min
Lab File: BP013874.D
Acq: 22 Feb 2023 14:59

Tgt Ion:169 Resp: 12
Ion Ratio Lower Upper
169 100
168 44.8 53.5 80.3#
167 79.3 29.8 44.6#









#68
4-Bromophenyl-phenylether
Concen: 0.000 ng/ul
RT: 16.681 min Scan# 21
Delta R.T. 0.006 min
Lab File: BP013874.D
Acq: 22 Feb 2023 14:59




Instrument :
BNA_P
ClientSampleId :
SBLK982


Tgt Ion:248	Resp:	4	
Ion	Ratio	Lower	Upper
248	100		
250	140.0	76.0	114.0#
141	270.0	56.0	84.0#



#69
Hexachlorobenzene
Concen: 0.004 ng/ul
RT: 16.804 min Scan# 2315
Delta R.T. 0.006 min
Lab File: BP013874.D
Acq: 22 Feb 2023 14:59

Tgt Ion:284	Resp:	61	
Ion	Ratio	Lower	Upper
284	100		
142	0.0	25.0	37.4#
249	16.7	24.9	37.3#





#70

Atrazine

Concen: 0.001 ng/ul

RT: 16.939 min Scan# 21

Delta R.T. -0.006 min

Lab File: BP013874.D

Acq: 22 Feb 2023 14:59

Instrument :

BNA_P

ClientSampleId :

SBLK982





Tgt Ion:200 Resp: 13

Ion Ratio Lower Upper

200 100

173 230.0 20.4 30.6#

215 130.0 39.3 58.9#



#71

Pentachlorophenol

Concen: 0.005 ng/ul

RT: 17.157 min Scan# 2375

Delta R.T. 0.018 min

Lab File: BP013874.D

Acq: 22 Feb 2023 14:59




Tgt Ion:266 Resp: 47


Ion Ratio Lower Upper

266 100

264 69.5 50.5 75.7

268 62.7 51.0 76.6





#72

Phenanthrene

Concen: 0.006 ng/ul

RT: 17.545 min Scan# 2440

Delta R.T. 0.006 min

Lab File: BP013874.D

Acq: 22 Feb 2023 14:59

Instrument :

BNA_P

ClientSampleId :

SBLK982





Tgt Ion:178 Resp: 343

Ion Ratio Lower Upper

178 100

179 25.0 12.2 18.2#

176 22.6 15.2 22.8



#74

Anthracene

Concen: 0.001 ng/ul

RT: 17.639 min Scan# 2457

Delta R.T. 0.006 min

Lab File: BP013874.D

Acq: 22 Feb 2023 14:59




Tgt Ion:178 Resp: 56


Ion Ratio Lower Upper

178 100

179 87.6 12.2 18.4#

176 0.0 15.0 22.4#





#75

1,2,3,4-Tetrachlorobenzene

Concen: 0.001 ng/uL

RT: 13.557 min Scan# 1

Delta R.T. 0.012 min

Lab File: BP013874.D

Acq: 22 Feb 2023 14:59

Instrument :

BNA_P

ClientSampleId :

SBLK982





Tgt Ion:216 Resp: 17

Ion Ratio Lower Upper

216 100

214 63.3 62.6 94.0

218 43.3 39.0 58.4



#76

Pentachlorobenzene

Concen: 0.001 ng/uL

RT: 15.039 min Scan# 2015

Delta R.T. -0.024 min

Lab File: BP013874.D

Acq: 22 Feb 2023 14:59




Tgt Ion:250 Resp: 12


Ion Ratio Lower Upper

250 100

248 60.0 51.1 76.7

252 66.7 51.0 76.6









#77
 Carbazole
 Concen: 0.003 ng/ul
 RT: 17.916 min Scan# 2504
 Delta R.T. 0.018 min
 Lab File: BP013874.D
 Acq: 22 Feb 2023 14:59




Instrument :
 BNA_P
 ClientSampleId :
 SBLK982


Tgt Ion:167 Resp: 144
 Ion Ratio Lower Upper
 167 100
 166 36.2 17.3 25.9#
 139 13.8 10.4 15.6



#78
 Di-n-butylphthalate
 Concen: 0.011 ng/ul
 RT: 18.427 min Scan# 2591
 Delta R.T. -0.000 min
 Lab File: BP013874.D
 Acq: 22 Feb 2023 14:59

Tgt Ion:149 Resp: 674
 Ion Ratio Lower Upper
 149 100
 150 4.6 7.1 10.7#
 104 20.4 4.1 6.1#





#80

Fluoranthene

Concen: 0.010 ng/ul

RT: 19.498 min Scan# 21

Delta R.T. 0.006 min

Lab File: BP013874.D

Acq: 22 Feb 2023 14:59

Instrument :

BNA_P

ClientSampleId :

SBLK982





Tgt Ion:202 Resp: 614

Ion Ratio Lower Upper

202 100

101 12.3 8.6 12.8

100 3.6 6.3 9.5#



#82

Pyrene

Concen: 0.011 ng/ul

RT: 19.845 min Scan# 2832

Delta R.T. -0.000 min

Lab File: BP013874.D

Acq: 22 Feb 2023 14:59




Tgt Ion:202 Resp: 707


Ion Ratio Lower Upper

202 100

101 64.5 10.3 15.5#

100 74.2 8.5 12.7#





#83

Butylbenzylphthalate

Concen: 0.014 ng/ul

RT: 20.716 min Scan# 2976

Delta R.T. 0.018 min

Lab File: BP013874.D

Acq: 22 Feb 2023 14:59

Instrument :

BNA_P

ClientSampleId :

SBLK982





Tgt Ion:149 Resp: 292

Ion Ratio Lower Upper

149 100

91 25.8 57.7 86.5#

206 0.0 16.7 25.1#



#84

3,3'-Dichlorobenzidine

Concen: 0.005 ng/ul

RT: 21.480 min Scan# 3110

Delta R.T. -0.000 min

Lab File: BP013874.D

Acq: 22 Feb 2023 14:59




Tgt Ion:252 Resp: 112


Ion Ratio Lower Upper

252 100

254 27.6 51.8 77.6#

126 44.4 7.9 11.9#









#85
Benzo(a)anthracene
Concen: 0.027 ng/ul
RT: 21.616 min Scan# 3123
Delta R.T. 0.059 min
Lab File: BP013874.D
Acq: 22 Feb 2023 14:59




Instrument :
BNA_P
ClientSampleId :
SBLK982


Tgt Ion:	228	Resp:	1658
Ion Ratio	Lower	Upper	
228	100		
229	19.9	15.4	23.2
226	28.2	21.8	32.6



#86
Bis(2-ethylhexyl)phthalate
Concen: 0.076 ng/ul
RT: 21.457 min Scan# 3106
Delta R.T. 0.006 min
Lab File: BP013874.D
Acq: 22 Feb 2023 14:59

Tgt Ion:	149	Resp:	2495
Ion Ratio	Lower	Upper	
149	100		
167	17.8	21.5	32.3#
279	28.2	3.2	4.8#





#87

Chrysene

Concen: 0.029 ng/ul

RT: 21.616 min Scan# 3132

Delta R.T. -0.000 min

Lab File: BP013874.D

Acq: 22 Feb 2023 14:59

Instrument :

BNA_P

ClientSampleId :

SBLK982





Tgt Ion:228 Resp: 1700

Ion Ratio Lower Upper

228 100

226 28.2 25.0 37.6

229 19.9 16.4 24.6



#89

Di-n-octyl phthalate

Concen: 0.006 ng/ul




RT: 22.433 min Scan# 3272


Delta R.T. 0.006 min

Lab File: BP013874.D

Acq: 22 Feb 2023 14:59

Tgt Ion:149 Resp: 387









#90
Benzo(b)fluoranthene
Concen: 0.025 ng/ul
RT: 23.345 min Scan# 3426
Delta R.T. 0.006 min
Lab File: BP013874.D
Acq: 22 Feb 2023 14:59




Instrument :
BNA_P
ClientSampleId :
SBLK982


Tgt Ion:252 Resp: 1526
Ion Ratio Lower Upper
252 100
253 59.8 17.5 26.3#
125 66.6 7.8 11.6#



#91
Benzo(k)fluoranthene
Concen: 0.019 ng/ul
RT: 23.392 min Scan# 3435
Delta R.T. -0.000 min
Lab File: BP013874.D
Acq: 22 Feb 2023 14:59

Tgt Ion:252 Resp: 1117
Ion Ratio Lower Upper
252 100
253 64.0 17.1 25.7#
125 93.3 7.2 10.8#









#93
Benzo(a)pyrene
Concen: 0.005 ng/ul
RT: 24.051 min Scan# 31
Delta R.T. 0.035 min
Lab File: BP013874.D
Acq: 22 Feb 2023 14:59




Instrument :
BNA_P
ClientSampleId :
SBLK982


Tgt Ion:252 Resp: 258
Ion Ratio Lower Upper
252 100
253 97.0 17.4 26.0#
125 68.0 8.2 12.2#



#94
Indeno(1,2,3-cd)pyrene
Concen: 0.003 ng/ul
RT: 26.827 min Scan# 4019
Delta R.T. 0.006 min
Lab File: BP013874.D
Acq: 22 Feb 2023 14:59

Tgt Ion:276 Resp: 180
Ion Ratio Lower Upper
276 100
138 31.1 18.4 27.6#
277 113.0 19.7 29.5#





#95

Dibenzo(a,h)anthracene

Concen: 0.005 ng/ul

RT: 26.862 min Scan# 4020

Delta R.T. 0.023 min

Lab File: BP013874.D

Acq: 22 Feb 2023 14:59

Instrument :

BNA_P

ClientSampleId :

SBLK982





Tgt Ion:278 Resp: 259

Ion Ratio Lower Upper

278 100

139 116.7 12.0 18.0#

279 96.0 18.9 28.3#



#96

Benzo(g,h,i)perylene

Concen: 0.004 ng/ul

RT: 27.656 min Scan# 4160

Delta R.T. -0.000 min

Lab File: BP013874.D

Acq: 22 Feb 2023 14:59

Tgt Ion:276 Resp: 203

Ion Ratio Lower Upper

276 100

138 25.2 15.9 23.9#

277 91.6 19.0 28.6#

