

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP022423\
 Data File : BP013884.D
 Acq On : 24 Feb 2023 12:56
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 LabSampleId :
 SSTDCCC040

Quant Time: Feb 24 16:42:08 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP021623.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Feb 24 16:37:47 2023
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	66	0.00
2	1,4-Dioxane	40.000	41.724	-4.3	73	0.00
3	Pyridine	40.000	43.574	-8.9	74	0.00
4	n-Nitrosodimethylamine	40.000	47.152	-17.9	80	0.00
5 S	2-Fluorophenol	80.000	85.041	-6.3	70	0.00
6	Aniline	40.000	43.036	-7.6	68	0.00
7 S	Phenol-d6	80.000	85.505	-6.9	67	0.00
8	2-Chlorophenol	40.000	42.088	-5.2	70	0.00
9	Benzaldehyde	40.000	40.629	-1.6	70	0.00
10 C	Phenol	40.000	42.225	-5.6	67	0.00
11	bis(2-Chloroethyl)ether	40.000	42.162	-5.4	73	0.00
12	1,3-Dichlorobenzene	40.000	42.144	-5.4	67	0.00
13 C	1,4-Dichlorobenzene	40.000	42.327	-5.8	70	0.00
14	1,2-Dichlorobenzene	40.000	42.983	-7.5	69	0.00
15	Benzyl Alcohol	40.000	47.548	-18.9	73	0.00
16	2,2'-oxybis(1-Chloropropane	40.000	41.321	-3.3	64	0.00
17	2-Methylphenol	40.000	45.009	-12.5	69	0.00
18	Hexachloroethane	40.000	43.610	-9.0	70	0.00
19 P	n-Nitroso-di-n-propylamine	40.000	47.460	-18.7	70	0.00
20	3+4-Methylphenols	40.000	44.924	-12.3	68	0.00
21 I	Naphthalene-d8	20.000	20.000	0.0	63	0.00
22	Acetophenone	40.000	47.238	-18.1	69	0.00
23 S	Nitrobenzene-d5	80.000	97.354	-21.7	72	0.00
24	Nitrobenzene	40.000	48.099	-20.2	70	0.00
25	Isophorone	40.000	46.839	-17.1	69	0.00
26 C	2-Nitrophenol	40.000	46.072	-15.2	72	0.00
27	2,4-Dimethylphenol	40.000	46.092	-15.2	67	0.00
28	bis(2-Chloroethoxy)methane	40.000	43.403	-8.5	65	0.00
29 C	2,4-Dichlorophenol	40.000	47.357	-18.4	69	0.00
30	1,2,4-Trichlorobenzene	40.000	44.316	-10.8	69	0.00
31	Naphthalene	40.000	42.801	-7.0	67	0.00
32	Benzoic acid	40.000	43.794	-9.5	71	0.00
33	4-Chloroaniline	40.000	45.324	-13.3	67	0.00
34 C	Hexachlorobutadiene	40.000	46.734	-16.8	73	0.00
35	Caprolactam	40.000	45.272	-13.2	69	0.00
36 C	4-Chloro-3-methylphenol	40.000	46.570	-16.4	70	0.00
37	2-Methylnaphthalene	40.000	43.508	-8.8	67	0.00
38	1-Methylnaphthalene	40.000	43.442	-8.6	67	0.00
39 I	Acenaphthene-d10	20.000	20.000	0.0	65	0.00
40	1,2,4,5-Tetrachlorobenzene	40.000	44.038	-10.1	71	0.00
41 P	Hexachlorocyclopentadiene	40.000	46.411	-16.0	71	0.00
42 S	2,4,6-Tribromophenol	80.000	99.845	-24.8	76	0.00
43 C	2,4,6-Trichlorophenol	40.000	46.430	-16.1	71	0.00
44	2,4,5-Trichlorophenol	40.000	45.385	-13.5	70	0.00
45 S	2-Fluorobiphenyl	80.000	86.435	-8.0	70	0.00
46	1,1'-Biphenyl	40.000	42.647	-6.6	68	0.00
47	2-Chloronaphthalene	40.000	43.012	-7.5	69	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP022423\
 Data File : BP013884.D
 Acq On : 24 Feb 2023 12:56
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 LabSampleId :
 SSTDCCC040

Quant Time: Feb 24 16:42:08 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP021623.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Feb 24 16:37:47 2023
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48	2-Nitroaniline	40.000	46.042	-15.1	76	0.00
49	Acenaphthylene	40.000	44.052	-10.1	69	0.00
50	Dimethylphthalate	40.000	44.505	-11.3	70	0.00
51	2,6-Dinitrotoluene	40.000	43.578	-8.9	70	0.00
52 C	Acenaphthene	40.000	43.683	-9.2	68	0.00
53	3-Nitroaniline	40.000	43.328	-8.3	70	0.00
54 P	2,4-Dinitrophenol	40.000	45.533	-13.8	76	0.00
55	Dibenzofuran	40.000	43.559	-8.9	69	0.00
56 P	4-Nitrophenol	40.000	45.156	-12.9	69	0.00
57	2,4-Dinitrotoluene	40.000	45.020	-12.6	73	0.00
58	Fluorene	40.000	44.065	-10.2	70	0.00
59	2,3,4,6-Tetrachlorophenol	40.000	46.846	-17.1	72	0.00
60	Diethylphthalate	40.000	45.381	-13.5	71	0.00
61	4-Chlorophenyl-phenylether	40.000	44.560	-11.4	70	0.00
62	4-Nitroaniline	40.000	44.424	-11.1	72	0.00
63	Azobenzene	40.000	45.472	-13.7	69	0.00
64 I	Phenanthrene-d10	20.000	20.000	0.0	66	0.00
65	4,6-Dinitro-2-methylphenol	40.000	44.259	-10.6	76	0.00
66 c	n-Nitrosodiphenylamine	40.000	43.352	-8.4	70	0.00
67	4-Bromophenyl-phenylether	40.000	44.207	-10.5	71	0.00
68	Hexachlorobenzene	40.000	44.924	-12.3	74	0.00
69	Atrazine	40.000	46.187	-15.5	73	0.00
70 C	Pentachlorophenol	40.000	47.631	-19.1	74	0.00
71	Phenanthrene	40.000	42.989	-7.5	71	0.00
72	Anthracene	40.000	43.728	-9.3	70	0.00
73	Carbazole	40.000	43.676	-9.2	70	0.00
74	Di-n-butylphthalate	40.000	45.640	-14.1	71	0.00
75 C	Fluoranthene	40.000	43.351	-8.4	71	0.00
76 I	Chrysene-d12	20.000	20.000	0.0	68	0.00
77	Benzydine	40.000	44.373	-10.9	79	0.00
78	Pyrene	40.000	44.051	-10.1	71	0.00
79 S	Terphenyl-d14	80.000	90.734	-13.4	75	0.00
80	Butylbenzylphthalate	40.000	48.275	-20.7	75	0.00
81	Benzo(a)anthracene	40.000	43.638	-9.1	72	0.00
82	3,3'-Dichlorobenzidine	40.000	46.448	-16.1	73	0.00
83	Chrysene	40.000	43.002	-7.5	71	0.00
84	Bis(2-ethylhexyl)phthalate	40.000	46.874	-17.2	73	0.00
85 c	Di-n-octyl phthalate	40.000	47.130	-17.8	76	0.00
86 I	Perylene-d12	20.000	20.000	0.0	71	0.00
87	Indeno(1,2,3-cd)pyrene	40.000	43.751	-9.4	76	0.00
88	Benzo(b)fluoranthene	40.000	43.545	-8.9	76	0.00
89	Benzo(k)fluoranthene	40.000	43.425	-8.6	73	0.00
90 C	Benzo(a)pyrene	40.000	43.844	-9.6	75	0.00
91	Dibenzo(a,h)anthracene	40.000	43.519	-8.8	75	0.00
92	Benzo(g,h,i)perylene	40.000	43.640	-9.1	75	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP022423\
Data File : BP013884.D
Acq On : 24 Feb 2023 12:56
Operator : CG/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_P
LabSampleId :
SSTDCCC040

Quant Time: Feb 24 16:42:08 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP021623.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Feb 24 16:37:47 2023
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range SPCC's out = 0 CCC's out = 0