

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP022522\
 Data File : BP009234.D
 Acq On : 25 Feb 2022 14:40
 Operator : CG/JU
 Sample : SST00550
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SST005601

Quant Time: Feb 25 23:45:33 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP022522.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Fri Feb 25 23:39:41 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.828	152	2215	20.000	ng/ul	0.00
20) Naphthalene-d8	10.628	136	7716	20.000	ng/ul	0.00
38) Acenaphthene-d10	14.469	164	6655	20.000	ng/ul	0.00
64) Phenanthrene-d10	17.228	188	16183	20.000	ng/ul	# 0.00
79) Chrysene-d12	21.339	240	20615	20.000	ng/ul	# 0.00
88) Perylene-d12	23.751	264	22632	20.000	ng/ul	# 0.00

System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.346	96	12	0.244	ng/uL	0.02
4) Pyridine-d5	3.728	84	843	6.130	ng/ul	0.00
7) Phenol-d5	6.993	99	1112	6.293	ng/ul	0.00
9) Bis-(2-Chloroethyl)eth...	7.210	67	14	0.139	ng/ul	0.05
11) 2-Chlorophenol-d4	7.358	132	457	3.251	ng/ul	0.00
15) 4-Methylphenol-d8	8.534	113	850	5.996	ng/ul	0.00
21) Nitrobenzene-d5	8.987	128	325	5.952	ng/ul	0.01
24) 2-Nitrophenol-d4	9.699	143	212	4.235	ng/ul	0.00
28) 2,4-Dichlorophenol-d3	10.240	165	883	7.185	ng/ul	0.00
31) 4-Chloroaniline-d4	10.751	131	989	5.830	ng/ul	0.00
46) Dimethylphthalate-d6	13.881	166	2505	5.094	ng/ul	0.00
49) Acenaphthylene-d8	14.163	160	3123	5.079	ng/ul	0.00
54) 4-Nitrophenol-d4	14.645	143	337	4.324	ng/ul	0.00
60) Fluorene-d10	15.463	176	2481	5.811	ng/ul	0.00
65) 4,6-Dinitro-2-methylph...	15.581	200	585	8.212	ng/ul	0.00
73) Anthracene-d10	17.328	188	3882	5.062	ng/ul	0.00
81) Pyrene-d10	19.575	212	4704	3.808	ng/ul	0.00
92) Benzo(a)pyrene-d12	23.592	264	5280	4.533	ng/ul	0.00

Target Compounds				Qvalue		
2) 1,4-Dioxane	3.358	88	70	1.318	ng/uL#	59
5) Pyridine	3.758	79	536	3.799	ng/ul	100
6) Benzaldehyde	6.787	77	6	0.060	ng/ul	91
8) Phenol	7.011	94	578	3.183	ng/ul#	84
10) Bis(2-Chloroethyl)ether	7.258	93	923	6.352	ng/ul#	79
12) 2-Chlorophenol	7.405	128	782	5.330	ng/ul#	70
13) 2-Methylphenol	8.310	108	6	0.043	ng/ul	88
14) 2,2'-oxybis(1-Chloropr...	8.363	45	304	1.816	ng/ul#	68
16) Acetophenone	8.646	105	1049	4.784	ng/ul	98
17) N-Nitroso-di-n-propyla...	8.634	70	405	3.711	ng/ul#	72
18) 4-Methylphenol	8.599	108	1003	6.660	ng/ul	83
19) Hexachloroethane	8.916	117	392	6.564	ng/ul#	83
22) Nitrobenzene	9.022	77	1056	8.382	ng/ul#	70
23) Isophorone	9.552	82	1973	7.468	ng/ul#	88
25) 2-Nitrophenol	9.599	139	6	0.104	ng/ul	87
26) 2,4-Dimethylphenol	9.799	107	947	6.955	ng/ul#	83
27) Bis(2-Chloroethoxy)met...	10.040	93	736	4.457	ng/ul#	97
29) 2,4-Dichlorophenol	10.263	162	867	7.034	ng/ul#	90
30) Naphthalene	10.675	128	2466	6.038	ng/ul#	93
32) 4-Chloroaniline	10.775	127	981	5.703	ng/ul	99
33) Hexachlorobutadiene	10.846	225	15	0.176	ng/ul	96
34) Caprolactam	11.710	113	7	0.199	ng/ul	84
35) 4-Chloro-3-methylphenol	11.910	107	1304	10.790	ng/ul#	87
36) 2-Methylnaphthalene	12.293	142	1503	5.201	ng/ul	88

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP022522\
 Data File : BP009234.D
 Acq On : 25 Feb 2022 14:40
 Operator : CG/JU
 Sample : SSTD00550
 Misc :
 ALS Vial : 2 Sample Multiplier: 1


Instrument :
 BNA_P
 ClientSampleId :
 SSTD005601


Quant Time: Feb 25 23:45:33 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP022522.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Fri Feb 25 23:39:41 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.516	142	1682	5.717	ng/ul	99
39) 1,2,4,5-Tetrachloroben...	12.657	216	1064	4.953	ng/ul#	90
40) Hexachlorocyclopentadiene	12.651	237	977	6.796	ng/ul	90
41) 2,4,6-Trichlorophenol	12.898	196	715	5.862	ng/ul#	69
42) 2,4,5-Trichlorophenol	12.963	196	800	6.027	ng/ul#	67
43) 1,1'-Biphenyl	13.304	154	2154	4.175	ng/ul#	74
44) 2-Chloronaphthalene	13.251	162	19	0.048	ng/ul	94
45) 2-Nitroaniline	13.493	65	5	0.063	ng/ul	88
47) Dimethylphthalate	13.928	163	2481	5.057	ng/ul#	85
48) 2,6-Dinitrotoluene	14.040	165	505	5.791	ng/ul#	76
50) Acenaphthylene	14.187	152	2780	4.419	ng/ul	97
51) 3-Nitroaniline	14.357	138	532	5.920	ng/ul#	95
52) Acenaphthene	14.534	153	1976	4.810	ng/ul	91
53) 2,4-Dinitrophenol	14.569	184	425	11.547	ng/ul#	83
55) 4-Nitrophenol	14.604	109	6	0.111	ng/ul	92
56) Dibenzofuran	14.869	168	2784	4.635	ng/ul	91
57) 2,4-Dinitrotoluene	14.828	165	706	5.851	ng/ul#	99
58) 2,3,4,6-Tetrachlorophenol	15.098	232	752	7.293	ng/ul#	89
59) Diethylphthalate	15.304	149	2418	5.046	ng/ul#	88
61) Fluorene	15.522	166	2336	4.881	ng/ul	96
62) 4-Chlorophenyl-phenyle...	15.528	204	1481	5.834	ng/ul#	87
63) 4-Nitroaniline	15.522	138	583	7.078	ng/ul	94
66) 4,6-Dinitro-2-methylph...	15.592	198	442	5.864	ng/ul#	85
67) N-Nitrosodiphenylamine	15.728	169	2141	4.326	ng/ul#	85
68) 4-Bromophenyl-phenylether	16.475	248	6	0.032	ng/ul	92
69) Hexachlorobenzene	16.528	284	501	2.318	ng/ul#	85
70) Atrazine	16.692	200	1040	5.538	ng/ul#	71
71) Pentachlorophenol	16.881	266	820	7.424	ng/ul#	81
72) Phenanthrene	17.369	178	4338	4.867	ng/ul	96
74) Anthracene	17.369	178	4338	4.854	ng/ul	96
75) 1,2,3,4-Tetrachloroben...	13.263	216	1190	4.335	ng/ul#	69
76) Pentachlorobenzene	14.787	250	1205	4.782	ng/ul	96
77) Carbazole	17.633	167	3668	4.816	ng/ul#	90
78) Di-n-butylphthalate	18.210	149	4578	5.148	ng/ul#	96
80) Fluoranthene	19.245	202	5384	3.666	ng/ul#	82
82) Pyrene	19.604	202	5779	3.905	ng/ul#	92
83) Butylbenzylphthalate	20.498	149	2303	4.591	ng/ul#	94
84) 3,3'-Dichlorobenzidine	21.257	252	2563	5.870	ng/ul#	99
85) Benzo(a)anthracene	21.322	228	6281	4.719	ng/ul	93
86) Bis(2-ethylhexyl)phtha...	21.274	149	3499	4.596	ng/ul#	96
87) Chrysene	21.374	228	5829	4.536	ng/ul#	91
89) Di-n-octyl phthalate	22.210	149	5594	3.882	ng/ul	100
90) Benzo(b)fluoranthene	23.016	252	6853	4.570	ng/ul#	95
91) Benzo(k)fluoranthene	23.016	252	6837	4.724	ng/ul	96
93) Benzo(a)pyrene	23.645	252	6578	4.582	ng/ul	96
94) Indeno(1,2,3-cd)pyrene	26.257	276	5104	3.224	ng/ul#	88
95) Dibenzo(a,h)anthracene	26.286	278	6833	4.974	ng/ul#	84
96) Benzo(g,h,i)perylene	27.021	276	4504	3.442	ng/ul#	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Quant Time: Feb 25 23:45:33 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP022522.M
Quant Title : SVOA CALIBRATION
QLast Update : Fri Feb 25 23:39:41 2022
Response via : Initial Calibration









#2
1,4-Dioxane
Concen: 1.318 ng/uL
RT: 3.358 min Scan# 61
Delta R.T. -0.000 min
Lab File: BP009234.D
Acq: 25 Feb 2022 14:40




Instrument :
BNA_P
ClientSampleId :
SSTD005601


Tgt Ion: 88 Resp: 70
Ion Ratio Lower Upper
88 100
43 29.3 6.8 10.2#
58 57.3 28.4 42.6#



#5
Pyridine
Concen: 3.799 ng/uL
RT: 3.758 min Scan# 131
Delta R.T. 0.018 min
Lab File: BP009234.D
Acq: 25 Feb 2022 14:40

Tgt Ion: 79 Resp: 536
Ion Ratio Lower Upper
79 100
52 44.5 35.5 53.3
51 25.8 20.2 30.4









#6
Benzaldehyde
Concen: 0.060 ng/ul
RT: 6.787 min Scan# 64
Delta R.T. -0.177 min
Lab File: BP009234.D
Acq: 25 Feb 2022 14:40




Instrument :
BNA_P
ClientSampleId :
SSTD005601


Tgt Ion: 77 Resp: 6
Ion Ratio Lower Upper
77 100
105 87.5 77.0 115.4
106 87.5 77.0 115.4



#8
Phenol
Concen: 3.183 ng/ul
RT: 7.011 min Scan# 684
Delta R.T. -0.006 min
Lab File: BP009234.D
Acq: 25 Feb 2022 14:40

Tgt Ion: 94 Resp: 578
Ion Ratio Lower Upper
94 100
65 36.0 19.0 28.6#
66 50.3 34.6 51.8









#10
Bis(2-Chloroethyl)ether
Concen: 6.352 ng/ul
RT: 7.258 min Scan# 71
Delta R.T. 0.006 min
Lab File: BP009234.D
Acq: 25 Feb 2022 14:40




Instrument :
BNA_P
ClientSampleId :
SSTD005601


Tgt Ion: 93 Resp: 923
Ion Ratio Lower Upper
93 100
63 40.8 45.0 67.4#
95 21.4 26.3 39.5#



#12
2-Chlorophenol
Concen: 5.330 ng/ul
RT: 7.405 min Scan# 751
Delta R.T. 0.018 min
Lab File: BP009234.D
Acq: 25 Feb 2022 14:40

Tgt Ion: 128 Resp: 782
Ion Ratio Lower Upper
128 100
64 36.7 45.3 67.9#
130 15.7 29.0 43.6#





#13

2-Methylphenol

Concen: 0.043 ng/ul

RT: 8.310 min Scan# 905

Delta R.T. 0.041 min

Lab File: BP009234.D

Acq: 25 Feb 2022 14:40

Instrument :

BNA_P

ClientSampleId :





SSTD005601

Tgt Ion:108 Resp: 6

Ion Ratio Lower Upper

108 100

107 100.0 70.8 106.2



#14

2,2'-oxybis(1-Chloropropane)

Concen: 1.816 ng/ul

RT: 8.363 min Scan# 914

Delta R.T. 0.006 min

Lab File: BP009234.D

Acq: 25 Feb 2022 14:40




Tgt Ion: 45 Resp: 304


Ion Ratio Lower Upper

45 100

77 27.1 17.1 25.7#

79 41.2 13.2 19.8#









#16
Acetophenone
Concen: 4.784 ng/ul
RT: 8.646 min Scan# 910
Delta R.T. -0.000 min
Lab File: BP009234.D
Acq: 25 Feb 2022 14:40




Instrument :
BNA_P
ClientSampleId :
SSTD005601


Tgt Ion:105 Resp: 1049
Ion Ratio Lower Upper
105 100
77 64.9 53.4 80.2
51 17.7 13.8 20.8



#17
N-Nitroso-di-n-propylamine
Concen: 3.711 ng/ul
RT: 8.634 min Scan# 960
Delta R.T. -0.006 min
Lab File: BP009234.D
Acq: 25 Feb 2022 14:40

Tgt Ion: 70 Resp: 405
Ion Ratio Lower Upper
70 100
42 22.4 31.8 47.8#
101 0.0 15.6 23.4#
130 12.9 13.7 20.5#









#18
4-Methylphenol
Concen: 6.660 ng/ul
RT: 8.599 min Scan# 91
Delta R.T. -0.000 min
Lab File: BP009234.D
Acq: 25 Feb 2022 14:40




Instrument :
BNA_P
ClientSampleId :
SSTD005601


Tgt Ion:108 Resp: 1003
Ion Ratio Lower Upper
108 100
107 87.1 83.6 125.4



#19
Hexachloroethane
Concen: 6.564 ng/ul
RT: 8.916 min Scan# 1008
Delta R.T. 0.006 min
Lab File: BP009234.D
Acq: 25 Feb 2022 14:40

Tgt Ion:117 Resp: 392
Ion Ratio Lower Upper
117 100
201 150.2 106.5 159.7
199 91.9 59.0 88.6#









#22
Nitrobenzene
Concen: 8.382 ng/ul
RT: 9.022 min Scan# 1026
Delta R.T. -0.000 min
Lab File: BP009234.D
Acq: 25 Feb 2022 14:40




Instrument :
BNA_P
ClientSampleId :
SSTD005601


Tgt Ion: 77 Resp: 1056
Ion Ratio Lower Upper
77 100
123 30.4 42.3 63.5#
65 12.1 5.8 8.8#



#23
Isophorone
Concen: 7.468 ng/ul
RT: 9.552 min Scan# 1116
Delta R.T. -0.000 min
Lab File: BP009234.D
Acq: 25 Feb 2022 14:40

Tgt Ion: 82 Resp: 1973
Ion Ratio Lower Upper
82 100
95 10.0 7.8 11.8
138 8.0 12.9 19.3#









#25
2-Nitrophenol
Concen: 0.104 ng/ul
RT: 9.599 min Scan# 1124
Delta R.T. -0.141 min
Lab File: BP009234.D
Acq: 25 Feb 2022 14:40




Instrument :
BNA_P
ClientSampleId :
SSTD005601


Tgt Ion:139	Resp:	6	
Ion	Ratio	Lower	Upper
139	100		
65	72.7	49.4	74.2
109	63.6	44.7	67.1



#26
2,4-Dimethylphenol
Concen: 6.955 ng/ul
RT: 9.799 min Scan# 1158
Delta R.T. -0.000 min
Lab File: BP009234.D
Acq: 25 Feb 2022 14:40

Tgt Ion:107	Resp:	947	
Ion	Ratio	Lower	Upper
107	100		
121	48.3	34.1	51.1
122	75.5	46.3	69.5#









#27
 Bis(2-Chloroethoxy)methane
 Concen: 4.457 ng/ul
 RT: 10.040 min Scan# 1199
 Delta R.T. -0.000 min
 Lab File: BP009234.D
 Acq: 25 Feb 2022 14:40




Instrument :
 BNA_P
 ClientSampleId :
 SSTD005601


Tgt Ion: 93 Resp: 736
 Ion Ratio Lower Upper
 93 100
 95 32.7 26.2 39.4
 123 3.2 6.2 9.4#



#29
 2,4-Dichlorophenol
 Concen: 7.034 ng/ul
 RT: 10.263 min Scan# 1237
 Delta R.T. -0.000 min
 Lab File: BP009234.D
 Acq: 25 Feb 2022 14:40

Tgt Ion:162 Resp: 867
 Ion Ratio Lower Upper
 162 100
 164 35.5 37.0 55.6#
 98 40.1 31.0 46.6









#30
Naphthalene
Concen: 6.038 ng/ul
RT: 10.675 min Scan# 1307
Delta R.T. -0.000 min
Lab File: BP009234.D
Acq: 25 Feb 2022 14:40




Instrument :
BNA_P
ClientSampleId :
SSTD005601


Tgt Ion:128 Resp: 2466
Ion Ratio Lower Upper
128 100
129 15.2 8.6 13.0#
127 16.7 12.2 18.4



#32
4-Chloroaniline
Concen: 5.703 ng/ul
RT: 10.775 min Scan# 1324
Delta R.T. -0.006 min
Lab File: BP009234.D
Acq: 25 Feb 2022 14:40


Tgt Ion:127 Resp: 981
Ion Ratio Lower Upper
127 100
129 30.2 24.6 37.0








#33
Hexachlorobutadiene
Concen: 0.176 ng/ul
RT: 10.846 min Scan# 11
Delta R.T. -0.135 min
Lab File: BP009234.D
Acq: 25 Feb 2022 14:40

Instrument :
BNA_P
ClientSampleId :
SSTD005601







Tgt Ion:225 Resp: 15
Ion Ratio Lower Upper
225 100
223 61.5 53.4 80.0
227 69.2 53.8 80.8



#34
Caprolactam
Concen: 0.199 ng/ul
RT: 11.710 min Scan# 1483
Delta R.T. 0.147 min
Lab File: BP009234.D
Acq: 25 Feb 2022 14:40

Tgt Ion:113 Resp: 7
Ion Ratio Lower Upper
113 100
55 60.0 58.2 87.2
56 70.0 68.6 103.0









#35
4-Chloro-3-methylphenol
Concen: 10.790 ng/ul
RT: 11.910 min Scan# 1517
Delta R.T. -0.000 min
Lab File: BP009234.D
Acq: 25 Feb 2022 14:40




Instrument :
BNA_P
ClientSampleId :
SSTD005601


Tgt Ion:107 Resp: 1304
Ion Ratio Lower Upper
107 100
144 12.2 15.8 23.8#
142 40.6 38.9 58.3



#36
2-Methylnaphthalene
Concen: 5.201 ng/ul
RT: 12.293 min Scan# 1582
Delta R.T. 0.006 min
Lab File: BP009234.D
Acq: 25 Feb 2022 14:40

Tgt Ion:142 Resp: 1503
Ion Ratio Lower Upper
142 100
141 91.4 64.3 96.5









#37
1-Methylnaphthalene
Concen: 5.717 ng/ul
RT: 12.516 min Scan# 1620
Delta R.T. 0.012 min
Lab File: BP009234.D
Acq: 25 Feb 2022 14:40




Instrument :
BNA_P
ClientSampleId :
SSTD005601


Tgt Ion:142 Resp: 1682
Ion Ratio Lower Upper
142 100
141 88.1 69.8 104.8
116 7.6 5.6 8.4



#39
1,2,4,5-Tetrachlorobenzene
Concen: 4.953 ng/ul
RT: 12.657 min Scan# 1644
Delta R.T. -0.000 min
Lab File: BP009234.D
Acq: 25 Feb 2022 14:40

Tgt Ion:216 Resp: 1064
Ion Ratio Lower Upper
216 100
214 77.2 58.2 87.2
179 30.4 14.2 21.2#
108 23.0 15.3 22.9#









#40
Hexachlorocyclopentadiene
Concen: 6.796 ng/ul
RT: 12.651 min Scan# 1643
Delta R.T. 0.006 min
Lab File: BP009234.D
Acq: 25 Feb 2022 14:40




Instrument :
BNA_P
ClientSampleId :
SSTD005601


Tgt Ion:	237	Resp:	977
Ion	Ratio	Lower	Upper
237	100		
235	57.5	53.2	79.8
272	14.3	10.8	16.2



#41
2,4,6-Trichlorophenol
Concen: 5.862 ng/ul
RT: 12.898 min Scan# 1685
Delta R.T. -0.000 min
Lab File: BP009234.D
Acq: 25 Feb 2022 14:40

Tgt Ion:	196	Resp:	715
Ion	Ratio	Lower	Upper
196	100		
198	46.3	64.1	96.1#
200	29.2	19.9	29.9









#42
2,4,5-Trichlorophenol
Concen: 6.027 ng/ul
RT: 12.963 min Scan# 1695
Delta R.T. 0.006 min
Lab File: BP009234.D
Acq: 25 Feb 2022 14:40




Instrument :
BNA_P
ClientSampleId :
SSTD005601


Tgt Ion:196 Resp: 800
Ion Ratio Lower Upper
196 100
198 62.3 68.6 103.0#
200 10.4 32.8 49.2#



#43
1,1'-Biphenyl
Concen: 4.175 ng/ul
RT: 13.304 min Scan# 1754
Delta R.T. 0.006 min
Lab File: BP009234.D
Acq: 25 Feb 2022 14:40

Tgt Ion:154 Resp: 2154
Ion Ratio Lower Upper
154 100
153 51.0 28.6 43.0#
76 21.2 8.5 12.7#









#44
2-Chloronaphthalene
Concen: 0.048 ng/ul
RT: 13.251 min Scan# 1785
Delta R.T. -0.088 min
Lab File: BP009234.D
Acq: 25 Feb 2022 14:40




Instrument :
BNA_P
ClientSampleId :
SSTD005601


Tgt Ion:	162	Resp:	19
Ion	Ratio	Lower	Upper
162	100		
164	38.9	27.5	41.3
127	50.0	37.1	55.7



#45
2-Nitroaniline
Concen: 0.063 ng/ul
RT: 13.493 min Scan# 1786
Delta R.T. -0.041 min
Lab File: BP009234.D
Acq: 25 Feb 2022 14:40

Tgt Ion:	65	Resp:	5
Ion	Ratio	Lower	Upper
65	100		
92	87.5	77.9	116.9
138	87.5	81.5	122.3









#47
Dimethylphthalate
Concen: 5.057 ng/ul
RT: 13.928 min Scan# 1860
Delta R.T. -0.000 min
Lab File: BP009234.D
Acq: 25 Feb 2022 14:40




Instrument :
BNA_P
ClientSampleId :
SSTD005601


Tgt Ion:163 Resp: 2481
Ion Ratio Lower Upper
163 100
194 4.0 4.1 6.1#
164 21.7 11.3 16.9#



#48
2,6-Dinitrotoluene
Concen: 5.791 ng/ul
RT: 14.040 min Scan# 1879
Delta R.T. 0.006 min
Lab File: BP009234.D
Acq: 25 Feb 2022 14:40

Tgt Ion:165 Resp: 505
Ion Ratio Lower Upper
165 100
89 44.4 51.7 77.5#
121 12.8 18.3 27.5#





#50

Acenaphthylene

Concen: 4.419 ng/ul

RT: 14.187 min Scan# 1904

Delta R.T. -0.000 min

Lab File: BP009234.D

Acq: 25 Feb 2022 14:40

Instrument :

BNA_P

ClientSampleId :

SSTD005601





Tgt Ion:152 Resp: 2780

Ion Ratio Lower Upper

152 100

151 24.0 18.3 27.5

153 15.3 11.0 16.6



#51

3-Nitroaniline

Concen: 5.920 ng/ul

RT: 14.357 min Scan# 1933

Delta R.T. -0.006 min

Lab File: BP009234.D

Acq: 25 Feb 2022 14:40




Tgt Ion:138 Resp: 532


Ion Ratio Lower Upper

138 100

108 37.8 15.8 23.8

92 147.2 117.1 175.7









#52
Acenaphthene
Concen: 4.810 ng/ul
RT: 14.534 min Scan# 1963
Delta R.T. 0.006 min
Lab File: BP009234.D
Acq: 25 Feb 2022 14:40




Instrument :
BNA_P
ClientSampleId :
SSTD005601


Tgt Ion:153 Resp: 1976
Ion Ratio Lower Upper
153 100
152 39.7 34.3 51.5
154 76.7 70.0 105.0



#53
2,4-Dinitrophenol
Concen: 11.547 ng/ul
RT: 14.569 min Scan# 1969
Delta R.T. -0.000 min
Lab File: BP009234.D
Acq: 25 Feb 2022 14:40

Tgt Ion:184 Resp: 425
Ion Ratio Lower Upper
184 100
63 44.8 34.6 52.0
154 50.2 57.7 86.5









#55
4-Nitrophenol
Concen: 0.111 ng/ul
RT: 14.604 min Scan# 1985
Delta R.T. -0.059 min
Lab File: BP009234.D
Acq: 25 Feb 2022 14:40




Instrument :
BNA_P
ClientSampleId :
SSTD005601


Tgt Ion:109 Resp: 6
Ion Ratio Lower Upper
109 100
139 66.7 61.4 92.2
65 77.8 65.3 97.9



#56
Dibenzofuran
Concen: 4.635 ng/ul
RT: 14.869 min Scan# 2020
Delta R.T. -0.000 min
Lab File: BP009234.D
Acq: 25 Feb 2022 14:40

Tgt Ion:168 Resp: 2784
Ion Ratio Lower Upper
168 100
139 32.0 30.1 45.1









#57
2,4-Dinitrotoluene
Concen: 5.851 ng/ul
RT: 14.828 min Scan# 2012
Delta R.T. 0.006 min
Lab File: BP009234.D
Acq: 25 Feb 2022 14:40




Instrument :
BNA_P
ClientSampleId :
SSTD005601


Tgt Ion:165 Resp: 706
Ion Ratio Lower Upper
165 100
63 36.6 29.5 44.3
182 4.5 2.8 4.2#



#58
2,3,4,6-Tetrachlorophenol
Concen: 7.293 ng/ul
RT: 15.098 min Scan# 2059
Delta R.T. 0.006 min
Lab File: BP009234.D
Acq: 25 Feb 2022 14:40

Tgt Ion:232 Resp: 752
Ion Ratio Lower Upper
232 100
131 28.4 25.4 38.2
130 0.0 4.2 6.2#
166 26.7 28.1 42.1#









#59
Diethylphthalate
Concen: 5.046 ng/ul
RT: 15.304 min Scan# 2094
Delta R.T. -0.000 min
Lab File: BP009234.D
Acq: 25 Feb 2022 14:40




Instrument :
BNA_P
ClientSampleId :
SSTD005601


Tgt Ion:149 Resp: 2418
Ion Ratio Lower Upper
149 100
177 22.0 22.0 33.0#
150 18.3 10.2 15.2#



#61
Fluorene
Concen: 4.881 ng/ul
RT: 15.522 min Scan# 2131
Delta R.T. -0.000 min
Lab File: BP009234.D
Acq: 25 Feb 2022 14:40

Tgt Ion:166 Resp: 2336
Ion Ratio Lower Upper
166 100
165 91.4 70.3 105.5
167 12.4 9.5 14.3









#62
4-Chlorophenyl-phenylether
Concen: 5.834 ng/ul
RT: 15.528 min Scan# 2132
Delta R.T. 0.006 min
Lab File: BP009234.D
Acq: 25 Feb 2022 14:40




Instrument :
BNA_P
ClientSampleId :
SSTD005601


Tgt Ion:204 Resp: 1481
Ion Ratio Lower Upper
204 100
206 35.6 27.6 41.4
141 43.3 46.6 69.8#



#63
4-Nitroaniline
Concen: 7.078 ng/ul
RT: 15.522 min Scan# 2131
Delta R.T. -0.000 min
Lab File: BP009234.D
Acq: 25 Feb 2022 14:40

Tgt Ion:138 Resp: 583
Ion Ratio Lower Upper
138 100
92 54.5 41.6 62.4
108 114.5 97.6 146.4









#66
 4,6-Dinitro-2-methylphenol
 Concen: 5.864 ng/ul
 RT: 15.592 min Scan# 2143
 Delta R.T. 0.006 min
 Lab File: BP009234.D
 Acq: 25 Feb 2022 14:40




Instrument :
 BNA_P
 ClientSampleId :
 SSTD005601


Tgt Ion:198 Resp: 442
 Ion Ratio Lower Upper
 198 100
 182 7.4 6.0 9.0
 77 31.8 17.7 26.5#



#67
 N-Nitrosodiphenylamine
 Concen: 4.326 ng/ul
 RT: 15.728 min Scan# 2166
 Delta R.T. -0.006 min
 Lab File: BP009234.D
 Acq: 25 Feb 2022 14:40

Tgt Ion:169 Resp: 2141
 Ion Ratio Lower Upper
 169 100
 168 67.8 48.3 72.5
 167 23.3 29.6 44.4#









#68
4-Bromophenyl-phenylether
Concen: 0.032 ng/ul
RT: 16.475 min Scan# 21
Delta R.T. 0.059 min
Lab File: BP009234.D
Acq: 25 Feb 2022 14:40




Instrument :
BNA_P
ClientSampleId :
SSTD005601


Tgt Ion:248	Resp:	6	
Ion	Ratio	Lower	Upper
248	100		
250	90.0	79.4	119.0
141	90.0	67.2	100.8



#69
Hexachlorobenzene
Concen: 2.318 ng/ul
RT: 16.528 min Scan# 2302
Delta R.T. -0.006 min
Lab File: BP009234.D
Acq: 25 Feb 2022 14:40

Tgt	Ion:284	Resp:	501
Ion	Ratio	Lower	Upper
284	100		
142	24.7	19.2	28.8
249	48.5	27.4	41.2#









#70
 Atrazine
 Concen: 5.538 ng/ul
 RT: 16.692 min Scan# 21
 Delta R.T. 0.006 min
 Lab File: BP009234.D
 Acq: 25 Feb 2022 14:40




Instrument :
 BNA_P
 ClientSampleId :
 SSTD005601


Tgt Ion:200 Resp: 1040
 Ion Ratio Lower Upper
 200 100
 173 22.8 19.2 28.8
 215 67.7 32.2 48.4#



#71
 Pentachlorophenol
 Concen: 7.424 ng/ul
 RT: 16.881 min Scan# 2362
 Delta R.T. 0.006 min
 Lab File: BP009234.D
 Acq: 25 Feb 2022 14:40

Tgt Ion:266 Resp: 820
 Ion Ratio Lower Upper
 266 100
 264 58.1 57.3 85.9
 268 62.7 65.7 98.5#









#72
Phenanthrene
Concen: 4.867 ng/ul
RT: 17.369 min Scan# 2428
Delta R.T. 0.100 min
Lab File: BP009234.D
Acq: 25 Feb 2022 14:40




Instrument :
BNA_P
ClientSampleId :
SSTD005601


Tgt Ion:178 Resp: 4338
Ion Ratio Lower Upper
178 100
179 13.7 11.7 17.5
176 15.4 14.5 21.7



#74
Anthracene
Concen: 4.854 ng/ul
RT: 17.369 min Scan# 2445
Delta R.T. 0.006 min
Lab File: BP009234.D
Acq: 25 Feb 2022 14:40

Tgt Ion:178 Resp: 4338
Ion Ratio Lower Upper
178 100
179 13.7 12.3 18.5
176 15.4 13.4 20.2









#75
1,2,3,4-Tetrachlorobenzene
Concen: 4.335 ng/uL
RT: 13.263 min Scan# 11
Delta R.T. -0.000 min
Lab File: BP009234.D
Acq: 25 Feb 2022 14:40




Instrument :
BNA_P
ClientSampleId :
SSTD005601


Tgt Ion:216 Resp: 1190
Ion Ratio Lower Upper
216 100
214 82.1 50.4 75.6#
218 70.4 35.2 52.8#



#76
Pentachlorobenzene
Concen: 4.782 ng/uL
RT: 14.787 min Scan# 2006
Delta R.T. -0.000 min
Lab File: BP009234.D
Acq: 25 Feb 2022 14:40

Tgt Ion:250 Resp: 1205
Ion Ratio Lower Upper
250 100
248 56.0 43.5 65.3
252 54.1 46.5 69.7









#77
Carbazole
Concen: 4.816 ng/ul
RT: 17.633 min Scan# 2489
Delta R.T. 0.006 min
Lab File: BP009234.D
Acq: 25 Feb 2022 14:40




Instrument :
BNA_P
ClientSampleId :
SSTD005601


Tgt Ion:167 Resp: 3668
Ion Ratio Lower Upper
167 100
166 26.2 15.8 23.8#
139 11.6 8.3 12.5



#78
Di-n-butylphthalate
Concen: 5.148 ng/ul
RT: 18.210 min Scan# 2588
Delta R.T. 0.006 min
Lab File: BP009234.D
Acq: 25 Feb 2022 14:40

Tgt Ion:149 Resp: 4578
Ion Ratio Lower Upper
149 100
150 10.9 9.6 14.4
104 5.7 6.2 9.4#









#80
Fluoranthene
Concen: 3.666 ng/ul
RT: 19.245 min Scan# 21
Delta R.T. -0.000 min
Lab File: BP009234.D
Acq: 25 Feb 2022 14:40




Instrument :
BNA_P
ClientSampleId :
SSTD005601


Tgt Ion:202 Resp: 5384
Ion Ratio Lower Upper
202 100
101 8.1 9.6 14.4#
100 18.3 6.3 9.5#



#82
Pyrene
Concen: 3.905 ng/ul
RT: 19.604 min Scan# 2825
Delta R.T. 0.006 min
Lab File: BP009234.D
Acq: 25 Feb 2022 14:40

Tgt Ion:202 Resp: 5779
Ion Ratio Lower Upper
202 100
101 10.1 10.6 15.8#
100 9.0 9.8 14.6#





#83

Butylbenzylphthalate

Concen: 4.591 ng/ul

RT: 20.498 min Scan# 2976

Delta R.T. 0.006 min

Lab File: BP009234.D

Acq: 25 Feb 2022 14:40

Instrument :

BNA_P

ClientSampleId :

SSTD005601





Tgt Ion:149 Resp: 2303

Ion Ratio Lower Upper

149 100

91 91.3 71.1 106.7

206 45.4 28.8 43.2#



#84

3,3'-Dichlorobenzidine

Concen: 5.870 ng/ul

RT: 21.257 min Scan# 3106

Delta R.T. 0.006 min

Lab File: BP009234.D

Acq: 25 Feb 2022 14:40




Tgt Ion:252 Resp: 2563


Ion Ratio Lower Upper

252 100

254 63.0 49.8 74.6

126 6.6 4.3 6.5#









#85
Benzo(a)anthracene
Concen: 4.719 ng/ul
RT: 21.322 min Scan# 3117
Delta R.T. -0.000 min
Lab File: BP009234.D
Acq: 25 Feb 2022 14:40




Instrument :
BNA_P
ClientSampleId :
SSTD005601


Tgt Ion:	228	Resp:	6281
Ion Ratio	Lower	Upper	
228	100		
229	24.0	18.2	27.2
226	23.3	22.8	34.2



#86
Bis(2-ethylhexyl)phthalate
Concen: 4.596 ng/ul
RT: 21.274 min Scan# 3109
Delta R.T. 0.006 min
Lab File: BP009234.D
Acq: 25 Feb 2022 14:40

Tgt Ion:	149	Resp:	3499
Ion Ratio	Lower	Upper	
149	100		
167	40.6	32.8	49.2
279	15.9	5.6	8.4





#87

Chrysene

Concen: 4.536 ng/ul

RT: 21.374 min Scan# 3125

Delta R.T. 0.006 min

Lab File: BP009234.D

Acq: 25 Feb 2022 14:40

Instrument :

BNA_P

ClientSampleId :

SSTD005601





Tgt Ion:228 Resp: 5829

Ion Ratio Lower Upper

228 100

226 31.8 22.7 34.1

229 25.1 15.2 22.8#



#89

Di-n-octyl phthalate

Concen: 3.882 ng/ul




RT: 22.210 min Scan# 3268


Delta R.T. 0.006 min

Lab File: BP009234.D

Acq: 25 Feb 2022 14:40

Tgt Ion:149 Resp: 5594









#90
Benzo(b)fluoranthene
Concen: 4.570 ng/ul
RT: 23.016 min Scan# 3404
Delta R.T. 0.006 min
Lab File: BP009234.D
Acq: 25 Feb 2022 14:40




Instrument :
BNA_P
ClientSampleId :
SSTD005601


Tgt Ion:252 Resp: 6853
Ion Ratio Lower Upper
252 100
253 19.8 17.4 26.0
125 6.2 7.3 10.9#



#91
Benzo(k)fluoranthene
Concen: 4.724 ng/ul
RT: 23.016 min Scan# 3405
Delta R.T. -0.047 min
Lab File: BP009234.D
Acq: 25 Feb 2022 14:40

Tgt Ion:252 Resp: 6837
Ion Ratio Lower Upper
252 100
253 19.8 17.9 26.9
125 6.2 5.1 7.7









#93
Benzo(a)pyrene
Concen: 4.582 ng/ul
RT: 23.645 min Scan# 3512
Delta R.T. 0.006 min
Lab File: BP009234.D
Acq: 25 Feb 2022 14:40




Instrument :
BNA_P
ClientSampleId :
SSTD005601


Tgt Ion:252 Resp: 6578
Ion Ratio Lower Upper
252 100
253 24.2 17.5 26.3
125 7.8 6.4 9.6



#94
Indeno(1,2,3-cd)pyrene
Concen: 3.224 ng/ul
RT: 26.257 min Scan# 3956
Delta R.T. -0.000 min
Lab File: BP009234.D
Acq: 25 Feb 2022 14:40

Tgt Ion:276 Resp: 5104
Ion Ratio Lower Upper
276 100
138 8.6 12.3 18.5#
277 28.7 19.3 28.9









#95
 Dibenzo(a,h)anthracene
 Concen: 4.974 ng/ul
 RT: 26.286 min Scan# 3961
 Delta R.T. 0.006 min
 Lab File: BP009234.D
 Acq: 25 Feb 2022 14:40

Instrument :
 BNA_P
 ClientSampleId :
 SSTD005601

Tgt Ion: 278 Resp: 6833
 Ion Ratio Lower Upper
 278 100
 139 0.6 6.6 10.0#
 279 16.0 18.4 27.6#



#96
 Benzo(g,h,i)perylene
 Concen: 3.442 ng/ul
 RT: 27.021 min Scan# 4086
 Delta R.T. -0.000 min
 Lab File: BP009234.D
 Acq: 25 Feb 2022 14:40

Tgt Ion: 276 Resp: 4504
 Ion Ratio Lower Upper
 276 100
 138 9.9 11.6 17.4#
 277 16.5 15.9 23.9

