

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP022522\
 Data File : BP009238.D
 Acq On : 25 Feb 2022 16:55
 Operator : CG/JU
 Sample : SSTD08054
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SSTD080605

Quant Time: Feb 25 23:46:30 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP022522.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Fri Feb 25 23:39:41 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.834	152	3264	20.000	ng/u1	0.00
20) Naphthalene-d8	10.628	136	11927	20.000	ng/u1	# 0.00
38) Acenaphthene-d10	14.469	164	9224	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.234	188	21264	20.000	ng/u1	# 0.00
79) Chrysene-d12	21.339	240	22500	20.000	ng/u1	0.00
88) Perylene-d12	23.757	264	25180	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.323	96	2926	40.407	ng/uL	0.00
4) Pyridine-d5	3.723	84	17328	85.506	ng/u1	0.00
7) Phenol-d5	6.999	99	26610	102.194	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.164	67	11598	78.421	ng/u1	0.00
11) 2-Chlorophenol-d4	7.364	132	14650	70.726	ng/u1	0.00
15) 4-Methylphenol-d8	8.540	113	20779	99.476	ng/u1	0.00
21) Nitrobenzene-d5	8.987	128	7854	93.061	ng/u1	0.01
24) 2-Nitrophenol-d4	9.705	143	9447	122.080	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.246	165	17975	94.627	ng/u1	0.00
31) 4-Chloroaniline-d4	10.763	131	20244	77.205	ng/u1	0.00
46) Dimethylphthalate-d6	13.887	166	56530	82.947	ng/u1	0.00
49) Acenaphthylene-d8	14.163	160	65706	77.104	ng/u1	0.00
54) 4-Nitrophenol-d4	14.669	143	9188	85.049	ng/u1	0.02
60) Fluorene-d10	15.469	176	48388	81.776	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.587	200	12498	133.520	ng/u1	0.01
73) Anthracene-d10	17.334	188	79409	78.806	ng/u1	0.00
81) Pyrene-d10	19.575	212	98242	72.867	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.604	264	103028	79.499	ng/u1	0.01
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.358	88	3055	39.022	ng/uL#	81
5) Pyridine	3.740	79	18346	88.231	ng/u1	98
6) Benzaldehyde	6.969	77	15399	103.754	ng/u1	98
8) Phenol	7.022	94	26602	99.428	ng/u1	90
10) Bis(2-Chloroethyl)ether	7.258	93	19907	92.967	ng/u1	95
12) 2-Chlorophenol	7.399	128	15693	72.590	ng/u1	88
13) 2-Methylphenol	8.269	108	20015	96.553	ng/u1	100
14) 2,2'-oxybis(1-Chloropr...	8.363	45	11131	45.129	ng/u1#	86
16) Acetophenone	8.658	105	34016	105.270	ng/u1	97
17) N-Nitroso-di-n-propyla...	8.658	70	14337	89.143	ng/u1#	88
18) 4-Methylphenol	8.611	108	21897	98.664	ng/u1	94
19) Hexachloroethane	8.911	117	8586	97.560	ng/u1#	76
22) Nitrobenzene	9.028	77	22598	116.044	ng/u1#	80
23) Isophorone	9.569	82	44988	110.159	ng/u1	98
25) 2-Nitrophenol	9.740	139	8931	100.285	ng/u1	97
26) 2,4-Dimethylphenol	9.805	107	28079	133.415	ng/u1	95
27) Bis(2-Chloroethoxy)met...	10.046	93	26847	105.173	ng/u1#	93
29) 2,4-Dichlorophenol	10.275	162	17568	92.203	ng/u1#	82
30) Naphthalene	10.675	128	49751	78.804	ng/u1	99
32) 4-Chloroaniline	10.787	127	21770	81.876	ng/u1	92
33) Hexachlorobutadiene	10.975	225	14732	111.512	ng/u1	92
34) Caprolactam	11.652	113	674	12.415	ng/u1	87
35) 4-Chloro-3-methylphenol	11.916	107	26062	139.517	ng/u1	97
36) 2-Methylnaphthalene	12.293	142	35565	79.612	ng/u1#	81


Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP022522\
 Data File : BP009238.D
 Acq On : 25 Feb 2022 16:55
 Operator : CG/JU
 Sample : SST08054
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SST080605

Quant Time: Feb 25 23:46:30 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP022522.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Fri Feb 25 23:39:41 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.510	142	35208	77.416	ng/ul#	91
39) 1,2,4,5-Tetrachloroben...	12.663	216	25569	85.880	ng/ul#	87
40) Hexachlorocyclopentadiene	12.646	237	20485	102.808	ng/ul#	92
41) 2,4,6-Trichlorophenol	12.899	196	16681	98.666	ng/ul#	81
42) 2,4,5-Trichlorophenol	12.969	196	17236	93.684	ng/ul	85
43) 1,1'-Biphenyl	13.310	154	51070	71.423	ng/ul#	92
44) 2-Chloronaphthalene	13.346	162	40767	73.947	ng/ul	96
45) 2-Nitroaniline	13.540	65	12624	113.928	ng/ul	91
47) Dimethylphthalate	13.934	163	55087	81.018	ng/ul#	91
48) 2,6-Dinitrotoluene	14.046	165	11364	94.022	ng/ul	87
50) Acenaphthylene	14.193	152	62862	72.094	ng/ul	97
51) 3-Nitroaniline	14.369	138	10426	83.709	ng/ul	90
52) Acenaphthene	14.534	153	41480	72.852	ng/ul	93
53) 2,4-Dinitrophenol	14.575	184	8930	175.042	ng/ul#	94
55) 4-Nitrophenol	14.687	109	12799	170.401	ng/ul	98
56) Dibenzofuran	14.875	168	62440	74.997	ng/ul	97
57) 2,4-Dinitrotoluene	14.834	165	17272	103.283	ng/ul	94
58) 2,3,4,6-Tetrachlorophenol	15.098	232	16293	114.003	ng/ul#	93
59) Diethylphthalate	15.310	149	53363	80.348	ng/ul	97
61) Fluorene	15.528	166	49833	75.127	ng/ul	93
62) 4-Chlorophenyl-phenyle...	15.522	204	29524	83.913	ng/ul	95
63) 4-Nitroaniline	15.545	138	11279	98.801	ng/ul#	80
66) 4,6-Dinitro-2-methylph...	15.604	198	12006	121.221	ng/ul#	94
67) N-Nitrosodiphenylamine	15.734	169	45658	70.213	ng/ul	90
68) 4-Bromophenyl-phenylether	16.422	248	20186	82.771	ng/ul	99
69) Hexachlorobenzene	16.534	284	22658	79.799	ng/ul	94
70) Atrazine	16.704	200	20975	85.004	ng/ul#	89
71) Pentachlorophenol	16.881	266	16917	116.565	ng/ul#	84
72) Phenanthrene	17.275	178	84047	71.764	ng/ul	99
74) Anthracene	17.369	178	87523	74.528	ng/ul	98
75) 1,2,3,4-Tetrachloroben...	13.269	216	27274	75.612	ng/ul#	85
76) Pentachlorobenzene	14.793	250	26156	78.989	ng/ul#	86
77) Carbazole	17.634	167	74201	74.147	ng/ul#	90
78) Di-n-butylphthalate	18.210	149	94792	81.126	ng/ul#	95
80) Fluoranthene	19.251	202	109593	68.367	ng/ul	98
82) Pyrene	19.604	202	109496	67.788	ng/ul	97
83) Butylbenzylphthalate	20.492	149	44261	80.846	ng/ul#	90
84) 3,3'-Dichlorobenzidine	21.257	252	43955	92.234	ng/ul#	99
85) Benzo(a)anthracene	21.322	228	114658	78.935	ng/ul	95
86) Bis(2-ethylhexyl)phtha...	21.269	149	65864	79.262	ng/ul	94
87) Chrysene	21.375	228	109616	78.159	ng/ul	98
89) Di-n-octyl phthalate	22.204	149	121296	75.647	ng/ul	100
90) Benzo(b)fluoranthene	23.022	252	127669	76.515	ng/ul#	95
91) Benzo(k)fluoranthene	23.069	252	113450	70.450	ng/ul	97
93) Benzo(a)pyrene	23.657	252	118744	74.347	ng/ul#	95
94) Indeno(1,2,3-cd)pyrene	26.274	276	139871	79.414	ng/ul	95
95) Dibenzo(a,h)anthracene	26.298	278	120606	78.916	ng/ul	98
96) Benzo(g,h,i)perylene	27.045	276	116004	79.689	ng/ul#	91





(#) = qualifier out of range (m) = manual integration (+) = signals summed



#2
1,4-Dioxane
Concen: 39.022 ng/uL
RT: 3.358 min Scan# 61
Delta R.T. 0.000 min
Lab File: BP009238.D
Acq: 25 Feb 2022 16:55




Instrument :
BNA_P
ClientSampleId :
SSTD080605


Tgt Ion: 88 Resp: 3055
Ion Ratio Lower Upper
88 100
43 20.1 6.8 10.2#
58 44.6 28.4 42.6#



#5
Pyridine
Concen: 88.231 ng/uL
RT: 3.740 min Scan# 128
Delta R.T. 0.000 min
Lab File: BP009238.D
Acq: 25 Feb 2022 16:55

Tgt Ion: 79 Resp: 18346
Ion Ratio Lower Upper
79 100
52 45.9 35.5 53.3
51 24.9 20.2 30.4









#6
Benzaldehyde
Concen: 103.754 ng/ul
RT: 6.969 min Scan# 61
Delta R.T. 0.006 min
Lab File: BP009238.D
Acq: 25 Feb 2022 16:55




Instrument :
BNA_P
ClientSampleId :
SSTD080605


Tgt Ion: 77 Resp: 15399
Ion Ratio Lower Upper
77 100
105 97.1 77.0 115.4
106 94.0 77.0 115.4



#8
Phenol
Concen: 99.428 ng/ul
RT: 7.022 min Scan# 686
Delta R.T. 0.006 min
Lab File: BP009238.D
Acq: 25 Feb 2022 16:55

Tgt Ion: 94 Resp: 26602
Ion Ratio Lower Upper
94 100
65 21.7 19.0 28.6
66 34.9 34.6 51.8









#10
Bis(2-Chloroethyl)ether
Concen: 92.967 ng/ul
RT: 7.258 min Scan# 71
Delta R.T. 0.006 min
Lab File: BP009238.D
Acq: 25 Feb 2022 16:55




Instrument :
BNA_P
ClientSampleId :
SSTD080605


Tgt Ion: 93 Resp: 19907
Ion Ratio Lower Upper
93 100
63 51.6 45.0 67.4
95 30.7 26.3 39.5



#12
2-Chlorophenol
Concen: 72.590 ng/ul
RT: 7.399 min Scan# 750
Delta R.T. 0.012 min
Lab File: BP009238.D
Acq: 25 Feb 2022 16:55

Tgt Ion: 128 Resp: 15693
Ion Ratio Lower Upper
128 100
64 47.1 45.3 67.9
130 29.3 29.0 43.6





#13

2-Methylphenol

Concen: 96.553 ng/ul

RT: 8.269 min Scan# 898

Delta R.T. 0.000 min

Lab File: BP009238.D

Acq: 25 Feb 2022 16:55

Instrument :

BNA_P

ClientSampleId :





SSTD080605

Tgt Ion:108 Resp: 20015

Ion Ratio Lower Upper

108 100

107 88.6 70.8 106.2



#14

2,2'-oxybis(1-Chloropropane)

Concen: 45.129 ng/ul

RT: 8.363 min Scan# 914

Delta R.T. 0.006 min

Lab File: BP009238.D

Acq: 25 Feb 2022 16:55




Tgt Ion: 45 Resp: 11131


Ion Ratio Lower Upper

45 100

77 24.7 17.1 25.7

79 26.6 13.2 19.8#









#16
Acetophenone
Concen: 105.270 ng/ul
RT: 8.658 min Scan# 90
Delta R.T. 0.012 min
Lab File: BP009238.D
Acq: 25 Feb 2022 16:55




Instrument :
BNA_P
ClientSampleId :
SSTD080605


Tgt Ion:105 Resp: 34016
Ion Ratio Lower Upper
105 100
77 69.1 53.4 80.2
51 15.5 13.8 20.8



#17
N-Nitroso-di-n-propylamine
Concen: 89.143 ng/ul
RT: 8.658 min Scan# 964
Delta R.T. 0.018 min
Lab File: BP009238.D
Acq: 25 Feb 2022 16:55

Tgt Ion: 70 Resp: 14337
Ion Ratio Lower Upper
70 100
42 32.0 31.8 47.8
101 12.2 15.6 23.4#
130 19.3 13.7 20.5









#18
4-Methylphenol
Concen: 98.664 ng/ul
RT: 8.611 min Scan# 91
Delta R.T. 0.012 min
Lab File: BP009238.D
Acq: 25 Feb 2022 16:55




Instrument :
BNA_P
ClientSampleId :
SSTD080605


Tgt Ion:108 Resp: 21897
Ion Ratio Lower Upper
108 100
107 110.5 83.6 125.4



#19
Hexachloroethane
Concen: 97.560 ng/ul
RT: 8.911 min Scan# 1007
Delta R.T. 0.000 min
Lab File: BP009238.D
Acq: 25 Feb 2022 16:55

Tgt Ion:117 Resp: 8586
Ion Ratio Lower Upper
117 100
201 94.2 106.5 159.7#
199 67.9 59.0 88.6









#22
Nitrobenzene
Concen: 116.044 ng/ul
RT: 9.028 min Scan# 1109
Delta R.T. 0.006 min
Lab File: BP009238.D
Acq: 25 Feb 2022 16:55




Instrument :
BNA_P
ClientSampleId :
SSTD080605


Tgt Ion: 77 Resp: 22598
Ion Ratio Lower Upper
77 100
123 38.9 42.3 63.5#
65 13.8 5.8 8.8#



#23
Isophorone
Concen: 110.159 ng/ul
RT: 9.569 min Scan# 1119
Delta R.T. 0.018 min
Lab File: BP009238.D
Acq: 25 Feb 2022 16:55

Tgt Ion: 82 Resp: 44988
Ion Ratio Lower Upper
82 100
95 9.7 7.8 11.8
138 14.5 12.9 19.3









#25
2-Nitrophenol
Concen: 100.285 ng/ul
RT: 9.740 min Scan# 1148
Delta R.T. 0.000 min
Lab File: BP009238.D
Acq: 25 Feb 2022 16:55




Instrument :
BNA_P
ClientSampleId :
SSTD080605


Tgt Ion:139	Resp:	8931
Ion Ratio	Lower	Upper
139	100	
65	57.1	49.4 74.2
109	55.7	44.7 67.1



#26
2,4-Dimethylphenol
Concen: 133.415 ng/ul
RT: 9.805 min Scan# 1159
Delta R.T. 0.006 min
Lab File: BP009238.D
Acq: 25 Feb 2022 16:55

Tgt Ion:107	Resp:	28079
Ion Ratio	Lower	Upper
107	100	
121	39.7	34.1 51.1
122	61.3	46.3 69.5









#27
Bis(2-Chloroethoxy)methane
Concen: 105.173 ng/ul
RT: 10.046 min Scan# 1199
Delta R.T. 0.006 min
Lab File: BP009238.D
Acq: 25 Feb 2022 16:55




Instrument :
BNA_P
ClientSampleId :
SSTD080605


Tgt Ion: 93 Resp: 26847
Ion Ratio Lower Upper
93 100
95 37.0 26.2 39.4
123 9.5 6.2 9.4#



#29
2,4-Dichlorophenol
Concen: 92.203 ng/ul
RT: 10.275 min Scan# 1239
Delta R.T. 0.012 min
Lab File: BP009238.D
Acq: 25 Feb 2022 16:55

Tgt Ion:162 Resp: 17568
Ion Ratio Lower Upper
162 100
164 63.5 37.0 55.6#
98 43.9 31.0 46.6









#30
Naphthalene
Concen: 78.804 ng/ul
RT: 10.675 min Scan# 1307
Delta R.T. 0.000 min
Lab File: BP009238.D
Acq: 25 Feb 2022 16:55




Instrument :
BNA_P
ClientSampleId :
SSTD080605


Tgt Ion:128 Resp: 49751
Ion Ratio Lower Upper
128 100
129 9.8 8.6 13.0
127 15.1 12.2 18.4



#32
4-Chloroaniline
Concen: 81.876 ng/ul
RT: 10.787 min Scan# 1326
Delta R.T. 0.006 min
Lab File: BP009238.D
Acq: 25 Feb 2022 16:55

Tgt Ion:127 Resp: 21770
Ion Ratio Lower Upper
127 100
129 35.0 24.6 37.0





#33

Hexachlorobutadiene

Concen: 111.512 ng/ul

RT: 10.975 min Scan# 1473

Delta R.T. -0.006 min

Lab File: BP009238.D

Acq: 25 Feb 2022 16:55

Instrument :

BNA_P

ClientSampleId :

SSTD080605





Tgt Ion:225 Resp: 14732

Ion Ratio Lower Upper

225 100

223 59.6 53.4 80.0

227 62.2 53.8 80.8



#34

Caprolactam

Concen: 12.415 ng/ul

RT: 11.652 min Scan# 1473

Delta R.T. 0.088 min

Lab File: BP009238.D

Acq: 25 Feb 2022 16:55




Tgt Ion:113 Resp: 674


Ion Ratio Lower Upper

113 100

55 85.4 58.2 87.2

56 95.2 68.6 103.0









#35
4-Chloro-3-methylphenol
Concen: 139.517 ng/ul
RT: 11.916 min Scan# 1518
Delta R.T. 0.006 min
Lab File: BP009238.D
Acq: 25 Feb 2022 16:55




Instrument :
BNA_P
ClientSampleId :
SSTD080605


Tgt Ion:107 Resp: 26062
Ion Ratio Lower Upper
107 100
144 19.9 15.8 23.8
142 51.8 38.9 58.3



#36
2-Methylnaphthalene
Concen: 79.612 ng/ul
RT: 12.293 min Scan# 1582
Delta R.T. 0.006 min
Lab File: BP009238.D
Acq: 25 Feb 2022 16:55

Tgt Ion:142 Resp: 35565
Ion Ratio Lower Upper
142 100
141 97.1 64.3 96.5#









#37
1-Methylnaphthalene
Concen: 77.416 ng/ul
RT: 12.510 min Scan# 1619
Delta R.T. 0.006 min
Lab File: BP009238.D
Acq: 25 Feb 2022 16:55




Instrument :
BNA_P
ClientSampleId :
SSTD080605


Tgt Ion:142 Resp: 35208
Ion Ratio Lower Upper
142 100
141 96.2 69.8 104.8
116 5.0 5.6 8.4#



#39
1,2,4,5-Tetrachlorobenzene
Concen: 85.880 ng/ul
RT: 12.663 min Scan# 1645
Delta R.T. 0.006 min
Lab File: BP009238.D
Acq: 25 Feb 2022 16:55

Tgt Ion:216 Resp: 25569
Ion Ratio Lower Upper
216 100
214 84.5 58.2 87.2
179 25.4 14.2 21.2#
108 18.2 15.3 22.9









#40
Hexachlorocyclopentadiene
Concen: 102.808 ng/ul
RT: 12.646 min Scan# 1642
Delta R.T. 0.000 min
Lab File: BP009238.D
Acq: 25 Feb 2022 16:55




Instrument :
BNA_P
ClientSampleId :
SSTD080605


Tgt Ion:237 Resp: 20485
Ion Ratio Lower Upper
237 100
235 60.5 53.2 79.8
272 9.2 10.8 16.2#



#41
2,4,6-Trichlorophenol
Concen: 98.666 ng/ul
RT: 12.899 min Scan# 1685
Delta R.T. 0.000 min
Lab File: BP009238.D
Acq: 25 Feb 2022 16:55

Tgt Ion:196 Resp: 16681
Ion Ratio Lower Upper
196 100
198 100.0 64.1 96.1#
200 29.5 19.9 29.9









#42
2,4,5-Trichlorophenol
Concen: 93.684 ng/ul
RT: 12.969 min Scan# 1695
Delta R.T. 0.012 min
Lab File: BP009238.D
Acq: 25 Feb 2022 16:55




Instrument :
BNA_P
ClientSampleId :
SSTD080605


Tgt Ion:196 Resp: 17236
Ion Ratio Lower Upper
196 100
198 102.0 68.6 103.0
200 35.7 32.8 49.2



#43
1,1'-Biphenyl
Concen: 71.423 ng/ul
RT: 13.310 min Scan# 1755
Delta R.T. 0.012 min
Lab File: BP009238.D
Acq: 25 Feb 2022 16:55

Tgt Ion:154 Resp: 51070
Ion Ratio Lower Upper
154 100
153 39.9 28.6 43.0
76 14.4 8.5 12.7#









#44
2-Chloronaphthalene
Concen: 73.947 ng/ul
RT: 13.346 min Scan# 1760
Delta R.T. 0.006 min
Lab File: BP009238.D
Acq: 25 Feb 2022 16:55




Instrument :
BNA_P
ClientSampleId :
SSTD080605


Tgt Ion:162 Resp: 40767
Ion Ratio Lower Upper
162 100
164 32.9 27.5 41.3
127 42.8 37.1 55.7



#45
2-Nitroaniline
Concen: 113.928 ng/ul
RT: 13.540 min Scan# 1794
Delta R.T. 0.006 min
Lab File: BP009238.D
Acq: 25 Feb 2022 16:55

Tgt Ion: 65 Resp: 12624
Ion Ratio Lower Upper
65 100
92 91.8 77.9 116.9
138 90.3 81.5 122.3









#47
Dimethylphthalate
Concen: 81.018 ng/ul
RT: 13.934 min Scan# 1861
Delta R.T. 0.006 min
Lab File: BP009238.D
Acq: 25 Feb 2022 16:55




Instrument :
BNA_P
ClientSampleId :
SSTD080605


Tgt Ion:163 Resp: 55087
Ion Ratio Lower Upper
163 100
194 4.2 4.1 6.1
164 9.7 11.3 16.9#



#48
2,6-Dinitrotoluene
Concen: 94.022 ng/ul
RT: 14.046 min Scan# 1880
Delta R.T. 0.012 min
Lab File: BP009238.D
Acq: 25 Feb 2022 16:55

Tgt Ion:165 Resp: 11364
Ion Ratio Lower Upper
165 100
89 76.3 51.7 77.5
121 18.6 18.3 27.5









#50
Acenaphthylene
Concen: 72.094 ng/ul
RT: 14.193 min Scan# 1905
Delta R.T. 0.006 min
Lab File: BP009238.D
Acq: 25 Feb 2022 16:55




Instrument :
BNA_P
ClientSampleId :
SSTD080605


Tgt Ion:152 Resp: 62862
Ion Ratio Lower Upper
152 100
151 21.7 18.3 27.5
153 14.9 11.0 16.6



#51
3-Nitroaniline
Concen: 83.709 ng/ul
RT: 14.369 min Scan# 1935
Delta R.T. 0.006 min
Lab File: BP009238.D
Acq: 25 Feb 2022 16:55

Tgt Ion:138 Resp: 10426
Ion Ratio Lower Upper
138 100
108 23.1 15.8 23.8
92 159.4 117.1 175.7









#52
Acenaphthene
Concen: 72.852 ng/ul
RT: 14.534 min Scan# 1963
Delta R.T. 0.006 min
Lab File: BP009238.D
Acq: 25 Feb 2022 16:55




Instrument :
BNA_P
ClientSampleId :
SSTD080605


Tgt Ion:153 Resp: 41480
Ion Ratio Lower Upper
153 100
152 40.8 34.3 51.5
154 79.6 70.0 105.0



#53
2,4-Dinitrophenol
Concen: 175.042 ng/ul
RT: 14.575 min Scan# 1970
Delta R.T. 0.006 min
Lab File: BP009238.D
Acq: 25 Feb 2022 16:55

Tgt Ion:184 Resp: 8930
Ion Ratio Lower Upper
184 100
63 52.2 34.6 52.0#
154 72.7 57.7 86.5









#55
4-Nitrophenol
Concen: 170.401 ng/ul
RT: 14.687 min Scan# 1989
Delta R.T. 0.024 min
Lab File: BP009238.D
Acq: 25 Feb 2022 16:55




Instrument :
BNA_P
ClientSampleId :
SSTD080605


Tgt Ion:109 Resp: 12799
Ion Ratio Lower Upper
109 100
139 76.3 61.4 92.2
65 84.0 65.3 97.9



#56
Dibenzofuran
Concen: 74.997 ng/ul
RT: 14.875 min Scan# 2021
Delta R.T. 0.006 min
Lab File: BP009238.D
Acq: 25 Feb 2022 16:55

Tgt Ion:168 Resp: 62440
Ion Ratio Lower Upper
168 100
139 39.4 30.1 45.1









#57
2,4-Dinitrotoluene
Concen: 103.283 ng/ul
RT: 14.834 min Scan# 2012
Delta R.T. 0.012 min
Lab File: BP009238.D
Acq: 25 Feb 2022 16:55




Instrument :
BNA_P
ClientSampleId :
SSTD080605


Tgt Ion:165 Resp: 17272
Ion Ratio Lower Upper
165 100
63 40.9 29.5 44.3
182 4.0 2.8 4.2



#58
2,3,4,6-Tetrachlorophenol
Concen: 114.003 ng/ul
RT: 15.098 min Scan# 2059
Delta R.T. 0.006 min
Lab File: BP009238.D
Acq: 25 Feb 2022 16:55

Tgt Ion:232 Resp: 16293
Ion Ratio Lower Upper
232 100
131 33.9 25.4 38.2
130 2.7 4.2 6.2#
166 28.9 28.1 42.1









#59
Diethylphthalate
Concen: 80.348 ng/ul
RT: 15.310 min Scan# 2095
Delta R.T. 0.006 min
Lab File: BP009238.D
Acq: 25 Feb 2022 16:55




Instrument :
BNA_P
ClientSampleId :
SSTD080605


Tgt Ion:149 Resp: 53363
Ion Ratio Lower Upper
149 100
177 26.1 22.0 33.0
150 11.2 10.2 15.2



#61
Fluorene
Concen: 75.127 ng/ul
RT: 15.528 min Scan# 2132
Delta R.T. 0.006 min
Lab File: BP009238.D
Acq: 25 Feb 2022 16:55

Tgt Ion:166 Resp: 49833
Ion Ratio Lower Upper
166 100
165 94.9 70.3 105.5
167 13.7 9.5 14.3









#62
4-Chlorophenyl-phenylether
Concen: 83.913 ng/ul
RT: 15.522 min Scan# 2131
Delta R.T. 0.000 min
Lab File: BP009238.D
Acq: 25 Feb 2022 16:55




Instrument :
BNA_P
ClientSampleId :
SSTD080605


Tgt Ion:204 Resp: 29524
Ion Ratio Lower Upper
204 100
206 30.1 27.6 41.4
141 55.7 46.6 69.8



#63
4-Nitroaniline
Concen: 98.801 ng/ul
RT: 15.545 min Scan# 2135
Delta R.T. 0.024 min
Lab File: BP009238.D
Acq: 25 Feb 2022 16:55

Tgt Ion:138 Resp: 11279
Ion Ratio Lower Upper
138 100
92 62.8 41.6 62.4#
108 147.2 97.6 146.4#









#66
4,6-Dinitro-2-methylphenol
Concen: 121.221 ng/ul
RT: 15.604 min Scan# 2145
Delta R.T. 0.018 min
Lab File: BP009238.D
Acq: 25 Feb 2022 16:55




Instrument :
BNA_P
ClientSampleId :
SSTD080605


Tgt Ion:198 Resp: 12006
Ion Ratio Lower Upper
198 100
182 3.1 6.0 9.0#
77 20.1 17.7 26.5



#67
N-Nitrosodiphenylamine
Concen: 70.213 ng/ul
RT: 15.734 min Scan# 2167
Delta R.T. 0.000 min
Lab File: BP009238.D
Acq: 25 Feb 2022 16:55

Tgt Ion:169 Resp: 45658
Ion Ratio Lower Upper
169 100
168 68.7 48.3 72.5
167 32.1 29.6 44.4









#68
4-Bromophenyl-phenylether
Concen: 82.771 ng/ul
RT: 16.422 min Scan# 21
Delta R.T. 0.006 min
Lab File: BP009238.D
Acq: 25 Feb 2022 16:55




Instrument :
BNA_P
ClientSampleId :
SSTD080605


Tgt Ion:248 Resp: 20186
Ion Ratio Lower Upper
248 100
250 98.4 79.4 119.0
141 82.7 67.2 100.8



#69
Hexachlorobenzene
Concen: 79.799 ng/ul
RT: 16.534 min Scan# 2303
Delta R.T. 0.000 min
Lab File: BP009238.D
Acq: 25 Feb 2022 16:55

Tgt Ion:284 Resp: 22658
Ion Ratio Lower Upper
284 100
142 27.0 19.2 28.8
249 37.9 27.4 41.2









#70
Atrazine
Concen: 85.004 ng/ul
RT: 16.704 min Scan# 21
Delta R.T. 0.018 min
Lab File: BP009238.D
Acq: 25 Feb 2022 16:55




Instrument :
BNA_P
ClientSampleId :
SSTD080605


Tgt Ion:200 Resp: 20975
Ion Ratio Lower Upper
200 100
173 24.5 19.2 28.8
215 50.9 32.2 48.4#



#71
Pentachlorophenol
Concen: 116.565 ng/ul
RT: 16.881 min Scan# 2362
Delta R.T. 0.006 min
Lab File: BP009238.D
Acq: 25 Feb 2022 16:55

Tgt Ion:266 Resp: 16917
Ion Ratio Lower Upper
266 100
264 57.2 57.3 85.9#
268 69.5 65.7 98.5





#72

Phenanthrene

Concen: 71.764 ng/ul

RT: 17.275 min Scan# 2429

Delta R.T. 0.006 min

Lab File: BP009238.D

Acq: 25 Feb 2022 16:55

Instrument :

BNA_P

ClientSampleId :

SSTD080605





Tgt Ion:178 Resp: 84047

Ion Ratio Lower Upper

178 100

179 14.8 11.7 17.5

176 18.3 14.5 21.7



#74

Anthracene

Concen: 74.528 ng/ul

RT: 17.369 min Scan# 2445

Delta R.T. 0.006 min

Lab File: BP009238.D

Acq: 25 Feb 2022 16:55




Tgt Ion:178 Resp: 87523


Ion Ratio Lower Upper

178 100

179 14.5 12.3 18.5

176 17.4 13.4 20.2









#75
1,2,3,4-Tetrachlorobenzene
Concen: 75.612 ng/uL
RT: 13.269 min Scan# 11
Delta R.T. 0.006 min
Lab File: BP009238.D
Acq: 25 Feb 2022 16:55




Instrument :
BNA_P
ClientSampleId :
SSTD080605


Tgt Ion:216 Resp: 27274
Ion Ratio Lower Upper
216 100
214 78.9 50.4 75.6#
218 49.3 35.2 52.8



#76
Pentachlorobenzene
Concen: 78.989 ng/uL
RT: 14.793 min Scan# 2007
Delta R.T. 0.006 min
Lab File: BP009238.D
Acq: 25 Feb 2022 16:55

Tgt Ion:250 Resp: 26156
Ion Ratio Lower Upper
250 100
248 73.4 43.5 65.3#
252 60.0 46.5 69.7









#77
 Carbazole
 Concen: 74.147 ng/ul
 RT: 17.634 min Scan# 2489
 Delta R.T. 0.006 min
 Lab File: BP009238.D
 Acq: 25 Feb 2022 16:55




Instrument :
 BNA_P
 ClientSampleId :
 SSTD080605


Tgt Ion:167 Resp: 74201
 Ion Ratio Lower Upper
 167 100
 166 24.9 15.8 23.8#
 139 13.8 8.3 12.5#



#78
 Di-n-butylphthalate
 Concen: 81.126 ng/ul
 RT: 18.210 min Scan# 2588
 Delta R.T. 0.006 min
 Lab File: BP009238.D
 Acq: 25 Feb 2022 16:55

Tgt Ion:149 Resp: 94792
 Ion Ratio Lower Upper
 149 100
 150 9.0 9.6 14.4#
 104 7.7 6.2 9.4









#80
Fluoranthene
Concen: 68.367 ng/u1
RT: 19.251 min Scan# 21
Delta R.T. 0.006 min
Lab File: BP009238.D
Acq: 25 Feb 2022 16:55




Instrument :
BNA_P
ClientSampleId :
SSTD080605


Tgt Ion:202 Resp: 109593
Ion Ratio Lower Upper
202 100
101 11.1 9.6 14.4
100 7.6 6.3 9.5



#82
Pyrene
Concen: 67.788 ng/u1
RT: 19.604 min Scan# 2825
Delta R.T. 0.006 min
Lab File: BP009238.D
Acq: 25 Feb 2022 16:55

Tgt Ion:202 Resp: 109496
Ion Ratio Lower Upper
202 100
101 12.5 10.6 15.8
100 10.6 9.8 14.6





#83

Butylbenzylphthalate

Concen: 80.846 ng/ul

RT: 20.492 min Scan# 2976

Delta R.T. 0.000 min

Lab File: BP009238.D

Acq: 25 Feb 2022 16:55

Instrument :

BNA_P

ClientSampleId :

SSTD080605





Tgt Ion:149 Resp: 44261

Ion Ratio Lower Upper

149 100

91 95.0 71.1 106.7

206 25.9 28.8 43.2#



#84

3,3'-Dichlorobenzidine

Concen: 92.234 ng/ul

RT: 21.257 min Scan# 3106

Delta R.T. 0.006 min

Lab File: BP009238.D

Acq: 25 Feb 2022 16:55




Tgt Ion:252 Resp: 43955


Ion Ratio Lower Upper

252 100

254 62.4 49.8 74.6

126 7.5 4.3 6.5#





#85

Benzo(a)anthracene

Concen: 78.935 ng/ul

RT: 21.322 min Scan# 3117

Delta R.T. 0.000 min

Lab File: BP009238.D

Acq: 25 Feb 2022 16:55

Instrument :

BNA_P

ClientSampleId :

SSTD080605





Tgt Ion:228 Resp: 114658

Ion Ratio Lower Upper

228 100

229 18.5 18.2 27.2

226 27.0 22.8 34.2



#86

Bis(2-ethylhexyl)phthalate

Concen: 79.262 ng/ul

RT: 21.269 min Scan# 3108

Delta R.T. 0.000 min

Lab File: BP009238.D

Acq: 25 Feb 2022 16:55




Tgt Ion:149 Resp: 65864


Ion Ratio Lower Upper

149 100

167 37.3 32.8 49.2

279 5.8 5.6 8.4





#87

Chrysene

Concen: 78.159 ng/ul

RT: 21.375 min Scan# 3125

Delta R.T. 0.006 min

Lab File: BP009238.D

Acq: 25 Feb 2022 16:55

Instrument :

BNA_P

ClientSampleId :

SSTD080605





Tgt Ion:228 Resp: 109616

Ion Ratio Lower Upper

228 100

226 26.9 22.7 34.1

229 18.9 15.2 22.8



#89

Di-n-octyl phthalate

Concen: 75.647 ng/ul




RT: 22.204 min Scan# 3267


Delta R.T. 0.000 min

Lab File: BP009238.D

Acq: 25 Feb 2022 16:55

Tgt Ion:149 Resp: 121296





#90

Benzo(b)fluoranthene

Concen: 76.515 ng/ul

RT: 23.022 min Scan# 3406

Delta R.T. 0.012 min

Lab File: BP009238.D

Acq: 25 Feb 2022 16:55

Instrument :

BNA_P

ClientSampleId :

SSTD080605





Tgt Ion:252 Resp: 127669

Ion Ratio Lower Upper

252 100

253 20.8 17.4 26.0

125 5.1 7.3 10.9#



#91

Benzo(k)fluoranthene

Concen: 70.450 ng/ul

RT: 23.069 min Scan# 3414

Delta R.T. 0.006 min

Lab File: BP009238.D

Acq: 25 Feb 2022 16:55




Tgt Ion:252 Resp: 113450


Ion Ratio Lower Upper

252 100

253 24.5 17.9 26.9

125 6.4 5.1 7.7





#93

Benzo(a)pyrene

Concen: 74.347 ng/ul

RT: 23.657 min Scan# 31

Delta R.T. 0.018 min

Lab File: BP009238.D

Acq: 25 Feb 2022 16:55

Instrument :

BNA_P

ClientSampleId :

SSTD080605





Tgt Ion:252 Resp: 118744

Ion Ratio Lower Upper

252 100

253 19.6 17.5 26.3

125 6.1 6.4 9.6#



#94

Indeno(1,2,3-cd)pyrene

Concen: 79.414 ng/ul

RT: 26.274 min Scan# 3959

Delta R.T. 0.018 min

Lab File: BP009238.D

Acq: 25 Feb 2022 16:55




Tgt Ion:276 Resp: 139871


Ion Ratio Lower Upper

276 100

138 12.9 12.3 18.5

277 26.2 19.3 28.9









#95
 Dibenzo(a,h)anthracene
 Concen: 78.916 ng/ul
 RT: 26.298 min Scan# 3963
 Delta R.T. 0.018 min
 Lab File: BP009238.D
 Acq: 25 Feb 2022 16:55

Instrument :
 BNA_P
 ClientSampleId :
 SSTD080605

Tgt Ion:278 Resp: 120606
 Ion Ratio Lower Upper
 278 100
 139 9.6 6.6 10.0
 279 24.0 18.4 27.6



#96
 Benzo(g,h,i)perylene
 Concen: 79.689 ng/ul
 RT: 27.045 min Scan# 4090
 Delta R.T. 0.024 min
 Lab File: BP009238.D
 Acq: 25 Feb 2022 16:55

Tgt Ion:276 Resp: 116004
 Ion Ratio Lower Upper
 276 100
 138 11.7 11.6 17.4
 277 24.9 15.9 23.9#

