

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP022522\
 Data File : BP009240.D
 Acq On : 25 Feb 2022 19:09
 Operator : CG/JU
 Sample : SSTDICV020
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :

Quant Time: Feb 28 04:28:02 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP022522.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Mon Feb 28 04:25:37 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.828	152	2415	20.000	ng/ul	0.00
20) Naphthalene-d8	10.628	136	8386	20.000	ng/ul	0.00
38) Acenaphthene-d10	14.469	164	6648	20.000	ng/ul	0.00
64) Phenanthrene-d10	17.222	188	15042	20.000	ng/ul	# 0.00
79) Chrysene-d12	21.333	240	18399	20.000	ng/ul	# 0.00
88) Perylene-d12	23.745	264	20390	20.000	ng/ul	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.275	96	9	0.106	ng/uL	-0.05
4) Pyridine-d5	3.722	84	3430	20.785	ng/ul	0.00
7) Phenol-d5	6.987	99	4726	19.357	ng/ul	0.00
9) Bis-(2-Chloroethyl)eth...	7.158	67	2159	19.702	ng/ul	0.00
11) 2-Chlorophenol-d4	7.363	132	2751	18.768	ng/ul	0.00
15) 4-Methylphenol-d8	8.528	113	3695	19.622	ng/ul	0.00
21) Nitrobenzene-d5	8.969	128	1482	21.989	ng/ul	0.00
24) 2-Nitrophenol-d4	9.699	143	1885	23.084	ng/ul	0.00
28) 2,4-Dichlorophenol-d3	10.075	165	9	0.052	ng/ul	-0.16
31) 4-Chloroaniline-d4	10.751	131	3399	18.456	ng/ul	0.00
46) Dimethylphthalate-d6	13.881	166	10253	19.904	ng/ul	0.00
49) Acenaphthylene-d8	14.157	160	12305	20.040	ng/ul	0.00
54) 4-Nitrophenol-d4	14.645	143	1709	19.761	ng/ul	0.00
60) Fluorene-d10	15.463	176	8707	18.686	ng/ul	0.00
65) 4,6-Dinitro-2-methylph...	15.575	200	2055	19.210	ng/ul	0.00
73) Anthracene-d10	17.322	188	14576	20.189	ng/ul	0.00
81) Pyrene-d10	19.569	212	17483	18.672	ng/ul	0.00
92) Benzo(a)pyrene-d12	23.592	264	20581	19.770	ng/ul	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.352	88	611	8.058	ng/uL#	64
5) Pyridine	3.740	79	2797	17.254	ng/ul#	85
6) Benzaldehyde	6.869	77	5	0.036	ng/ul	90
8) Phenol	7.016	94	4717	19.164	ng/ul	95
10) Bis(2-Chloroethyl)ether	7.252	93	3707	19.643	ng/ul	93
12) 2-Chlorophenol	7.393	128	2873	18.458	ng/ul#	87
13) 2-Methylphenol	8.263	108	3568	18.583	ng/ul	82
14) 2,2'-oxybis(1-Chloropr...	8.357	45	2116	20.790	ng/ul#	55
16) Acetophenone	8.646	105	6198	19.610	ng/ul	92
17) N-Nitroso-di-n-propyla...	8.640	70	2626	19.055	ng/ul#	91
18) 4-Methylphenol	8.593	108	3901	19.269	ng/ul	93
19) Hexachloroethane	8.916	117	1613	19.070	ng/ul#	80
22) Nitrobenzene	9.016	77	3934	19.552	ng/ul#	94
23) Isophorone	9.552	82	8162	20.094	ng/ul#	95
25) 2-Nitrophenol	9.734	139	1705	21.666	ng/ul#	87
26) 2,4-Dimethylphenol	9.963	107	6	0.025	ng/ul	91
27) Bis(2-Chloroethoxy)met...	10.034	93	4785	19.394	ng/ul#	94
29) 2,4-Dichlorophenol	10.269	162	2840	18.045	ng/ul#	74
30) Naphthalene	10.675	128	8869	19.082	ng/ul	95
32) 4-Chloroaniline	10.775	127	3895	20.562	ng/ul#	82
33) Hexachlorobutadiene	10.975	225	2780	19.839	ng/ul	90
34) Caprolactam	11.457	113	4	0.073	ng/ul	93
35) 4-Chloro-3-methylphenol	11.904	107	4642	19.106	ng/ul#	95
36) 2-Methylnaphthalene	12.287	142	6321	20.101	ng/ul	99

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP022522\
 Data File : BP009240.D
 Acq On : 25 Feb 2022 19:09
 Operator : CG/JU
 Sample : SSTDICV020
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :

Quant Time: Feb 28 04:28:02 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP022522.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Mon Feb 28 04:25:37 2022
 Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.504	142	6327	19.565	ng/ul#	99
39) 1,2,4,5-Tetrachloroben...	12.657	216	4799	20.064	ng/ul	91
40) Hexachlorocyclopentadiene	12.645	237	3880	20.683	ng/ul#	79
41) 2,4,6-Trichlorophenol	12.898	196	3025	19.402	ng/ul	93
42) 2,4,5-Trichlorophenol	12.957	196	3021	19.100	ng/ul#	74
43) 1,1'-Biphenyl	13.298	154	9477	20.648	ng/ul#	95
44) 2-Chloronaphthalene	13.340	162	7145	19.606	ng/ul	90
45) 2-Nitroaniline	13.528	65	2180	20.577	ng/ul#	68
47) Dimethylphthalate	13.928	163	9542	19.202	ng/ul#	97
48) 2,6-Dinitrotoluene	14.040	165	1988	19.470	ng/ul#	85
50) Acenaphthylene	14.187	152	11138	19.248	ng/ul	97
51) 3-Nitroaniline	14.357	138	1835	19.320	ng/ul#	64
52) Acenaphthene	14.534	153	7738	19.967	ng/ul	91
53) 2,4-Dinitrophenol	14.563	184	1567	20.201	ng/ul#	83
55) 4-Nitrophenol	14.663	109	2147	18.290	ng/ul	93
56) Dibenzofuran	14.869	168	11438	19.599	ng/ul	95
57) 2,4-Dinitrotoluene	14.822	165	3251	21.563	ng/ul#	88
58) 2,3,4,6-Tetrachlorophenol	15.092	232	3224	21.322	ng/ul#	94
59) Diethylphthalate	15.298	149	9267	18.988	ng/ul#	94
61) Fluorene	15.516	166	9506	20.073	ng/ul	100
62) 4-Chlorophenyl-phenyle...	15.522	204	5601	19.687	ng/ul	98
63) 4-Nitroaniline	15.522	138	1996	20.311	ng/ul	92
66) 4,6-Dinitro-2-methylph...	15.587	198	2047	20.324	ng/ul#	86
67) N-Nitrosodiphenylamine	15.728	169	8640	21.200	ng/ul	90
68) 4-Bromophenyl-phenylether	16.422	248	3851	20.338	ng/ul	91
69) Hexachlorobenzene	16.534	284	4337	20.840	ng/ul	93
70) Atrazine	16.692	200	3479	18.570	ng/ul	95
71) Pentachlorophenol	16.869	266	2900	19.799	ng/ul#	74
72) Phenanthrene	17.269	178	15334	19.935	ng/ul	98
74) Anthracene	17.357	178	15909	19.918	ng/ul	95
75) 1,2,3,4-Tetrachloroben...	13.263	216	4646	18.930	ng/ul	91
76) Pentachlorobenzene	14.792	250	4937	20.493	ng/ul	97
77) Carbazole	17.628	167	14294	21.196	ng/ul#	94
78) Di-n-butylphthalate	18.204	149	16878	19.556	ng/ul	96
80) Fluoranthene	19.245	202	19570	18.665	ng/ul	96
82) Pyrene	19.598	202	20397	19.179	ng/ul#	96
83) Butylbenzylphthalate	20.492	149	7835	18.283	ng/ul	86
84) 3,3'-Dichlorobenzidine	21.251	252	9070	20.250	ng/ul	94
85) Benzo(a)anthracene	21.316	228	23812	20.880	ng/ul	92
86) Bis(2-ethylhexyl)phtha...	21.263	149	11670	18.134	ng/ul	95
87) Chrysene	21.369	228	20848	18.712	ng/ul	98
89) Di-n-octyl phthalate	22.198	149	21861	18.290	ng/ul	100
90) Benzo(b)fluoranthene	23.010	252	24956	19.466	ng/ul#	95
91) Benzo(k)fluoranthene	23.057	252	23420	20.288	ng/ul	97
93) Benzo(a)pyrene	23.639	252	23860	19.821	ng/ul	94
94) Indeno(1,2,3-cd)pyrene	26.257	276	20425	13.992	ng/ul	93
95) Dibenzo(a,h)anthracene	26.274	278	23716	19.182	ng/ul#	96
96) Benzo(g,h,i)perylene	27.021	276	23428	19.261	ng/ul#	93


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP022522\
 Data File : BP009240.D
 Acq On : 25 Feb 2022 19:09
 Operator : CG/JU
 Sample : SSTDICV020
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :

Quant Time: Feb 28 04:28:02 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP022522.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Mon Feb 28 04:25:37 2022
 Response via : Initial Calibration









#2
1,4-Dioxane
Concen: 8.058 ng/uL
RT: 3.352 min Scan# 61
Delta R.T. -0.006 min
Lab File: BP009240.D
Acq: 25 Feb 2022 19:09




Instrument :
BNA_P
ClientSampleId :


Tgt Ion: 88 Resp: 611
Ion Ratio Lower Upper
88 100
43 16.9 6.8 10.2#
58 58.0 28.4 42.6#



#5
Pyridine
Concen: 17.254 ng/uL
RT: 3.740 min Scan# 128
Delta R.T. -0.000 min
Lab File: BP009240.D
Acq: 25 Feb 2022 19:09


Tgt Ion: 79 Resp: 2797
Ion Ratio Lower Upper
79 100
52 58.4 35.5 53.3#
51 24.0 20.2 30.4








#6
Benzaldehyde
Concen: 0.036 ng/ul
RT: 6.869 min Scan# 60
Delta R.T. -0.094 min
Lab File: BP009240.D
Acq: 25 Feb 2022 19:09

Instrument :
BNA_P
ClientSampleId :







Tgt Ion: 77 Resp: 5
Ion Ratio Lower Upper
77 100
105 112.5 77.0 115.4
106 100.0 77.0 115.4



#8
Phenol
Concen: 19.164 ng/ul
RT: 7.016 min Scan# 685
Delta R.T. -0.000 min
Lab File: BP009240.D
Acq: 25 Feb 2022 19:09


Tgt Ion: 94 Resp: 4717
Ion Ratio Lower Upper
94 100
65 19.8 19.0 28.6
66 41.1 34.6 51.8








#10
 Bis(2-Chloroethyl)ether
 Concen: 19.643 ng/ul
 RT: 7.252 min Scan# 71
 Delta R.T. -0.000 min
 Lab File: BP009240.D
 Acq: 25 Feb 2022 19:09

Instrument :
 BNA_P
 ClientSampleId :



Tgt Ion: 93 Resp: 3707




Ion	Ratio	Lower	Upper
93	100		
63	61.0	45.0	67.4
95	29.3	26.3	39.5




#12
 2-Chlorophenol
 Concen: 18.458 ng/ul
 RT: 7.393 min Scan# 749
 Delta R.T. 0.006 min
 Lab File: BP009240.D
 Acq: 25 Feb 2022 19:09

Tgt Ion: 128 Resp: 2873

Ion	Ratio	Lower	Upper
128	100		
64	53.8	45.3	67.9
130	19.9	29.0	43.6





#13

2-Methylphenol

Concen: 18.583 ng/ul

RT: 8.263 min Scan# 898

Delta R.T. -0.006 min

Lab File: BP009240.D

Acq: 25 Feb 2022 19:09

Instrument :

BNA_P





ClientSampleId :

Tgt Ion:108 Resp: 3568

Ion Ratio Lower Upper

108 100

107 71.9 70.8 106.2



#14

2,2'-oxybis(1-Chloropropane)

Concen: 20.790 ng/ul

RT: 8.357 min Scan# 913

Delta R.T. -0.000 min

Lab File: BP009240.D

Acq: 25 Feb 2022 19:09




Tgt Ion: 45 Resp: 2116


Ion Ratio Lower Upper

45 100

77 53.8 17.1 25.7#

79 22.5 13.2 19.8#









#16
Acetophenone
Concen: 19.610 ng/ul
RT: 8.646 min Scan# 962
Delta R.T. -0.000 min
Lab File: BP009240.D
Acq: 25 Feb 2022 19:09




Instrument :
BNA_P
ClientSampleId :


Tgt Ion:105 Resp: 6198
Ion Ratio Lower Upper
105 100
77 74.0 53.4 80.2
51 19.6 13.8 20.8



#17
N-Nitroso-di-n-propylamine
Concen: 19.055 ng/ul
RT: 8.640 min Scan# 961
Delta R.T. -0.000 min
Lab File: BP009240.D
Acq: 25 Feb 2022 19:09

Tgt Ion: 70 Resp: 2626
Ion Ratio Lower Upper
70 100
42 41.1 31.8 47.8
101 10.1 15.6 23.4#
130 12.8 13.7 20.5#





#18

4-Methylphenol

Concen: 19.269 ng/ul

RT: 8.593 min Scan# 91

Delta R.T. -0.006 min

Lab File: BP009240.D

Acq: 25 Feb 2022 19:09

Instrument :

BNA_P





ClientSampleId :

Tgt Ion:108 Resp: 3901

Ion Ratio Lower Upper

108 100

107 111.4 83.6 125.4



#19

Hexachloroethane

Concen: 19.070 ng/ul

RT: 8.916 min Scan# 1008

Delta R.T. 0.006 min

Lab File: BP009240.D

Acq: 25 Feb 2022 19:09




Tgt Ion:117 Resp: 1613


Ion Ratio Lower Upper

117 100

201 103.4 106.5 159.7#

199 63.6 59.0 88.6









#22
Nitrobenzene
Concen: 19.552 ng/ul
RT: 9.016 min Scan# 1026
Delta R.T. -0.006 min
Lab File: BP009240.D
Acq: 25 Feb 2022 19:09




Instrument :
BNA_P
ClientSampleId :


Tgt Ion: 77 Resp: 3934
Ion Ratio Lower Upper
77 100
123 55.5 42.3 63.5
65 13.8 5.8 8.8#



#23
Isophorone
Concen: 20.094 ng/ul
RT: 9.552 min Scan# 1116
Delta R.T. -0.000 min
Lab File: BP009240.D
Acq: 25 Feb 2022 19:09

Tgt Ion: 82 Resp: 8162
Ion Ratio Lower Upper
82 100
95 13.1 7.8 11.8#
138 14.9 12.9 19.3









#25
2-Nitrophenol
Concen: 21.666 ng/ul
RT: 9.734 min Scan# 1148
Delta R.T. -0.006 min
Lab File: BP009240.D
Acq: 25 Feb 2022 19:09




Instrument :
BNA_P
ClientSampleId :


Tgt Ion:139 Resp: 1705
Ion Ratio Lower Upper
139 100
65 42.8 49.4 74.2#
109 55.6 44.7 67.1



#26
2,4-Dimethylphenol
Concen: 0.025 ng/ul
RT: 9.963 min Scan# 1186
Delta R.T. 0.165 min
Lab File: BP009240.D
Acq: 25 Feb 2022 19:09

Tgt Ion:107 Resp: 6
Ion Ratio Lower Upper
107 100
121 43.8 34.1 51.1
122 68.8 46.3 69.5









#27
 Bis(2-Chloroethoxy)methane
 Concen: 19.394 ng/ul
 RT: 10.034 min Scan# 1199
 Delta R.T. -0.006 min
 Lab File: BP009240.D
 Acq: 25 Feb 2022 19:09




Instrument :
 BNA_P
 ClientSampleId :


Tgt Ion: 93 Resp: 4785
 Ion Ratio Lower Upper
 93 100
 95 31.2 26.2 39.4
 123 14.3 6.2 9.4#



#29
 2,4-Dichlorophenol
 Concen: 18.045 ng/ul
 RT: 10.269 min Scan# 1238
 Delta R.T. 0.006 min
 Lab File: BP009240.D
 Acq: 25 Feb 2022 19:09

Tgt Ion:162 Resp: 2840
 Ion Ratio Lower Upper
 162 100
 164 65.3 37.0 55.6#
 98 53.0 31.0 46.6#









#30
Naphthalene
Concen: 19.082 ng/ul
RT: 10.675 min Scan# 1307
Delta R.T. -0.000 min
Lab File: BP009240.D
Acq: 25 Feb 2022 19:09




Instrument :
BNA_P
ClientSampleId :


Tgt Ion:128 Resp: 8869
Ion Ratio Lower Upper
128 100
129 11.4 8.6 13.0
127 12.4 12.2 18.4



#32
4-Chloroaniline
Concen: 20.562 ng/ul
RT: 10.775 min Scan# 1324
Delta R.T. -0.006 min
Lab File: BP009240.D
Acq: 25 Feb 2022 19:09

Tgt Ion:127 Resp: 3895
Ion Ratio Lower Upper
127 100
129 40.6 24.6 37.0#





#33

Hexachlorobutadiene

Concen: 19.839 ng/ul

RT: 10.975 min Scan# 11

Delta R.T. -0.006 min

Lab File: BP009240.D





Acq: 25 Feb 2022 19:09

Instrument :

BNA_P

ClientSampleId :

Tgt Ion:	225	Resp:	2780
Ion	Ratio	Lower	Upper
225	100		
223	72.8	53.4	80.0
227	57.8	53.8	80.8



#34

Caprolactam

Concen: 0.073 ng/ul




RT: 11.457 min Scan# 1440


Delta R.T. -0.106 min

Lab File: BP009240.D

Acq: 25 Feb 2022 19:09

Tgt Ion:	113	Resp:	4
Ion	Ratio	Lower	Upper
113	100		
55	80.0	58.2	87.2
56	80.0	68.6	103.0









#35
4-Chloro-3-methylphenol
Concen: 19.106 ng/ul
RT: 11.904 min Scan# 1517
Delta R.T. -0.006 min
Lab File: BP009240.D
Acq: 25 Feb 2022 19:09




Instrument :
BNA_P
ClientSampleId :


Tgt Ion:107 Resp: 4642
Ion Ratio Lower Upper
107 100
144 13.1 15.8 23.8#
142 48.1 38.9 58.3



#36
2-Methylnaphthalene
Concen: 20.101 ng/ul
RT: 12.287 min Scan# 1581
Delta R.T. -0.000 min
Lab File: BP009240.D
Acq: 25 Feb 2022 19:09

Tgt Ion:142 Resp: 6321
Ion Ratio Lower Upper
142 100
141 81.3 64.3 96.5









#37
1-Methylnaphthalene
Concen: 19.565 ng/ul
RT: 12.504 min Scan# 1618
Delta R.T. -0.000 min
Lab File: BP009240.D
Acq: 25 Feb 2022 19:09




Instrument :
BNA_P
ClientSampleId :


Tgt Ion:142 Resp: 6327
Ion Ratio Lower Upper
142 100
141 87.4 69.8 104.8
116 4.4 5.6 8.4#



#39
1,2,4,5-Tetrachlorobenzene
Concen: 20.064 ng/ul
RT: 12.657 min Scan# 1644
Delta R.T. -0.000 min
Lab File: BP009240.D
Acq: 25 Feb 2022 19:09

Tgt Ion:216 Resp: 4799
Ion Ratio Lower Upper
216 100
214 81.4 58.2 87.2
179 19.6 14.2 21.2
108 16.4 15.3 22.9









#40
Hexachlorocyclopentadiene
Concen: 20.683 ng/ul
RT: 12.645 min Scan# 1
Delta R.T. -0.000 min
Lab File: BP009240.D
Acq: 25 Feb 2022 19:09




Instrument :
BNA_P
ClientSampleId :


Tgt Ion:237 Resp: 3880
Ion Ratio Lower Upper
237 100
235 48.8 53.2 79.8#
272 20.1 10.8 16.2#



#41
2,4,6-Trichlorophenol
Concen: 19.402 ng/ul
RT: 12.898 min Scan# 1685
Delta R.T. -0.000 min
Lab File: BP009240.D
Acq: 25 Feb 2022 19:09

Tgt Ion:196 Resp: 3025
Ion Ratio Lower Upper
196 100
198 85.4 64.1 96.1
200 29.2 19.9 29.9









#42
2,4,5-Trichlorophenol
Concen: 19.100 ng/ul
RT: 12.957 min Scan# 1695
Delta R.T. -0.000 min
Lab File: BP009240.D
Acq: 25 Feb 2022 19:09




Instrument :
BNA_P
ClientSampleId :


Tgt Ion:196 Resp: 3021
Ion Ratio Lower Upper
196 100
198 116.6 68.6 103.0#
200 35.6 32.8 49.2



#43
1,1'-Biphenyl
Concen: 20.648 ng/ul
RT: 13.298 min Scan# 1753
Delta R.T. -0.000 min
Lab File: BP009240.D
Acq: 25 Feb 2022 19:09

Tgt Ion:154 Resp: 9477
Ion Ratio Lower Upper
154 100
153 37.8 28.6 43.0
76 14.8 8.5 12.7#









#44
2-Chloronaphthalene
Concen: 19.606 ng/ul
RT: 13.340 min Scan# 1760
Delta R.T. -0.000 min
Lab File: BP009240.D
Acq: 25 Feb 2022 19:09




Instrument :
BNA_P
ClientSampleId :


Tgt Ion:	162	Resp:	7145
Ion Ratio	Lower	Upper	
162	100		
164	39.5	27.5	41.3
127	54.0	37.1	55.7



#45
2-Nitroaniline
Concen: 20.577 ng/ul
RT: 13.528 min Scan# 1792
Delta R.T. -0.006 min
Lab File: BP009240.D
Acq: 25 Feb 2022 19:09

Tgt Ion:	65	Resp:	2180
Ion Ratio	Lower	Upper	
65	100		
92	74.3	77.9	116.9#
138	60.6	81.5	122.3#









#47
Dimethylphthalate
Concen: 19.202 ng/ul
RT: 13.928 min Scan# 1860
Delta R.T. -0.000 min
Lab File: BP009240.D
Acq: 25 Feb 2022 19:09




Instrument :
BNA_P
ClientSampleId :


Tgt Ion:163 Resp: 9542
Ion Ratio Lower Upper
163 100
194 3.1 4.1 6.1#
164 13.4 11.3 16.9



#48
2,6-Dinitrotoluene
Concen: 19.470 ng/ul
RT: 14.040 min Scan# 1879
Delta R.T. 0.006 min
Lab File: BP009240.D
Acq: 25 Feb 2022 19:09

Tgt Ion:165 Resp: 1988
Ion Ratio Lower Upper
165 100
89 76.5 51.7 77.5
121 30.5 18.3 27.5#









#50
Acenaphthylene
Concen: 19.248 ng/ul
RT: 14.187 min Scan# 1904
Delta R.T. -0.000 min
Lab File: BP009240.D
Acq: 25 Feb 2022 19:09




Instrument :
BNA_P
ClientSampleId :


Tgt Ion:152	Resp:	11138
Ion Ratio	Lower	Upper
152	100	
151	21.6	18.3 27.5
153	12.3	11.0 16.6



#51
3-Nitroaniline
Concen: 19.320 ng/ul
RT: 14.357 min Scan# 1933
Delta R.T. -0.006 min
Lab File: BP009240.D
Acq: 25 Feb 2022 19:09

Tgt Ion:138	Resp:	1835
Ion Ratio	Lower	Upper
138	100	
108	29.7	15.8 23.8#
92	194.0	117.1 175.7#





#52

Acenaphthene

Concen: 19.967 ng/ul

RT: 14.534 min Scan# 1962

Delta R.T. 0.006 min

Lab File: BP009240.D

Acq: 25 Feb 2022 19:09

Instrument :

BNA_P

ClientSampleId :





Tgt Ion:153 Resp: 7738

Ion Ratio Lower Upper

153 100

152 44.4 34.3 51.5

154 75.9 70.0 105.0



#53

2,4-Dinitrophenol

Concen: 20.201 ng/ul

RT: 14.563 min Scan# 1968

Delta R.T. -0.006 min

Lab File: BP009240.D

Acq: 25 Feb 2022 19:09




Tgt Ion:184 Resp: 1567


Ion Ratio Lower Upper

184 100

63 60.6 34.6 52.0#

154 81.6 57.7 86.5









#55
4-Nitrophenol
Concen: 18.290 ng/ul
RT: 14.663 min Scan# 1985
Delta R.T. -0.000 min
Lab File: BP009240.D
Acq: 25 Feb 2022 19:09




Instrument :
BNA_P
ClientSampleId :


Tgt Ion:109 Resp: 2147
Ion Ratio Lower Upper
109 100
139 81.2 61.4 92.2
65 73.3 65.3 97.9



#56
Dibenzofuran
Concen: 19.599 ng/ul
RT: 14.869 min Scan# 2020
Delta R.T. -0.000 min
Lab File: BP009240.D
Acq: 25 Feb 2022 19:09

Tgt Ion:168 Resp: 11438
Ion Ratio Lower Upper
168 100
139 34.6 30.1 45.1









#57
2,4-Dinitrotoluene
Concen: 21.563 ng/ul
RT: 14.822 min Scan# 2012
Delta R.T. -0.000 min
Lab File: BP009240.D
Acq: 25 Feb 2022 19:09




Instrument :
BNA_P
ClientSampleId :


Tgt Ion:165 Resp: 3251
Ion Ratio Lower Upper
165 100
63 44.7 29.5 44.3#
182 2.6 2.8 4.2#



#58
2,3,4,6-Tetrachlorophenol
Concen: 21.322 ng/ul
RT: 15.092 min Scan# 2058
Delta R.T. -0.000 min
Lab File: BP009240.D
Acq: 25 Feb 2022 19:09

Tgt Ion:232 Resp: 3224
Ion Ratio Lower Upper
232 100
131 35.8 25.4 38.2
130 1.0 4.2 6.2#
166 37.1 28.1 42.1









#59
Diethylphthalate
Concen: 18.988 ng/ul
RT: 15.298 min Scan# 2094
Delta R.T. -0.006 min
Lab File: BP009240.D
Acq: 25 Feb 2022 19:09




Instrument :
BNA_P
ClientSampleId :


Tgt Ion:149 Resp: 9267
Ion Ratio Lower Upper
149 100
177 25.0 22.0 33.0
150 15.9 10.2 15.2#



#61
Fluorene
Concen: 20.073 ng/ul
RT: 15.516 min Scan# 2130
Delta R.T. -0.006 min
Lab File: BP009240.D
Acq: 25 Feb 2022 19:09

Tgt Ion:166 Resp: 9506
Ion Ratio Lower Upper
166 100
165 87.7 70.3 105.5
167 11.5 9.5 14.3









#62
4-Chlorophenyl-phenylether
Concen: 19.687 ng/ul
RT: 15.522 min Scan# 2131
Delta R.T. -0.000 min
Lab File: BP009240.D
Acq: 25 Feb 2022 19:09




Instrument :
BNA_P
ClientSampleId :


Tgt Ion:204 Resp: 5601
Ion Ratio Lower Upper
204 100
206 36.4 27.6 41.4
141 57.3 46.6 69.8



#63
4-Nitroaniline
Concen: 20.311 ng/ul
RT: 15.522 min Scan# 2131
Delta R.T. -0.000 min
Lab File: BP009240.D
Acq: 25 Feb 2022 19:09

Tgt Ion:138 Resp: 1996
Ion Ratio Lower Upper
138 100
92 41.9 41.6 62.4
108 115.8 97.6 146.4









#66
4,6-Dinitro-2-methylphenol
Concen: 20.324 ng/ul
RT: 15.587 min Scan# 2142
Delta R.T. -0.000 min
Lab File: BP009240.D
Acq: 25 Feb 2022 19:09




Instrument :
BNA_P
ClientSampleId :


Tgt Ion:198 Resp: 2047
Ion Ratio Lower Upper
198 100
182 4.5 6.0 9.0#
77 29.7 17.7 26.5#



#67
N-Nitrosodiphenylamine
Concen: 21.200 ng/ul
RT: 15.728 min Scan# 2166
Delta R.T. -0.006 min
Lab File: BP009240.D
Acq: 25 Feb 2022 19:09

Tgt Ion:169 Resp: 8640
Ion Ratio Lower Upper
169 100
168 69.7 48.3 72.5
167 41.0 29.6 44.4









#68
4-Bromophenyl-phenylether
Concen: 20.338 ng/ul
RT: 16.422 min Scan# 21
Delta R.T. 0.006 min
Lab File: BP009240.D
Acq: 25 Feb 2022 19:09




Instrument :
BNA_P
ClientSampleId :


Tgt Ion:248 Resp: 3851
Ion Ratio Lower Upper
248 100
250 92.2 79.4 119.0
141 73.2 67.2 100.8



#69
Hexachlorobenzene
Concen: 20.840 ng/ul
RT: 16.534 min Scan# 2303
Delta R.T. -0.000 min
Lab File: BP009240.D
Acq: 25 Feb 2022 19:09

Tgt Ion:284 Resp: 4337
Ion Ratio Lower Upper
284 100
142 23.2 19.2 28.8
249 40.8 27.4 41.2





#70

Atrazine

Concen: 18.570 ng/ul

RT: 16.692 min Scan# 21

Delta R.T. 0.006 min

Lab File: BP009240.D

Acq: 25 Feb 2022 19:09

Instrument :

BNA_P

ClientSampleId :





Tgt Ion:200 Resp: 3479

Ion Ratio Lower Upper

200 100

173 27.6 19.2 28.8

215 42.5 32.2 48.4



#71

Pentachlorophenol

Concen: 19.799 ng/ul

RT: 16.869 min Scan# 2360

Delta R.T. -0.006 min

Lab File: BP009240.D

Acq: 25 Feb 2022 19:09




Tgt Ion:266 Resp: 2900


Ion Ratio Lower Upper

266 100

264 53.4 57.3 85.9#

268 55.8 65.7 98.5#





#72

Phenanthrene

Concen: 19.935 ng/ul

RT: 17.269 min Scan# 2428

Delta R.T. -0.000 min

Lab File: BP009240.D





Acq: 25 Feb 2022 19:09

Instrument :

BNA_P

ClientSampleId :

Tgt Ion:178	Resp:	15334
Ion Ratio	Lower	Upper
178	100	
179	14.7	11.7 17.5
176	16.5	14.5 21.7



#74

Anthracene

Concen: 19.918 ng/ul




RT: 17.357 min Scan# 2443


Delta R.T. -0.006 min

Lab File: BP009240.D

Acq: 25 Feb 2022 19:09

Tgt Ion:178	Resp:	15909
Ion Ratio	Lower	Upper
178	100	
179	14.2	12.3 18.5
176	19.9	13.4 20.2









#75
1,2,3,4-Tetrachlorobenzene
Concen: 18.930 ng/uL
RT: 13.263 min Scan# 1
Delta R.T. -0.000 min
Lab File: BP009240.D
Acq: 25 Feb 2022 19:09




Instrument :
BNA_P
ClientSampleId :


Tgt Ion:216 Resp: 4646
Ion Ratio Lower Upper
216 100
214 72.7 50.4 75.6
218 40.9 35.2 52.8



#76
Pentachlorobenzene
Concen: 20.493 ng/uL
RT: 14.792 min Scan# 2007
Delta R.T. 0.006 min
Lab File: BP009240.D
Acq: 25 Feb 2022 19:09

Tgt Ion:250 Resp: 4937
Ion Ratio Lower Upper
250 100
248 57.7 43.5 65.3
252 57.4 46.5 69.7









#77
Carbazole
Concen: 21.196 ng/ul
RT: 17.628 min Scan# 2489
Delta R.T. -0.000 min
Lab File: BP009240.D
Acq: 25 Feb 2022 19:09




Instrument :
BNA_P
ClientSampleId :


Tgt Ion:167 Resp: 14294
Ion Ratio Lower Upper
167 100
166 22.6 15.8 23.8
139 12.9 8.3 12.5#



#78
Di-n-butylphthalate
Concen: 19.556 ng/ul
RT: 18.204 min Scan# 2587
Delta R.T. -0.000 min
Lab File: BP009240.D
Acq: 25 Feb 2022 19:09

Tgt Ion:149 Resp: 16878
Ion Ratio Lower Upper
149 100
150 10.1 9.6 14.4
104 6.9 6.2 9.4









#80
Fluoranthene
Concen: 18.665 ng/ul
RT: 19.245 min Scan# 2764
Delta R.T. -0.000 min
Lab File: BP009240.D
Acq: 25 Feb 2022 19:09




Instrument :
BNA_P
ClientSampleId :


Tgt Ion:202 Resp: 19570
Ion Ratio Lower Upper
202 100
101 9.9 9.6 14.4
100 9.0 6.3 9.5



#82
Pyrene
Concen: 19.179 ng/ul
RT: 19.598 min Scan# 2824
Delta R.T. -0.000 min
Lab File: BP009240.D
Acq: 25 Feb 2022 19:09

Tgt Ion:202 Resp: 20397
Ion Ratio Lower Upper
202 100
101 13.0 10.6 15.8
100 9.1 9.8 14.6#





#83

Butylbenzylphthalate

Concen: 18.283 ng/ul

RT: 20.492 min Scan# 2976

Delta R.T. -0.000 min

Lab File: BP009240.D

Acq: 25 Feb 2022 19:09

Instrument :

BNA_P

ClientSampleId :





Tgt Ion:149 Resp: 7835

Ion Ratio Lower Upper

149 100

91 75.0 71.1 106.7

206 29.9 28.8 43.2



#84

3,3'-Dichlorobenzidine

Concen: 20.250 ng/ul

RT: 21.251 min Scan# 3105

Delta R.T. -0.000 min

Lab File: BP009240.D

Acq: 25 Feb 2022 19:09




Tgt Ion:252 Resp: 9070


Ion Ratio Lower Upper

252 100

254 57.2 49.8 74.6

126 6.2 4.3 6.5









#85
Benzo(a)anthracene
Concen: 20.880 ng/ul
RT: 21.316 min Scan# 3116
Delta R.T. -0.006 min
Lab File: BP009240.D
Acq: 25 Feb 2022 19:09




Instrument :
BNA_P
ClientSampleId :


Tgt Ion:228 Resp: 23812
Ion Ratio Lower Upper
228 100
229 18.2 18.2 27.2
226 25.2 22.8 34.2



#86
Bis(2-ethylhexyl)phthalate
Concen: 18.134 ng/ul
RT: 21.263 min Scan# 3107
Delta R.T. -0.006 min
Lab File: BP009240.D
Acq: 25 Feb 2022 19:09

Tgt Ion:149 Resp: 11670
Ion Ratio Lower Upper
149 100
167 37.4 32.8 49.2
279 6.6 5.6 8.4









#87
Chrysene
Concen: 18.712 ng/ul
RT: 21.369 min Scan# 3125
Delta R.T. -0.000 min
Lab File: BP009240.D
Acq: 25 Feb 2022 19:09




Instrument :
BNA_P
ClientSampleId :


Tgt Ion:228 Resp: 20848
Ion Ratio Lower Upper
228 100
226 29.6 22.7 34.1
229 19.6 15.2 22.8



#89
Di-n-octyl phthalate
Concen: 18.290 ng/ul
RT: 22.198 min Scan# 3266
Delta R.T. -0.006 min
Lab File: BP009240.D
Acq: 25 Feb 2022 19:09

Tgt Ion:149 Resp: 21861









#90
Benzo(b)fluoranthene
Concen: 19.466 ng/ul
RT: 23.010 min Scan# 3404
Delta R.T. -0.000 min
Lab File: BP009240.D
Acq: 25 Feb 2022 19:09




Instrument :
BNA_P
ClientSampleId :


Tgt Ion:252 Resp: 24956
Ion Ratio Lower Upper
252 100
253 20.1 17.4 26.0
125 6.1 7.3 10.9#



#91
Benzo(k)fluoranthene
Concen: 20.288 ng/ul
RT: 23.057 min Scan# 3412
Delta R.T. -0.006 min
Lab File: BP009240.D
Acq: 25 Feb 2022 19:09

Tgt Ion:252 Resp: 23420
Ion Ratio Lower Upper
252 100
253 23.9 17.9 26.9
125 6.7 5.1 7.7





#93

Benzo(a)pyrene

Concen: 19.821 ng/ul

RT: 23.639 min Scan# 3511

Delta R.T. -0.000 min

Lab File: BP009240.D

Acq: 25 Feb 2022 19:09

Instrument :

BNA_P

ClientSampleId :





Tgt Ion:252 Resp: 23860

Ion Ratio Lower Upper

252 100

253 18.6 17.5 26.3

125 8.7 6.4 9.6



#94

Indeno(1,2,3-cd)pyrene

Concen: 13.992 ng/ul

RT: 26.257 min Scan# 3956

Delta R.T. -0.000 min

Lab File: BP009240.D

Acq: 25 Feb 2022 19:09




Tgt Ion:276 Resp: 20425


Ion Ratio Lower Upper

276 100

138 13.1 12.3 18.5

277 28.3 19.3 28.9





#95

Dibenzo(a,h)anthracene

Concen: 19.182 ng/ul

RT: 26.274 min Scan# 3959

Delta R.T. -0.006 min

Lab File: BP009240.D





Acq: 25 Feb 2022 19:09

Instrument :

BNA_P

ClientSampleId :

Tgt Ion:	278	Resp:	23716
Ion Ratio	Lower	Upper	
278	100		
139	11.0	6.6	10.0#
279	24.5	18.4	27.6



#96

Benzo(g,h,i)perylene

Concen: 19.261 ng/ul

RT: 27.021 min Scan# 4086

Delta R.T. -0.000 min

Lab File: BP009240.D

Acq: 25 Feb 2022 19:09

Tgt Ion:	276	Resp:	23428
Ion Ratio	Lower	Upper	
276	100		
138	8.5	11.6	17.4#
277	19.3	15.9	23.9

