

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP022621\
 Data File : BP004866.D
 Acq On : 26 Feb 2021 17:43
 Operator : CG/JU
 Sample : SSTDICV040
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 ICVBP022621

Quant Time: Feb 26 18:14:26 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP022621.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Feb 26 17:41:20 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	100	0.00
2	1,4-Dioxane	40.000	39.165	2.1	99	0.00
3	Pyridine	40.000	41.420	-3.6	101	0.00
4	n-Nitrosodimethylamine	40.000	39.332	1.7	99	0.00
5 S	2-Fluorophenol	80.000	79.437	0.7	101	0.00
6	Aniline	40.000	40.663	-1.7	102	0.00
7 S	Phenol-d6	80.000	80.902	-1.1	102	0.00
8	2-Chlorophenol	40.000	39.766	0.6	101	0.00
9	Benzaldehyde	40.000	42.794	-7.0	103	0.00
10 C	Phenol	40.000	39.841	0.4	100	0.00
11	bis(2-Chloroethyl)ether	40.000	40.691	-1.7	104	0.00
12	1,3-Dichlorobenzene	40.000	38.921	2.7	99	0.00
13 C	1,4-Dichlorobenzene	40.000	39.852	0.4	100	0.00
14	1,2-Dichlorobenzene	40.000	39.600	1.0	100	0.00
15	Benzyl Alcohol	40.000	40.561	-1.4	101	0.00
16	2,2'-oxybis(1-Chloropropane	40.000	39.696	0.8	104	0.00
17	2-Methylphenol	40.000	40.794	-2.0	102	0.00
18	Hexachloroethane	40.000	39.398	1.5	99	0.00
19 P	n-Nitroso-di-n-propylamine	40.000	40.919	-2.3	102	0.00
20	3+4-Methylphenols	40.000	40.147	-0.4	100	0.01
21 I	Naphthalene-d8	20.000	20.000	0.0	99	0.00
22	Acetophenone	40.000	40.343	-0.9	100	0.00
23 S	Nitrobenzene-d5	80.000	79.726	0.3	100	0.01
24	Nitrobenzene	40.000	40.298	-0.7	101	0.00
25	Isophorone	40.000	41.286	-3.2	103	0.00
26 C	2-Nitrophenol	40.000	41.454	-3.6	102	0.00
27	2,4-Dimethylphenol	40.000	40.413	-1.0	101	0.01
28	bis(2-Chloroethoxy)methane	40.000	40.604	-1.5	102	0.00
29 C	2,4-Dichlorophenol	40.000	40.800	-2.0	100	0.00
30	1,2,4-Trichlorobenzene	40.000	39.621	0.9	100	0.00
31	Naphthalene	40.000	39.909	0.2	101	0.01
32	Benzoic acid	40.000	43.224	-8.1	107	0.00
33	4-Chloroaniline	40.000	40.513	-1.3	100	0.00
34 C	Hexachlorobutadiene	40.000	39.111	2.2	98	0.00
35	Caprolactam	40.000	41.979	-4.9	105	0.01
36 C	4-Chloro-3-methylphenol	40.000	40.099	-0.2	101	0.01
37	2-Methylnaphthalene	40.000	39.740	0.6	101	0.01
38	1-Methylnaphthalene	40.000	39.775	0.6	101	0.01
39 I	Acenaphthene-d10	20.000	20.000	0.0	102	0.00
40	1,2,4,5-Tetrachlorobenzene	40.000	40.678	-1.7	100	0.01
41 P	Hexachlorocyclopentadiene	40.000	41.086	-2.7	100	0.00
42 S	2,4,6-Tribromophenol	80.000	79.967	0.0	99	0.00
43 C	2,4,6-Trichlorophenol	40.000	41.326	-3.3	103	0.01
44	2,4,5-Trichlorophenol	40.000	41.239	-3.1	103	0.00
45 S	2-Fluorobiphenyl	80.000	80.041	-0.1	100	0.00
46	1,1'-Biphenyl	40.000	40.118	-0.3	101	0.00
47	2-Chloronaphthalene	40.000	39.750	0.6	100	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP022621\
 Data File : BP004866.D
 Acq On : 26 Feb 2021 17:43
 Operator : CG/JU
 Sample : SSTDICV040
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 ICSVBP022621

Quant Time: Feb 26 18:14:26 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP022621.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Feb 26 17:41:20 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48	2-Nitroaniline	40.000	41.490	-3.7	102	0.00
49	Acenaphthylene	40.000	40.281	-0.7	100	0.01
50	Dimethylphthalate	40.000	40.499	-1.2	102	0.00
51	2,6-Dinitrotoluene	40.000	41.097	-2.7	102	0.00
52 C	Acenaphthene	40.000	40.433	-1.1	103	0.00
53	3-Nitroaniline	40.000	40.988	-2.5	102	0.01
54 P	2,4-Dinitrophenol	40.000	42.461	-6.2	109	0.00
55	Dibenzofuran	40.000	39.774	0.6	102	0.01
56 P	4-Nitrophenol	40.000	44.001	-10.0	109	0.01
57	2,4-Dinitrotoluene	40.000	41.090	-2.7	102	0.00
58	Fluorene	40.000	40.571	-1.4	103	0.01
59	2,3,4,6-Tetrachlorophenol	40.000	39.970	0.1	101	0.01
60	Diethylphthalate	40.000	40.089	-0.2	103	0.01
61	4-Chlorophenyl-phenylether	40.000	40.365	-0.9	102	0.01
62	4-Nitroaniline	40.000	41.677	-4.2	104	0.00
63	Azobenzene	40.000	40.313	-0.8	102	0.01
64 I	Phenanthrene-d10	20.000	20.000	0.0	102	0.01
65	4,6-Dinitro-2-methylphenol	40.000	42.913	-7.3	104	0.01
66 c	n-Nitrosodiphenylamine	40.000	40.455	-1.1	102	0.01
67	4-Bromophenyl-phenylether	40.000	40.060	-0.2	100	0.01
68	Hexachlorobenzene	40.000	40.006	-0.0	102	0.00
69	Atrazine	40.000	41.520	-3.8	105	0.01
70 C	Pentachlorophenol	40.000	42.478	-6.2	102	0.01
71	Phenanthrene	40.000	39.874	0.3	101	0.01
72	Anthracene	40.000	40.623	-1.6	103	0.01
73	Carbazole	40.000	40.543	-1.4	103	0.01
74	Di-n-butylphthalate	40.000	41.796	-4.5	105	0.00
75 C	Fluoranthene	40.000	39.921	0.2	103	0.01
76 I	Chrysene-d12	20.000	20.000	0.0	104	0.00
77	Benzidine	40.000	43.889	-9.7	107	0.00
78	Pyrene	40.000	40.807	-2.0	103	0.01
79 S	Terphenyl-d14	80.000	81.293	-1.6	102	0.01
80	Butylbenzylphthalate	40.000	42.523	-6.3	104	0.01
81	Benzo(a)anthracene	40.000	40.125	-0.3	103	0.01
82	3,3'-Dichlorobenzidine	40.000	40.602	-1.5	103	0.01
83	Chrysene	40.000	40.057	-0.1	103	0.01
84	Bis(2-ethylhexyl)phthalate	40.000	43.003	-7.5	107	0.00
85 c	Di-n-octyl phthalate	40.000	41.986	-5.0	104	0.01
86 I	Perylene-d12	20.000	20.000	0.0	102	0.02
87	Indeno(1,2,3-cd)pyrene	40.000	40.435	-1.1	103	0.02
88	Benzo(b)fluoranthene	40.000	40.175	-0.4	102	0.01
89	Benzo(k)fluoranthene	40.000	40.480	-1.2	103	0.01
90 C	Benzo(a)pyrene	40.000	40.398	-1.0	102	0.01
91	Dibenzo(a,h)anthracene	40.000	40.775	-1.9	104	0.02
92	Benzo(g,h,i)perylene	40.000	40.608	-1.5	104	0.02

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP022621\
Data File : BP004866.D
Acq On : 26 Feb 2021 17:43
Operator : CG/JU
Sample : SSTDICV040
Misc :
ALS Vial : 9 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
ICVBP022621

Quant Time: Feb 26 18:14:26 2021
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP022621.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Feb 26 17:41:20 2021
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range SPCC's out = 0 CCC's out = 0