

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP022825\
 Data File : BP024131.D
 Acq On : 28 Feb 2025 12:09
 Operator : RC/JU
 Sample : PB166914BL
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 PB166914BL

Quant Time: Feb 28 12:36:03 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP021725.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Feb 18 02:25:07 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.763	152	401774	20.000	ng	0.00
21) Naphthalene-d8	10.540	136	1583690	20.000	ng	0.00
39) Acenaphthene-d10	14.398	164	978288	20.000	ng	0.00
64) Phenanthrene-d10	17.210	188	1805232	20.000	ng	0.02
76) Chrysene-d12	21.651	240	1521502	20.000	ng	0.00
86) Perylene-d12	25.051	264	1635731	20.000	ng	0.02
System Monitoring Compounds						
5) 2-Fluorophenol	5.381	112	3696274	143.551	ng	0.00
7) Phenol-d6	6.946	99	4622913	139.690	ng	0.00
23) Nitrobenzene-d5	8.910	82	2712345	95.548	ng	0.00
42) 2,4,6-Tribromophenol	15.922	330	1851511	154.512	ng	0.02
45) 2-Fluorobiphenyl	13.004	172	6217222	88.846	ng	0.00
79) Terphenyl-d14	19.927	244	8313708	108.122	ng	0.01

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP022825\
Data File : BP024131.D
Acq On : 28 Feb 2025 12:09
Operator : RC/JU
Sample : PB166914BL
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
PB166914BL

Quant Time: Feb 28 12:36:03 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP021725.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Feb 18 02:25:07 2025
Response via : Initial Calibration

