

Data Path : Z:\SVOASRV\HPCHEM1\BNA\_P\DATA\BP022920\  
 Data File : BP001971.D  
 Acq On : 29 Feb 2020 22:27  
 Operator : CG/JU  
 Sample : L1615-19  
 Misc :  
 ALS Vial : 19 Sample Multiplier: 1

Instrument :  
 BNA\_P  
 ClientSampleId :

Quant Time: Mar 02 00:31:54 2020  
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA\_P\METHODS\SOM-EPA-BP022720MA.M  
 Quant Title : SVOA CALIBRATION  
 QLast Update : Thu Feb 27 14:59:37 2020  
 Response via : Initial Calibration

| Internal Standards        | R.T.  | QIon | Response | Conc  | Units | Dev(Min) |
|---------------------------|-------|------|----------|-------|-------|----------|
| 1) 1,4-Dichlorobenzene-d4 | 7.64  | 152  | 234693   | 20.00 | ng/ul | 0.00     |
| 18) Naphthalene-d8        | 10.41 | 136  | 947076   | 20.00 | ng/ul | 0.00     |
| 36) Acenaphthene-d10      | 14.27 | 164  | 606878   | 20.00 | ng/ul | 0.00     |
| 62) Phenanthrene-d10      | 17.03 | 188  | 1333609  | 20.00 | ng/ul | 0.00     |
| 78) Chrysene-d12          | 21.12 | 240  | 1328505  | 20.00 | ng/ul | 0.00     |
| 86) Perylene-d12          | 23.33 | 264  | 1527417  | 20.00 | ng/ul | -0.01    |

| System Monitoring Compounds    | R.T.  | QIon | Response | Conc  | Units | Dev(Min) |
|--------------------------------|-------|------|----------|-------|-------|----------|
| 3) 1,4-Dioxane-d8              | 3.17  | 96   | 13161    | 2.41  | ng/uL | 0.00     |
| 5) Phenol-d5                   | 6.82  | 99   | 243556   | 13.60 | ng/ul | 0.00     |
| 7) Bis-(2-Chloroethyl)ether-d  | 6.98  | 67   | 146311   | 15.43 | ng/ul | 0.00     |
| 9) 2-Chlorophenol-d4           | 7.18  | 132  | 220154   | 14.22 | ng/ul | 0.00     |
| 13) 4-Methylphenol-d8          | 8.35  | 113  | 161280   | 10.96 | ng/ul | 0.00     |
| 19) Nitrobenzene-d5            | 8.78  | 128  | 97121    | 13.35 | ng/ul | 0.00     |
| 22) 2-Nitrophenol-d4           | 9.50  | 143  | 103167   | 12.60 | ng/ul | 0.00     |
| 26) 2,4-Dichlorophenol-d3      | 10.04 | 165  | 208444   | 13.22 | ng/ul | 0.00     |
| 29) 4-Chloroaniline-d4         | 10.55 | 131  | 192649   | 11.15 | ng/ul | 0.00     |
| 44) Dimethylphthalate-d6       | 13.69 | 166  | 702861   | 15.30 | ng/ul | 0.00     |
| 47) Acenaphthylene-d8          | 13.96 | 160  | 824173   | 15.00 | ng/ul | 0.00     |
| 52) 4-Nitrophenol-d4           | 14.48 | 143  | 66127    | 7.53  | ng/ul | 0.00     |
| 58) Fluorene-d10               | 15.27 | 176  | 643538   | 16.19 | ng/ul | 0.00     |
| 63) 4,6-Dinitro-2-methylphenol | 15.39 | 200  | 59980    | 8.16  | ng/ul | 0.00     |
| 71) Anthracene-d10             | 17.13 | 188  | 943691   | 15.74 | ng/ul | 0.00     |
| 79) Pyrene-d10                 | 19.37 | 212  | 1163288  | 16.73 | ng/ul | 0.00     |
| 90) Benzo(a)pyrene-d12         | 23.19 | 264  | 1211748  | 15.51 | ng/ul | -0.01    |

| Target Compounds | Ovalue |
|------------------|--------|
| -----            |        |

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA P\DATA\BP022920\  
Data File : BP001971.D  
Acq On : 29 Feb 2020 22:27  
Operator : CG/JU  
Sample : L1615-19  
Misc :  
ALS Vial : 19 Sample Multiplier: 1

Instrument :  
BNA\_P  
ClientSampleId :

Quant Time: Mar 02 00:31:54 2020  
Quant Method : Z:\SVOASRV\HPCHEM1\BNA\_P\METHODS\SOM-EPA-BP022720MA.M  
Quant Title : SVOA CALIBRATION  
QLast Update : Thu Feb 27 14:59:37 2020  
Response via : Initial Calibration

