

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP030222\
 Data File : BP009248.D
 Acq On : 02 Mar 2022 16:39
 Operator : CG/JU
 Sample : SSTD02010
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SSTD020614

Quant Time: Mar 02 17:37:09 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP030222.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed Mar 02 17:32:26 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.828	152	349859	20.000	ng/ul	# 0.00
20) Naphthalene-d8	10.628	136	1357567	20.000	ng/ul	# 0.00
38) Acenaphthene-d10	14.475	164	791554	20.000	ng/ul	0.00
64) Phenanthrene-d10	17.233	188	1553347	20.000	ng/ul	# 0.00
79) Chrysene-d12	21.339	240	1553653	20.000	ng/ul	# 0.00
88) Perylene-d12	23.756	264	1674106	20.000	ng/ul	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.463	96	304	0.036	ng/uL	0.14
4) Pyridine-d5	3.722	84	430087	19.842	ng/ul	0.00
7) Phenol-d5	6.993	99	556179	19.654	ng/ul	0.00
9) Bis-(2-Chloroethyl)eth...	7.163	67	333688	19.570	ng/ul	0.00
11) 2-Chlorophenol-d4	7.363	132	441596	19.913	ng/ul	0.00
15) 4-Methylphenol-d8	8.534	113	424412	19.532	ng/ul	0.00
21) Nitrobenzene-d5	8.981	128	172526	22.423	ng/ul	0.00
24) 2-Nitrophenol-d4	9.704	143	167791	23.264	ng/ul	0.00
28) 2,4-Dichlorophenol-d3	10.245	165	411805	20.690	ng/ul	0.00
31) 4-Chloroaniline-d4	10.757	131	568856	20.142	ng/ul	0.00
46) Dimethylphthalate-d6	13.886	166	1120465	20.212	ng/ul	0.00
49) Acenaphthylene-d8	14.163	160	1465999	19.957	ng/ul	0.00
54) 4-Nitrophenol-d4	14.657	143	176283	20.861	ng/ul	0.00
60) Fluorene-d10	15.469	176	971027	20.193	ng/ul	0.00
65) 4,6-Dinitro-2-methylph...	15.580	200	116672	19.341	ng/ul	0.00
73) Anthracene-d10	17.333	188	1473143	20.202	ng/ul	0.00
81) Pyrene-d10	19.574	212	1698737	19.777	ng/ul	0.00
92) Benzo(a)pyrene-d12	23.598	264	1664196	19.616	ng/ul	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.352	88	68769	7.938	ng/uL#	49
5) Pyridine	3.740	79	462033	20.177	ng/ul#	75
6) Benzaldehyde	6.969	77	390600	23.306	ng/ul	99
8) Phenol	7.022	94	563044	19.750	ng/ul	90
10) Bis(2-Chloroethyl)ether	7.257	93	445854	19.517	ng/ul#	86
12) 2-Chlorophenol	7.393	128	453227	20.225	ng/ul#	88
13) 2-Methylphenol	8.269	108	411592	19.205	ng/ul	99
14) 2,2'-oxybis(1-Chloropr...	8.363	45	540033	19.591	ng/ul#	91
16) Acetophenone	8.651	105	658607	19.704	ng/ul#	87
17) N-Nitroso-di-n-propyla...	8.640	70	326829	19.603	ng/ul#	83
18) 4-Methylphenol	8.598	108	443375	19.718	ng/ul	92
19) Hexachloroethane	8.910	117	181343	19.770	ng/ul#	77
22) Nitrobenzene	9.022	77	451571	22.243	ng/ul#	85
23) Isophorone	9.557	82	902096	19.640	ng/ul#	95
25) 2-Nitrophenol	9.734	139	197969	23.177	ng/ul#	75
26) 2,4-Dimethylphenol	9.804	107	479174	20.211	ng/ul#	71
27) Bis(2-Chloroethoxy)met...	10.040	93	558198	19.952	ng/ul#	97
29) 2,4-Dichlorophenol	10.269	162	403569	20.551	ng/ul#	83
30) Naphthalene	10.675	128	1377254	19.982	ng/ul	96
32) 4-Chloroaniline	10.781	127	569715	20.435	ng/ul	98
33) Hexachlorobutadiene	10.981	225	282386	19.809	ng/ul	95
34) Caprolactam	11.545	113	108461	19.133	ng/ul#	46
35) 4-Chloro-3-methylphenol	12.016	107	800	0.040	ng/ul	91
36) 2-Methylnaphthalene	12.292	142	917869	19.818	ng/ul	91


Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP030222\
 Data File : BP009248.D
 Acq On : 02 Mar 2022 16:39
 Operator : CG/JU
 Sample : SST02010
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SST020614

Quant Time: Mar 02 17:37:09 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP030222.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed Mar 02 17:32:26 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.510	142	929478	20.088	ng/ul#	96
39) 1,2,4,5-Tetrachloroben...	12.663	216	501536	19.905	ng/ul	94
40) Hexachlorocyclopentadiene	12.651	237	303738	19.365	ng/ul	99
41) 2,4,6-Trichlorophenol	12.898	196	300296	20.628	ng/ul#	83
42) 2,4,5-Trichlorophenol	12.969	196	316465	20.544	ng/ul#	86
43) 1,1'-Biphenyl	13.310	154	1201463	19.836	ng/ul#	94
44) 2-Chloronaphthalene	13.345	162	920098	19.982	ng/ul	92
45) 2-Nitroaniline	13.539	65	176338	22.640	ng/ul#	85
47) Dimethylphthalate	13.933	163	1086929	20.309	ng/ul#	92
48) 2,6-Dinitrotoluene	14.039	165	162350	22.826	ng/ul	97
50) Acenaphthylene	14.192	152	1471574	20.270	ng/ul	95
51) 3-Nitroaniline	14.369	138	188033	20.981	ng/ul	80
52) Acenaphthene	14.533	153	940411	19.915	ng/ul	98
53) 2,4-Dinitrophenol	14.575	184	120607	31.880	ng/ul#	82
55) 4-Nitrophenol	14.675	109	139122	20.591	ng/ul#	34
56) Dibenzofuran	14.875	168	1343838	20.146	ng/ul	98
57) 2,4-Dinitrotoluene	14.828	165	243243	24.073	ng/ul	91
58) 2,3,4,6-Tetrachlorophenol	15.098	232	254492	21.347	ng/ul#	86
59) Diethylphthalate	15.304	149	1056971	20.438	ng/ul	93
61) Fluorene	15.527	166	1065665	20.436	ng/ul	90
62) 4-Chlorophenyl-phenyle...	15.527	204	535513	20.220	ng/ul	97
63) 4-Nitroaniline	15.533	138	195721	22.664	ng/ul#	71
66) 4,6-Dinitro-2-methylph...	15.598	198	127031	20.072	ng/ul#	94
67) N-Nitrosodiphenylamine	15.733	169	919059	20.090	ng/ul	94
68) 4-Bromophenyl-phenylether	16.422	248	321835	19.722	ng/ul	94
69) Hexachlorobenzene	16.539	284	368591	19.768	ng/ul#	89
70) Atrazine	16.698	200	319133	19.696	ng/ul#	90
71) Pentachlorophenol	16.880	266	207058	18.976	ng/ul#	85
72) Phenanthrene	17.274	178	1660516	20.083	ng/ul	98
74) Anthracene	17.369	178	1674069	20.288	ng/ul	98
75) 1,2,3,4-Tetrachloroben...	13.269	216	510172	19.183	ng/ul#	87
76) Pentachlorobenzene	14.798	250	452698	19.176	ng/ul	88
77) Carbazole	17.639	167	1510366	20.629	ng/ul#	96
78) Di-n-butylphthalate	18.210	149	1696015	20.797	ng/ul#	92
80) Fluoranthene	19.251	202	1960632	19.793	ng/ul	98
82) Pyrene	19.604	202	1990821	19.695	ng/ul	98
83) Butylbenzylphthalate	20.498	149	694101	21.538	ng/ul#	80
84) 3,3'-Dichlorobenzidine	21.257	252	665204	20.599	ng/ul#	96
85) Benzo(a)anthracene	21.327	228	1966064	20.033	ng/ul	95
86) Bis(2-ethylhexyl)phtha...	21.274	149	1089987	21.147	ng/ul#	80
87) Chrysene	21.380	228	1905850	20.058	ng/ul	99
89) Di-n-octyl phthalate	22.204	149	1735059	19.176	ng/ul	100
90) Benzo(b)fluoranthene	23.021	252	2068435	19.577	ng/ul	99
91) Benzo(k)fluoranthene	23.068	252	1918275	20.137	ng/ul#	96
93) Benzo(a)pyrene	23.651	252	2003944	19.820	ng/ul#	98
94) Indeno(1,2,3-cd)pyrene	26.268	276	2353257	19.686	ng/ul#	92
95) Dibenzo(a,h)anthracene	26.292	278	2002060	19.823	ng/ul#	92
96) Benzo(g,h,i)perylene	27.039	276	1998891	19.478	ng/ul#	86





(#) = qualifier out of range (m) = manual integration (+) = signals summed



#2
1,4-Dioxane
Concen: 7.938 ng/uL
RT: 3.352 min Scan# 61
Delta R.T. 0.000 min
Lab File: BP009248.D
Acq: 02 Mar 2022 16:39




Instrument :
BNA_P
ClientSampleId :
SSTD020614


Tgt Ion: 88 Resp: 68769
Ion Ratio Lower Upper
88 100
43 26.0 6.8 10.2#
58 65.9 28.4 42.6#



#5
Pyridine
Concen: 20.177 ng/uL
RT: 3.740 min Scan# 128
Delta R.T. 0.000 min
Lab File: BP009248.D
Acq: 02 Mar 2022 16:39

Tgt Ion: 79 Resp: 462033
Ion Ratio Lower Upper
79 100
52 66.1 35.5 53.3#
51 30.8 20.2 30.4#









#6
Benzaldehyde
Concen: 23.306 ng/ul
RT: 6.969 min Scan# 61
Delta R.T. 0.000 min
Lab File: BP009248.D
Acq: 02 Mar 2022 16:39




Instrument :
BNA_P
ClientSampleId :
SSTD020614


Tgt Ion: 77 Resp: 390600
Ion Ratio Lower Upper
77 100
105 96.5 77.0 115.4
106 94.7 77.0 115.4



#8
Phenol
Concen: 19.750 ng/ul
RT: 7.022 min Scan# 686
Delta R.T. 0.000 min
Lab File: BP009248.D
Acq: 02 Mar 2022 16:39

Tgt Ion: 94 Resp: 563044
Ion Ratio Lower Upper
94 100
65 26.7 19.0 28.6
66 35.7 34.6 51.8









#10
 Bis(2-Chloroethyl)ether
 Concen: 19.517 ng/ul
 RT: 7.257 min Scan# 726
 Delta R.T. 0.000 min
 Lab File: BP009248.D
 Acq: 02 Mar 2022 16:39




Instrument :
 BNA_P
 ClientSampleId :
 SSTD020614


Tgt Ion: 93 Resp: 445854
 Ion Ratio Lower Upper
 93 100
 63 72.9 45.0 67.4#
 95 32.9 26.3 39.5



#12
 2-Chlorophenol
 Concen: 20.225 ng/ul
 RT: 7.393 min Scan# 749
 Delta R.T. 0.000 min
 Lab File: BP009248.D
 Acq: 02 Mar 2022 16:39

Tgt Ion:128 Resp: 453227
 Ion Ratio Lower Upper
 128 100
 64 45.2 45.3 67.9#
 130 32.2 29.0 43.6





#13

2-Methylphenol

Concen: 19.205 ng/ul

RT: 8.269 min Scan# 898

Delta R.T. 0.000 min

Lab File: BP009248.D

Acq: 02 Mar 2022 16:39

Instrument :

BNA_P

ClientSampleId :





SSTD020614

Tgt Ion:108 Resp: 411592

Ion Ratio Lower Upper

108 100

107 89.1 70.8 106.2



#14

2,2'-oxybis(1-Chloropropane)

Concen: 19.591 ng/ul

RT: 8.363 min Scan# 914

Delta R.T. 0.000 min

Lab File: BP009248.D

Acq: 02 Mar 2022 16:39




Tgt Ion: 45 Resp: 540033


Ion Ratio Lower Upper

45 100

77 16.3 17.1 25.7#

79 13.7 13.2 19.8









#16
 Acetophenone
 Concen: 19.704 ng/ul
 RT: 8.651 min Scan# 963
 Delta R.T. 0.000 min
 Lab File: BP009248.D
 Acq: 02 Mar 2022 16:39




Instrument :
 BNA_P
 ClientSampleId :
 SSTD020614


Tgt Ion:105 Resp: 658607
 Ion Ratio Lower Upper
 105 100
 77 77.1 53.4 80.2
 51 23.7 13.8 20.8#



#17
 N-Nitroso-di-n-propylamine
 Concen: 19.603 ng/ul
 RT: 8.640 min Scan# 961
 Delta R.T. 0.000 min
 Lab File: BP009248.D
 Acq: 02 Mar 2022 16:39

Tgt Ion: 70 Resp: 326829
 Ion Ratio Lower Upper
 70 100
 42 51.3 31.8 47.8#
 101 9.6 15.6 23.4#
 130 20.2 13.7 20.5









#18
4-Methylphenol
Concen: 19.718 ng/ul
RT: 8.598 min Scan# 91
Delta R.T. 0.000 min
Lab File: BP009248.D
Acq: 02 Mar 2022 16:39




Instrument :
BNA_P
ClientSampleId :
SSTD020614


Tgt Ion:108 Resp: 443375
Ion Ratio Lower Upper
108 100
107 112.3 83.6 125.4



#19
Hexachloroethane
Concen: 19.770 ng/ul
RT: 8.910 min Scan# 1007
Delta R.T. 0.000 min
Lab File: BP009248.D
Acq: 02 Mar 2022 16:39

Tgt Ion:117 Resp: 181343
Ion Ratio Lower Upper
117 100
201 101.5 106.5 159.7#
199 60.4 59.0 88.6









#22
Nitrobenzene
Concen: 22.243 ng/ul
RT: 9.022 min Scan# 1026
Delta R.T. 0.000 min
Lab File: BP009248.D
Acq: 02 Mar 2022 16:39




Instrument :
BNA_P
ClientSampleId :
SSTD020614


Tgt Ion: 77 Resp: 451571
Ion Ratio Lower Upper
77 100
123 42.7 42.3 63.5
65 12.9 5.8 8.8#



#23
Isophorone
Concen: 19.640 ng/ul
RT: 9.557 min Scan# 1117
Delta R.T. 0.000 min
Lab File: BP009248.D
Acq: 02 Mar 2022 16:39

Tgt Ion: 82 Resp: 902096
Ion Ratio Lower Upper
82 100
95 7.0 7.8 11.8#
138 17.5 12.9 19.3









#25
2-Nitrophenol
Concen: 23.177 ng/ul
RT: 9.734 min Scan# 1147
Delta R.T. 0.000 min
Lab File: BP009248.D
Acq: 02 Mar 2022 16:39




Instrument :
BNA_P
ClientSampleId :
SSTD020614


Tgt Ion:139 Resp: 197969
Ion Ratio Lower Upper
139 100
65 50.2 49.4 74.2
109 29.3 44.7 67.1#



#26
2,4-Dimethylphenol
Concen: 20.211 ng/ul
RT: 9.804 min Scan# 1159
Delta R.T. 0.000 min
Lab File: BP009248.D
Acq: 02 Mar 2022 16:39

Tgt Ion:107 Resp: 479174
Ion Ratio Lower Upper
107 100
121 50.9 34.1 51.1
122 87.6 46.3 69.5#









#27
 Bis(2-Chloroethoxy)methane
 Concen: 19.952 ng/ul
 RT: 10.040 min Scan# 1199
 Delta R.T. 0.000 min
 Lab File: BP009248.D
 Acq: 02 Mar 2022 16:39




Instrument :
 BNA_P
 ClientSampleId :
 SSTD020614


Tgt Ion: 93 Resp: 558198
 Ion Ratio Lower Upper
 93 100
 95 32.1 26.2 39.4
 123 11.0 6.2 9.4#



#29
 2,4-Dichlorophenol
 Concen: 20.551 ng/ul
 RT: 10.269 min Scan# 1238
 Delta R.T. 0.000 min
 Lab File: BP009248.D
 Acq: 02 Mar 2022 16:39

Tgt Ion:162 Resp: 403569
 Ion Ratio Lower Upper
 162 100
 164 63.5 37.0 55.6#
 98 35.1 31.0 46.6









#30
Naphthalene
Concen: 19.982 ng/ul
RT: 10.675 min Scan# 1307
Delta R.T. 0.000 min
Lab File: BP009248.D
Acq: 02 Mar 2022 16:39




Instrument :
BNA_P
ClientSampleId :
SSTD020614


Tgt Ion:128 Resp: 1377254
Ion Ratio Lower Upper
128 100
129 11.0 8.6 13.0
127 12.7 12.2 18.4



#32
4-Chloroaniline
Concen: 20.435 ng/ul
RT: 10.781 min Scan# 1325
Delta R.T. 0.000 min
Lab File: BP009248.D
Acq: 02 Mar 2022 16:39

Tgt Ion:127 Resp: 569715
Ion Ratio Lower Upper
127 100
129 32.1 24.6 37.0









#33
Hexachlorobutadiene
Concen: 19.809 ng/ul
RT: 10.981 min Scan# 1359
Delta R.T. 0.000 min
Lab File: BP009248.D
Acq: 02 Mar 2022 16:39




Instrument :
BNA_P
ClientSampleId :
SSTD020614


Tgt Ion:225 Resp: 282386
Ion Ratio Lower Upper
225 100
223 62.2 53.4 80.0
227 64.0 53.8 80.8



#34
Caprolactam
Concen: 19.133 ng/ul
RT: 11.545 min Scan# 1455
Delta R.T. 0.000 min
Lab File: BP009248.D
Acq: 02 Mar 2022 16:39

Tgt Ion:113 Resp: 108461
Ion Ratio Lower Upper
113 100
55 146.1 58.2 87.2#
56 108.5 68.6 103.0#









#35
4-Chloro-3-methylphenol
Concen: 0.040 ng/ul
RT: 12.016 min Scan# 11
Delta R.T. 0.106 min
Lab File: BP009248.D
Acq: 02 Mar 2022 16:39




Instrument :
BNA_P
ClientSampleId :
SSTD020614


Tgt Ion:107 Resp: 800
Ion Ratio Lower Upper
107 100
144 23.0 15.8 23.8
142 55.5 38.9 58.3



#36
2-Methylnaphthalene
Concen: 19.818 ng/ul
RT: 12.292 min Scan# 1582
Delta R.T. 0.000 min
Lab File: BP009248.D
Acq: 02 Mar 2022 16:39

Tgt Ion:142 Resp: 917869
Ion Ratio Lower Upper
142 100
141 88.2 64.3 96.5









#37
1-Methylnaphthalene
Concen: 20.088 ng/ul
RT: 12.510 min Scan# 1619
Delta R.T. 0.000 min
Lab File: BP009248.D
Acq: 02 Mar 2022 16:39




Instrument :
BNA_P
ClientSampleId :
SSTD020614


Tgt Ion:142 Resp: 929478
Ion Ratio Lower Upper
142 100
141 90.8 69.8 104.8
116 3.8 5.6 8.4#



#39
1,2,4,5-Tetrachlorobenzene
Concen: 19.905 ng/ul
RT: 12.663 min Scan# 1645
Delta R.T. 0.000 min
Lab File: BP009248.D
Acq: 02 Mar 2022 16:39

Tgt Ion:216 Resp: 501536
Ion Ratio Lower Upper
216 100
214 77.2 58.2 87.2
179 20.9 14.2 21.2
108 16.9 15.3 22.9









#40
Hexachlorocyclopentadiene
Concen: 19.365 ng/ul
RT: 12.651 min Scan# 1643
Delta R.T. 0.000 min
Lab File: BP009248.D
Acq: 02 Mar 2022 16:39




Instrument :
BNA_P
ClientSampleId :
SSTD020614


Tgt Ion:237 Resp: 303738
Ion Ratio Lower Upper
237 100
235 65.6 53.2 79.8
272 12.5 10.8 16.2



#41
2,4,6-Trichlorophenol
Concen: 20.628 ng/ul
RT: 12.898 min Scan# 1685
Delta R.T. 0.000 min
Lab File: BP009248.D
Acq: 02 Mar 2022 16:39

Tgt Ion:196 Resp: 300296
Ion Ratio Lower Upper
196 100
198 96.4 64.1 96.1#
200 30.3 19.9 29.9#









#42
2,4,5-Trichlorophenol
Concen: 20.544 ng/ul
RT: 12.969 min Scan# 1697
Delta R.T. 0.000 min
Lab File: BP009248.D
Acq: 02 Mar 2022 16:39




Instrument :
BNA_P
ClientSampleId :
SSTD020614


Tgt Ion:196 Resp: 316465
Ion Ratio Lower Upper
196 100
198 97.2 68.6 103.0
200 31.0 32.8 49.2#



#43
1,1'-Biphenyl
Concen: 19.836 ng/ul
RT: 13.310 min Scan# 1755
Delta R.T. 0.000 min
Lab File: BP009248.D
Acq: 02 Mar 2022 16:39

Tgt Ion:154 Resp: 1201463
Ion Ratio Lower Upper
154 100
153 39.3 28.6 43.0
76 13.3 8.5 12.7#









#44
2-Chloronaphthalene
Concen: 19.982 ng/ul
RT: 13.345 min Scan# 1761
Delta R.T. 0.000 min
Lab File: BP009248.D
Acq: 02 Mar 2022 16:39




Instrument :
BNA_P
ClientSampleId :
SSTD020614


Tgt Ion:162 Resp: 920098
Ion Ratio Lower Upper
162 100
164 32.7 27.5 41.3
127 38.9 37.1 55.7



#45
2-Nitroaniline
Concen: 22.640 ng/ul
RT: 13.539 min Scan# 1794
Delta R.T. 0.000 min
Lab File: BP009248.D
Acq: 02 Mar 2022 16:39

Tgt Ion: 65 Resp: 176338
Ion Ratio Lower Upper
65 100
92 72.4 77.9 116.9#
138 107.3 81.5 122.3









#47
Dimethylphthalate
Concen: 20.309 ng/ul
RT: 13.933 min Scan# 1861
Delta R.T. 0.000 min
Lab File: BP009248.D
Acq: 02 Mar 2022 16:39




Instrument :
BNA_P
ClientSampleId :
SSTD020614


Tgt Ion:163 Resp: 1086929
Ion Ratio Lower Upper
163 100
194 4.1 4.1 6.1#
164 9.9 11.3 16.9#



#48
2,6-Dinitrotoluene
Concen: 22.826 ng/ul
RT: 14.039 min Scan# 1879
Delta R.T. 0.000 min
Lab File: BP009248.D
Acq: 02 Mar 2022 16:39

Tgt Ion:165 Resp: 162350
Ion Ratio Lower Upper
165 100
89 61.3 51.7 77.5
121 23.4 18.3 27.5









#50
Acenaphthylene
Concen: 20.270 ng/ul
RT: 14.192 min Scan# 1905
Delta R.T. 0.000 min
Lab File: BP009248.D
Acq: 02 Mar 2022 16:39




Instrument :
BNA_P
ClientSampleId :
SSTD020614


Tgt Ion:152 Resp: 1471574
Ion Ratio Lower Upper
152 100
151 19.5 18.3 27.5
153 12.7 11.0 16.6



#51
3-Nitroaniline
Concen: 20.981 ng/ul
RT: 14.369 min Scan# 1935
Delta R.T. 0.000 min
Lab File: BP009248.D
Acq: 02 Mar 2022 16:39

Tgt Ion:138 Resp: 188033
Ion Ratio Lower Upper
138 100
108 17.8 15.8 23.8
92 118.4 117.1 175.7









#52
Acenaphthene
Concen: 19.915 ng/ul
RT: 14.533 min Scan# 1963
Delta R.T. 0.000 min
Lab File: BP009248.D
Acq: 02 Mar 2022 16:39




Instrument :
BNA_P
ClientSampleId :
SSTD020614


Tgt Ion:153 Resp: 940411
Ion Ratio Lower Upper
153 100
152 47.1 34.3 51.5
154 87.1 70.0 105.0



#53
2,4-Dinitrophenol
Concen: 31.880 ng/ul
RT: 14.575 min Scan# 1970
Delta R.T. 0.000 min
Lab File: BP009248.D
Acq: 02 Mar 2022 16:39

Tgt Ion:184 Resp: 120607
Ion Ratio Lower Upper
184 100
63 56.6 34.6 52.0#
154 58.6 57.7 86.5









#55
4-Nitrophenol
Concen: 20.591 ng/ul
RT: 14.675 min Scan# 1987
Delta R.T. 0.000 min
Lab File: BP009248.D
Acq: 02 Mar 2022 16:39




Instrument :
BNA_P
ClientSampleId :
SSTD020614


Tgt Ion:109 Resp: 139122
Ion Ratio Lower Upper
109 100
139 144.6 61.4 92.2#
65 129.2 65.3 97.9#



#56
Dibenzofuran
Concen: 20.146 ng/ul
RT: 14.875 min Scan# 2021
Delta R.T. 0.000 min
Lab File: BP009248.D
Acq: 02 Mar 2022 16:39

Tgt Ion:168 Resp: 1343838
Ion Ratio Lower Upper
168 100
139 36.5 30.1 45.1









#57
2,4-Dinitrotoluene
Concen: 24.073 ng/ul
RT: 14.828 min Scan# 2013
Delta R.T. 0.000 min
Lab File: BP009248.D
Acq: 02 Mar 2022 16:39




Instrument :
BNA_P
ClientSampleId :
SSTD020614


Tgt Ion:165 Resp: 243243
Ion Ratio Lower Upper
165 100
63 42.6 29.5 44.3
182 2.8 2.8 4.2



#58
2,3,4,6-Tetrachlorophenol
Concen: 21.347 ng/ul
RT: 15.098 min Scan# 2059
Delta R.T. 0.000 min
Lab File: BP009248.D
Acq: 02 Mar 2022 16:39

Tgt Ion:232 Resp: 254492
Ion Ratio Lower Upper
232 100
131 43.1 25.4 38.2#
130 2.3 4.2 6.2#
166 30.1 28.1 42.1









#59
Diethylphthalate
Concen: 20.438 ng/ul
RT: 15.304 min Scan# 2094
Delta R.T. 0.000 min
Lab File: BP009248.D
Acq: 02 Mar 2022 16:39




Instrument :
BNA_P
ClientSampleId :
SSTD020614


Tgt Ion:149 Resp: 1056971
Ion Ratio Lower Upper
149 100
177 22.5 22.0 33.0
150 12.2 10.2 15.2



#61
Fluorene
Concen: 20.436 ng/ul
RT: 15.527 min Scan# 2132
Delta R.T. 0.000 min
Lab File: BP009248.D
Acq: 02 Mar 2022 16:39

Tgt Ion:166 Resp: 1065665
Ion Ratio Lower Upper
166 100
165 98.0 70.3 105.5
167 13.4 9.5 14.3









#62
4-Chlorophenyl-phenylether
Concen: 20.220 ng/ul
RT: 15.527 min Scan# 2132
Delta R.T. 0.000 min
Lab File: BP009248.D
Acq: 02 Mar 2022 16:39




Instrument :
BNA_P
ClientSampleId :
SSTD020614


Tgt Ion:204 Resp: 535513
Ion Ratio Lower Upper
204 100
206 33.1 27.6 41.4
141 55.9 46.6 69.8



#63
4-Nitroaniline
Concen: 22.664 ng/ul
RT: 15.533 min Scan# 2133
Delta R.T. 0.000 min
Lab File: BP009248.D
Acq: 02 Mar 2022 16:39

Tgt Ion:138 Resp: 195721
Ion Ratio Lower Upper
138 100
92 52.9 41.6 62.4
108 76.6 97.6 146.4#









#66
4,6-Dinitro-2-methylphenol
Concen: 20.072 ng/ul
RT: 15.598 min Scan# 2144
Delta R.T. 0.000 min
Lab File: BP009248.D
Acq: 02 Mar 2022 16:39




Instrument :
BNA_P
ClientSampleId :
SSTD020614


Tgt Ion:198 Resp: 127031
Ion Ratio Lower Upper
198 100
182 5.1 6.0 9.0#
77 24.9 17.7 26.5



#67
N-Nitrosodiphenylamine
Concen: 20.090 ng/ul
RT: 15.733 min Scan# 2167
Delta R.T. 0.000 min
Lab File: BP009248.D
Acq: 02 Mar 2022 16:39

Tgt Ion:169 Resp: 919059
Ion Ratio Lower Upper
169 100
168 65.9 48.3 72.5
167 35.2 29.6 44.4









#68
4-Bromophenyl-phenylether
Concen: 19.722 ng/ul
RT: 16.422 min Scan# 21
Delta R.T. 0.000 min
Lab File: BP009248.D
Acq: 02 Mar 2022 16:39




Instrument :
BNA_P
ClientSampleId :
SSTD020614


Tgt Ion:248 Resp: 321835
Ion Ratio Lower Upper
248 100
250 97.2 79.4 119.0
141 74.2 67.2 100.8



#69
Hexachlorobenzene
Concen: 19.768 ng/ul
RT: 16.539 min Scan# 2304
Delta R.T. 0.000 min
Lab File: BP009248.D
Acq: 02 Mar 2022 16:39

Tgt Ion:284 Resp: 368591
Ion Ratio Lower Upper
284 100
142 33.6 19.2 28.8#
249 31.1 27.4 41.2









#70
Atrazine
Concen: 19.696 ng/ul
RT: 16.698 min Scan# 21
Delta R.T. 0.000 min
Lab File: BP009248.D
Acq: 02 Mar 2022 16:39




Instrument :
BNA_P
ClientSampleId :
SSTD020614


Tgt Ion:200 Resp: 319133
Ion Ratio Lower Upper
200 100
173 26.6 19.2 28.8
215 48.7 32.2 48.4#



#71
Pentachlorophenol
Concen: 18.976 ng/ul
RT: 16.880 min Scan# 2362
Delta R.T. 0.000 min
Lab File: BP009248.D
Acq: 02 Mar 2022 16:39

Tgt Ion:266 Resp: 207058
Ion Ratio Lower Upper
266 100
264 62.9 57.3 85.9
268 64.7 65.7 98.5#





#72

Phenanthrene

Concen: 20.083 ng/u1

RT: 17.274 min Scan# 2429

Delta R.T. 0.000 min

Lab File: BP009248.D

Acq: 02 Mar 2022 16:39

Instrument :

BNA_P

ClientSampleId :

SSTD020614





Tgt Ion:178 Resp: 1660516

Ion Ratio Lower Upper

178 100

179 15.4 11.7 17.5

176 18.8 14.5 21.7



#74

Anthracene

Concen: 20.288 ng/u1

RT: 17.369 min Scan# 2445

Delta R.T. 0.000 min

Lab File: BP009248.D

Acq: 02 Mar 2022 16:39




Tgt Ion:178 Resp: 1674069


Ion Ratio Lower Upper

178 100

179 15.1 12.3 18.5

176 18.2 13.4 20.2





#75

1,2,3,4-Tetrachlorobenzene

Concen: 19.183 ng/uL

RT: 13.269 min Scan# 1

Delta R.T. 0.000 min

Lab File: BP009248.D

Acq: 02 Mar 2022 16:39

Instrument :

BNA_P

ClientSampleId :

SSTD020614





Tgt Ion:216 Resp: 510172

Ion Ratio Lower Upper

216 100

214 77.1 50.4 75.6#

218 48.1 35.2 52.8



#76

Pentachlorobenzene

Concen: 19.176 ng/uL

RT: 14.798 min Scan# 2008

Delta R.T. 0.000 min

Lab File: BP009248.D

Acq: 02 Mar 2022 16:39




Tgt Ion:250 Resp: 452698


Ion Ratio Lower Upper

250 100

248 64.9 43.5 65.3

252 64.6 46.5 69.7









#77
 Carbazole
 Concen: 20.629 ng/ul
 RT: 17.639 min Scan# 2491
 Delta R.T. 0.000 min
 Lab File: BP009248.D
 Acq: 02 Mar 2022 16:39




Instrument :
 BNA_P
 ClientSampleId :
 SSTD020614


Tgt Ion:167 Resp: 1510366
 Ion Ratio Lower Upper
 167 100
 166 21.1 15.8 23.8
 139 12.6 8.3 12.5#



#78
 Di-n-butylphthalate
 Concen: 20.797 ng/ul
 RT: 18.210 min Scan# 2588
 Delta R.T. 0.000 min
 Lab File: BP009248.D
 Acq: 02 Mar 2022 16:39

Tgt Ion:149 Resp: 1696015
 Ion Ratio Lower Upper
 149 100
 150 9.0 9.6 14.4#
 104 5.1 6.2 9.4#









#80
Fluoranthene
Concen: 19.793 ng/u1
RT: 19.251 min Scan# 21
Delta R.T. 0.000 min
Lab File: BP009248.D
Acq: 02 Mar 2022 16:39




Instrument :
BNA_P
ClientSampleId :
SSTD020614


Tgt Ion:202 Resp: 1960632
Ion Ratio Lower Upper
202 100
101 11.4 9.6 14.4
100 8.8 6.3 9.5



#82
Pyrene
Concen: 19.695 ng/u1
RT: 19.604 min Scan# 2825
Delta R.T. 0.000 min
Lab File: BP009248.D
Acq: 02 Mar 2022 16:39

Tgt Ion:202 Resp: 1990821
Ion Ratio Lower Upper
202 100
101 14.0 10.6 15.8
100 11.2 9.8 14.6





#83

Butylbenzylphthalate

Concen: 21.538 ng/ul

RT: 20.498 min Scan# 2977

Delta R.T. 0.000 min

Lab File: BP009248.D

Acq: 02 Mar 2022 16:39

Instrument :

BNA_P

ClientSampleId :

SSTD020614





Tgt Ion:149 Resp: 694101

Ion Ratio Lower Upper

149 100

91 71.9 71.1 106.7

206 21.3 28.8 43.2#



#84

3,3'-Dichlorobenzidine

Concen: 20.599 ng/ul

RT: 21.257 min Scan# 3106

Delta R.T. 0.000 min

Lab File: BP009248.D

Acq: 02 Mar 2022 16:39




Tgt Ion:252 Resp: 665204


Ion Ratio Lower Upper

252 100

254 64.3 49.8 74.6

126 11.4 4.3 6.5#









#85
Benzo(a)anthracene
Concen: 20.033 ng/ul
RT: 21.327 min Scan# 3118
Delta R.T. 0.000 min
Lab File: BP009248.D
Acq: 02 Mar 2022 16:39




Instrument :
BNA_P
ClientSampleId :
SSTD020614


Tgt Ion:228 Resp: 1966064
Ion Ratio Lower Upper
228 100
229 19.3 18.2 27.2
226 26.5 22.8 34.2



#86
Bis(2-ethylhexyl)phthalate
Concen: 21.147 ng/ul
RT: 21.274 min Scan# 3109
Delta R.T. 0.000 min
Lab File: BP009248.D
Acq: 02 Mar 2022 16:39

Tgt Ion:149 Resp: 1089987
Ion Ratio Lower Upper
149 100
167 27.4 32.8 49.2#
279 4.4 5.6 8.4#









#87
Chrysene
Concen: 20.058 ng/ul
RT: 21.380 min Scan# 3127
Delta R.T. 0.000 min
Lab File: BP009248.D
Acq: 02 Mar 2022 16:39

Instrument : BNA_P
ClientSampleId : SSTD020614




Tgt Ion:228 Resp: 1905850


Ion	Ratio	Lower	Upper
228	100		
226	29.6	22.7	34.1
229	19.1	15.2	22.8



#89
Di-n-octyl phthalate
Concen: 19.176 ng/ul
RT: 22.204 min Scan# 3267
Delta R.T. 0.000 min
Lab File: BP009248.D
Acq: 02 Mar 2022 16:39

Tgt Ion:149 Resp: 1735059









#90
Benzo(b)fluoranthene
Concen: 19.577 ng/ul
RT: 23.021 min Scan# 3406
Delta R.T. 0.000 min
Lab File: BP009248.D
Acq: 02 Mar 2022 16:39




Instrument :
BNA_P
ClientSampleId :
SSTD020614


Tgt Ion:252 Resp: 2068435
Ion Ratio Lower Upper
252 100
253 21.9 17.4 26.0
125 10.3 7.3 10.9



#91
Benzo(k)fluoranthene
Concen: 20.137 ng/ul
RT: 23.068 min Scan# 3414
Delta R.T. 0.000 min
Lab File: BP009248.D
Acq: 02 Mar 2022 16:39

Tgt Ion:252 Resp: 1918275
Ion Ratio Lower Upper
252 100
253 21.7 17.9 26.9
125 10.0 5.1 7.7#









#93
Benzo(a)pyrene
Concen: 19.820 ng/ul
RT: 23.651 min Scan# 3513
Delta R.T. 0.000 min
Lab File: BP009248.D
Acq: 02 Mar 2022 16:39




Instrument :
BNA_P
ClientSampleId :
SSTD020614


Tgt Ion:252 Resp: 2003944
Ion Ratio Lower Upper
252 100
253 21.9 17.5 26.3
125 11.0 6.4 9.6#



#94
Indeno(1,2,3-cd)pyrene
Concen: 19.686 ng/ul
RT: 26.268 min Scan# 3958
Delta R.T. 0.000 min
Lab File: BP009248.D
Acq: 02 Mar 2022 16:39

Tgt Ion:276 Resp: 2353257
Ion Ratio Lower Upper
276 100
138 22.8 12.3 18.5#
277 24.7 19.3 28.9





#95

Dibenzo(a,h)anthracene

Concen: 19.823 ng/ul

RT: 26.292 min Scan# 3962

Delta R.T. 0.000 min

Lab File: BP009248.D

Acq: 02 Mar 2022 16:39

Instrument :

BNA_P

ClientSampleId :

SSTD020614





Tgt Ion:278 Resp: 2002060

Ion Ratio Lower Upper

278 100

139 16.1 6.6 10.0#

279 24.3 18.4 27.6



#96

Benzo(g,h,i)perylene

Concen: 19.478 ng/ul

RT: 27.039 min Scan# 4089

Delta R.T. 0.000 min

Lab File: BP009248.D

Acq: 02 Mar 2022 16:39

Tgt Ion:276 Resp: 1998891

Ion Ratio Lower Upper

276 100

138 22.5 11.6 17.4#

277 24.2 15.9 23.9#

