

Data Path : Z:\SVOASRV\HPCHEM1\BNA_P\DATA\BP030521\
 Data File : BP004920.D
 Acq On : 05 Mar 2021 17:17
 Operator : CG/JU
 Sample : PB134930BL
 Misc : SOM-MDL-S-BL-05
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SBLK30

Quant Time: Mar 05 17:49:23 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_P\METHODS\SOM-EPA-BP030521MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Fri Mar 05 13:02:07 2021
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.73	152	32554	20.00	ng/ul	0.00
18) Naphthalene-d8	10.53	136	127272	20.00	ng/ul	0.00
36) Acenaphthene-d10	14.39	164	86184	20.00	ng/ul	0.00
62) Phenanthrene-d10	17.15	188	195212	20.00	ng/ul	0.00
78) Chrysene-d12	21.24	240	221985	20.00	ng/ul	0.00
86) Perylene-d12	23.54	264	228851	20.00	ng/ul	0.00

System Monitoring Compounds

3) 1,4-Dioxane-d8	3.23	96	5062	6.31	ng/uL	0.00
5) Phenol-d5	6.91	99	83097	33.39	ng/ul	0.00
7) Bis-(2-Chloroethyl)ether-d	7.08	67	46216	36.21	ng/ul	0.00
9) 2-Chlorophenol-d4	7.27	132	67265	33.90	ng/ul	0.00
13) 4-Methylphenol-d8	8.45	113	68860	34.24	ng/ul	0.00
19) Nitrobenzene-d5	8.89	128	32211	35.25	ng/ul	0.00
22) 2-Nitrophenol-d4	9.62	143	38026	36.43	ng/ul	0.00
26) 2,4-Dichlorophenol-d3	10.15	165	65328	31.81	ng/ul	0.00
29) 4-Chloroaniline-d4	10.67	131	90164	40.21	ng/ul	0.00
44) Dimethylphthalate-d6	13.80	166	243607	39.29	ng/ul	0.00
47) Acenaphthylene-d8	14.08	160	273953	35.91	ng/ul	0.00
52) 4-Nitrophenol-d4	14.60	143	37135	34.11	ng/ul	0.00
58) Fluorene-d10	15.39	176	199576	36.12	ng/ul	0.00
63) 4,6-Dinitro-2-methylphenol	15.51	200	41417	32.41	ng/ul	0.00
71) Anthracene-d10	17.25	188	333249	38.02	ng/ul	0.00
79) Pyrene-d10	19.49	212	406210	40.59	ng/ul	0.00
90) Benzo(a)pyrene-d12	23.39	264	465629	40.03	ng/ul	0.00

Target Compounds Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA P\DATA\BP030521\
Data File : BP004920.D
Acq On : 05 Mar 2021 17:17
Operator : CG/JU
Sample : PB134930BL
Misc : SOM-MDL-S-BL-05
ALS Vial : 15 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
SBLK30

Quant Time: Mar 05 17:49:23 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_P\METHODS\SOM-EPA-BP030521MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Fri Mar 05 13:02:07 2021
Response via : Initial Calibration

